

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061020\
 Data File : BN010821.D
 Acq On : 10 Jun 2020 17:27
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 10 18:29:26 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 10 15:14:26 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	115	0.00
2	1,4-Dioxane	0.392	0.418	-6.6	117	0.00
3	n-Nitrosodimethylamine	0.228	0.195	14.5	112	0.00
4 S	2-Fluorophenol	0.772	0.813	-5.3	118	0.00
5 S	Phenol-d6	0.910	0.958	-5.3	121	0.00
6	bis(2-Chloroethyl)ether	0.887	0.890	-0.3	115	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	113	0.00
8 S	Nitrobenzene-d5	0.296	0.314	-6.1	116	0.00
9	Naphthalene	1.005	1.070	-6.5	115	0.00
10	Hexachlorobutadiene	0.248	0.261	-5.2	110	0.00
11 SURR	2-Methylnaphthalene-d10	0.620	0.658	-6.1	115	0.00
12	2-Methylnaphthalene	0.673	0.698	-3.7	114	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	116	0.00
14 S	2,4,6-Tribromophenol	0.143	0.136	4.9	114	0.00
15 S	2-Fluorobiphenyl	1.578	1.665	-5.5	115	0.00
16	Acenaphthylene	1.639	1.638	0.1	115	0.00
17	Acenaphthene	1.117	1.141	-2.1	114	0.00
18	Fluorene	1.441	1.467	-1.8	115	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	114	0.00
20	4-Bromophenyl-phenylether	0.240	0.247	-2.9	115	0.00
21	Hexachlorobenzene	0.251	0.261	-4.0	113	0.00
22	Pentachlorophenol	0.084	0.080	4.8	112	0.00
23	Phenanthrene	1.093	1.127	-3.1	115	0.00
24	Anthracene	0.912	0.917	-0.5	116	0.00
25 SURR	Fluoranthene-d10	1.104	1.118	-1.3	115	0.00
26	Fluoranthene	1.240	1.239	0.1	114	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	111	0.00
28	Pvrene	1.323	1.415	-7.0	115	0.00
29 S	Terphenyl-d14	0.936	1.001	-6.9	114	0.00
30	Benzo(a)anthracene	1.176	1.161	1.3	111	0.00
31	Chrysene	1.370	1.475	-7.7	111	0.01
32	Bis(2-ethylhexyl)phthalate	0.392	0.416	-6.1	121	0.01
33	Indeno(1,2,3-cd)pyrene	1.531	1.586	-3.6	110	0.00
34 I	Perylene-d12	1.000	1.000	0.0	113	0.00
35	Benzo(b)fluoranthene	1.360	1.380	-1.5	110	0.00
36	Benzo(k)fluoranthene	1.449	1.486	-2.6	110	0.00
37 C	Benzo(a)pyrene	1.167	1.168	-0.1	112	0.00
38	Dibenzo(a,h)anthracene	1.330	1.332	-0.2	111	0.00
39	Benzo(a,h,i)perylene	1.391	1.399	-0.6	107	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061020\
Data File : BN010821.D
Acq On : 10 Jun 2020 17:27
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 10 18:29:26 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 10 15:14:26 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			