

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061020\
 Data File : BN010821.D
 Acq On : 10 Jun 2020 17:27
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 10 18:29:26 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 10 15:14:26 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	115	0.00
2	1,4-Dioxane	0.400	0.427	-6.7	117	0.00
3	n-Nitrosodimethylamine	0.400	0.343	14.2	112	0.00
4 S	2-Fluorophenol	0.400	0.421	-5.2	118	0.00
5 S	Phenol-d6	0.400	0.421	-5.2	121	0.00
6	bis(2-Chloroethyl)ether	0.400	0.402	-0.5	115	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	113	0.00
8 S	Nitrobenzene-d5	0.400	0.425	-6.2	116	0.00
9	Naphthalene	0.400	0.426	-6.5	115	0.00
10	Hexachlorobutadiene	0.400	0.420	-5.0	110	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.424	-6.0	115	0.00
12	2-Methylnaphthalene	0.400	0.415	-3.7	114	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	116	0.00
14 S	2,4,6-Tribromophenol	0.400	0.379	5.3	114	0.00
15 S	2-Fluorobiphenyl	0.400	0.422	-5.5	115	0.00
16	Acenaphthylene	0.400	0.400	0.0	115	0.00
17	Acenaphthene	0.400	0.408	-2.0	114	0.00
18	Fluorene	0.400	0.407	-1.7	115	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	114	0.00
20	4-Bromophenyl-phenylether	0.400	0.411	-2.7	115	0.00
21	Hexachlorobenzene	0.400	0.416	-4.0	113	0.00
22	Pentachlorophenol	0.400	0.380	5.0	112	0.00
23	Phenanthrene	0.400	0.413	-3.2	115	0.00
24	Anthracene	0.400	0.402	-0.5	116	0.00
25 SURR	Fluoranthene-d10	0.400	0.405	-1.3	115	0.00
26	Fluoranthene	0.400	0.400	0.0	114	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	111	0.00
28	Pvrene	0.400	0.428	-7.0	115	0.00
29 S	Terphenyl-d14	0.400	0.428	-7.0	114	0.00
30	Benzo(a)anthracene	0.400	0.395	1.3	111	0.00
31	Chrysene	0.400	0.431	-7.7	111	0.01
32	Bis(2-ethylhexyl)phthalate	0.400	0.424	-6.0	121	0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.414	-3.5	110	0.00
34 I	Perylene-d12	0.400	0.400	0.0	113	0.00
35	Benzo(b)fluoranthene	0.400	0.406	-1.5	110	0.00
36	Benzo(k)fluoranthene	0.400	0.410	-2.5	110	0.00
37 C	Benzo(a)pyrene	0.400	0.401	-0.3	112	0.00
38	Dibenzo(a,h)anthracene	0.400	0.400	0.0	111	0.00
39	Benzo(a,h,i)perylene	0.400	0.402	-0.5	107	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061020\
Data File : BN010821.D
Acq On : 10 Jun 2020 17:27
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 10 18:29:26 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 10 15:14:26 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				