

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061020\
 Data File : BN010832.D
 Acq On : 11 Jun 2020 00:03
 Operator : CG/JU
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDC00.4

Quant Time: Jun 11 00:52:24 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 10 15:14:26 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	150#	0.00
2	1,4-Dioxane	0.392	0.412	-5.1	151#	0.00
3	n-Nitrosodimethylamine	0.228	0.241	-5.7	181#	0.00
4 S	2-Fluorophenol	0.772	1.042	-35.0#	198#	0.00
5 S	Phenol-d6	0.910	1.239	-36.2#	205#	0.00
6	bis(2-Chloroethyl)ether	0.887	1.071	-20.7	182#	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	165#	0.00
8 S	Nitrobenzene-d5	0.296	0.313	-5.7	168#	0.00
9	Naphthalene	1.005	1.056	-5.1	166#	0.00
10	Hexachlorobutadiene	0.248	0.243	2.0	151#	0.00
11 SURR	2-Methylnaphthalene-d10	0.620	0.658	-6.1	167#	0.00
12	2-Methylnaphthalene	0.673	0.715	-6.2	171#	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	173#	0.00
14 S	2,4,6-Tribromophenol	0.143	0.167	-16.8	209#	0.00
15 S	2-Fluorobiphenyl	1.578	1.568	0.6	161#	0.00
16	Acenaphthylene	1.639	1.764	-7.6	184#	0.00
17	Acenaphthene	1.117	1.160	-3.8	172#	0.00
18	Fluorene	1.441	1.477	-2.5	172#	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	171#	-0.01
20	4-Bromophenyl-phenylether	0.240	0.231	3.7	161#	0.00
21	Hexachlorobenzene	0.251	0.258	-2.8	168#	0.00
22	Pentachlorophenol	0.084	0.102	-21.4	214#	0.00
23	Phenanthrene	1.093	1.148	-5.0	176#	0.00
24	Anthracene	0.912	1.006	-10.3	191#	0.00
25 SURR	Fluoranthene-d10	1.104	1.254	-13.6	193#	0.00
26	Fluoranthene	1.240	1.376	-11.0	190#	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	196#	0.00
28	Pvrene	1.323	1.394	-5.4	198#	0.00
29 S	Terphenyl-d14	0.936	0.938	-0.2	187#	0.00
30	Benzo(a)anthracene	1.176	1.339	-13.9	225#	0.00
31	Chrysene	1.370	1.411	-3.0	186#	0.00
32	Bis(2-ethylhexyl)phthalate	0.392	0.667	-70.2#	341#	0.01
33	Indeno(1,2,3-cd)pyrene	1.531	1.634	-6.7	200#	0.00
34 I	Perylene-d12	1.000	1.000	0.0	220#	0.00
35	Benzo(b)fluoranthene	1.360	1.304	4.1	204#	0.00
36	Benzo(k)fluoranthene	1.449	1.247	13.9	180#	0.00
37 C	Benzo(a)pyrene	1.167	1.204	-3.2	226#	0.00
38	Dibenzo(a,h)anthracene	1.330	1.255	5.6	204#	0.00
39	Benzo(a,h,i)perylene	1.391	1.269	8.8	189#	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061020\
Data File : BN010832.D
Acq On : 11 Jun 2020 00:03
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 11 00:52:24 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 10 15:14:26 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			