

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061020\
 Data File : BN010814.D
 Acq On : 10 Jun 2020 12:14
 Operator : CG/JU
 Sample : SSTDICC8.0
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC8.0

Manual Integrations
 APPROVED

mohammad
 6/11/2020 1:49:11 PM

Quant Time: Jun 10 13:47:27 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 10 13:41:15 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.56	152	2296	0.40	ng	0.00
7) Naphthalene-d8	10.30	136	7743	0.40	ng	0.00
13) Acenaphthene-d10	14.18	164	4261	0.40	ng	0.00
19) Phenanthrene-d10	16.92	188	8821	0.40	ng	-0.01
27) Chrysene-d12	21.13	240	8113	0.40	ng	0.00
34) Perylene-d12	23.28	264	7667	0.40	ng	0.00

System Monitoring Compounds						
4) 2-Fluorophenol	5.21	112	3344	0.73	ng	0.00
5) Phenol-d6	6.75	99	3960	0.73	ng	0.00
8) Nitrobenzene-d5	8.69	82	4376	0.77	ng	0.00
11) 2-Methylnaphthalene-d10	11.90	152	9099	0.77	ng	0.00
14) 2,4,6-Tribromophenol	15.67	330	1141	0.77	ng	-0.01
15) 2-Fluorobiphenyl	12.80	172	12985	0.78	ng	0.00
25) Fluoranthene-d10	18.96	212	18342	0.75	ng	0.00
29) Terphenyl-d14	19.58	244	14475	0.80	ng	0.00

Target Compounds					Qvalue	
2) 1,4-Dioxane	3.22	88	1741	0.767	ng	99
3) n-Nitrosodimethylamine	3.55	42	979	0.741	ng	# 93
6) bis(2-Chloroethyl)ether	7.00	93	4101	0.795	ng	95
9) Naphthalene	10.35	128	14651	0.760	ng	98
10) Hexachlorobutadiene	10.66	225	3617	0.768	ng	# 99
12) 2-Methylnaphthalene	11.98	142	9886	0.773	ng	99
16) Acenaphthylene	13.89	152	13135	0.761	ng	100
17) Acenaphthene	14.24	154	9074	0.775	ng	99
18) Fluorene	15.23	166	11553	0.761	ng	98
20) 4-Bromophenyl-phenylether	16.12	248	4017	0.781	ng	# 79
21) Hexachlorobenzene	16.24	284	4260	0.784	ng	98
22) Pentachlorophenol	16.59	266	1291	0.698	ng	96
23) Phenanthrene	16.96	178	18388	0.771	ng	99
24) Anthracene	17.06	178	14972	0.741	ng	100
26) Fluoranthene	18.99	202	20655	0.748	ng	99
28) Pyrene	19.36	202	20216	0.786	ng	100
30) Benzo(a)anthracene	21.12	228	17382	0.730	ng	99
31) Chrysene	21.16	228	21151	0.782	ng	98
32) Bis(2-ethylhexyl)phthalate	21.07	149	5901	0.750	ng	98
33) Indeno(1,2,3-cd)pyrene	25.39	276	22995	0.708	ng	99
35) Benzo(b)fluoranthene	22.65	252	19641m	0.787	ng	
36) Benzo(k)fluoranthene	22.69	252	21264	0.774	ng	# 95
37) Benzo(a)pyrene	23.19	252	16485	0.740	ng	# 90
38) Dibenzo(a,h)anthracene	25.41	278	19073	0.760	ng	95
39) Benzo(g,h,i)perylene	26.03	276	19586	0.754	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061020\
Data File : BN010814.D
Acq On : 10 Jun 2020 12:14
Operator : CG/JU
Sample : SSTDICC8.0
Misc :
ALS Vial : 5 Sample Multiplier: 1


Instrument :
BNA_N
ClientSampleId :
SSTDICC8.0

Manual Integrations
APPROVED

mohammad
6/11/2020 1:49:11 PM

Quant Time: Jun 10 13:47:27 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 10 13:41:15 2020
Response via : Initial Calibration





#1


1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.56 min Scan# 587
Delta R.T. 0.00 min
Lab File: BN010814.D
Acq: 10 Jun 2020 12:14

Instrument :
BNA_N
ClientSampleId :
SSTDICC8.0

Tot Ion	Ratio	Lower	Upper
152	100		
150	156.5	123.8	185.6
115	60.8	49.2	73.8

Manual Integrations
APPROVED

mohammad
6/11/2020 1:49:11 PM






Abundance

Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E


Ion 115.00 (114.70 to 115.70): E



#2

1,4-Dioxane
Concen: 0.767 ng
RT: 3.22 min Scan# 49
Delta R.T. 0.00 min
Lab File: BN010814.D
Acq: 10 Jun 2020 12:14

Tgt Ion	Ratio	Lower	Upper
88	100		
58	48.5	38.2	57.4
43	18.2	15.1	22.7








Abundance

Ion 88.00 (87.70 to 88.70): BNC

Ion 58.00 (57.70 to 58.70): BNC

Ion 43.00 (42.70 to 43.70): BNC









#3
n-Nitrosodimethylamine
Concen: 0.741 ng
RT: 3.55 min Scan# 89
Delta R.T. -0.02 min
Lab File: BN010814.D
Acq: 10 Jun 2020 12:14

Tot Ion:	42	Resp:	979
Ion	Ratio	Lower	Upper
42	100		
74	166.4	140.7	211.1
44	5.1	1.4	2.2#

Instrument :
BNA_N
ClientSampleId :
SSTDICC8.0


Manual Integrations
APPROVED




mohammad
6/11/2020 1:49:11 PM



#4
2-Fluorophenol
Concen: 0.730 ng
RT: 5.21 min Scan# 295
Delta R.T. 0.00 min
Lab File: BN010814.D
Acq: 10 Jun 2020 12:14

Tgt Ion:	112	Resp:	3344
Ion	Ratio	Lower	Upper
112	100		
64	46.5	37.4	56.0
63	26.7	22.2	33.4





#5

Phenol-d6

Concen: 0.732 ng

RT: 6.75 min Scan# 487

Delta R.T. 0.00 min

Lab File: BN010814.D

Acq: 10 Jun 2020 12:14

Tot Ion:	99	Resp:	3960
Ion	Ratio	Lower	Upper
99	100		
42	14.7	11.1	16.7
71	39.1	31.4	47.2

Instrument :





BNA_N

ClientSampled :

SSTDICC8.0

Manual Integrations
APPROVED

mohammad
6/11/2020 1:49:11 PM



#6

bis(2-Chloroethyl)ether

Concen: 0.795 ng


RT: 7.00 min Scan# 518


Delta R.T. 0.00 min

Lab File: BN010814.D

Acq: 10 Jun 2020 12:14

Tgt Ion:	93	Resp:	4101
Ion	Ratio	Lower	Upper
93	100		
63	57.4	49.4	74.2
95	31.5	26.7	40.1





#7

Naphthalene-d8

Concen: 0.400 ng

RT: 10.30 min Scan# 838

Delta R.T. 0.00 min

Lab File: BN010814.D

Acq: 10 Jun 2020 12:14

Instrument :

BNA_N


ClientSampleId :

SSTDICC8.0

Tot Ion:	136	Resp:	7743
Ion	Ratio	Lower	Upper
136	100		
137	12.0	9.7	14.5
54	5.3	4.9	7.3
68	5.9	4.7	7.1

Manual Integrations
APPROVED

mohammad
6/11/2020 1:49:11 PM






Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BNC

Ion 68.00 (67.70 to 68.70): BNC



#8

Nitrobenzene-d5

Concen: 0.771 ng


RT: 8.69 min Scan# 711

Delta R.T. 0.00 min

Lab File: BN010814.D

Acq: 10 Jun 2020 12:14



Tgt Ion:	82	Resp:	4376
Ion	Ratio	Lower	Upper
82	100		
128	45.1	35.4	53.0
54	31.6	28.2	42.4




Abundance Ion 82.00 (81.70 to 82.70): BNC

Ion 128.00 (127.70 to 128.70): E

Ion 54.00 (53.70 to 54.70): BNC





#9

Naphthalene

Concen: 0.760 ng

RT: 10.35 min Scan# 842

Delta R.T. 0.00 min

Lab File: BN010814.D

Acq: 10 Jun 2020 12:14

Instrument :

BNA_N





ClientSampleId :

SSTDICC8.0

Tot Ion:	128	Resp:	14651
Ion Ratio	Lower	Upper	
128	100		
129	11.5	9.6	14.4
127	13.4	11.4	17.2

Manual Integrations
APPROVED

mohammad
6/11/2020 1:49:11 PM



#10

Hexachlorobutadiene

Concen: 0.768 ng




RT: 10.66 min Scan# 866


Delta R.T. 0.00 min

Lab File: BN010814.D

Acq: 10 Jun 2020 12:14

Tgt Ion:	225	Resp:	3617
Ion Ratio	Lower	Upper	
225	100		
223	0.0	0.0	0.0
227	64.1	50.5	75.7









#11
2-Methylnaphthalene-d10
Concen: 0.775 ng
RT: 11.90 min Scan# 1022
Delta R.T. -0.00 min
Lab File: BN010814.D
Acq: 10 Jun 2020 12:14

Instrument :
BNA_N
ClientSampleId :
SSTDICC8.0

Tot Ion	Ratio	Lower	Upper
152	100		
151	19.3	15.3	22.9




Manual Integrations
APPROVED


mohammad
6/11/2020 1:49:11 PM



#12
2-Methylnaphthalene
Concen: 0.773 ng
RT: 11.98 min Scan# 1039
Delta R.T. -0.00 min
Lab File: BN010814.D
Acq: 10 Jun 2020 12:14

Tgt Ion	Ratio	Lower	Upper
142	100		
141	88.2	70.8	106.2
115	38.6	31.5	47.3









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.18 min Scan# 1308
Delta R.T. 0.00 min
Lab File: BN010814.D
Acq: 10 Jun 2020 12:14

Instrument :
BNA_N
ClientSampleId :
SSTDICC8.0

Tot Ion	Ratio	Lower	Upper
164	100		
162	100.3	81.0	121.6
160	48.6	38.9	58.3




Manual Integrations
APPROVED


mohammad
6/11/2020 1:49:11 PM



#14
2,4,6-Tribromophenol
Concen: 0.766 ng
RT: 15.67 min Scan# 1442
Delta R.T. -0.01 min
Lab File: BN010814.D
Acq: 10 Jun 2020 12:14

Tgt Ion	Ratio	Lower	Upper
330	100		
332	0.0	0.0	0.0
141	39.5	33.2	49.8









#15
2-Fluorobiphenyl
Concen: 0.782 ng
RT: 12.80 min Scan# 1184
Delta R.T. 0.00 min
Lab File: BN010814.D
Acq: 10 Jun 2020 12:14

Instrument :
BNA_N
ClientSampled :
SSTDICC8.0

Tot Ion	Ratio	Lower	Upper
172	100		
171	35.2	28.5	42.7
170	23.2	19.7	29.5




Manual Integrations
APPROVED


mohammad
6/11/2020 1:49:11 PM



#16
Acenaphthylene
Concen: 0.761 ng
RT: 13.89 min Scan# 1282
Delta R.T. 0.00 min
Lab File: BN010814.D
Acq: 10 Jun 2020 12:14

Tgt Ion	Ratio	Lower	Upper
152	100		
151	19.5	15.6	23.4
153	13.0	10.6	16.0









#17
Acenaphthene
Concen: 0.775 ng
RT: 14.24 min Scan# 1313
Delta R.T. 0.00 min
Lab File: BN010814.D
Acq: 10 Jun 2020 12:14

Instrument :
BNA_N
ClientSampleId :
SSTDICC8.0

Tot Ion	Ratio	Lower	Upper
154	100		
153	113.9	91.8	137.6
152	58.7	47.6	71.4




Manual Integrations
APPROVED


mohammad
6/11/2020 1:49:11 PM



#18
Fluorene
Concen: 0.761 ng
RT: 15.23 min Scan# 1402
Delta R.T. -0.01 min
Lab File: BN010814.D
Acq: 10 Jun 2020 12:14

Tgt Ion	Ratio	Lower	Upper
166	100		
165	98.3	77.2	115.8
167	13.4	10.7	16.1





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 16.92 min Scan# 1558

Delta R.T. -0.01 min

Lab File: BN010814.D

Acq: 10 Jun 2020 12:14

Instrument :

BNA_N





ClientSampleId :

SSTDICC8.0

Tot Ion:	188	Resp:	8821
Ion Ratio	Lower	Upper	
188	100		
94	0.0	0.0	0.0
80	8.7	4.5	6.7#

Manual Integrations
APPROVED

mohammad
6/11/2020 1:49:11 PM



#20

4-Bromophenyl-phenylether

Concen: 0.781 ng




RT: 16.12 min Scan# 1483


Delta R.T. -0.01 min

Lab File: BN010814.D

Acq: 10 Jun 2020 12:14

Tgt Ion:	248	Resp:	4017
Ion Ratio	Lower	Upper	
248	100		
250	91.7	81.8	122.8
141	81.9	42.6	64.0#





#21

Hexachlorobenzene

Concen: 0.784 ng

RT: 16.24 min Scan# 1494

Delta R.T. 0.00 min

Lab File: BN010814.D

Acq: 10 Jun 2020 12:14

Instrument :

BNA_N





ClientSampleId :

SSTDICC8.0

Tot Ion:	284	Resp:	4260
Ion Ratio	Lower	Upper	
284	100		
142	40.1	31.0	46.4
249	32.5	27.1	40.7

Manual Integrations
APPROVED

mohammad
6/11/2020 1:49:11 PM



#22

Pentachlorophenol

Concen: 0.698 ng




RT: 16.59 min Scan# 1527


Delta R.T. 0.00 min

Lab File: BN010814.D

Acq: 10 Jun 2020 12:14

Tgt Ion:	266	Resp:	1291
Ion Ratio	Lower	Upper	
266	100		
264	62.9	54.8	82.2
268	69.0	55.9	83.9









#23
Phenanthrene
Concen: 0.771 ng
RT: 16.96 min Scan# 1562
Delta R.T. 0.00 min
Lab File: BN010814.D
Acq: 10 Jun 2020 12:14

Instrument :
BNA_N
ClientSampleId :
SSTDICC8.0

Tot Ion	Ratio	Lower	Upper
178	100		
176	19.0	15.3	22.9
179	15.5	12.6	19.0




Manual Integrations
APPROVED


mohammad
6/11/2020 1:49:11 PM



#24
Anthracene
Concen: 0.741 ng
RT: 17.06 min Scan# 1571
Delta R.T. 0.00 min
Lab File: BN010814.D
Acq: 10 Jun 2020 12:14

Tgt Ion	Ratio	Lower	Upper
178	100		
176	18.3	14.6	22.0
179	15.4	12.3	18.5





#25

Fluoranthene-d10

Concen: 0.749 ng

RT: 18.96 min Scan# 1853

Delta R.T. 0.00 min

Lab File: BN010814.D

Acq: 10 Jun 2020 12:14

Instrument :

BNA_N


ClientSampleId :

SSTDICC8.0




Tot Ion:	212	Resp:	18342
Ion Ratio	Lower	Upper	
212	100		
106	8.3	6.6	10.0
104	4.8	4.0	6.0

Manual Integrations
APPROVED

mohammad
6/11/2020 1:49:11 PM



Abundance Ion 212.00 (211.70 to 212.70): E
Ion 106.00 (105.70 to 106.70): E
Ion 104.00 (103.70 to 104.70): E



#26

Fluoranthene

Concen: 0.748 ng

RT: 18.99 min Scan# 1859




Delta R.T. -0.00 min


Lab File: BN010814.D

Acq: 10 Jun 2020 12:14

Tgt Ion:	202	Resp:	20655
Ion Ratio	Lower	Upper	
202	100		
101	7.5	5.8	8.6
203	17.4	13.7	20.5

Abundance Ion 202.00 (201.70 to 202.70): E
Ion 101.00 (100.70 to 101.70): E
Ion 203.00 (202.70 to 203.70): E





#27

Chrysene-d12

Concen: 0.400 ng

RT: 21.13 min Scan# 2169

Delta R.T. 0.00 min

Lab File: BN010814.D

Acq: 10 Jun 2020 12:14

Instrument :

BNA_N





ClientSampleId :

SSTDICC8.0

Tot Ion	Ratio	Lower	Upper
240	100		
120	7.5	7.5	11.3#
236	27.9	22.3	33.5

Manual Integrations
APPROVED

mohammad
6/11/2020 1:49:11 PM



#28

Pyrene

Concen: 0.786 ng




RT: 19.36 min Scan# 1932


Delta R.T. -0.00 min

Lab File: BN010814.D

Acq: 10 Jun 2020 12:14

Tgt Ion	Ratio	Lower	Upper
202	100		
200	20.4	16.4	24.6
203	17.7	14.2	21.4









#29
 Terphenyl-d14
 Concen: 0.800 ng
 RT: 19.58 min Scan# 1977
 Delta R.T. 0.00 min
 Lab File: BN010814.D
 Acq: 10 Jun 2020 12:14

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC8.0

Tot Ion	Ratio	Lower	Upper
244	100		
212	8.4	7.4	11.2
122	8.4	7.5	11.3




Manual Integrations
 APPROVED


mohammad
 6/11/2020 1:49:11 PM



#30
 Benzo(a)anthracene
 Concen: 0.730 ng
 RT: 21.12 min Scan# 2168
 Delta R.T. 0.00 min
 Lab File: BN010814.D
 Acq: 10 Jun 2020 12:14

Tgt Ion	Ratio	Lower	Upper
228	100		
226	26.0	21.3	31.9
229	20.2	16.6	24.8





#31

Chrysene

Concen: 0.782 ng

RT: 21.16 min Scan# 2172

Delta R.T. 0.00 min

Lab File: BN010814.D

Acq: 10 Jun 2020 12:14

Instrument :

BNA_N





ClientSampleId :

SSTDICC8.0

Tot Ion:	228	Resp:	21151
Ion Ratio	Lower	Upper	
228	100		
226	30.3	25.0	37.6
229	19.1	16.1	24.1

Manual Integrations
APPROVED

mohammad
6/11/2020 1:49:11 PM



#32

Bis(2-ethylhexyl)phthalate

Concen: 0.750 ng




RT: 21.07 min Scan# 2163


Delta R.T. 0.00 min

Lab File: BN010814.D

Acq: 10 Jun 2020 12:14

Tgt Ion:	149	Resp:	5901
Ion Ratio	Lower	Upper	
149	100		
167	28.6	24.1	36.1
279	5.5	4.6	7.0









#33
Indeno(1,2,3-cd)pyrene
Concen: 0.708 ng
RT: 25.39 min Scan# 3162
Delta R.T. 0.00 min
Lab File: BN010814.D
Acq: 10 Jun 2020 12:14

Instrument :
BNA_N
ClientSampleId :
SSTDICC8.0

Tot Ion	Ratio	Lower	Upper
276	100		
138	17.2	13.0	19.6
277	24.6	19.7	29.5




Manual Integrations
APPROVED


mohammad
6/11/2020 1:49:11 PM



#34
Perylene-d12
Concen: 0.400 ng
RT: 23.28 min Scan# 2545
Delta R.T. 0.00 min
Lab File: BN010814.D
Acq: 10 Jun 2020 12:14

Tgt Ion	Ratio	Lower	Upper
264	100		
260	25.5	20.4	30.6
265	83.5	75.0	112.4









#35
Benzo(b)fluoranthene
Concen: 0.787 ng m
RT: 22.65 min Scan# 2359
Delta R.T. 0.00 min
Lab File: BN010814.D
Acq: 10 Jun 2020 12:14

Instrument :
BNA_N
ClientSampleId :
SSTDICC8.0

Tot Ion	Ratio	Lower	Upper
252	100		
253	24.8	22.4	33.6
125	9.3	9.4	14.0#




Manual Integrations
APPROVED


mohammad
6/11/2020 1:49:11 PM



#36
Benzo(k)fluoranthene
Concen: 0.774 ng
RT: 22.69 min Scan# 2371
Delta R.T. 0.00 min
Lab File: BN010814.D
Acq: 10 Jun 2020 12:14

Tgt Ion	Ratio	Lower	Upper
252	100		
253	25.2	22.2	33.4
125	9.7	9.8	14.8#









#37
Benzo(a)pyrene
Concen: 0.740 ng
RT: 23.19 min Scan# 2518
Delta R.T. 0.00 min
Lab File: BN010814.D
Acq: 10 Jun 2020 12:14

Instrument :
BNA_N
ClientSampled :
SSTDICC8.0

Tot Ion	Ratio	Lower	Upper
252	100		
253	26.5	25.8	38.8
125	11.1	12.4	18.6




Manual Integrations
APPROVED


mohammad
6/11/2020 1:49:11 PM



#38
Dibenzo(a,h)anthracene
Concen: 0.760 ng
RT: 25.41 min Scan# 3166
Delta R.T. 0.00 min
Lab File: BN010814.D
Acq: 10 Jun 2020 12:14

Tgt Ion	Ratio	Lower	Upper
278	100		
139	12.2	10.6	15.8
279	26.2	23.6	35.4





#39

Benzo(a,h,i)perylene

Concen: 0.754 ng

RT: 26.03 min Scan# 3348

Delta R.T. -0.00 min

Lab File: BN010814.D

Acq: 10 Jun 2020 12:14

Instrument :

BNA_N


ClientSampleId :

SSTDICC8.0

Tot Ion:	276	Resp:	19586
Ion Ratio		Lower	Upper
276	100		
277	24.7	20.5	30.7
138	16.0	13.7	20.5

Manual Integrations
APPROVED

mohammad
6/11/2020 1:49:11 PM



Abundance Ion 276.00 (275.70 to 276.70): E
Ion 277.00 (276.70 to 277.70): E
Ion 138.00 (137.70 to 138.70): E

