

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061024\
 Data File : BN031885.D
 Acq On : 10 Jun 2024 19:39
 Operator : MA/JU
 Sample : P2771-06
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BP-VPB-196-GW-200-202

Quant Time: Jun 11 00:11:06 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN061024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 10 15:47:32 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.668	152	8750	0.400	ng	0.00
7) Naphthalene-d8	10.453	136	27901	0.400	ng	0.00
13) Acenaphthene-d10	14.306	164	18254	0.400	ng	-0.01
19) Phenanthrene-d10	17.054	188	41851	0.400	ng	#-0.01
29) Chrysene-d12	21.256	240	38012	0.400	ng	# 0.00
35) Perylene-d12	23.493	264	39810	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.270	112	3885	0.156	ng	0.00
5) Phenol-d6	6.852	99	3128	0.096	ng	0.00
8) Nitrobenzene-d5	8.820	82	8249	0.356	ng	0.00
11) 2-Methylnaphthalene-d10	12.047	152	15932	0.358	ng	0.00
14) 2,4,6-Tribromophenol	15.812	330	2777	0.317	ng	0.00
15) 2-Fluorobiphenyl	12.927	172	25862	0.373	ng	0.00
27) Fluoranthene-d10	19.096	212	46710	0.410	ng	0.00
31) Terphenyl-d14	19.700	244	42289	0.455	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.219	88	4228	0.395	ng	# 82
33) Chrysene	21.292	228	2980	0.022	ng	# 86
34) Bis(2-ethylhexyl)phtha...	21.175	149	3902	0.040	ng	# 93
37) Benzo(b)fluoranthene	22.823	252	3973	0.027	ng	# 2

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061024\
Data File : BN031885.D
Acq On : 10 Jun 2024 19:39
Operator : MA/JU
Sample : P2771-06
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BP-VPB-196-GW-200-202

Quant Time: Jun 11 00:11:06 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN061024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 10 15:47:32 2024
Response via : Initial Calibration

