

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061024\
 Data File : BN031879.D
 Acq On : 10 Jun 2024 15:55
 Operator : MA/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN061024

Quant Time: Jun 10 16:22:08 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN061024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 10 15:47:32 2024
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.668	152	11424	0.400	ng	0.00
7) Naphthalene-d8	10.453	136	37516	0.400	ng	0.00
13) Acenaphthene-d10	14.306	164	23549	0.400	ng	-0.01
19) Phenanthrene-d10	17.066	188	54192	0.400	ng	0.00
29) Chrysene-d12	21.265	240	50575	0.400	ng	0.00
35) Perylene-d12	23.505	264	54922	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.263	112	11376	0.350	ng	0.00
5) Phenol-d6	6.844	99	14613	0.343	ng	0.00
8) Nitrobenzene-d5	8.819	82	10629	0.341	ng	0.00
11) 2-Methylnaphthalene-d10	12.047	152	20652	0.346	ng	0.00
14) 2,4,6-Tribromophenol	15.812	330	3441	0.305	ng	0.00
15) 2-Fluorobiphenyl	12.927	172	31001	0.347	ng	0.00
27) Fluoranthene-d10	19.100	212	49742	0.337	ng	0.00
31) Terphenyl-d14	19.704	244	39139	0.317	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.219	88	4822	0.345	ng	96
3) n-Nitrosodimethylamine	3.522	42	4864	0.366	ng	97
6) bis(2-Chloroethyl)ether	7.097	93	12890	0.351	ng	99
9) Naphthalene	10.496	128	35462	0.351	ng	100
10) Hexachlorobutadiene	10.773	225	6219	0.352	ng	# 99
12) 2-Methylnaphthalene	12.119	142	24007	0.348	ng	100
16) Acenaphthylene	14.028	152	35791	0.330	ng	100
17) Acenaphthene	14.370	154	23802	0.341	ng	100
18) Fluorene	15.365	166	31848	0.341	ng	100
20) 4,6-Dinitro-2-methylph...	15.461	198	2373	0.382	ng	94
21) 4-Bromophenyl-phenylether	16.259	248	10317	0.336	ng	96
22) Hexachlorobenzene	16.358	284	10937	0.343	ng	99
23) Atrazine	16.532	200	8851	0.314	ng	99
24) Pentachlorophenol	16.718	266	4427	0.389	ng	99
25) Phenanthrene	17.103	178	51200	0.340	ng	100
26) Anthracene	17.190	178	45624	0.327	ng	100
28) Fluoranthene	19.128	202	61907	0.340	ng	100
30) Pyrene	19.495	202	64114	0.319	ng	100
32) Benzo(a)anthracene	21.247	228	61459	0.323	ng	100
33) Chrysene	21.301	228	60132	0.334	ng	100
34) Bis(2-ethylhexyl)phtha...	21.175	149	41277	0.317	ng	99
36) Indeno(1,2,3-cd)pyrene	25.767	276	72082	0.352	ng	99
37) Benzo(b)fluoranthene	22.829	252	65261	0.326	ng	98
38) Benzo(k)fluoranthene	22.876	252	62062	0.328	ng	98
39) Benzo(a)pyrene	23.405	252	57539	0.333	ng	99
40) Dibenzo(a,h)anthracene	25.782	278	57109	0.354	ng	97
41) Benzo(g,h,i)perylene	26.457	276	61668	0.355	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061024\
Data File : BN031879.D
Acq On : 10 Jun 2024 15:55
Operator : MA/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
ICVBN061024


Quant Time: Jun 10 16:22:08 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN061024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 10 15:47:32 2024
Response via : Initial Calibration








#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.668 min Scan# 64
 Delta R.T. -0.007 min
 Lab File: BN031879.D
 Acq: 10 Jun 2024 15:55


Instrument :
 BNA_N
 ClientSampleId :
 ICVBN061024





Tgt Ion:152 Resp: 11424
 Ion Ratio Lower Upper
 152 100
 150 154.6 121.0 181.4
 115 62.1 49.4 74.0




#2
 1,4-Dioxane
 Concen: 0.345 ng
 RT: 3.219 min Scan# 33
 Delta R.T. 0.000 min
 Lab File: BN031879.D
 Acq: 10 Jun 2024 15:55



Tgt Ion: 88 Resp: 4822
 Ion Ratio Lower Upper
 88 100
 43 26.3 22.7 34.1
 58 67.1 51.3 76.9





#3

n-Nitrosodimethylamine

Concen: 0.366 ng

RT: 3.522 min Scan# 71

Delta R.T. -0.007 min

Lab File: BN031879.D

Acq: 10 Jun 2024 15:55

Instrument :

BNA_N

ClientSampleId :

ICVBN061024





Tgt Ion: 42 Resp: 4864

Ion Ratio Lower Upper

42 100

74 152.6 125.2 187.8

44 8.4 6.9 10.3



#4

2-Fluorophenol

Concen: 0.350 ng

RT: 5.263 min Scan# 316

Delta R.T. -0.007 min

Lab File: BN031879.D

Acq: 10 Jun 2024 15:55




Tgt Ion: 112 Resp: 11376


Ion Ratio Lower Upper

112 100

64 53.5 42.4 63.6

63 27.9 22.1 33.1









#5
Phenol-d6
Concen: 0.343 ng
RT: 6.844 min Scan# 51
Delta R.T. -0.007 min
Lab File: BN031879.D
Acq: 10 Jun 2024 15:55




Instrument :
BNA_N
ClientSampleId :
ICVBN061024


Tgt Ion: 99 Resp: 14613
Ion Ratio Lower Upper
99 100
42 16.0 12.8 19.2
71 32.3 25.4 38.2



#6
bis(2-Chloroethyl)ether
Concen: 0.351 ng
RT: 7.097 min Scan# 570
Delta R.T. -0.007 min
Lab File: BN031879.D
Acq: 10 Jun 2024 15:55

Tgt Ion: 93 Resp: 12890
Ion Ratio Lower Upper
93 100
63 69.7 56.2 84.2
95 32.7 26.4 39.6









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.453 min Scan# 91
Delta R.T. 0.000 min
Lab File: BN031879.D
Acq: 10 Jun 2024 15:55




Instrument :
BNA_N
ClientSampleId :
ICVBN061024


Tgt Ion:	136	Resp:	37516
Ion	Ratio	Lower	Upper
136	100		
137	11.2	8.9	13.3
54	5.9	5.6	8.4
68	5.2	5.0	7.4



#8
Nitrobenzene-d5
Concen: 0.341 ng
RT: 8.819 min Scan# 786
Delta R.T. 0.000 min
Lab File: BN031879.D
Acq: 10 Jun 2024 15:55

Tgt Ion:	82	Resp:	10629
Ion	Ratio	Lower	Upper
82	100		
128	44.2	32.1	48.1
54	44.6	39.7	59.5









#9
Naphthalene
Concen: 0.351 ng
RT: 10.496 min Scan# 944
Delta R.T. -0.011 min
Lab File: BN031879.D
Acq: 10 Jun 2024 15:55




Instrument :
BNA_N
ClientSampleId :
ICVBN061024


Tgt Ion:128 Resp: 35462
Ion Ratio Lower Upper
128 100
129 11.5 9.2 13.8
127 13.3 10.5 15.7



#10
Hexachlorobutadiene
Concen: 0.352 ng
RT: 10.773 min Scan# 969
Delta R.T. -0.011 min
Lab File: BN031879.D
Acq: 10 Jun 2024 15:55

Tgt Ion:225 Resp: 6219
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 62.8 50.6 75.8









#11
2-Methylnaphthalene-d10
Concen: 0.346 ng
RT: 12.047 min Scan# 1180
Delta R.T. -0.004 min
Lab File: BN031879.D
Acq: 10 Jun 2024 15:55




Instrument :
BNA_N
ClientSampleId :
ICVBN061024


Tgt Ion:152 Resp: 20652
Ion Ratio Lower Upper
152 100
151 20.4 16.3 24.5



#12
2-Methylnaphthalene
Concen: 0.348 ng
RT: 12.119 min Scan# 1199
Delta R.T. -0.004 min
Lab File: BN031879.D
Acq: 10 Jun 2024 15:55


Tgt Ion:142 Resp: 24007
Ion Ratio Lower Upper
142 100
141 89.9 72.2 108.2
115 35.4 28.6 43.0








#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.306 min Scan# 1487
Delta R.T. -0.011 min
Lab File: BN031879.D
Acq: 10 Jun 2024 15:55

Instrument :
BNA_N
ClientSampleId :
ICVBN061024







Tgt Ion:164 Resp: 23549
Ion Ratio Lower Upper
164 100
162 104.1 80.7 121.1
160 48.8 36.2 54.4



#14
2,4,6-Tribromophenol
Concen: 0.305 ng
RT: 15.812 min Scan# 1624
Delta R.T. 0.000 min
Lab File: BN031879.D
Acq: 10 Jun 2024 15:55

Tgt Ion:330 Resp: 3441
Ion Ratio Lower Upper
330 100
332 97.6 77.5 116.3
141 42.9 34.7 52.1









#15
2-Fluorobiphenyl
Concen: 0.347 ng
RT: 12.927 min Scan# 1358
Delta R.T. 0.000 min
Lab File: BN031879.D
Acq: 10 Jun 2024 15:55




Instrument :
BNA_N
ClientSampleId :
ICVBN061024


Tgt Ion:172 Resp: 31001
Ion Ratio Lower Upper
172 100
171 35.3 28.6 42.8
170 23.8 19.2 28.8



#16
Acenaphthylene
Concen: 0.330 ng
RT: 14.028 min Scan# 1461
Delta R.T. 0.000 min
Lab File: BN031879.D
Acq: 10 Jun 2024 15:55

Tgt Ion:152 Resp: 35791
Ion Ratio Lower Upper
152 100
151 19.4 15.7 23.5
153 13.0 10.5 15.7





#17

Acenaphthene

Concen: 0.341 ng

RT: 14.370 min Scan# 1493

Delta R.T. 0.000 min

Lab File: BN031879.D

Acq: 10 Jun 2024 15:55

Instrument :

BNA_N

ClientSampleId :

ICVBN061024





Tgt Ion:154 Resp: 23802

Ion Ratio Lower Upper

154 100

153 113.0 90.8 136.2

152 55.1 44.3 66.5



#18

Fluorene

Concen: 0.341 ng

RT: 15.365 min Scan# 1586

Delta R.T. 0.000 min

Lab File: BN031879.D

Acq: 10 Jun 2024 15:55




Tgt Ion:166 Resp: 31848


Ion Ratio Lower Upper

166 100

165 98.3 78.3 117.5

167 13.3 10.7 16.1









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.066 min Scan# 11
Delta R.T. 0.000 min
Lab File: BN031879.D
Acq: 10 Jun 2024 15:55




Instrument :
BNA_N
ClientSampleId :
ICVBN061024


Tgt Ion:	188	Resp:	54192
Ion Ratio	Lower	Upper	
188	100		
94	0.0	0.0	0.0
80	7.4	6.4	9.6



#20
4,6-Dinitro-2-methylphenol
Concen: 0.382 ng
RT: 15.461 min Scan# 1595
Delta R.T. -0.000 min
Lab File: BN031879.D
Acq: 10 Jun 2024 15:55

Tgt Ion:	198	Resp:	2373
Ion Ratio	Lower	Upper	
198	100		
51	50.3	42.7	64.1
105	46.9	41.8	62.8









#21
4-Bromophenyl-phenylether
Concen: 0.336 ng
RT: 16.259 min Scan# 1660
Delta R.T. 0.000 min
Lab File: BN031879.D
Acq: 10 Jun 2024 15:55




Instrument :
BNA_N
ClientSampleId :
ICVBN061024


Tgt Ion:248 Resp: 10317
Ion Ratio Lower Upper
248 100
250 99.5 78.5 117.7
141 47.9 43.1 64.7



#22
Hexachlorobenzene
Concen: 0.343 ng
RT: 16.358 min Scan# 1668
Delta R.T. -0.012 min
Lab File: BN031879.D
Acq: 10 Jun 2024 15:55

Tgt Ion:284 Resp: 10937
Ion Ratio Lower Upper
284 100
142 39.1 30.7 46.1
249 30.7 25.0 37.4









#23
 Atrazine
 Concen: 0.314 ng
 RT: 16.532 min Scan# 1682
 Delta R.T. 0.000 min
 Lab File: BN031879.D
 Acq: 10 Jun 2024 15:55




Instrument :
 BNA_N
 ClientSampleId :
 ICVBN061024


Tgt Ion:	200	Resp:	8851
Ion	Ratio	Lower	Upper
200	100		
173	27.6	22.0	33.0
215	48.6	37.9	56.9



#24
 Pentachlorophenol
 Concen: 0.389 ng
 RT: 16.718 min Scan# 1697
 Delta R.T. 0.000 min
 Lab File: BN031879.D
 Acq: 10 Jun 2024 15:55

Tgt Ion:	266	Resp:	4427
Ion	Ratio	Lower	Upper
266	100		
264	62.1	50.0	75.0
268	64.9	51.0	76.6





#25

Phenanthrene

Concen: 0.340 ng

RT: 17.103 min Scan# 1728

Delta R.T. 0.000 min

Lab File: BN031879.D

Acq: 10 Jun 2024 15:55

Instrument :

BNA_N

ClientSampleId :

ICVBN061024





Tgt Ion:178 Resp: 51200

Ion Ratio Lower Upper

178 100

176 19.1 15.4 23.0

179 15.4 12.2 18.2



#26

Anthracene

Concen: 0.327 ng

RT: 17.190 min Scan# 1735

Delta R.T. 0.000 min

Lab File: BN031879.D

Acq: 10 Jun 2024 15:55




Tgt Ion:178 Resp: 45624


Ion Ratio Lower Upper

178 100

176 18.6 14.8 22.2

179 15.3 12.2 18.2









#27
Fluoranthene-d10
Concen: 0.337 ng
RT: 19.100 min Scan# 1969
Delta R.T. 0.000 min
Lab File: BN031879.D
Acq: 10 Jun 2024 15:55




Instrument :
BNA_N
ClientSampleId :
ICVBN061024


Tgt Ion:212 Resp: 49742
Ion Ratio Lower Upper
212 100
106 15.0 11.8 17.8
104 8.4 6.6 10.0



#28
Fluoranthene
Concen: 0.340 ng
RT: 19.128 min Scan# 1975
Delta R.T. 0.000 min
Lab File: BN031879.D
Acq: 10 Jun 2024 15:55

Tgt Ion:202 Resp: 61907
Ion Ratio Lower Upper
202 100
101 12.2 9.8 14.6
203 17.3 13.7 20.5





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.265 min Scan# 2408

Delta R.T. 0.000 min

Lab File: BN031879.D

Acq: 10 Jun 2024 15:55

Instrument :

BNA_N

ClientSampleId :

ICVBN061024





Tgt Ion:240 Resp: 50575

Ion Ratio Lower Upper

240 100

120 11.3 8.2 12.4

236 27.1 21.8 32.8



#30

Pyrene

Concen: 0.319 ng

RT: 19.495 min Scan# 2054

Delta R.T. 0.000 min

Lab File: BN031879.D

Acq: 10 Jun 2024 15:55




Tgt Ion:202 Resp: 64114


Ion Ratio Lower Upper

202 100

200 20.9 16.7 25.1

203 18.0 14.3 21.5









#31
 Terphenyl-d14
 Concen: 0.317 ng
 RT: 19.704 min Scan# 2099
 Delta R.T. 0.000 min
 Lab File: BN031879.D
 Acq: 10 Jun 2024 15:55




Instrument :
 BNA_N
 ClientSampleId :
 ICVBN061024


Tgt Ion:244 Resp: 39139
 Ion Ratio Lower Upper
 244 100
 212 8.1 6.5 9.7
 122 13.9 11.4 17.0



#32
 Benzo(a)anthracene
 Concen: 0.323 ng
 RT: 21.247 min Scan# 2406
 Delta R.T. 0.000 min
 Lab File: BN031879.D
 Acq: 10 Jun 2024 15:55

Tgt Ion:228 Resp: 61459
 Ion Ratio Lower Upper
 228 100
 226 27.1 21.8 32.6
 229 19.7 15.8 23.8





#33

Chrysene

Concen: 0.334 ng

RT: 21.301 min Scan# 2412

Delta R.T. 0.000 min

Lab File: BN031879.D

Acq: 10 Jun 2024 15:55

Instrument :

BNA_N

ClientSampleId :

ICVBN061024





Tgt Ion:228 Resp: 60132

Ion Ratio Lower Upper

228 100

226 29.3 23.6 35.4

229 19.5 15.8 23.6



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.317 ng

RT: 21.175 min Scan# 2398

Delta R.T. 0.000 min

Lab File: BN031879.D

Acq: 10 Jun 2024 15:55




Tgt Ion:149 Resp: 41277


Ion Ratio Lower Upper

149 100

167 27.5 22.5 33.7

279 3.6 2.9 4.3









#35
Perylene-d12
Concen: 0.400 ng
RT: 23.505 min Scan# 2910
Delta R.T. 0.003 min
Lab File: BN031879.D
Acq: 10 Jun 2024 15:55




Instrument :
BNA_N
ClientSampleId :
ICVBN061024


Tgt Ion:	264	Resp:	54922
Ion	Ratio	Lower	Upper
264	100		
260	25.3	20.5	30.7
265	45.2	39.6	59.4



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.352 ng
RT: 25.767 min Scan# 3684
Delta R.T. 0.000 min
Lab File: BN031879.D
Acq: 10 Jun 2024 15:55


Tgt Ion:	276	Resp:	72082
Ion	Ratio	Lower	Upper
276	100		
138	27.4	21.5	32.3
277	24.8	19.8	29.8








#37
Benzo(b)fluoranthene
Concen: 0.326 ng
RT: 22.829 min Scan# 2679
Delta R.T. 0.000 min
Lab File: BN031879.D
Acq: 10 Jun 2024 15:55

Instrument :
BNA_N
ClientSampleId :
ICVBN061024







Tgt Ion:252 Resp: 65261
Ion Ratio Lower Upper
252 100
253 24.3 20.2 30.2
125 16.9 14.2 21.4



#38
Benzo(k)fluoranthene
Concen: 0.328 ng
RT: 22.876 min Scan# 2695
Delta R.T. 0.003 min
Lab File: BN031879.D
Acq: 10 Jun 2024 15:55


Tgt Ion:252 Resp: 62062
Ion Ratio Lower Upper
252 100
253 24.1 19.9 29.9
125 15.1 13.0 19.4








#39
Benzo(a)pyrene
Concen: 0.333 ng
RT: 23.405 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN031879.D
Acq: 10 Jun 2024 15:55

Instrument :
BNA_N
ClientSampleId :
ICVBN061024







Tgt Ion:252 Resp: 57539
Ion Ratio Lower Upper
252 100
253 24.9 20.5 30.7
125 19.8 16.4 24.6



#40
Dibenzo(a,h)anthracene
Concen: 0.354 ng
RT: 25.782 min Scan# 3689
Delta R.T. 0.003 min
Lab File: BN031879.D
Acq: 10 Jun 2024 15:55

Tgt Ion:278 Resp: 57109
Ion Ratio Lower Upper
278 100
139 19.9 17.0 25.4
279 25.4 21.4 32.0





#41
Benzo(g,h,i)perylene
Concen: 0.355 ng
RT: 26.457 min Scan# 3920
Delta R.T. -0.006 min
Lab File: BN031879.D
Acq: 10 Jun 2024 15:55

Instrument :
BNA_N
ClientSampleId :
ICVBN061024

Tgt Ion:276 Resp: 61668
Ion Ratio Lower Upper
276 100
277 24.4 20.2 30.2
138 25.3 20.6 31.0

