

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061120\
 Data File : BN010834.D
 Acq On : 11 Jun 2020 11:33
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 11 12:53:23 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 10 15:14:26 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	157#	0.00
2	1,4-Dioxane	0.392	0.405	-3.3	155#	0.00
3	n-Nitrosodimethylamine	0.228	0.210	7.9	165#	0.00
4 S	2-Fluorophenol	0.772	0.973	-26.0#	194#	0.00
5 S	Phenol-d6	0.910	1.153	-26.7#	199#	0.00
6	bis(2-Chloroethyl)ether	0.887	1.086	-22.4	192#	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	172#	0.00
8 S	Nitrobenzene-d5	0.296	0.302	-2.0	169#	0.00
9	Naphthalene	1.005	1.070	-6.5	174#	0.00
10	Hexachlorobutadiene	0.248	0.243	2.0	156#	-0.01
11 SURR	2-Methylnaphthalene-d10	0.620	0.659	-6.3	174#	0.00
12	2-Methylnaphthalene	0.673	0.722	-7.3	179#	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	176#	0.00
14 S	2,4,6-Tribromophenol	0.143	0.143	0.0	182#	0.00
15 S	2-Fluorobiphenyl	1.578	1.618	-2.5	170#	0.00
16	Acenaphthylene	1.639	1.720	-4.9	184#	0.00
17	Acenaphthene	1.117	1.156	-3.5	175#	0.00
18	Fluorene	1.441	1.466	-1.7	175#	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	175#	0.00
20	4-Bromophenyl-phenylether	0.240	0.231	3.7	164#	0.00
21	Hexachlorobenzene	0.251	0.255	-1.6	169#	0.00
22	Pentachlorophenol	0.084	0.081	3.6	173#	0.00
23	Phenanthrene	1.093	1.128	-3.2	176#	0.00
24	Anthracene	0.912	0.958	-5.0	185#	0.00
25 SURR	Fluoranthene-d10	1.104	1.163	-5.3	182#	0.00
26	Fluoranthene	1.240	1.279	-3.1	180#	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	183#	0.00
28	Pvrene	1.323	1.386	-4.8	185#	0.00
29 S	Terphenyl-d14	0.936	0.960	-2.6	179#	0.00
30	Benzo(a)anthracene	1.176	1.294	-10.0	204#	0.00
31	Chrysene	1.370	1.391	-1.5	172#	0.00
32	Bis(2-ethylhexyl)phthalate	0.392	0.552	-40.8#	264#	0.00
33	Indeno(1,2,3-cd)pyrene	1.531	1.633	-6.7	187#	0.00
34 I	Perylene-d12	1.000	1.000	0.0	203#	0.00
35	Benzo(b)fluoranthene	1.360	1.326	2.5	192#	0.00
36	Benzo(k)fluoranthene	1.449	1.338	7.7	178#	0.00
37 C	Benzo(a)pyrene	1.167	1.190	-2.0	207#	0.00
38	Dibenzo(a,h)anthracene	1.330	1.238	6.9	186#	0.00
39	Benzo(a,h,i)perylene	1.391	1.283	7.8	176#	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061120\
Data File : BN010834.D
Acq On : 11 Jun 2020 11:33
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 11 12:53:23 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 10 15:14:26 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			