

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061120\
 Data File : BN010834.D
 Acq On : 11 Jun 2020 11:33
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 11 12:53:23 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 10 15:14:26 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	157	0.00
2	1,4-Dioxane	0.400	0.413	-3.2	155	0.00
3	n-Nitrosodimethylamine	0.400	0.369	7.8	165	0.00
4 S	2-Fluorophenol	0.400	0.504	-26.0#	194	0.00
5 S	Phenol-d6	0.400	0.507	-26.7#	199	0.00
6	bis(2-Chloroethyl)ether	0.400	0.490	-22.5	192	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	172	0.00
8 S	Nitrobenzene-d5	0.400	0.408	-2.0	169	0.00
9	Naphthalene	0.400	0.426	-6.5	174	0.00
10	Hexachlorobutadiene	0.400	0.391	2.3	156	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.425	-6.2	174	0.00
12	2-Methylnaphthalene	0.400	0.429	-7.2	179	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	176	0.00
14 S	2,4,6-Tribromophenol	0.400	0.399	0.3	182	0.00
15 S	2-Fluorobiphenyl	0.400	0.410	-2.5	170	0.00
16	Acenaphthylene	0.400	0.420	-5.0	184	0.00
17	Acenaphthene	0.400	0.414	-3.5	175	0.00
18	Fluorene	0.400	0.407	-1.7	175	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	175	0.00
20	4-Bromophenyl-phenylether	0.400	0.384	4.0	164	0.00
21	Hexachlorobenzene	0.400	0.407	-1.7	169	0.00
22	Pentachlorophenol	0.400	0.385	3.8	173	0.00
23	Phenanthrene	0.400	0.413	-3.2	176	0.00
24	Anthracene	0.400	0.420	-5.0	185	0.00
25 SURR	Fluoranthene-d10	0.400	0.421	-5.2	182	0.00
26	Fluoranthene	0.400	0.413	-3.2	180	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	183	0.00
28	Pvrene	0.400	0.419	-4.7	185	0.00
29 S	Terphenyl-d14	0.400	0.410	-2.5	179	0.00
30	Benzo(a)anthracene	0.400	0.440	-10.0	204	0.00
31	Chrysene	0.400	0.406	-1.5	172	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.563	-40.7#	264	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.426	-6.5	187	0.00
34 I	Perylene-d12	0.400	0.400	0.0	203	0.00
35	Benzo(b)fluoranthene	0.400	0.390	2.5	192	0.00
36	Benzo(k)fluoranthene	0.400	0.369	7.8	178	0.00
37 C	Benzo(a)pyrene	0.400	0.408	-2.0	207	0.00
38	Dibenzo(a,h)anthracene	0.400	0.372	7.0	186	0.00
39	Benzo(a,h,i)perylene	0.400	0.369	7.8	176	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061120\
Data File : BN010834.D
Acq On : 11 Jun 2020 11:33
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 11 12:53:23 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 10 15:14:26 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				