

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061122\
 Data File : BN020126.D
 Acq On : 11 Jun 2022 07:09
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020317

Quant Time: Jun 13 00:43:18 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN060622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jun 06 15:14:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.816	152	163786	20.000	ng/u1	0.00
20) Naphthalene-d8	10.605	136	769686	20.000	ng/u1	-0.01
38) Acenaphthene-d10	14.446	164	501557	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.187	188	1032328	20.000	ng/u1	0.00
79) Chrysene-d12	21.375	240	925066	20.000	ng/u1	0.00
88) Perylene-d12	23.698	264	853435	20.000	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.281	96	28306	7.485	ng/uL	0.00
4) Pyridine-d5	3.687	84	201578	19.503	ng/u1	0.00
7) Phenol-d5	6.987	99	269139	19.859	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.146	67	169410	20.007	ng/u1	0.00
11) 2-Chlorophenol-d4	7.352	132	224510	20.388	ng/u1	0.00
15) 4-Methylphenol-d8	8.522	113	238520	20.342	ng/u1	0.00
21) Nitrobenzene-d5	8.969	128	116258	19.854	ng/u1	0.00
24) 2-Nitrophenol-d4	9.687	143	130771	21.043	ng/u1	-0.01
28) 2,4-Dichlorophenol-d3	10.228	165	224504	20.694	ng/u1	0.00
31) 4-Chloroaniline-d4	10.740	131	351258	19.994	ng/u1	0.00
46) Dimethylphthalate-d6	13.857	166	709787	19.449	ng/u1	0.00
49) Acenaphthylene-d8	14.134	160	847617	20.028	ng/u1	-0.01
54) 4-Nitrophenol-d4	14.640	143	137564	18.811	ng/u1	0.00
60) Fluorene-d10	15.434	176	595744	19.727	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.557	200	109420	18.298	ng/u1	0.00
73) Anthracene-d10	17.281	188	926032	20.554	ng/u1	-0.01
81) Pyrene-d10	19.575	212	1030237	20.543	ng/u1	-0.01
92) Benzo(a)pyrene-d12	23.551	264	840291	20.124	ng/u1	-0.01
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.317	88	29130	7.297	ng/uL	92
5) Pyridine	3.705	79	211218	19.493	ng/u1	94
6) Benzaldehyde	6.952	77	160419	25.121	ng/u1	94
8) Phenol	7.011	94	290316	20.117	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.240	93	224532	19.878	ng/u1	99
12) 2-Chlorophenol	7.381	128	232889	20.233	ng/u1	99
13) 2-Methylphenol	8.258	108	222519	20.056	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.346	45	354677	20.024	ng/u1	96
16) Acetophenone	8.634	105	370071	20.754	ng/u1	97
17) N-Nitroso-di-n-propyla...	8.622	70	185118	20.322	ng/u1	93
18) 4-Methylphenol	8.587	108	248076	20.259	ng/u1	93
19) Hexachloroethane	8.893	117	90628	19.568	ng/u1	89
22) Nitrobenzene	9.011	77	270907	19.953	ng/u1	96
23) Isophorone	9.534	82	529600	19.756	ng/u1	98
25) 2-Nitrophenol	9.722	139	141496	20.605	ng/u1	98
26) 2,4-Dimethylphenol	9.787	107	285059	20.037	ng/u1	96
27) Bis(2-Chloroethoxy)met...	10.016	93	323446	19.717	ng/u1	99
29) 2,4-Dichlorophenol	10.258	162	232358	20.706	ng/u1	97
30) Naphthalene	10.658	128	787633	19.745	ng/u1	100
32) 4-Chloroaniline	10.763	127	350745	19.960	ng/u1	99
33) Hexachlorobutadiene	10.952	225	128415	19.931	ng/u1	98
34) Caprolactam	11.510	113	82916	18.921	ng/u1	93
35) 4-Chloro-3-methylphenol	11.893	107	273758	20.660	ng/u1	100
36) 2-Methylnaphthalene	12.269	142	547332	20.295	ng/u1	99


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061122\
 Data File : BN020126.D
 Acq On : 11 Jun 2022 07:09
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020317

Quant Time: Jun 13 00:43:18 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN060622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jun 06 15:14:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.487	142	556035	20.204	ng/ul	97
39) 1,2,4,5-Tetrachloroben...	12.640	216	255040	19.938	ng/ul	99
40) Hexachlorocyclopentadiene	12.622	237	139191	17.888	ng/ul	99
41) 2,4,6-Trichlorophenol	12.881	196	178174	20.200	ng/ul	96
42) 2,4,5-Trichlorophenol	12.951	196	197748	20.243	ng/ul	98
43) 1,1'-Biphenyl	13.281	154	730816	19.678	ng/ul	98
44) 2-Chloronaphthalene	13.316	162	556146	19.739	ng/ul	97
45) 2-Nitroaniline	13.522	65	182080	20.773	ng/ul	94
47) Dimethylphthalate	13.904	163	715597	19.478	ng/ul	100
48) 2,6-Dinitrotoluene	14.022	165	147425	20.524	ng/ul	94
50) Acenaphthylene	14.163	152	930779	19.864	ng/ul	99
51) 3-Nitroaniline	14.346	138	162150	21.353	ng/ul	96
52) Acenaphthene	14.510	153	602014	19.669	ng/ul	99
53) 2,4-Dinitrophenol	14.551	184	62279	13.542	ng/ul	91
55) 4-Nitrophenol	14.651	109	117886	18.885	ng/ul	97
56) Dibenzofuran	14.840	168	848948	19.688	ng/ul	98
57) 2,4-Dinitrotoluene	14.804	165	216758	20.380	ng/ul	97
58) 2,3,4,6-Tetrachlorophenol	15.069	232	153136	19.970	ng/ul	91
59) Diethylphthalate	15.269	149	742976	19.643	ng/ul	97
61) Fluorene	15.493	166	669280	19.947	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.487	204	307940	19.929	ng/ul	94
63) 4-Nitroaniline	15.504	138	155537	21.919	ng/ul	96
66) 4,6-Dinitro-2-methylph...	15.569	198	110654	18.251	ng/ul#	93
67) N-Nitrosodiphenylamine	15.698	169	590716	19.960	ng/ul	98
68) 4-Bromophenyl-phenylether	16.381	248	180730	20.103	ng/ul#	91
69) Hexachlorobenzene	16.498	284	202504	19.596	ng/ul#	89
70) Atrazine	16.651	200	211173	20.450	ng/ul	93
71) Pentachlorophenol	16.840	266	113846	17.688	ng/ul	98
72) Phenanthrene	17.228	178	1085511	19.925	ng/ul	99
74) Anthracene	17.316	178	1095736	20.002	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.246	216	271244	19.621	ng/uL	100
76) Pentachlorobenzene	14.769	250	247317	20.242	ng/uL	97
77) Carbazole	17.587	167	1014775	20.362	ng/ul	97
78) Di-n-butylphthalate	18.157	149	1275532	20.641	ng/ul	100
80) Fluoranthene	19.245	202	1235656	20.401	ng/ul	98
82) Pyrene	19.604	202	1262409	20.316	ng/ul#	93
83) Butylbenzylphthalate	20.510	149	582989	20.847	ng/ul	96
84) 3,3'-Dichlorobenzidine	21.292	252	360146	20.384	ng/ul#	97
85) Benzo(a)anthracene	21.357	228	1168438	20.002	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.298	149	842874	21.362	ng/ul	98
87) Chrysene	21.410	228	1121670	19.627	ng/ul	99
89) Di-n-octyl phthalate	22.198	149	1468683	21.144	ng/ul	100
90) Benzo(b)fluoranthene	22.992	252	1099986	20.536	ng/ul#	95
91) Benzo(k)fluoranthene	23.039	252	1057692	20.744	ng/ul	96
93) Benzo(a)pyrene	23.598	252	1035756	19.977	ng/ul#	98
94) Indeno(1,2,3-cd)pyrene	26.092	276	1006864	18.875	ng/ul	92
95) Dibenzo(a,h)anthracene	26.104	278	856022	18.788	ng/ul	95
96) Benzo(g,h,i)perylene	26.821	276	829253	18.408	ng/ul	95





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 7.297 ng/uL
RT: 3.317 min Scan# 39
Delta R.T. 0.001 min
Lab File: BN020126.D
Acq: 11 Jun 2022 07:09




Instrument :
BNA_N
ClientSampleId :
SSTD020317


Tgt Ion: 88	Resp: 29130
Ion Ratio	Lower Upper
88 100	
43 32.3	22.4 33.6
58 83.6	61.0 91.4



#5
Pyridine
Concen: 19.493 ng/uL
RT: 3.705 min Scan# 105
Delta R.T. -0.006 min
Lab File: BN020126.D
Acq: 11 Jun 2022 07:09

Tgt Ion: 79	Resp: 211218
Ion Ratio	Lower Upper
79 100	
52 88.7	65.0 97.4
51 40.8	31.6 47.4









#6
Benzaldehyde
Concen: 25.121 ng/ul
RT: 6.952 min Scan# 61
Delta R.T. -0.011 min
Lab File: BN020126.D
Acq: 11 Jun 2022 07:09




Instrument :
BNA_N
ClientSampleId :
SSTD020317


Tgt Ion: 77 Resp: 160419
Ion Ratio Lower Upper
77 100
105 98.3 74.7 112.1
106 94.2 70.2 105.4



#8
Phenol
Concen: 20.117 ng/ul
RT: 7.011 min Scan# 667
Delta R.T. -0.006 min
Lab File: BN020126.D
Acq: 11 Jun 2022 07:09

Tgt Ion: 94 Resp: 290316
Ion Ratio Lower Upper
94 100
65 29.2 23.6 35.4
66 39.6 31.2 46.8









#10
 Bis(2-Chloroethyl)ether
 Concen: 19.878 ng/ul
 RT: 7.240 min Scan# 706
 Delta R.T. -0.006 min
 Lab File: BN020126.D
 Acq: 11 Jun 2022 07:09

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020317

Tgt Ion: 93 Resp: 224532




Ion	Ratio	Lower	Upper
93	100		
63	80.8	63.4	95.0
95	32.7	26.2	39.2




#12
 2-Chlorophenol
 Concen: 20.233 ng/ul
 RT: 7.381 min Scan# 730
 Delta R.T. -0.006 min
 Lab File: BN020126.D
 Acq: 11 Jun 2022 07:09

Tgt Ion: 128 Resp: 232889

Ion	Ratio	Lower	Upper
128	100		
64	47.8	38.6	57.8
130	32.0	25.0	37.4









#13
2-Methylphenol
Concen: 20.056 ng/ul
RT: 8.258 min Scan# 81
Delta R.T. -0.006 min
Lab File: BN020126.D
Acq: 11 Jun 2022 07:09




Instrument :
BNA_N
ClientSampleId :
SSTD020317


Tgt Ion:108 Resp: 222519
Ion Ratio Lower Upper
108 100
107 90.6 71.5 107.3



#14
2,2'-oxybis(1-Chloropropane)
Concen: 20.024 ng/ul
RT: 8.346 min Scan# 894
Delta R.T. -0.011 min
Lab File: BN020126.D
Acq: 11 Jun 2022 07:09

Tgt Ion: 45 Resp: 354677
Ion Ratio Lower Upper
45 100
77 11.7 10.9 16.3
79 9.5 8.6 13.0









#16
Acetophenone
Concen: 20.754 ng/ul
RT: 8.634 min Scan# 941
Delta R.T. -0.006 min
Lab File: BN020126.D
Acq: 11 Jun 2022 07:09




Instrument :
BNA_N
ClientSampleId :
SSTD020317


Tgt Ion:105 Resp: 370071
Ion Ratio Lower Upper
105 100
77 76.7 63.5 95.3
51 30.4 23.5 35.3



#17
N-Nitroso-di-n-propylamine
Concen: 20.322 ng/ul
RT: 8.622 min Scan# 941
Delta R.T. -0.006 min
Lab File: BN020126.D
Acq: 11 Jun 2022 07:09

Tgt Ion: 70 Resp: 185118
Ion Ratio Lower Upper
70 100
42 66.4 47.8 71.6
101 10.2 7.8 11.6
130 22.4 15.8 23.8









#18
4-Methylphenol
Concen: 20.259 ng/ul
RT: 8.587 min Scan# 911
Delta R.T. -0.006 min
Lab File: BN020126.D
Acq: 11 Jun 2022 07:09




Instrument :
BNA_N
ClientSampleId :
SSTD020317


Tgt Ion:108 Resp: 248076
Ion Ratio Lower Upper
108 100
107 117.3 88.0 132.0



#19
Hexachloroethane
Concen: 19.568 ng/ul
RT: 8.893 min Scan# 987
Delta R.T. -0.011 min
Lab File: BN020126.D
Acq: 11 Jun 2022 07:09

Tgt Ion:117 Resp: 90628
Ion Ratio Lower Upper
117 100
201 91.5 83.2 124.8
199 56.5 50.2 75.4









#22
Nitrobenzene
Concen: 19.953 ng/ul
RT: 9.011 min Scan# 1007
Delta R.T. -0.006 min
Lab File: BN020126.D
Acq: 11 Jun 2022 07:09




Instrument :
BNA_N
ClientSampleId :
SSTD020317


Tgt Ion: 77 Resp: 270907
Ion Ratio Lower Upper
77 100
123 46.2 34.5 51.7
65 14.9 11.1 16.7



#23
Isophorone
Concen: 19.756 ng/ul
RT: 9.534 min Scan# 1096
Delta R.T. -0.006 min
Lab File: BN020126.D
Acq: 11 Jun 2022 07:09

Tgt Ion: 82 Resp: 529600
Ion Ratio Lower Upper
82 100
95 7.8 5.8 8.8
138 18.2 13.8 20.8









#25
2-Nitrophenol
Concen: 20.605 ng/ul
RT: 9.722 min Scan# 1128
Delta R.T. -0.006 min
Lab File: BN020126.D
Acq: 11 Jun 2022 07:09




Instrument :
BNA_N
ClientSampleId :
SSTD020317


Tgt Ion:139 Resp: 141496
Ion Ratio Lower Upper
139 100
65 52.9 43.7 65.5
109 33.3 25.8 38.6



#26
2,4-Dimethylphenol
Concen: 20.037 ng/ul
RT: 9.787 min Scan# 1139
Delta R.T. -0.006 min
Lab File: BN020126.D
Acq: 11 Jun 2022 07:09

Tgt Ion:107 Resp: 285059
Ion Ratio Lower Upper
107 100
121 52.8 38.3 57.5
122 85.2 66.3 99.5









#27
 Bis(2-Chloroethoxy)methane
 Concen: 19.717 ng/ul
 RT: 10.016 min Scan# 1178
 Delta R.T. -0.012 min
 Lab File: BN020126.D
 Acq: 11 Jun 2022 07:09




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020317


Tgt Ion: 93 Resp: 323446
 Ion Ratio Lower Upper
 93 100
 95 32.1 25.4 38.0
 123 11.5 8.3 12.5



#29
 2,4-Dichlorophenol
 Concen: 20.706 ng/ul
 RT: 10.258 min Scan# 1219
 Delta R.T. -0.006 min
 Lab File: BN020126.D
 Acq: 11 Jun 2022 07:09

Tgt Ion: 162 Resp: 232358
 Ion Ratio Lower Upper
 162 100
 164 63.1 52.9 79.3
 98 37.5 29.8 44.8









#30
Naphthalene
Concen: 19.745 ng/ul
RT: 10.658 min Scan# 1287
Delta R.T. -0.006 min
Lab File: BN020126.D
Acq: 11 Jun 2022 07:09




Instrument :
BNA_N
ClientSampleId :
SSTD020317


Tgt Ion:128 Resp: 787633
Ion Ratio Lower Upper
128 100
129 11.2 8.7 13.1
127 13.3 10.6 16.0



#32
4-Chloroaniline
Concen: 19.960 ng/ul
RT: 10.763 min Scan# 1305
Delta R.T. -0.006 min
Lab File: BN020126.D
Acq: 11 Jun 2022 07:09

Tgt Ion:127 Resp: 350745
Ion Ratio Lower Upper
127 100
129 32.2 26.1 39.1









#33
Hexachlorobutadiene
Concen: 19.931 ng/ul
RT: 10.952 min Scan# 11
Delta R.T. -0.006 min
Lab File: BN020126.D
Acq: 11 Jun 2022 07:09




Instrument :
BNA_N
ClientSampleId :
SSTD020317


Tgt Ion:225 Resp: 128415
Ion Ratio Lower Upper
225 100
223 64.5 51.8 77.6
227 59.6 50.6 76.0



#34
Caprolactam
Concen: 18.921 ng/ul
RT: 11.510 min Scan# 1432
Delta R.T. -0.006 min
Lab File: BN020126.D
Acq: 11 Jun 2022 07:09

Tgt Ion:113 Resp: 82916
Ion Ratio Lower Upper
113 100
55 189.6 145.5 218.3
56 144.5 106.5 159.7









#35
4-Chloro-3-methylphenol
Concen: 20.660 ng/ul
RT: 11.893 min Scan# 1497
Delta R.T. -0.006 min
Lab File: BN020126.D
Acq: 11 Jun 2022 07:09




Instrument :
BNA_N
ClientSampleId :
SSTD020317


Tgt Ion:107 Resp: 273758
Ion Ratio Lower Upper
107 100
144 25.0 20.6 31.0
142 77.8 62.2 93.4



#36
2-Methylnaphthalene
Concen: 20.295 ng/ul
RT: 12.269 min Scan# 1561
Delta R.T. -0.006 min
Lab File: BN020126.D
Acq: 11 Jun 2022 07:09

Tgt Ion:142 Resp: 547332
Ion Ratio Lower Upper
142 100
141 87.3 70.6 105.8









#37
1-Methylnaphthalene
Concen: 20.204 ng/ul
RT: 12.487 min Scan# 1598
Delta R.T. -0.006 min
Lab File: BN020126.D
Acq: 11 Jun 2022 07:09




Instrument :
BNA_N
ClientSampleId :
SSTD020317


Tgt Ion:142 Resp: 556035
Ion Ratio Lower Upper
142 100
141 91.2 70.9 106.3
116 3.8 3.1 4.7



#39
1,2,4,5-Tetrachlorobenzene
Concen: 19.938 ng/ul
RT: 12.640 min Scan# 1624
Delta R.T. -0.011 min
Lab File: BN020126.D
Acq: 11 Jun 2022 07:09

Tgt Ion:216 Resp: 255040
Ion Ratio Lower Upper
216 100
214 79.0 63.4 95.0
179 23.3 17.8 26.8
108 20.8 15.1 22.7









#40
Hexachlorocyclopentadiene
Concen: 17.888 ng/ul
RT: 12.622 min Scan# 1622
Delta R.T. -0.011 min
Lab File: BN020126.D
Acq: 11 Jun 2022 07:09




Instrument :
BNA_N
ClientSampleId :
SSTD020317


Tgt Ion:237 Resp: 139191
Ion Ratio Lower Upper
237 100
235 61.3 48.8 73.2
272 12.4 9.3 13.9



#41
2,4,6-Trichlorophenol
Concen: 20.200 ng/ul
RT: 12.881 min Scan# 1665
Delta R.T. -0.006 min
Lab File: BN020126.D
Acq: 11 Jun 2022 07:09

Tgt Ion:196 Resp: 178174
Ion Ratio Lower Upper
196 100
198 94.6 79.8 119.8
200 31.3 25.0 37.6









#42
2,4,5-Trichlorophenol
Concen: 20.243 ng/ul
RT: 12.951 min Scan# 1677
Delta R.T. -0.006 min
Lab File: BN020126.D
Acq: 11 Jun 2022 07:09




Instrument :
BNA_N
ClientSampleId :
SSTD020317


Tgt Ion:196 Resp: 197748
Ion Ratio Lower Upper
196 100
198 94.0 77.6 116.4
200 30.1 24.4 36.6



#43
1,1'-Biphenyl
Concen: 19.678 ng/ul
RT: 13.281 min Scan# 1733
Delta R.T. -0.006 min
Lab File: BN020126.D
Acq: 11 Jun 2022 07:09

Tgt Ion:154 Resp: 730816
Ion Ratio Lower Upper
154 100
153 41.1 31.5 47.3
76 12.8 10.5 15.7









#44
2-Chloronaphthalene
Concen: 19.739 ng/ul
RT: 13.316 min Scan# 1739
Delta R.T. -0.011 min
Lab File: BN020126.D
Acq: 11 Jun 2022 07:09




Instrument :
BNA_N
ClientSampleId :
SSTD020317


Tgt Ion:162 Resp: 556146
Ion Ratio Lower Upper
162 100
164 32.7 25.8 38.6
127 42.4 31.4 47.2



#45
2-Nitroaniline
Concen: 20.773 ng/ul
RT: 13.522 min Scan# 1774
Delta R.T. -0.006 min
Lab File: BN020126.D
Acq: 11 Jun 2022 07:09

Tgt Ion: 65 Resp: 182080
Ion Ratio Lower Upper
65 100
92 65.9 54.1 81.1
138 103.9 77.0 115.4









#47
Dimethylphthalate
Concen: 19.478 ng/ul
RT: 13.904 min Scan# 1839
Delta R.T. -0.006 min
Lab File: BN020126.D
Acq: 11 Jun 2022 07:09




Instrument :
BNA_N
ClientSampleId :
SSTD020317


Tgt Ion:163 Resp: 715597
Ion Ratio Lower Upper
163 100
194 4.3 3.4 5.0
164 9.8 7.7 11.5



#48
2,6-Dinitrotoluene
Concen: 20.524 ng/ul
RT: 14.022 min Scan# 1859
Delta R.T. -0.006 min
Lab File: BN020126.D
Acq: 11 Jun 2022 07:09

Tgt Ion:165 Resp: 147425
Ion Ratio Lower Upper
165 100
89 57.3 50.6 75.8
121 21.5 17.9 26.9









#50
Acenaphthylene
Concen: 19.864 ng/ul
RT: 14.163 min Scan# 1883
Delta R.T. -0.006 min
Lab File: BN020126.D
Acq: 11 Jun 2022 07:09




Instrument :
BNA_N
ClientSampleId :
SSTD020317


Tgt Ion:152 Resp: 930779
Ion Ratio Lower Upper
152 100
151 20.3 16.0 24.0
153 13.1 10.7 16.1



#51
3-Nitroaniline
Concen: 21.353 ng/ul
RT: 14.346 min Scan# 1914
Delta R.T. -0.006 min
Lab File: BN020126.D
Acq: 11 Jun 2022 07:09

Tgt Ion:138 Resp: 162150
Ion Ratio Lower Upper
138 100
108 11.4 8.6 12.8
92 111.9 92.9 139.3









#52
Acenaphthene
Concen: 19.669 ng/ul
RT: 14.510 min Scan# 1942
Delta R.T. -0.006 min
Lab File: BN020126.D
Acq: 11 Jun 2022 07:09




Instrument :
BNA_N
ClientSampleId :
SSTD020317


Tgt Ion:153 Resp: 602014
Ion Ratio Lower Upper
153 100
152 46.7 38.4 57.6
154 86.5 69.7 104.5



#53
2,4-Dinitrophenol
Concen: 13.542 ng/ul
RT: 14.551 min Scan# 1949
Delta R.T. -0.006 min
Lab File: BN020126.D
Acq: 11 Jun 2022 07:09

Tgt Ion:184 Resp: 62279
Ion Ratio Lower Upper
184 100
63 78.5 55.9 83.9
154 72.6 53.1 79.7









#55
4-Nitrophenol
Concen: 18.885 ng/ul
RT: 14.651 min Scan# 1966
Delta R.T. -0.006 min
Lab File: BN020126.D
Acq: 11 Jun 2022 07:09




Instrument :
BNA_N
ClientSampleId :
SSTD020317


Tgt Ion:109 Resp: 117886
Ion Ratio Lower Upper
109 100
139 120.1 100.1 150.1
65 127.9 104.6 157.0



#56
Dibenzofuran
Concen: 19.688 ng/ul
RT: 14.840 min Scan# 1998
Delta R.T. -0.011 min
Lab File: BN020126.D
Acq: 11 Jun 2022 07:09

Tgt Ion:168 Resp: 848948
Ion Ratio Lower Upper
168 100
139 40.4 31.1 46.7









#57
2,4-Dinitrotoluene
Concen: 20.380 ng/ul
RT: 14.804 min Scan# 1993
Delta R.T. -0.006 min
Lab File: BN020126.D
Acq: 11 Jun 2022 07:09




Instrument :
BNA_N
ClientSampleId :
SSTD020317


Tgt Ion:165 Resp: 216758
Ion Ratio Lower Upper
165 100
63 47.7 36.3 54.5
182 3.1 2.3 3.5



#58
2,3,4,6-Tetrachlorophenol
Concen: 19.970 ng/ul
RT: 15.069 min Scan# 2037
Delta R.T. -0.006 min
Lab File: BN020126.D
Acq: 11 Jun 2022 07:09

Tgt Ion:232 Resp: 153136
Ion Ratio Lower Upper
232 100
131 55.6 38.5 57.7
130 2.8 2.2 3.2
166 35.7 25.8 38.8









#59
Diethylphthalate
Concen: 19.643 ng/ul
RT: 15.269 min Scan# 2071
Delta R.T. -0.006 min
Lab File: BN020126.D
Acq: 11 Jun 2022 07:09




Instrument :
BNA_N
ClientSampleId :
SSTD020317


Tgt Ion:149 Resp: 742976
Ion Ratio Lower Upper
149 100
177 20.6 17.8 26.8
150 11.8 10.5 15.7



#61
Fluorene
Concen: 19.947 ng/ul
RT: 15.493 min Scan# 2109
Delta R.T. -0.006 min
Lab File: BN020126.D
Acq: 11 Jun 2022 07:09

Tgt Ion:166 Resp: 669280
Ion Ratio Lower Upper
166 100
165 99.1 77.7 116.5
167 13.1 11.1 16.7









#62
4-Chlorophenyl-phenylether
Concen: 19.929 ng/ul
RT: 15.487 min Scan# 2108
Delta R.T. -0.006 min
Lab File: BN020126.D
Acq: 11 Jun 2022 07:09




Instrument :
BNA_N
ClientSampleId :
SSTD020317


Tgt Ion:204 Resp: 307940
Ion Ratio Lower Upper
204 100
206 31.2 26.7 40.1
141 68.4 50.2 75.2



#63
4-Nitroaniline
Concen: 21.919 ng/ul
RT: 15.504 min Scan# 2111
Delta R.T. -0.006 min
Lab File: BN020126.D
Acq: 11 Jun 2022 07:09

Tgt Ion:138 Resp: 155537
Ion Ratio Lower Upper
138 100
92 53.1 41.0 61.6
108 85.6 65.3 97.9









#66
 4,6-Dinitro-2-methylphenol
 Concen: 18.251 ng/ul
 RT: 15.569 min Scan# 2122
 Delta R.T. -0.006 min
 Lab File: BN020126.D
 Acq: 11 Jun 2022 07:09




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020317


Tgt Ion:198 Resp: 110654
 Ion Ratio Lower Upper
 198 100
 182 4.5 3.0 4.4#
 77 29.2 20.1 30.1



#67
 N-Nitrosodiphenylamine
 Concen: 19.960 ng/ul
 RT: 15.698 min Scan# 2144
 Delta R.T. -0.006 min
 Lab File: BN020126.D
 Acq: 11 Jun 2022 07:09

Tgt Ion:169 Resp: 590716
 Ion Ratio Lower Upper
 169 100
 168 67.5 52.4 78.6
 167 34.5 27.5 41.3









#68
4-Bromophenyl-phenylether
Concen: 20.103 ng/ul
RT: 16.381 min Scan# 21
Delta R.T. -0.006 min
Lab File: BN020126.D
Acq: 11 Jun 2022 07:09




Instrument :
BNA_N
ClientSampleId :
SSTD020317


Tgt Ion:248 Resp: 180730
Ion Ratio Lower Upper
248 100
250 99.0 77.8 116.8
141 94.6 62.6 93.8#



#69
Hexachlorobenzene
Concen: 19.596 ng/ul
RT: 16.498 min Scan# 2280
Delta R.T. -0.006 min
Lab File: BN020126.D
Acq: 11 Jun 2022 07:09

Tgt Ion:284 Resp: 202504
Ion Ratio Lower Upper
284 100
142 45.3 27.8 41.8#
249 33.8 25.7 38.5





#70

Atrazine

Concen: 20.450 ng/ul

RT: 16.651 min Scan# 2111

Delta R.T. -0.006 min

Lab File: BN020126.D

Acq: 11 Jun 2022 07:09

Instrument :

BNA_N

ClientSampleId :

SSTD020317





Tgt Ion:200 Resp: 211173

Ion Ratio Lower Upper

200 100

173 26.9 20.5 30.7

215 42.6 38.9 58.3



#71

Pentachlorophenol

Concen: 17.688 ng/ul

RT: 16.840 min Scan# 2338

Delta R.T. -0.006 min

Lab File: BN020126.D

Acq: 11 Jun 2022 07:09




Tgt Ion:266 Resp: 113846


Ion Ratio Lower Upper

266 100

264 61.3 50.2 75.2

268 63.4 49.3 73.9









#72
Phenanthrene
Concen: 19.925 ng/ul
RT: 17.228 min Scan# 2419
Delta R.T. -0.006 min
Lab File: BN020126.D
Acq: 11 Jun 2022 07:09




Instrument :
BNA_N
ClientSampleId :
SSTD020317


Tgt Ion:178 Resp: 1085511
Ion Ratio Lower Upper
178 100
179 15.6 12.2 18.4
176 19.3 15.2 22.8



#74
Anthracene
Concen: 20.002 ng/ul
RT: 17.316 min Scan# 2419
Delta R.T. -0.011 min
Lab File: BN020126.D
Acq: 11 Jun 2022 07:09

Tgt Ion:178 Resp: 1095736
Ion Ratio Lower Upper
178 100
179 16.2 13.0 19.4
176 18.9 15.5 23.3









#75
1,2,3,4-Tetrachlorobenzene
Concen: 19.621 ng/uL
RT: 13.246 min Scan# 1727
Delta R.T. -0.006 min
Lab File: BN020126.D
Acq: 11 Jun 2022 07:09




Instrument :
BNA_N
ClientSampleId :
SSTD020317


Tgt Ion:216 Resp: 271244
Ion Ratio Lower Upper
216 100
214 79.1 63.2 94.8
218 47.1 37.9 56.9



#76
Pentachlorobenzene
Concen: 20.242 ng/uL
RT: 14.769 min Scan# 1986
Delta R.T. -0.006 min
Lab File: BN020126.D
Acq: 11 Jun 2022 07:09

Tgt Ion:250 Resp: 247317
Ion Ratio Lower Upper
250 100
248 68.6 53.0 79.4
252 65.0 53.4 80.2









#77
 Carbazole
 Concen: 20.362 ng/ul
 RT: 17.587 min Scan# 2465
 Delta R.T. -0.006 min
 Lab File: BN020126.D
 Acq: 11 Jun 2022 07:09




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020317


Tgt Ion:167 Resp: 1014775
 Ion Ratio Lower Upper
 167 100
 166 21.9 16.1 24.1
 139 13.7 10.2 15.2



#78
 Di-n-butylphthalate
 Concen: 20.641 ng/ul
 RT: 18.157 min Scan# 2562
 Delta R.T. -0.006 min
 Lab File: BN020126.D
 Acq: 11 Jun 2022 07:09

Tgt Ion:149 Resp: 1275532
 Ion Ratio Lower Upper
 149 100
 150 9.1 7.1 10.7
 104 5.2 4.1 6.1









#80
Fluoranthene
Concen: 20.401 ng/u1
RT: 19.245 min Scan# 21
Delta R.T. -0.006 min
Lab File: BN020126.D
Acq: 11 Jun 2022 07:09




Instrument :
BNA_N
ClientSampleId :
SSTD020317


Tgt Ion:202 Resp: 1235656
Ion Ratio Lower Upper
202 100
101 11.5 8.9 13.3
100 9.3 6.7 10.1



#82
Pyrene
Concen: 20.316 ng/u1
RT: 19.604 min Scan# 2808
Delta R.T. -0.012 min
Lab File: BN020126.D
Acq: 11 Jun 2022 07:09

Tgt Ion:202 Resp: 1262409
Ion Ratio Lower Upper
202 100
101 15.0 9.7 14.5#
100 12.5 8.0 12.0#









#83
Butylbenzylphthalate
Concen: 20.847 ng/ul
RT: 20.510 min Scan# 2962
Delta R.T. -0.012 min
Lab File: BN020126.D
Acq: 11 Jun 2022 07:09




Instrument :
BNA_N
ClientSampleId :
SSTD020317


Tgt Ion:149 Resp: 582989
Ion Ratio Lower Upper
149 100
91 70.9 58.8 88.2
206 19.0 18.1 27.1



#84
3,3'-Dichlorobenzidine
Concen: 20.384 ng/ul
RT: 21.292 min Scan# 3095
Delta R.T. -0.006 min
Lab File: BN020126.D
Acq: 11 Jun 2022 07:09

Tgt Ion:252 Resp: 360146
Ion Ratio Lower Upper
252 100
254 65.2 50.8 76.2
126 12.8 8.2 12.4#









#85
Benzo(a)anthracene
Concen: 20.002 ng/ul
RT: 21.357 min Scan# 3106
Delta R.T. -0.006 min
Lab File: BN020126.D
Acq: 11 Jun 2022 07:09




Instrument :
BNA_N
ClientSampleId :
SSTD020317


Tgt Ion:228 Resp: 1168438
Ion Ratio Lower Upper
228 100
229 20.5 16.3 24.5
226 26.9 23.2 34.8



#86
Bis(2-ethylhexyl)phthalate
Concen: 21.362 ng/ul
RT: 21.298 min Scan# 3096
Delta R.T. -0.006 min
Lab File: BN020126.D
Acq: 11 Jun 2022 07:09

Tgt Ion:149 Resp: 842874
Ion Ratio Lower Upper
149 100
167 27.0 22.5 33.7
279 3.9 3.5 5.3





#87

Chrysene

Concen: 19.627 ng/ul

RT: 21.410 min Scan# 3115

Delta R.T. -0.012 min

Lab File: BN020126.D

Acq: 11 Jun 2022 07:09

Instrument :

BNA_N

ClientSampleId :

SSTD020317





Tgt Ion:228 Resp: 1121670

Ion Ratio Lower Upper

228 100

226 31.0 24.2 36.2

229 19.5 16.2 24.2



#89

Di-n-octyl phthalate

Concen: 21.144 ng/ul




RT: 22.198 min Scan# 3249


Delta R.T. -0.012 min

Lab File: BN020126.D

Acq: 11 Jun 2022 07:09

Tgt Ion:149 Resp: 1468683









#90
Benzo(b)fluoranthene
Concen: 20.536 ng/ul
RT: 22.992 min Scan# 3384
Delta R.T. -0.011 min
Lab File: BN020126.D
Acq: 11 Jun 2022 07:09




Instrument :
BNA_N
ClientSampleId :
SSTD020317


Tgt Ion:252 Resp: 1099986
Ion Ratio Lower Upper
252 100
253 22.5 16.5 24.7
125 12.9 7.8 11.8#



#91
Benzo(k)fluoranthene
Concen: 20.744 ng/ul
RT: 23.039 min Scan# 3392
Delta R.T. -0.011 min
Lab File: BN020126.D
Acq: 11 Jun 2022 07:09

Tgt Ion:252 Resp: 1057692
Ion Ratio Lower Upper
252 100
253 21.5 18.2 27.4
125 12.3 8.2 12.4









#93
Benzo(a)pyrene
Concen: 19.977 ng/ul
RT: 23.598 min Scan# 3487
Delta R.T. -0.012 min
Lab File: BN020126.D
Acq: 11 Jun 2022 07:09




Instrument :
BNA_N
ClientSampleId :
SSTD020317


Tgt Ion:252 Resp: 1035756
Ion Ratio Lower Upper
252 100
253 23.3 18.7 28.1
125 13.9 9.2 13.8#



#94
Indeno(1,2,3-cd)pyrene
Concen: 18.875 ng/ul
RT: 26.092 min Scan# 3911
Delta R.T. -0.006 min
Lab File: BN020126.D
Acq: 11 Jun 2022 07:09

Tgt Ion:276 Resp: 1006864
Ion Ratio Lower Upper
276 100
138 29.0 19.5 29.3
277 23.3 21.4 32.0





#95

Dibenzo(a,h)anthracene

Concen: 18.788 ng/ul

RT: 26.104 min Scan# 3913

Delta R.T. -0.017 min

Lab File: BN020126.D

Acq: 11 Jun 2022 07:09

Instrument :

BNA_N

ClientSampleId :

SSTD020317





Tgt Ion:278 Resp: 856022

Ion Ratio Lower Upper

278 100

139 20.3 13.8 20.8

279 24.5 18.3 27.5



#96

Benzo(g,h,i)perylene

Concen: 18.408 ng/ul

RT: 26.821 min Scan# 4035

Delta R.T. -0.023 min

Lab File: BN020126.D

Acq: 11 Jun 2022 07:09

Tgt Ion:276 Resp: 829253

Ion Ratio Lower Upper

276 100

138 25.0 17.0 25.6

277 22.8 17.8 26.6

