

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061122\
 Data File : BN020143.D
 Acq On : 11 Jun 2022 17:52
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020318

Quant Time: Jun 13 00:45:57 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN060622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jun 06 15:14:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.816	152	153700	20.000	ng/u1	0.00
20) Naphthalene-d8	10.604	136	735688	20.000	ng/u1	-0.01
38) Acenaphthene-d10	14.445	164	481498	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.180	188	973474	20.000	ng/u1	-0.01
79) Chrysene-d12	21.374	240	883112	20.000	ng/u1	0.00
88) Perylene-d12	23.698	264	791987	20.000	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.281	96	27474	7.742	ng/uL	0.00
4) Pyridine-d5	3.687	84	193986	20.000	ng/u1	0.00
7) Phenol-d5	6.981	99	259170	20.378	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.146	67	163043	20.519	ng/u1	0.00
11) 2-Chlorophenol-d4	7.346	132	218003	21.096	ng/u1	-0.01
15) 4-Methylphenol-d8	8.522	113	229247	20.834	ng/u1	0.00
21) Nitrobenzene-d5	8.969	128	112589	20.116	ng/u1	0.00
24) 2-Nitrophenol-d4	9.687	143	124956	21.037	ng/u1	-0.01
28) 2,4-Dichlorophenol-d3	10.228	165	213758	20.614	ng/u1	0.00
31) 4-Chloroaniline-d4	10.740	131	338229	20.142	ng/u1	0.00
46) Dimethylphthalate-d6	13.851	166	679125	19.384	ng/u1	-0.01
49) Acenaphthylene-d8	14.134	160	828314	20.387	ng/u1	-0.01
54) 4-Nitrophenol-d4	14.639	143	128880	18.358	ng/u1	0.00
60) Fluorene-d10	15.434	176	581896	20.072	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.557	200	104642	18.557	ng/u1	0.00
73) Anthracene-d10	17.280	188	881944	20.759	ng/u1	-0.01
81) Pyrene-d10	19.580	212	979491	20.459	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.551	264	816670	21.075	ng/u1	-0.01
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.311	88	28385	7.577	ng/uL	94
5) Pyridine	3.705	79	203293	19.993	ng/u1	93
6) Benzaldehyde	6.952	77	147474	24.609	ng/u1	93
8) Phenol	7.010	94	278065	20.532	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.240	93	219491	20.707	ng/u1	100
12) 2-Chlorophenol	7.381	128	226492	20.968	ng/u1	100
13) 2-Methylphenol	8.257	108	218570	20.993	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.346	45	335094	20.160	ng/u1	98
16) Acetophenone	8.634	105	354407	21.179	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.622	70	179324	20.978	ng/u1	94
18) 4-Methylphenol	8.587	108	244809	21.304	ng/u1	100
19) Hexachloroethane	8.893	117	88062	20.261	ng/u1	86
22) Nitrobenzene	9.010	77	262099	20.197	ng/u1	98
23) Isophorone	9.534	82	503658	19.656	ng/u1	98
25) 2-Nitrophenol	9.722	139	137083	20.885	ng/u1	97
26) 2,4-Dimethylphenol	9.787	107	278299	20.466	ng/u1	96
27) Bis(2-Chloroethoxy)met...	10.016	93	307960	19.640	ng/u1	97
29) 2,4-Dichlorophenol	10.251	162	222820	20.773	ng/u1	99
30) Naphthalene	10.657	128	760475	19.945	ng/u1	100
32) 4-Chloroaniline	10.763	127	338988	20.183	ng/u1	99
33) Hexachlorobutadiene	10.951	225	123536	20.060	ng/u1	99
34) Caprolactam	11.510	113	57475	13.721	ng/u1	97
35) 4-Chloro-3-methylphenol	11.893	107	264724	20.901	ng/u1	100
36) 2-Methylnaphthalene	12.263	142	523804	20.321	ng/u1	100


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061122\
 Data File : BN020143.D
 Acq On : 11 Jun 2022 17:52
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020318

Quant Time: Jun 13 00:45:57 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN060622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jun 06 15:14:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.487	142	531766	20.215	ng/ul	94
39) 1,2,4,5-Tetrachloroben...	12.640	216	243861	19.858	ng/ul	98
40) Hexachlorocyclopentadiene	12.622	237	129157	17.289	ng/ul	95
41) 2,4,6-Trichlorophenol	12.875	196	174574	20.617	ng/ul	93
42) 2,4,5-Trichlorophenol	12.945	196	189436	20.200	ng/ul	99
43) 1,1'-Biphenyl	13.281	154	705366	19.784	ng/ul	99
44) 2-Chloronaphthalene	13.316	162	540963	20.000	ng/ul	97
45) 2-Nitroaniline	13.516	65	174918	20.787	ng/ul	96
47) Dimethylphthalate	13.898	163	690011	19.564	ng/ul	99
48) 2,6-Dinitrotoluene	14.016	165	141749	20.556	ng/ul	100
50) Acenaphthylene	14.163	152	902108	20.054	ng/ul	99
51) 3-Nitroaniline	14.345	138	149014	20.441	ng/ul	98
52) Acenaphthene	14.510	153	579366	19.718	ng/ul	99
53) 2,4-Dinitrophenol	14.551	184	60935	13.802	ng/ul	89
55) 4-Nitrophenol	14.651	109	111772	18.652	ng/ul	98
56) Dibenzofuran	14.839	168	816918	19.734	ng/ul	97
57) 2,4-Dinitrotoluene	14.804	165	209722	20.540	ng/ul	95
58) 2,3,4,6-Tetrachlorophenol	15.069	232	146610	19.915	ng/ul	89
59) Diethylphthalate	15.269	149	703793	19.382	ng/ul	99
61) Fluorene	15.492	166	639017	19.838	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.486	204	299949	20.220	ng/ul	93
63) 4-Nitroaniline	15.504	138	141501	20.772	ng/ul	97
66) 4,6-Dinitro-2-methylph...	15.569	198	106911	18.700	ng/ul	94
67) N-Nitrosodiphenylamine	15.698	169	578369	20.724	ng/ul	97
68) 4-Bromophenyl-phenylether	16.375	248	176154	20.779	ng/ul#	91
69) Hexachlorobenzene	16.498	284	197099	20.226	ng/ul#	88
70) Atrazine	16.651	200	204173	20.968	ng/ul	98
71) Pentachlorophenol	16.839	266	111859	18.430	ng/ul	95
72) Phenanthrene	17.228	178	1041639	20.276	ng/ul	99
74) Anthracene	17.316	178	1071594	20.744	ng/ul	97
75) 1,2,3,4-Tetrachloroben...	13.245	216	261099	20.029	ng/ul	98
76) Pentachlorobenzene	14.769	250	240840	20.904	ng/ul	96
77) Carbazole	17.586	167	990540	21.078	ng/ul	98
78) Di-n-butylphthalate	18.157	149	1199869	20.590	ng/ul	99
80) Fluoranthene	19.245	202	1194148	20.652	ng/ul	97
82) Pyrene	19.604	202	1215871	20.496	ng/ul#	94
83) Butylbenzylphthalate	20.510	149	558648	20.925	ng/ul#	97
84) 3,3'-Dichlorobenzidine	21.292	252	343444	20.362	ng/ul#	96
85) Benzo(a)anthracene	21.357	228	1124306	20.161	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.298	149	804572	21.360	ng/ul	100
87) Chrysene	21.410	228	1078763	19.773	ng/ul	98
89) Di-n-octyl phthalate	22.198	149	1407764	21.839	ng/ul	100
90) Benzo(b)fluoranthene	22.998	252	1099754	22.125	ng/ul	97
91) Benzo(k)fluoranthene	23.045	252	989533	20.913	ng/ul	95
93) Benzo(a)pyrene	23.598	252	1001263	20.810	ng/ul	97
94) Indeno(1,2,3-cd)pyrene	26.092	276	970134	19.597	ng/ul#	94
95) Dibenzo(a,h)anthracene	26.104	278	830541	19.643	ng/ul	96
96) Benzo(g,h,i)perylene	26.821	276	806401	19.290	ng/ul#	93





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 7.577 ng/uL
RT: 3.311 min Scan# 31
Delta R.T. -0.005 min
Lab File: BN020143.D
Acq: 11 Jun 2022 17:52




Instrument :
BNA_N
ClientSampleId :
SSTD020318


Tgt Ion: 88 Resp: 28385
Ion Ratio Lower Upper
88 100
43 31.3 22.4 33.6
58 70.7 61.0 91.4



#5
Pyridine
Concen: 19.993 ng/uL
RT: 3.705 min Scan# 105
Delta R.T. -0.006 min
Lab File: BN020143.D
Acq: 11 Jun 2022 17:52

Tgt Ion: 79 Resp: 203293
Ion Ratio Lower Upper
79 100
52 88.4 65.0 97.4
51 42.7 31.6 47.4









#6
Benzaldehyde
Concen: 24.609 ng/ul
RT: 6.952 min Scan# 61
Delta R.T. -0.012 min
Lab File: BN020143.D
Acq: 11 Jun 2022 17:52




Instrument :
BNA_N
ClientSampleId :
SSTD020318


Tgt Ion: 77 Resp: 147474
Ion Ratio Lower Upper
77 100
105 99.0 74.7 112.1
106 95.8 70.2 105.4



#8
Phenol
Concen: 20.532 ng/ul
RT: 7.010 min Scan# 667
Delta R.T. -0.006 min
Lab File: BN020143.D
Acq: 11 Jun 2022 17:52

Tgt Ion: 94 Resp: 278065
Ion Ratio Lower Upper
94 100
65 29.6 23.6 35.4
66 40.7 31.2 46.8









#10
 Bis(2-Chloroethyl)ether
 Concen: 20.707 ng/ul
 RT: 7.240 min Scan# 706
 Delta R.T. -0.006 min
 Lab File: BN020143.D
 Acq: 11 Jun 2022 17:52

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020318

Tgt Ion: 93 Resp: 219491




Ion	Ratio	Lower	Upper
93	100		
63	79.2	63.4	95.0
95	32.6	26.2	39.2




#12
 2-Chlorophenol
 Concen: 20.968 ng/ul
 RT: 7.381 min Scan# 730
 Delta R.T. -0.006 min
 Lab File: BN020143.D
 Acq: 11 Jun 2022 17:52

Tgt Ion: 128 Resp: 226492

Ion	Ratio	Lower	Upper
128	100		
64	47.9	38.6	57.8
130	31.2	25.0	37.4









#13
2-Methylphenol
Concen: 20.993 ng/ul
RT: 8.257 min Scan# 81
Delta R.T. -0.006 min
Lab File: BN020143.D
Acq: 11 Jun 2022 17:52




Instrument :
BNA_N
ClientSampleId :
SSTD020318


Tgt Ion:108 Resp: 218570
Ion Ratio Lower Upper
108 100
107 88.0 71.5 107.3



#14
2,2'-oxybis(1-Chloropropane)
Concen: 20.160 ng/ul
RT: 8.346 min Scan# 894
Delta R.T. -0.012 min
Lab File: BN020143.D
Acq: 11 Jun 2022 17:52

Tgt Ion: 45 Resp: 335094
Ion Ratio Lower Upper
45 100
77 12.6 10.9 16.3
79 10.1 8.6 13.0









#16
Acetophenone
Concen: 21.179 ng/ul
RT: 8.634 min Scan# 943
Delta R.T. -0.006 min
Lab File: BN020143.D
Acq: 11 Jun 2022 17:52




Instrument :
BNA_N
ClientSampleId :
SSTD020318


Tgt Ion:105 Resp: 354407
Ion Ratio Lower Upper
105 100
77 78.3 63.5 95.3
51 30.3 23.5 35.3



#17
N-Nitroso-di-n-propylamine
Concen: 20.978 ng/ul
RT: 8.622 min Scan# 941
Delta R.T. -0.006 min
Lab File: BN020143.D
Acq: 11 Jun 2022 17:52

Tgt Ion: 70 Resp: 179324
Ion Ratio Lower Upper
70 100
42 64.6 47.8 71.6
101 10.2 7.8 11.6
130 22.4 15.8 23.8









#18
4-Methylphenol
Concen: 21.304 ng/ul
RT: 8.587 min Scan# 911
Delta R.T. -0.006 min
Lab File: BN020143.D
Acq: 11 Jun 2022 17:52




Instrument :
BNA_N
ClientSampleId :
SSTD020318


Tgt Ion:108 Resp: 244809
Ion Ratio Lower Upper
108 100
107 109.7 88.0 132.0



#19
Hexachloroethane
Concen: 20.261 ng/ul
RT: 8.893 min Scan# 987
Delta R.T. -0.012 min
Lab File: BN020143.D
Acq: 11 Jun 2022 17:52

Tgt Ion:117 Resp: 88062
Ion Ratio Lower Upper
117 100
201 88.1 83.2 124.8
199 54.6 50.2 75.4









#22
Nitrobenzene
Concen: 20.197 ng/ul
RT: 9.010 min Scan# 1007
Delta R.T. -0.006 min
Lab File: BN020143.D
Acq: 11 Jun 2022 17:52




Instrument :
BNA_N
ClientSampleId :
SSTD020318


Tgt Ion: 77 Resp: 262099
Ion Ratio Lower Upper
77 100
123 45.0 34.5 51.7
65 13.5 11.1 16.7



#23
Isophorone
Concen: 19.656 ng/ul
RT: 9.534 min Scan# 1096
Delta R.T. -0.006 min
Lab File: BN020143.D
Acq: 11 Jun 2022 17:52

Tgt Ion: 82 Resp: 503658
Ion Ratio Lower Upper
82 100
95 7.6 5.8 8.8
138 18.3 13.8 20.8





#25

2-Nitrophenol

Concen: 20.885 ng/ul

RT: 9.722 min Scan# 1128

Delta R.T. -0.006 min

Lab File: BN020143.D

Acq: 11 Jun 2022 17:52

Instrument :

BNA_N

ClientSampleId :

SSTD020318





Tgt Ion:139 Resp: 137083

Ion Ratio Lower Upper

139 100

65 51.6 43.7 65.5

109 33.0 25.8 38.6



#26

2,4-Dimethylphenol

Concen: 20.466 ng/ul

RT: 9.787 min Scan# 1139

Delta R.T. -0.006 min

Lab File: BN020143.D

Acq: 11 Jun 2022 17:52




Tgt Ion:107 Resp: 278299


Ion Ratio Lower Upper

107 100

121 50.2 38.3 57.5

122 86.6 66.3 99.5









#27
Bis(2-Chloroethoxy)methane
Concen: 19.640 ng/ul
RT: 10.016 min Scan# 1178
Delta R.T. -0.012 min
Lab File: BN020143.D
Acq: 11 Jun 2022 17:52




Instrument :
BNA_N
ClientSampleId :
SSTD020318


Tgt Ion: 93 Resp: 307960
Ion Ratio Lower Upper
93 100
95 33.4 25.4 38.0
123 12.3 8.3 12.5



#29
2,4-Dichlorophenol
Concen: 20.773 ng/ul
RT: 10.251 min Scan# 1218
Delta R.T. -0.012 min
Lab File: BN020143.D
Acq: 11 Jun 2022 17:52

Tgt Ion: 162 Resp: 222820
Ion Ratio Lower Upper
162 100
164 65.9 52.9 79.3
98 38.6 29.8 44.8









#30
Naphthalene
Concen: 19.945 ng/ul
RT: 10.657 min Scan# 1287
Delta R.T. -0.006 min
Lab File: BN020143.D
Acq: 11 Jun 2022 17:52




Instrument :
BNA_N
ClientSampleId :
SSTD020318


Tgt Ion:128 Resp: 760475
Ion Ratio Lower Upper
128 100
129 11.0 8.7 13.1
127 13.5 10.6 16.0



#32
4-Chloroaniline
Concen: 20.183 ng/ul
RT: 10.763 min Scan# 1305
Delta R.T. -0.006 min
Lab File: BN020143.D
Acq: 11 Jun 2022 17:52

Tgt Ion:127 Resp: 338988
Ion Ratio Lower Upper
127 100
129 32.1 26.1 39.1









#33
Hexachlorobutadiene
Concen: 20.060 ng/ul
RT: 10.951 min Scan# 11
Delta R.T. -0.006 min
Lab File: BN020143.D
Acq: 11 Jun 2022 17:52




Instrument :
BNA_N
ClientSampleId :
SSTD020318


Tgt Ion:225 Resp: 123536
Ion Ratio Lower Upper
225 100
223 64.9 51.8 77.6
227 65.1 50.6 76.0



#34
Caprolactam
Concen: 13.721 ng/ul
RT: 11.510 min Scan# 1432
Delta R.T. -0.006 min
Lab File: BN020143.D
Acq: 11 Jun 2022 17:52

Tgt Ion:113 Resp: 57475
Ion Ratio Lower Upper
113 100
55 181.5 145.5 218.3
56 142.2 106.5 159.7









#35
4-Chloro-3-methylphenol
Concen: 20.901 ng/ul
RT: 11.893 min Scan# 1497
Delta R.T. -0.006 min
Lab File: BN020143.D
Acq: 11 Jun 2022 17:52




Instrument :
BNA_N
ClientSampleId :
SSTD020318


Tgt Ion:107 Resp: 264724
Ion Ratio Lower Upper
107 100
144 25.2 20.6 31.0
142 77.7 62.2 93.4



#36
2-Methylnaphthalene
Concen: 20.321 ng/ul
RT: 12.263 min Scan# 1560
Delta R.T. -0.012 min
Lab File: BN020143.D
Acq: 11 Jun 2022 17:52

Tgt Ion:142 Resp: 523804
Ion Ratio Lower Upper
142 100
141 88.1 70.6 105.8





#37

1-Methylnaphthalene

Concen: 20.215 ng/ul

RT: 12.487 min Scan# 1

Delta R.T. -0.006 min

Lab File: BN020143.D

Acq: 11 Jun 2022 17:52

Instrument :

BNA_N

ClientSampleId :

SSTD020318





Tgt Ion:142 Resp: 531766

Ion Ratio Lower Upper

142 100

141 94.0 70.9 106.3

116 4.0 3.1 4.7



#39

1,2,4,5-Tetrachlorobenzene

Concen: 19.858 ng/ul

RT: 12.640 min Scan# 1624

Delta R.T. -0.012 min

Lab File: BN020143.D

Acq: 11 Jun 2022 17:52

Tgt Ion:216 Resp: 243861




Ion Ratio Lower Upper


216 100

214 78.1 63.4 95.0

179 24.0 17.8 26.8

108 21.4 15.1 22.7









#40
Hexachlorocyclopentadiene
Concen: 17.289 ng/ul
RT: 12.622 min Scan# 1622
Delta R.T. -0.012 min
Lab File: BN020143.D
Acq: 11 Jun 2022 17:52




Instrument :
BNA_N
ClientSampleId :
SSTD020318


Tgt Ion:237 Resp: 129157
Ion Ratio Lower Upper
237 100
235 65.4 48.8 73.2
272 12.5 9.3 13.9



#41
2,4,6-Trichlorophenol
Concen: 20.617 ng/ul
RT: 12.875 min Scan# 1664
Delta R.T. -0.012 min
Lab File: BN020143.D
Acq: 11 Jun 2022 17:52

Tgt Ion:196 Resp: 174574
Ion Ratio Lower Upper
196 100
198 92.7 79.8 119.8
200 27.2 25.0 37.6









#42
2,4,5-Trichlorophenol
Concen: 20.200 ng/ul
RT: 12.945 min Scan# 1676
Delta R.T. -0.012 min
Lab File: BN020143.D
Acq: 11 Jun 2022 17:52




Instrument :
BNA_N
ClientSampleId :
SSTD020318


Tgt Ion:196 Resp: 189436
Ion Ratio Lower Upper
196 100
198 97.7 77.6 116.4
200 30.5 24.4 36.6



#43
1,1'-Biphenyl
Concen: 19.784 ng/ul
RT: 13.281 min Scan# 1733
Delta R.T. -0.006 min
Lab File: BN020143.D
Acq: 11 Jun 2022 17:52

Tgt Ion:154 Resp: 705366
Ion Ratio Lower Upper
154 100
153 39.3 31.5 47.3
76 12.0 10.5 15.7









#44
2-Chloronaphthalene
Concen: 20.000 ng/ul
RT: 13.316 min Scan# 1739
Delta R.T. -0.012 min
Lab File: BN020143.D
Acq: 11 Jun 2022 17:52




Instrument :
BNA_N
ClientSampleId :
SSTD020318


Tgt Ion:162 Resp: 540963
Ion Ratio Lower Upper
162 100
164 30.9 25.8 38.6
127 41.1 31.4 47.2



#45
2-Nitroaniline
Concen: 20.787 ng/ul
RT: 13.516 min Scan# 1773
Delta R.T. -0.012 min
Lab File: BN020143.D
Acq: 11 Jun 2022 17:52

Tgt Ion: 65 Resp: 174918
Ion Ratio Lower Upper
65 100
92 69.0 54.1 81.1
138 102.1 77.0 115.4









#47
Dimethylphthalate
Concen: 19.564 ng/ul
RT: 13.898 min Scan# 1838
Delta R.T. -0.012 min
Lab File: BN020143.D
Acq: 11 Jun 2022 17:52




Instrument :
BNA_N
ClientSampleId :
SSTD020318


Tgt Ion:163 Resp: 690011
Ion Ratio Lower Upper
163 100
194 3.9 3.4 5.0
164 10.2 7.7 11.5



#48
2,6-Dinitrotoluene
Concen: 20.556 ng/ul
RT: 14.016 min Scan# 1858
Delta R.T. -0.012 min
Lab File: BN020143.D
Acq: 11 Jun 2022 17:52

Tgt Ion:165 Resp: 141749
Ion Ratio Lower Upper
165 100
89 63.1 50.6 75.8
121 22.8 17.9 26.9









#50
Acenaphthylene
Concen: 20.054 ng/ul
RT: 14.163 min Scan# 1883
Delta R.T. -0.006 min
Lab File: BN020143.D
Acq: 11 Jun 2022 17:52




Instrument :
BNA_N
ClientSampleId :
SSTD020318


Tgt Ion:152 Resp: 902108
Ion Ratio Lower Upper
152 100
151 20.2 16.0 24.0
153 12.8 10.7 16.1



#51
3-Nitroaniline
Concen: 20.441 ng/ul
RT: 14.345 min Scan# 1914
Delta R.T. -0.006 min
Lab File: BN020143.D
Acq: 11 Jun 2022 17:52

Tgt Ion:138 Resp: 149014
Ion Ratio Lower Upper
138 100
108 11.0 8.6 12.8
92 114.3 92.9 139.3









#52
Acenaphthene
Concen: 19.718 ng/ul
RT: 14.510 min Scan# 1942
Delta R.T. -0.006 min
Lab File: BN020143.D
Acq: 11 Jun 2022 17:52




Instrument :
BNA_N
ClientSampleId :
SSTD020318


Tgt Ion:153 Resp: 579366
Ion Ratio Lower Upper
153 100
152 46.4 38.4 57.6
154 86.8 69.7 104.5



#53
2,4-Dinitrophenol
Concen: 13.802 ng/ul
RT: 14.551 min Scan# 1949
Delta R.T. -0.006 min
Lab File: BN020143.D
Acq: 11 Jun 2022 17:52

Tgt Ion:184 Resp: 60935
Ion Ratio Lower Upper
184 100
63 81.4 55.9 83.9
154 72.4 53.1 79.7









#55
4-Nitrophenol
Concen: 18.652 ng/ul
RT: 14.651 min Scan# 1966
Delta R.T. -0.006 min
Lab File: BN020143.D
Acq: 11 Jun 2022 17:52




Instrument :
BNA_N
ClientSampleId :
SSTD020318


Tgt Ion:109 Resp: 111772
Ion Ratio Lower Upper
109 100
139 128.5 100.1 150.1
65 129.6 104.6 157.0



#56
Dibenzofuran
Concen: 19.734 ng/ul
RT: 14.839 min Scan# 1998
Delta R.T. -0.012 min
Lab File: BN020143.D
Acq: 11 Jun 2022 17:52

Tgt Ion:168 Resp: 816918
Ion Ratio Lower Upper
168 100
139 41.0 31.1 46.7









#57
2,4-Dinitrotoluene
Concen: 20.540 ng/ul
RT: 14.804 min Scan# 1992
Delta R.T. -0.006 min
Lab File: BN020143.D
Acq: 11 Jun 2022 17:52




Instrument :
BNA_N
ClientSampleId :
SSTD020318


Tgt Ion:165 Resp: 209722
Ion Ratio Lower Upper
165 100
63 48.7 36.3 54.5
182 3.0 2.3 3.5



#58
2,3,4,6-Tetrachlorophenol
Concen: 19.915 ng/ul
RT: 15.069 min Scan# 2037
Delta R.T. -0.006 min
Lab File: BN020143.D
Acq: 11 Jun 2022 17:52

Tgt Ion:232 Resp: 146610
Ion Ratio Lower Upper
232 100
131 57.2 38.5 57.7
130 3.1 2.2 3.2
166 36.1 25.8 38.8









#59
Diethylphthalate
Concen: 19.382 ng/ul
RT: 15.269 min Scan# 2071
Delta R.T. -0.006 min
Lab File: BN020143.D
Acq: 11 Jun 2022 17:52




Instrument :
BNA_N
ClientSampleId :
SSTD020318


Tgt Ion:149 Resp: 703793
Ion Ratio Lower Upper
149 100
177 21.7 17.8 26.8
150 12.5 10.5 15.7



#61
Fluorene
Concen: 19.838 ng/ul
RT: 15.492 min Scan# 2109
Delta R.T. -0.006 min
Lab File: BN020143.D
Acq: 11 Jun 2022 17:52

Tgt Ion:166 Resp: 639017
Ion Ratio Lower Upper
166 100
165 99.3 77.7 116.5
167 13.2 11.1 16.7









#62
4-Chlorophenyl-phenylether
Concen: 20.220 ng/ul
RT: 15.486 min Scan# 2108
Delta R.T. -0.006 min
Lab File: BN020143.D
Acq: 11 Jun 2022 17:52




Instrument :
BNA_N
ClientSampleId :
SSTD020318


Tgt Ion:204 Resp: 299949
Ion Ratio Lower Upper
204 100
206 32.6 26.7 40.1
141 70.0 50.2 75.2



#63
4-Nitroaniline
Concen: 20.772 ng/ul
RT: 15.504 min Scan# 2111
Delta R.T. -0.006 min
Lab File: BN020143.D
Acq: 11 Jun 2022 17:52

Tgt Ion:138 Resp: 141501
Ion Ratio Lower Upper
138 100
92 49.6 41.0 61.6
108 85.3 65.3 97.9









#66
4,6-Dinitro-2-methylphenol
Concen: 18.700 ng/ul
RT: 15.569 min Scan# 2122
Delta R.T. -0.006 min
Lab File: BN020143.D
Acq: 11 Jun 2022 17:52




Instrument :
BNA_N
ClientSampleId :
SSTD020318


Tgt Ion:198 Resp: 106911
Ion Ratio Lower Upper
198 100
182 3.6 3.0 4.4
77 28.5 20.1 30.1



#67
N-Nitrosodiphenylamine
Concen: 20.724 ng/ul
RT: 15.698 min Scan# 2144
Delta R.T. -0.006 min
Lab File: BN020143.D
Acq: 11 Jun 2022 17:52

Tgt Ion:169 Resp: 578369
Ion Ratio Lower Upper
169 100
168 67.3 52.4 78.6
167 36.3 27.5 41.3









#68
4-Bromophenyl-phenylether
Concen: 20.779 ng/ul
RT: 16.375 min Scan# 21
Delta R.T. -0.012 min
Lab File: BN020143.D
Acq: 11 Jun 2022 17:52




Instrument :
BNA_N
ClientSampleId :
SSTD020318


Tgt Ion:248 Resp: 176154
Ion Ratio Lower Upper
248 100
250 96.7 77.8 116.8
141 94.5 62.6 93.8#



#69
Hexachlorobenzene
Concen: 20.226 ng/ul
RT: 16.498 min Scan# 2280
Delta R.T. -0.006 min
Lab File: BN020143.D
Acq: 11 Jun 2022 17:52

Tgt Ion:284 Resp: 197099
Ion Ratio Lower Upper
284 100
142 45.8 27.8 41.8#
249 34.2 25.7 38.5









#70
 Atrazine
 Concen: 20.968 ng/ul
 RT: 16.651 min Scan# 21
 Delta R.T. -0.006 min
 Lab File: BN020143.D
 Acq: 11 Jun 2022 17:52




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020318


Tgt Ion:200 Resp: 204173
 Ion Ratio Lower Upper
 200 100
 173 27.9 20.5 30.7
 215 47.7 38.9 58.3



#71
 Pentachlorophenol
 Concen: 18.430 ng/ul
 RT: 16.839 min Scan# 2338
 Delta R.T. -0.006 min
 Lab File: BN020143.D
 Acq: 11 Jun 2022 17:52

Tgt Ion:266 Resp: 111859
 Ion Ratio Lower Upper
 266 100
 264 64.9 50.2 75.2
 268 67.8 49.3 73.9









#72
Phenanthrene
Concen: 20.276 ng/ul
RT: 17.228 min Scan# 2419
Delta R.T. -0.006 min
Lab File: BN020143.D
Acq: 11 Jun 2022 17:52




Instrument :
BNA_N
ClientSampleId :
SSTD020318


Tgt Ion:178 Resp: 1041639
Ion Ratio Lower Upper
178 100
179 15.5 12.2 18.4
176 19.4 15.2 22.8



#74
Anthracene
Concen: 20.744 ng/ul
RT: 17.316 min Scan# 2419
Delta R.T. -0.012 min
Lab File: BN020143.D
Acq: 11 Jun 2022 17:52

Tgt Ion:178 Resp: 1071594
Ion Ratio Lower Upper
178 100
179 14.8 13.0 19.4
176 18.3 15.5 23.3









#75
1,2,3,4-Tetrachlorobenzene
Concen: 20.029 ng/uL
RT: 13.245 min Scan# 1727
Delta R.T. -0.006 min
Lab File: BN020143.D
Acq: 11 Jun 2022 17:52




Instrument :
BNA_N
ClientSampleId :
SSTD020318


Tgt Ion:216 Resp: 261099
Ion Ratio Lower Upper
216 100
214 76.1 63.2 94.8
218 47.3 37.9 56.9



#76
Pentachlorobenzene
Concen: 20.904 ng/uL
RT: 14.769 min Scan# 1986
Delta R.T. -0.006 min
Lab File: BN020143.D
Acq: 11 Jun 2022 17:52

Tgt Ion:250 Resp: 240840
Ion Ratio Lower Upper
250 100
248 64.1 53.0 79.4
252 62.8 53.4 80.2









#77
 Carbazole
 Concen: 21.078 ng/ul
 RT: 17.586 min Scan# 2465
 Delta R.T. -0.006 min
 Lab File: BN020143.D
 Acq: 11 Jun 2022 17:52




Instrument :
 BNA_N
 ClientSampleId :
 SST020318


Tgt Ion:167 Resp: 990540
 Ion Ratio Lower Upper
 167 100
 166 21.2 16.1 24.1
 139 13.7 10.2 15.2



#78
 Di-n-butylphthalate
 Concen: 20.590 ng/ul
 RT: 18.157 min Scan# 2562
 Delta R.T. -0.006 min
 Lab File: BN020143.D
 Acq: 11 Jun 2022 17:52

Tgt Ion:149 Resp: 1199869
 Ion Ratio Lower Upper
 149 100
 150 9.6 7.1 10.7
 104 5.2 4.1 6.1









#80
Fluoranthene
Concen: 20.652 ng/u1
RT: 19.245 min Scan# 21
Delta R.T. -0.006 min
Lab File: BN020143.D
Acq: 11 Jun 2022 17:52




Instrument :
BNA_N
ClientSampleId :
SSTD020318


Tgt Ion:202 Resp: 1194148
Ion Ratio Lower Upper
202 100
101 12.3 8.9 13.3
100 9.4 6.7 10.1



#82
Pyrene
Concen: 20.496 ng/u1
RT: 19.604 min Scan# 2808
Delta R.T. -0.012 min
Lab File: BN020143.D
Acq: 11 Jun 2022 17:52

Tgt Ion:202 Resp: 1215871
Ion Ratio Lower Upper
202 100
101 14.2 9.7 14.5
100 12.3 8.0 12.0#









#83
Butylbenzylphthalate
Concen: 20.925 ng/ul
RT: 20.510 min Scan# 2962
Delta R.T. -0.012 min
Lab File: BN020143.D
Acq: 11 Jun 2022 17:52




Instrument :
BNA_N
ClientSampleId :
SSTD020318


Tgt Ion:149 Resp: 558648
Ion Ratio Lower Upper
149 100
91 72.8 58.8 88.2
206 17.9 18.1 27.1#



#84
3,3'-Dichlorobenzidine
Concen: 20.362 ng/ul
RT: 21.292 min Scan# 3095
Delta R.T. -0.006 min
Lab File: BN020143.D
Acq: 11 Jun 2022 17:52

Tgt Ion:252 Resp: 343444
Ion Ratio Lower Upper
252 100
254 65.5 50.8 76.2
126 14.0 8.2 12.4#





#85

Benzo(a)anthracene

Concen: 20.161 ng/ul

RT: 21.357 min Scan# 3106

Delta R.T. -0.006 min

Lab File: BN020143.D

Acq: 11 Jun 2022 17:52

Instrument :

BNA_N

ClientSampleId :

SSTD020318





Tgt Ion:228 Resp: 1124306

Ion Ratio Lower Upper

228 100

229 20.2 16.3 24.5

226 29.2 23.2 34.8



#86

Bis(2-ethylhexyl)phthalate

Concen: 21.360 ng/ul

RT: 21.298 min Scan# 3096

Delta R.T. -0.006 min

Lab File: BN020143.D

Acq: 11 Jun 2022 17:52




Tgt Ion:149 Resp: 804572


Ion Ratio Lower Upper

149 100

167 28.0 22.5 33.7

279 4.0 3.5 5.3





#87

Chrysene

Concen: 19.773 ng/ul

RT: 21.410 min Scan# 3115

Delta R.T. -0.012 min

Lab File: BN020143.D

Acq: 11 Jun 2022 17:52

Instrument :

BNA_N

ClientSampleId :

SSTD020318





Tgt Ion:228 Resp: 1078763

Ion Ratio Lower Upper

228 100

226 31.5 24.2 36.2

229 21.2 16.2 24.2



#89

Di-n-octyl phthalate

Concen: 21.839 ng/ul




RT: 22.198 min Scan# 3249


Delta R.T. -0.012 min

Lab File: BN020143.D

Acq: 11 Jun 2022 17:52

Tgt Ion:149 Resp: 1407764









#90
Benzo(b)fluoranthene
Concen: 22.125 ng/ul
RT: 22.998 min Scan# 3385
Delta R.T. -0.006 min
Lab File: BN020143.D
Acq: 11 Jun 2022 17:52




Instrument :
BNA_N
ClientSampleId :
SSTD020318


Tgt Ion:252 Resp: 1099754
Ion Ratio Lower Upper
252 100
253 21.6 16.5 24.7
125 11.7 7.8 11.8



#91
Benzo(k)fluoranthene
Concen: 20.913 ng/ul
RT: 23.045 min Scan# 3393
Delta R.T. -0.006 min
Lab File: BN020143.D
Acq: 11 Jun 2022 17:52

Tgt Ion:252 Resp: 989533
Ion Ratio Lower Upper
252 100
253 20.0 18.2 27.4
125 11.4 8.2 12.4





#93

Benzo(a)pyrene

Concen: 20.810 ng/u1

RT: 23.598 min Scan# 3487

Delta R.T. -0.012 min

Lab File: BN020143.D

Acq: 11 Jun 2022 17:52

Instrument :

BNA_N

ClientSampleId :

SSTD020318





Tgt Ion:252 Resp: 1001263

Ion Ratio Lower Upper

252 100

253 21.8 18.7 28.1

125 12.7 9.2 13.8



#94

Indeno(1,2,3-cd)pyrene

Concen: 19.597 ng/u1

RT: 26.092 min Scan# 3911

Delta R.T. -0.006 min

Lab File: BN020143.D

Acq: 11 Jun 2022 17:52




Tgt Ion:276 Resp: 970134


Ion Ratio Lower Upper

276 100

138 29.9 19.5 29.3#

277 26.1 21.4 32.0





#95

Dibenzo(a,h)anthracene

Concen: 19.643 ng/ul

RT: 26.104 min Scan# 3914

Delta R.T. -0.018 min

Lab File: BN020143.D

Acq: 11 Jun 2022 17:52

Instrument :

BNA_N

ClientSampleId :

SSTD020318





Tgt Ion:278 Resp: 830541

Ion Ratio Lower Upper

278 100

139 19.0 13.8 20.8

279 20.7 18.3 27.5



#96

Benzo(g,h,i)perylene

Concen: 19.290 ng/ul

RT: 26.821 min Scan# 4035

Delta R.T. -0.023 min

Lab File: BN020143.D

Acq: 11 Jun 2022 17:52

Tgt Ion:276 Resp: 806401

Ion Ratio Lower Upper

276 100

138 26.6 17.0 25.6#

277 23.7 17.8 26.6

