

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061122\
 Data File : BNCCC126.D
 Acq On : 11 Jun 2022 07:44
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :

Quant Time: Jun 13 00:46:11 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN060622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jun 06 15:14:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.816	152	154802	20.000	ng/u1	0.00
20) Naphthalene-d8	10.605	136	713021	20.000	ng/u1	-0.01
38) Acenaphthene-d10	14.445	164	469797	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.186	188	975737	20.000	ng/u1	0.00
79) Chrysene-d12	21.375	240	872654	20.000	ng/u1	0.00
88) Perylene-d12	23.698	264	790855	20.000	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.281	96	27639	7.733	ng/uL	0.00
4) Pyridine-d5	3.687	84	195141	19.976	ng/u1	0.00
7) Phenol-d5	6.981	99	254251	19.849	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.146	67	157823	19.720	ng/u1	0.00
11) 2-Chlorophenol-d4	7.352	132	212425	20.410	ng/u1	0.00
15) 4-Methylphenol-d8	8.522	113	224208	20.231	ng/u1	0.00
21) Nitrobenzene-d5	8.969	128	109428	20.173	ng/u1	0.00
24) 2-Nitrophenol-d4	9.687	143	121012	21.020	ng/u1	-0.01
28) 2,4-Dichlorophenol-d3	10.228	165	208832	20.779	ng/u1	0.00
31) 4-Chloroaniline-d4	10.740	131	327859	20.145	ng/u1	0.00
46) Dimethylphthalate-d6	13.857	166	669029	19.572	ng/u1	0.00
49) Acenaphthylene-d8	14.134	160	800198	20.186	ng/u1	-0.01
54) 4-Nitrophenol-d4	14.640	143	127821	18.660	ng/u1	0.00
60) Fluorene-d10	15.434	176	568885	20.111	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.557	200	104228	18.441	ng/u1	0.00
73) Anthracene-d10	17.281	188	862000	20.242	ng/u1	-0.01
81) Pyrene-d10	19.575	212	977550	20.663	ng/u1	-0.01
92) Benzo(a)pyrene-d12	23.551	264	801057	20.702	ng/u1	-0.01
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.317	88	28871	7.652	ng/uL#	92
5) Pyridine	3.705	79	202780	19.801	ng/u1	95
6) Benzaldehyde	6.958	77	142705	23.644	ng/u1	93
8) Phenol	7.011	94	274524	20.126	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.240	93	212369	19.893	ng/u1	99
12) 2-Chlorophenol	7.381	128	221801	20.388	ng/u1	100
13) 2-Methylphenol	8.258	108	209931	20.020	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.346	45	329920	19.708	ng/u1	98
16) Acetophenone	8.634	105	347343	20.609	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.622	70	173296	20.129	ng/u1	94
18) 4-Methylphenol	8.587	108	237140	20.490	ng/u1	98
19) Hexachloroethane	8.899	117	88118	20.130	ng/u1	85
22) Nitrobenzene	9.010	77	254767	20.256	ng/u1	94
23) Isophorone	9.534	82	487634	19.636	ng/u1	99
25) 2-Nitrophenol	9.722	139	133463	20.980	ng/u1	99
26) 2,4-Dimethylphenol	9.787	107	270125	20.497	ng/u1	97
27) Bis(2-Chloroethoxy)met...	10.016	93	301543	19.842	ng/u1	98
29) 2,4-Dichlorophenol	10.257	162	218094	20.979	ng/u1	96
30) Naphthalene	10.657	128	740133	20.029	ng/u1	99
32) 4-Chloroaniline	10.763	127	334021	20.519	ng/u1	98
33) Hexachlorobutadiene	10.952	225	118543	19.861	ng/u1	100
34) Caprolactam	11.510	113	52025	12.815	ng/u1	95
35) 4-Chloro-3-methylphenol	11.893	107	256485	20.895	ng/u1	99
36) 2-Methylnaphthalene	12.269	142	506387	20.269	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061122\
 Data File : BNCCC126.D
 Acq On : 11 Jun 2022 07:44
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 35 Sample Multiplier: 1


Instrument :
 BNA_N
 ClientSampleId :


Quant Time: Jun 13 00:46:11 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN060622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jun 06 15:14:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.487	142	514694	20.189	ng/ul	95
39) 1,2,4,5-Tetrachloroben...	12.640	216	232741	19.424	ng/ul	98
40) Hexachlorocyclopentadiene	12.628	237	129550	17.774	ng/ul	98
41) 2,4,6-Trichlorophenol	12.875	196	166922	20.204	ng/ul	91
42) 2,4,5-Trichlorophenol	12.951	196	180305	19.705	ng/ul	97
43) 1,1'-Biphenyl	13.281	154	697661	20.055	ng/ul	98
44) 2-Chloronaphthalene	13.322	162	520582	19.726	ng/ul	96
45) 2-Nitroaniline	13.516	65	169117	20.598	ng/ul	96
47) Dimethylphthalate	13.898	163	675235	19.622	ng/ul	99
48) 2,6-Dinitrotoluene	14.016	165	138950	20.652	ng/ul	96
50) Acenaphthylene	14.163	152	871242	19.851	ng/ul	99
51) 3-Nitroaniline	14.345	138	146585	20.608	ng/ul	98
52) Acenaphthene	14.510	153	560099	19.537	ng/ul	98
53) 2,4-Dinitrophenol	14.551	184	58421	13.562	ng/ul	88
55) 4-Nitrophenol	14.651	109	109884	18.793	ng/ul	99
56) Dibenzofuran	14.840	168	797784	19.752	ng/ul	98
57) 2,4-Dinitrotoluene	14.804	165	206695	20.748	ng/ul	95
58) 2,3,4,6-Tetrachlorophenol	15.069	232	144152	20.069	ng/ul	94
59) Diethylphthalate	15.269	149	685969	19.362	ng/ul	99
61) Fluorene	15.492	166	626215	19.925	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.487	204	287182	19.842	ng/ul	92
63) 4-Nitroaniline	15.504	138	138270	20.803	ng/ul	99
66) 4,6-Dinitro-2-methylph...	15.569	198	104078	18.162	ng/ul#	94
67) N-Nitrosodiphenylamine	15.698	169	562861	20.122	ng/ul	97
68) 4-Bromophenyl-phenylether	16.381	248	170420	20.056	ng/ul	91
69) Hexachlorobenzene	16.498	284	193377	19.798	ng/ul#	86
70) Atrazine	16.651	200	198188	20.306	ng/ul	97
71) Pentachlorophenol	16.839	266	108401	17.819	ng/ul	97
72) Phenanthrene	17.228	178	1027713	19.958	ng/ul	99
74) Anthracene	17.316	178	1054193	20.360	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.246	216	249631	19.105	ng/ul	98
76) Pentachlorobenzene	14.769	250	233754	20.242	ng/ul	99
77) Carbazole	17.586	167	974859	20.696	ng/ul	97
78) Di-n-butylphthalate	18.157	149	1188877	20.354	ng/ul	100
80) Fluoranthene	19.245	202	1180601	20.663	ng/ul	98
82) Pyrene	19.604	202	1198961	20.454	ng/ul#	94
83) Butylbenzylphthalate	20.510	149	552659	20.949	ng/ul	97
84) 3,3'-Dichlorobenzidine	21.292	252	356380	21.382	ng/ul#	98
85) Benzo(a)anthracene	21.357	228	1124158	20.400	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.298	149	808314	21.716	ng/ul	99
87) Chrysene	21.410	228	1074248	19.926	ng/ul	99
89) Di-n-octyl phthalate	22.198	149	1427174	22.172	ng/ul	100
90) Benzo(b)fluoranthene	22.992	252	1064313	21.443	ng/ul#	97
91) Benzo(k)fluoranthene	22.992	252	1064313	22.525	ng/ul	96
93) Benzo(a)pyrene	23.598	252	972930	20.250	ng/ul	96
94) Indeno(1,2,3-cd)pyrene	26.086	276	958172	19.383	ng/ul#	93
95) Dibenzo(a,h)anthracene	26.098	278	824899	19.537	ng/ul	95
96) Benzo(g,h,i)perylene	26.821	276	796359	19.077	ng/ul#	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jun 13 00:46:11 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN060622.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Jun 06 15:14:59 2022
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 7.652 ng/uL
RT: 3.317 min Scan# 39
Delta R.T. 0.001 min
Lab File: BNCCC126.D
Acq: 11 Jun 2022 07:44




Instrument :
BNA_N
ClientSampleId :


Tgt Ion: 88 Resp: 28871
Ion Ratio Lower Upper
88 100
43 34.0 22.4 33.6#
58 81.8 61.0 91.4



#5
Pyridine
Concen: 19.801 ng/uL
RT: 3.705 min Scan# 105
Delta R.T. -0.006 min
Lab File: BNCCC126.D
Acq: 11 Jun 2022 07:44

Tgt Ion: 79 Resp: 202780
Ion Ratio Lower Upper
79 100
52 87.9 65.0 97.4
51 40.4 31.6 47.4









#6
Benzaldehyde
Concen: 23.644 ng/ul
RT: 6.958 min Scan# 61
Delta R.T. -0.006 min
Lab File: BNCCC126.D
Acq: 11 Jun 2022 07:44




Instrument :
BNA_N
ClientSampleId :


Tgt Ion: 77 Resp: 142705
Ion Ratio Lower Upper
77 100
105 99.0 74.7 112.1
106 94.7 70.2 105.4



#8
Phenol
Concen: 20.126 ng/ul
RT: 7.011 min Scan# 667
Delta R.T. -0.006 min
Lab File: BNCCC126.D
Acq: 11 Jun 2022 07:44

Tgt Ion: 94 Resp: 274524
Ion Ratio Lower Upper
94 100
65 30.2 23.6 35.4
66 40.6 31.2 46.8





#10

Bis(2-Chloroethyl)ether

Concen: 19.893 ng/ul

RT: 7.240 min Scan# 706

Delta R.T. -0.006 min


Lab File: BNCCC126.D

Acq: 11 Jun 2022 07:44

Instrument :

BNA_N

ClientSampleId :






Tgt Ion: 93 Resp: 212369

Ion Ratio Lower Upper

93 100

63 80.5 63.4 95.0

95 32.7 26.2 39.2



#12

2-Chlorophenol

Concen: 20.388 ng/ul

RT: 7.381 min Scan# 730

Delta R.T. -0.006 min

Lab File: BNCCC126.D

Acq: 11 Jun 2022 07:44




Tgt Ion:128 Resp: 221801


Ion Ratio Lower Upper

128 100

64 47.8 38.6 57.8

130 31.3 25.0 37.4









#13
2-Methylphenol
Concen: 20.020 ng/ul
RT: 8.258 min Scan# 81
Delta R.T. -0.006 min
Lab File: BNCCC126.D
Acq: 11 Jun 2022 07:44




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:108 Resp: 209931
Ion Ratio Lower Upper
108 100
107 89.4 71.5 107.3



#14
2,2'-oxybis(1-Chloropropane)
Concen: 19.708 ng/ul
RT: 8.346 min Scan# 894
Delta R.T. -0.012 min
Lab File: BNCCC126.D
Acq: 11 Jun 2022 07:44

Tgt Ion: 45 Resp: 329920
Ion Ratio Lower Upper
45 100
77 13.1 10.9 16.3
79 9.7 8.6 13.0









#16
Acetophenone
Concen: 20.609 ng/ul
RT: 8.634 min Scan# 941
Delta R.T. -0.006 min
Lab File: BNCCC126.D
Acq: 11 Jun 2022 07:44




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:105 Resp: 347343
Ion Ratio Lower Upper
105 100
77 77.9 63.5 95.3
51 30.1 23.5 35.3



#17
N-Nitroso-di-n-propylamine
Concen: 20.129 ng/ul
RT: 8.622 min Scan# 941
Delta R.T. -0.006 min
Lab File: BNCCC126.D
Acq: 11 Jun 2022 07:44

Tgt Ion: 70 Resp: 173296
Ion Ratio Lower Upper
70 100
42 65.5 47.8 71.6
101 10.1 7.8 11.6
130 22.1 15.8 23.8









#18
4-Methylphenol
Concen: 20.490 ng/ul
RT: 8.587 min Scan# 911
Delta R.T. -0.006 min
Lab File: BNCCC126.D
Acq: 11 Jun 2022 07:44




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:108 Resp: 237140
Ion Ratio Lower Upper
108 100
107 112.3 88.0 132.0



#19
Hexachloroethane
Concen: 20.130 ng/ul
RT: 8.899 min Scan# 988
Delta R.T. -0.006 min
Lab File: BNCCC126.D
Acq: 11 Jun 2022 07:44

Tgt Ion:117 Resp: 88118
Ion Ratio Lower Upper
117 100
201 86.3 83.2 124.8
199 54.7 50.2 75.4





#22

Nitrobenzene

Concen: 20.256 ng/ul

RT: 9.010 min Scan# 1007

Delta R.T. -0.006 min

Lab File: BNCCC126.D

Acq: 11 Jun 2022 07:44

Instrument :

BNA_N

ClientSampleId :





Tgt Ion: 77 Resp: 254767

Ion Ratio Lower Upper

77 100

123 47.6 34.5 51.7

65 14.4 11.1 16.7



#23

Isophorone

Concen: 19.636 ng/ul

RT: 9.534 min Scan# 1096

Delta R.T. -0.006 min

Lab File: BNCCC126.D

Acq: 11 Jun 2022 07:44




Tgt Ion: 82 Resp: 487634


Ion Ratio Lower Upper

82 100

95 7.5 5.8 8.8

138 17.7 13.8 20.8





#25

2-Nitrophenol

Concen: 20.980 ng/ul

RT: 9.722 min Scan# 1128

Delta R.T. -0.006 min

Lab File: BNCCC126.D

Acq: 11 Jun 2022 07:44

Instrument :

BNA_N

ClientSampleId :





Tgt Ion:139 Resp: 133463

Ion Ratio Lower Upper

139 100

65 54.7 43.7 65.5

109 33.3 25.8 38.6



#26

2,4-Dimethylphenol

Concen: 20.497 ng/ul

RT: 9.787 min Scan# 1139

Delta R.T. -0.006 min

Lab File: BNCCC126.D

Acq: 11 Jun 2022 07:44




Tgt Ion:107 Resp: 270125


Ion Ratio Lower Upper

107 100

121 49.8 38.3 57.5

122 85.4 66.3 99.5









#27
 Bis(2-Chloroethoxy)methane
 Concen: 19.842 ng/ul
 RT: 10.016 min Scan# 1178
 Delta R.T. -0.012 min
 Lab File: BNCCC126.D
 Acq: 11 Jun 2022 07:44




Instrument :
 BNA_N
 ClientSampleId :


Tgt Ion: 93 Resp: 301543
 Ion Ratio Lower Upper
 93 100
 95 31.1 25.4 38.0
 123 11.8 8.3 12.5



#29
 2,4-Dichlorophenol
 Concen: 20.979 ng/ul
 RT: 10.257 min Scan# 1219
 Delta R.T. -0.006 min
 Lab File: BNCCC126.D
 Acq: 11 Jun 2022 07:44

Tgt Ion: 162 Resp: 218094
 Ion Ratio Lower Upper
 162 100
 164 62.1 52.9 79.3
 98 38.0 29.8 44.8









#30
Naphthalene
Concen: 20.029 ng/ul
RT: 10.657 min Scan# 1287
Delta R.T. -0.006 min
Lab File: BNCCC126.D
Acq: 11 Jun 2022 07:44




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:128 Resp: 740133
Ion Ratio Lower Upper
128 100
129 10.9 8.7 13.1
127 12.4 10.6 16.0



#32
4-Chloroaniline
Concen: 20.519 ng/ul
RT: 10.763 min Scan# 1305
Delta R.T. -0.006 min
Lab File: BNCCC126.D
Acq: 11 Jun 2022 07:44

Tgt Ion:127 Resp: 334021
Ion Ratio Lower Upper
127 100
129 31.4 26.1 39.1









#33
Hexachlorobutadiene
Concen: 19.861 ng/ul
RT: 10.952 min Scan# 11
Delta R.T. -0.006 min
Lab File: BNCCC126.D
Acq: 11 Jun 2022 07:44




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:225 Resp: 118543
Ion Ratio Lower Upper
225 100
223 64.6 51.8 77.6
227 63.8 50.6 76.0



#34
Caprolactam
Concen: 12.815 ng/ul
RT: 11.510 min Scan# 1432
Delta R.T. -0.006 min
Lab File: BNCCC126.D
Acq: 11 Jun 2022 07:44


Tgt Ion:113 Resp: 52025
Ion Ratio Lower Upper
113 100
55 172.9 145.5 218.3
56 130.7 106.5 159.7








#35
4-Chloro-3-methylphenol
Concen: 20.895 ng/ul
RT: 11.893 min Scan# 1497
Delta R.T. -0.006 min
Lab File: BNCCC126.D
Acq: 11 Jun 2022 07:44


Instrument :
BNA_N
ClientSampleId :





Tgt Ion:107 Resp: 256485
Ion Ratio Lower Upper
107 100
144 25.5 20.6 31.0
142 79.0 62.2 93.4




#36
2-Methylnaphthalene
Concen: 20.269 ng/ul
RT: 12.269 min Scan# 1561
Delta R.T. -0.006 min
Lab File: BNCCC126.D
Acq: 11 Jun 2022 07:44



Tgt Ion:142 Resp: 506387
Ion Ratio Lower Upper
142 100
141 89.1 70.6 105.8









#37
1-Methylnaphthalene
Concen: 20.189 ng/ul
RT: 12.487 min Scan# 1598
Delta R.T. -0.006 min
Lab File: BNCCC126.D
Acq: 11 Jun 2022 07:44




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:142 Resp: 514694
Ion Ratio Lower Upper
142 100
141 93.1 70.9 106.3
116 4.1 3.1 4.7



#39
1,2,4,5-Tetrachlorobenzene
Concen: 19.424 ng/ul
RT: 12.640 min Scan# 1624
Delta R.T. -0.012 min
Lab File: BNCCC126.D
Acq: 11 Jun 2022 07:44

Tgt Ion:216 Resp: 232741
Ion Ratio Lower Upper
216 100
214 79.7 63.4 95.0
179 24.3 17.8 26.8
108 21.4 15.1 22.7





#40

Hexachlorocyclopentadiene

Concen: 17.774 ng/ul

RT: 12.628 min Scan# 1622

Delta R.T. -0.006 min


Lab File: BNCCC126.D

Acq: 11 Jun 2022 07:44

Instrument :

BNA_N

ClientSampleId :






Tgt Ion:237 Resp: 129550

Ion Ratio Lower Upper

237 100

235 62.6 48.8 73.2

272 12.4 9.3 13.9



#41

2,4,6-Trichlorophenol

Concen: 20.204 ng/ul

RT: 12.875 min Scan# 1664

Delta R.T. -0.012 min

Lab File: BNCCC126.D

Acq: 11 Jun 2022 07:44




Tgt Ion:196 Resp: 166922


Ion Ratio Lower Upper

196 100

198 88.7 79.8 119.8

200 29.6 25.0 37.6





#42

2,4,5-Trichlorophenol

Concen: 19.705 ng/ul

RT: 12.951 min Scan# 1

Delta R.T. -0.006 min


Lab File: BNCCC126.D

Acq: 11 Jun 2022 07:44




Instrument :

BNA_N

ClientSampleId :



Tgt Ion:196	Resp:	180305
Ion Ratio	Lower	Upper
196	100	
198	92.5	77.6 116.4
200	30.4	24.4 36.6



#43

1,1'-Biphenyl

Concen: 20.055 ng/ul




RT: 13.281 min Scan# 1733


Delta R.T. -0.006 min

Lab File: BNCCC126.D

Acq: 11 Jun 2022 07:44

Tgt Ion:154	Resp:	697661
Ion Ratio	Lower	Upper
154	100	
153	40.6	31.5 47.3
76	12.6	10.5 15.7





#44

2-Chloronaphthalene

Concen: 19.726 ng/ul

RT: 13.322 min Scan# 11

Delta R.T. -0.006 min


Lab File: BNCCC126.D

Acq: 11 Jun 2022 07:44

Instrument :

BNA_N

ClientSampleId :






Tgt Ion:162 Resp: 520582

Ion Ratio Lower Upper

162 100

164 33.2 25.8 38.6

127 43.0 31.4 47.2



#45

2-Nitroaniline

Concen: 20.598 ng/ul

RT: 13.516 min Scan# 1773

Delta R.T. -0.012 min

Lab File: BNCCC126.D

Acq: 11 Jun 2022 07:44




Tgt Ion: 65 Resp: 169117


Ion Ratio Lower Upper

65 100

92 67.4 54.1 81.1

138 101.9 77.0 115.4









#47
Dimethylphthalate
Concen: 19.622 ng/ul
RT: 13.898 min Scan# 1838
Delta R.T. -0.012 min
Lab File: BNCCC126.D
Acq: 11 Jun 2022 07:44




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:163 Resp: 675235
Ion Ratio Lower Upper
163 100
194 3.8 3.4 5.0
164 9.7 7.7 11.5



#48
2,6-Dinitrotoluene
Concen: 20.652 ng/ul
RT: 14.016 min Scan# 1858
Delta R.T. -0.012 min
Lab File: BNCCC126.D
Acq: 11 Jun 2022 07:44

Tgt Ion:165 Resp: 138950
Ion Ratio Lower Upper
165 100
89 59.0 50.6 75.8
121 22.2 17.9 26.9









#50
Acenaphthylene
Concen: 19.851 ng/ul
RT: 14.163 min Scan# 1883
Delta R.T. -0.006 min
Lab File: BNCCC126.D
Acq: 11 Jun 2022 07:44




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:152 Resp: 871242
Ion Ratio Lower Upper
152 100
151 19.8 16.0 24.0
153 13.1 10.7 16.1



#51
3-Nitroaniline
Concen: 20.608 ng/ul
RT: 14.345 min Scan# 1914
Delta R.T. -0.006 min
Lab File: BNCCC126.D
Acq: 11 Jun 2022 07:44


Tgt Ion:138 Resp: 146585
Ion Ratio Lower Upper
138 100
108 10.7 8.6 12.8
92 113.2 92.9 139.3








#52
Acenaphthene
Concen: 19.537 ng/ul
RT: 14.510 min Scan# 1942
Delta R.T. -0.006 min
Lab File: BNCCC126.D
Acq: 11 Jun 2022 07:44

Instrument :
BNA_N
ClientSampleId :







Tgt Ion:153 Resp: 560099
Ion Ratio Lower Upper
153 100
152 46.7 38.4 57.6
154 88.9 69.7 104.5



#53
2,4-Dinitrophenol
Concen: 13.562 ng/ul
RT: 14.551 min Scan# 1949
Delta R.T. -0.006 min
Lab File: BNCCC126.D
Acq: 11 Jun 2022 07:44

Tgt Ion:184 Resp: 58421
Ion Ratio Lower Upper
184 100
63 80.4 55.9 83.9
154 75.6 53.1 79.7









#55
4-Nitrophenol
Concen: 18.793 ng/ul
RT: 14.651 min Scan# 1966
Delta R.T. -0.006 min
Lab File: BNCCC126.D
Acq: 11 Jun 2022 07:44




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:109 Resp: 109884
Ion Ratio Lower Upper
109 100
139 124.2 100.1 150.1
65 130.1 104.6 157.0



#56
Dibenzofuran
Concen: 19.752 ng/ul
RT: 14.840 min Scan# 1998
Delta R.T. -0.012 min
Lab File: BNCCC126.D
Acq: 11 Jun 2022 07:44


Tgt Ion:168 Resp: 797784
Ion Ratio Lower Upper
168 100
139 40.3 31.1 46.7








#57
2,4-Dinitrotoluene
Concen: 20.748 ng/ul
RT: 14.804 min Scan# 1992
Delta R.T. -0.006 min
Lab File: BNCCC126.D
Acq: 11 Jun 2022 07:44

Instrument :
BNA_N
ClientSampleId :







Tgt Ion:165 Resp: 206695
Ion Ratio Lower Upper
165 100
63 49.1 36.3 54.5
182 2.9 2.3 3.5



#58
2,3,4,6-Tetrachlorophenol
Concen: 20.069 ng/ul
RT: 15.069 min Scan# 2037
Delta R.T. -0.006 min
Lab File: BNCCC126.D
Acq: 11 Jun 2022 07:44


Tgt Ion:232 Resp: 144152
Ion Ratio Lower Upper
232 100
131 52.9 38.5 57.7
130 3.0 2.2 3.2
166 35.5 25.8 38.8








#59
Diethylphthalate
Concen: 19.362 ng/ul
RT: 15.269 min Scan# 2071
Delta R.T. -0.006 min
Lab File: BNCCC126.D
Acq: 11 Jun 2022 07:44

Instrument :
BNA_N
ClientSampleId :







Tgt Ion:149 Resp: 685969
Ion Ratio Lower Upper
149 100
177 21.7 17.8 26.8
150 12.3 10.5 15.7



#61
Fluorene
Concen: 19.925 ng/ul
RT: 15.492 min Scan# 2109
Delta R.T. -0.006 min
Lab File: BNCCC126.D
Acq: 11 Jun 2022 07:44

Tgt Ion:166 Resp: 626215
Ion Ratio Lower Upper
166 100
165 98.3 77.7 116.5
167 13.3 11.1 16.7









#62
4-Chlorophenyl-phenylether
Concen: 19.842 ng/ul
RT: 15.487 min Scan# 2108
Delta R.T. -0.006 min
Lab File: BNCCC126.D
Acq: 11 Jun 2022 07:44




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:204 Resp: 287182
Ion Ratio Lower Upper
204 100
206 33.3 26.7 40.1
141 71.6 50.2 75.2



#63
4-Nitroaniline
Concen: 20.803 ng/ul
RT: 15.504 min Scan# 2111
Delta R.T. -0.006 min
Lab File: BNCCC126.D
Acq: 11 Jun 2022 07:44

Tgt Ion:138 Resp: 138270
Ion Ratio Lower Upper
138 100
92 49.9 41.0 61.6
108 81.4 65.3 97.9









#66
 4,6-Dinitro-2-methylphenol
 Concen: 18.162 ng/ul
 RT: 15.569 min Scan# 2122
 Delta R.T. -0.006 min
 Lab File: BNCCC126.D
 Acq: 11 Jun 2022 07:44




Instrument :
 BNA_N
 ClientSampleId :


Tgt Ion:198 Resp: 104078
 Ion Ratio Lower Upper
 198 100
 182 5.2 3.0 4.4#
 77 28.4 20.1 30.1



#67
 N-Nitrosodiphenylamine
 Concen: 20.122 ng/ul
 RT: 15.698 min Scan# 2144
 Delta R.T. -0.006 min
 Lab File: BNCCC126.D
 Acq: 11 Jun 2022 07:44


Tgt Ion:169 Resp: 562861
 Ion Ratio Lower Upper
 169 100
 168 67.5 52.4 78.6
 167 36.0 27.5 41.3








#68
4-Bromophenyl-phenylether
Concen: 20.056 ng/ul
RT: 16.381 min Scan# 21
Delta R.T. -0.006 min
Lab File: BNCCC126.D
Acq: 11 Jun 2022 07:44

Instrument :
BNA_N
ClientSampleId :







Tgt Ion:248 Resp: 170420
Ion Ratio Lower Upper
248 100
250 91.0 77.8 116.8
141 88.9 62.6 93.8



#69
Hexachlorobenzene
Concen: 19.798 ng/ul
RT: 16.498 min Scan# 2280
Delta R.T. -0.006 min
Lab File: BNCCC126.D
Acq: 11 Jun 2022 07:44

Tgt Ion:284 Resp: 193377
Ion Ratio Lower Upper
284 100
142 47.5 27.8 41.8#
249 35.3 25.7 38.5





#70

Atrazine

Concen: 20.306 ng/ul

RT: 16.651 min Scan# 21

Delta R.T. -0.006 min

Lab File: BNCCC126.D

Acq: 11 Jun 2022 07:44

Instrument :

BNA_N

ClientSampleId :





Tgt Ion:200 Resp: 198188

Ion Ratio Lower Upper

200 100

173 28.8 20.5 30.7

215 49.3 38.9 58.3



#71

Pentachlorophenol

Concen: 17.819 ng/ul

RT: 16.839 min Scan# 2338

Delta R.T. -0.006 min

Lab File: BNCCC126.D

Acq: 11 Jun 2022 07:44




Tgt Ion:266 Resp: 108401


Ion Ratio Lower Upper

266 100

264 61.7 50.2 75.2

268 64.5 49.3 73.9









#72
Phenanthrene
Concen: 19.958 ng/ul
RT: 17.228 min Scan# 2419
Delta R.T. -0.006 min
Lab File: BNCCC126.D
Acq: 11 Jun 2022 07:44




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:178 Resp: 1027713
Ion Ratio Lower Upper
178 100
179 15.2 12.2 18.4
176 18.0 15.2 22.8



#74
Anthracene
Concen: 20.360 ng/ul
RT: 17.316 min Scan# 2419
Delta R.T. -0.012 min
Lab File: BNCCC126.D
Acq: 11 Jun 2022 07:44

Tgt Ion:178 Resp: 1054193
Ion Ratio Lower Upper
178 100
179 16.0 13.0 19.4
176 18.8 15.5 23.3









#75
1,2,3,4-Tetrachlorobenzene
Concen: 19.105 ng/uL
RT: 13.246 min Scan# 1727
Delta R.T. -0.006 min
Lab File: BNCCC126.D
Acq: 11 Jun 2022 07:44




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:216 Resp: 249631
Ion Ratio Lower Upper
216 100
214 76.0 63.2 94.8
218 47.8 37.9 56.9



#76
Pentachlorobenzene
Concen: 20.242 ng/uL
RT: 14.769 min Scan# 1986
Delta R.T. -0.006 min
Lab File: BNCCC126.D
Acq: 11 Jun 2022 07:44

Tgt Ion:250 Resp: 233754
Ion Ratio Lower Upper
250 100
248 66.2 53.0 79.4
252 65.8 53.4 80.2









#77
 Carbazole
 Concen: 20.696 ng/ul
 RT: 17.586 min Scan# 2465
 Delta R.T. -0.006 min
 Lab File: BNCCC126.D
 Acq: 11 Jun 2022 07:44




Instrument :
 BNA_N
 ClientSampleId :


Tgt Ion:167 Resp: 974859
 Ion Ratio Lower Upper
 167 100
 166 22.1 16.1 24.1
 139 13.6 10.2 15.2



#78
 Di-n-butylphthalate
 Concen: 20.354 ng/ul
 RT: 18.157 min Scan# 2562
 Delta R.T. -0.006 min
 Lab File: BNCCC126.D
 Acq: 11 Jun 2022 07:44


Tgt Ion:149 Resp: 1188877
 Ion Ratio Lower Upper
 149 100
 150 8.9 7.1 10.7
 104 5.4 4.1 6.1








#80
Fluoranthene
Concen: 20.663 ng/u1
RT: 19.245 min Scan# 21
Delta R.T. -0.006 min
Lab File: BNCCC126.D
Acq: 11 Jun 2022 07:44

Instrument :
BNA_N
ClientSampleId :







Tgt Ion:202 Resp: 1180601
Ion Ratio Lower Upper
202 100
101 12.0 8.9 13.3
100 9.0 6.7 10.1



#82
Pyrene
Concen: 20.454 ng/u1
RT: 19.604 min Scan# 2808
Delta R.T. -0.012 min
Lab File: BNCCC126.D
Acq: 11 Jun 2022 07:44


Tgt Ion:202 Resp: 1198961
Ion Ratio Lower Upper
202 100
101 14.6 9.7 14.5#
100 12.3 8.0 12.0#








#83
Butylbenzylphthalate
Concen: 20.949 ng/ul
RT: 20.510 min Scan# 2962
Delta R.T. -0.012 min
Lab File: BNCCC126.D
Acq: 11 Jun 2022 07:44

Instrument :
BNA_N
ClientSampleId :







Tgt Ion:149 Resp: 552659
Ion Ratio Lower Upper
149 100
91 75.3 58.8 88.2
206 19.0 18.1 27.1



#84
3,3'-Dichlorobenzidine
Concen: 21.382 ng/ul
RT: 21.292 min Scan# 3095
Delta R.T. -0.006 min
Lab File: BNCCC126.D
Acq: 11 Jun 2022 07:44

Tgt Ion:252 Resp: 356380
Ion Ratio Lower Upper
252 100
254 62.9 50.8 76.2
126 13.1 8.2 12.4#





#85

Benzo(a)anthracene

Concen: 20.400 ng/ul

RT: 21.357 min Scan# 3106

Delta R.T. -0.006 min


Lab File: BNCCC126.D

Acq: 11 Jun 2022 07:44

Instrument :

BNA_N

ClientSampleId :






Tgt Ion:228 Resp: 1124158

Ion Ratio Lower Upper

228 100

229 21.0 16.3 24.5

226 27.5 23.2 34.8



#86

Bis(2-ethylhexyl)phthalate

Concen: 21.716 ng/ul

RT: 21.298 min Scan# 3096

Delta R.T. -0.006 min

Lab File: BNCCC126.D

Acq: 11 Jun 2022 07:44




Tgt Ion:149 Resp: 808314


Ion Ratio Lower Upper

149 100

167 27.7 22.5 33.7

279 4.0 3.5 5.3





#87

Chrysene

Concen: 19.926 ng/ul

RT: 21.410 min Scan# 3115

Delta R.T. -0.012 min

Lab File: BNCCC126.D

Acq: 11 Jun 2022 07:44

Instrument :

BNA_N

ClientSampleId :





Tgt Ion:228 Resp: 1074248

Ion Ratio Lower Upper

228 100

226 30.7 24.2 36.2

229 20.0 16.2 24.2



#89

Di-n-octyl phthalate

Concen: 22.172 ng/ul




RT: 22.198 min Scan# 3249


Delta R.T. -0.012 min

Lab File: BNCCC126.D

Acq: 11 Jun 2022 07:44

Tgt Ion:149 Resp: 1427174









#90
Benzo(b)fluoranthene
Concen: 21.443 ng/ul
RT: 22.992 min Scan# 3384
Delta R.T. -0.012 min
Lab File: BNCCC126.D
Acq: 11 Jun 2022 07:44




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:252 Resp: 1064313
Ion Ratio Lower Upper
252 100
253 21.3 16.5 24.7
125 12.2 7.8 11.8#



#91
Benzo(k)fluoranthene
Concen: 22.525 ng/ul
RT: 22.992 min Scan# 3384
Delta R.T. -0.059 min
Lab File: BNCCC126.D
Acq: 11 Jun 2022 07:44

Tgt Ion:252 Resp: 1064313
Ion Ratio Lower Upper
252 100
253 21.3 18.2 27.4
125 12.2 8.2 12.4





#93

Benzo(a)pyrene

Concen: 20.250 ng/ul

RT: 23.598 min Scan# 3487

Delta R.T. -0.012 min

Lab File: BNCCC126.D

Acq: 11 Jun 2022 07:44

Instrument :

BNA_N

ClientSampleId :





Tgt Ion:252 Resp: 972930

Ion Ratio Lower Upper

252 100

253 21.6 18.7 28.1

125 13.6 9.2 13.8



#94

Indeno(1,2,3-cd)pyrene

Concen: 19.383 ng/ul

RT: 26.086 min Scan# 3910

Delta R.T. -0.012 min

Lab File: BNCCC126.D

Acq: 11 Jun 2022 07:44




Tgt Ion:276 Resp: 958172


Ion Ratio Lower Upper

276 100

138 31.4 19.5 29.3#

277 26.4 21.4 32.0





#95

Dibenzo(a,h)anthracene

Concen: 19.537 ng/ul

RT: 26.098 min Scan# 3914

Delta R.T. -0.023 min

Lab File: BNCCC126.D

Acq: 11 Jun 2022 07:44

Instrument :

BNA_N

ClientSampleId :





Tgt Ion:278 Resp: 824899

Ion Ratio Lower Upper

278 100

139 20.2 13.8 20.8

279 21.4 18.3 27.5



#96

Benzo(g,h,i)perylene

Concen: 19.077 ng/ul

RT: 26.821 min Scan# 4035

Delta R.T. -0.023 min

Lab File: BNCCC126.D

Acq: 11 Jun 2022 07:44

Tgt Ion:276 Resp: 796359

Ion Ratio Lower Upper

276 100

138 25.8 17.0 25.6#

277 21.9 17.8 26.6

