

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061122\
 Data File : BN020106.D
 Acq On : 10 Jun 2022 18:32
 Operator : CG/JU
 Sample : N3250-14
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 EXYS7

Quant Time: Jun 10 22:46:21 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN060622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jun 06 15:14:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.816	152	141015	20.000	ng/u1	0.00
20) Naphthalene-d8	10.610	136	663563	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.445	164	441877	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.187	188	909188	20.000	ng/u1	0.00
79) Chrysene-d12	21.375	240	847826	20.000	ng/u1	0.00
88) Perylene-d12	23.698	264	717540	20.000	ng/u1	-0.01

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.281	96	13788	4.235	ng/uL	0.00
4) Pyridine-d5	3.693	84	45426	5.105	ng/u1	0.00
7) Phenol-d5	6.987	99	66013	5.657	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.152	67	223667	30.680	ng/u1	0.00
11) 2-Chlorophenol-d4	7.352	132	225048	23.737	ng/u1	0.00
15) 4-Methylphenol-d8	8.522	113	144016	14.266	ng/u1	0.00
21) Nitrobenzene-d5	8.969	128	152714	30.251	ng/u1	0.00
24) 2-Nitrophenol-d4	9.693	143	161465	30.138	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.228	165	254081	27.166	ng/u1	0.00
31) 4-Chloroaniline-d4	10.740	131	342191	22.593	ng/u1	0.00
46) Dimethylphthalate-d6	13.857	166	981143	30.516	ng/u1	0.00
49) Acenaphthylene-d8	14.140	160	1169600	31.368	ng/u1	0.00
54) 4-Nitrophenol-d4	14.640	143	27050	4.199	ng/u1	0.00
60) Fluorene-d10	15.440	176	867568	32.609	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.557	200	161169	30.602	ng/u1	0.00
73) Anthracene-d10	17.287	188	1617004	40.752	ng/u1	0.00
81) Pyrene-d10	19.581	212	1800009	39.163	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.551	264	1450501	41.316	ng/u1	-0.01

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061122\
Data File : BN020106.D
Acq On : 10 Jun 2022 18:32
Operator : CG/JU
Sample : N3250-14
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
EXYS7

Quant Time: Jun 10 22:46:21 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN060622.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Jun 06 15:14:59 2022
Response via : Initial Calibration

