

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061122\
 Data File : BN020099.D
 Acq On : 10 Jun 2022 14:17
 Operator : CG/JU
 Sample : N3250-04MS
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :

BNA_N

ClientSampleId :

EXYW2MSMS

Manual Integrations

APPROVED

Reviewed By :Jagrut
Upadhyay

06/13/2022

Supervised By :mohammad
ahmed

06/14/2022

Quant Time: Jun 10 22:45:35 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN060622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jun 06 15:14:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.822	152	153742	20.000	ng/u1	0.00
20) Naphthalene-d8	10.610	136	716095	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.451	164	450633	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.186	188	910796	20.000	ng/u1	0.00
79) Chrysene-d12	21.380	240	765672	20.000	ng/u1	0.00
88) Perylene-d12	23.710	264	698863	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.281	96	17875	5.036	ng/uL	0.00
4) Pyridine-d5	3.693	84	97812	10.082	ng/u1	0.00
7) Phenol-d5	6.987	99	97623	7.674	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.152	67	275170	34.620	ng/u1	0.00
11) 2-Chlorophenol-d4	7.352	132	282494	27.329	ng/u1	0.00
15) 4-Methylphenol-d8	8.528	113	198678	18.051	ng/u1	0.00
21) Nitrobenzene-d5	8.975	128	196641	36.095	ng/u1	0.00
24) 2-Nitrophenol-d4	9.693	143	208273	36.023	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.234	165	337406	33.429	ng/u1	0.00
31) 4-Chloroaniline-d4	10.746	131	492383	30.124	ng/u1	0.00
46) Dimethylphthalate-d6	13.863	166	1306467	39.845	ng/u1	0.00
49) Acenaphthylene-d8	14.139	160	1449902	38.130	ng/u1	0.00
54) 4-Nitrophenol-d4	14.645	143	54718	8.328	ng/u1	0.00
60) Fluorene-d10	15.439	176	1046610	38.574	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.563	200	213162	40.403	ng/u1	0.00
73) Anthracene-d10	17.286	188	1609769	40.498	ng/u1	0.00
81) Pyrene-d10	19.580	212	1742891	41.989	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.562	264	1425383	41.685	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.316	88	18215m	4.861	ng/uL	
5) Pyridine	3.711	79	110537	10.868	ng/u1	94
6) Benzaldehyde	6.957	77	283383m	47.275	ng/u1	
8) Phenol	7.016	94	109829	8.107	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.246	93	355080	33.490	ng/u1	99
12) 2-Chlorophenol	7.387	128	288302	26.683	ng/u1	99
13) 2-Methylphenol	8.263	108	221363	21.256	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.351	45	546652	32.879	ng/u1	98
16) Acetophenone	8.640	105	571572	34.148	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.628	70	296802	34.712	ng/u1	94
18) 4-Methylphenol	8.593	108	212764	18.511	ng/u1	100
19) Hexachloroethane	8.899	117	135629	31.197	ng/u1	90
22) Nitrobenzene	9.016	77	428165	33.896	ng/u1	95
23) Isophorone	9.540	82	877562	35.186	ng/u1	99
25) 2-Nitrophenol	9.722	139	219457	34.350	ng/u1	97
26) 2,4-Dimethylphenol	9.793	107	353518	26.709	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.022	93	516933	33.869	ng/u1	97
29) 2,4-Dichlorophenol	10.263	162	338390	32.411	ng/u1	98
30) Naphthalene	10.663	128	1245744	33.566	ng/u1	100
32) 4-Chloroaniline	10.769	127	472020	28.872	ng/u1	99
33) Hexachlorobutadiene	10.957	225	189290	31.578	ng/u1	97
34) Caprolactam	11.534	113	19355m	4.747	ng/u1	
35) 4-Chloro-3-methylphenol	11.893	107	393302	31.903	ng/u1	98
36) 2-Methylnaphthalene	12.275	142	869583	34.658	ng/u1	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061122\
 Data File : BN020099.D
 Acq On : 10 Jun 2022 14:17
 Operator : CG/JU
 Sample : N3250-04MS
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :

BNA_N

ClientSampleId :

EXYW2MSMS

Manual Integrations

APPROVED

 Reviewed By :Jagrut
 Upadhyay


06/13/2022

 Supervised By :mohammad
 ahmed

Quant Time: Jun 10 22:45:35 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN060622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jun 06 15:14:59 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.492	142	898493	35.091	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.645	216	401766	34.957	ng/ul	98
40) Hexachlorocyclopentadiene	12.628	237	184542	26.396	ng/ul	98
41) 2,4,6-Trichlorophenol	12.881	196	303360	38.280	ng/ul	95
42) 2,4,5-Trichlorophenol	12.951	196	333255	37.970	ng/ul	97
43) 1,1'-Biphenyl	13.287	154	1193544	35.769	ng/ul	98
44) 2-Chloronaphthalene	13.322	162	916940	36.222	ng/ul	96
45) 2-Nitroaniline	13.522	65	329138	41.794	ng/ul	97
47) Dimethylphthalate	13.910	163	1250889	37.896	ng/ul	98
48) 2,6-Dinitrotoluene	14.028	165	270291	41.881	ng/ul	95
50) Acenaphthylene	14.169	152	1528428	36.305	ng/ul	99
51) 3-Nitroaniline	14.351	138	273733	40.121	ng/ul	96
52) Acenaphthene	14.510	153	1005083	36.550	ng/ul	98
53) 2,4-Dinitrophenol	14.557	184	120961	29.275	ng/ul	91
55) 4-Nitrophenol	14.657	109	46864	8.356	ng/ul	96
56) Dibenzofuran	14.845	168	1422724	36.723	ng/ul	94
57) 2,4-Dinitrotoluene	14.810	165	396348	41.478	ng/ul	96
58) 2,3,4,6-Tetrachlorophenol	15.075	232	265714	38.567	ng/ul#	89
59) Diethylphthalate	15.275	149	1330248	39.144	ng/ul	97
61) Fluorene	15.498	166	1113958	36.951	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.492	204	503922	36.297	ng/ul	94
63) 4-Nitroaniline	15.510	138	280800	44.044	ng/ul	98
66) 4,6-Dinitro-2-methylph...	15.575	198	206340	38.575	ng/ul#	91
67) N-Nitrosodiphenylamine	15.704	169	1018381	39.002	ng/ul	98
68) 4-Bromophenyl-phenylether	16.386	248	326882	41.212	ng/ul	96
69) Hexachlorobenzene	16.504	284	356374	39.087	ng/ul#	88
70) Atrazine	16.657	200	367648	40.354	ng/ul	93
71) Pentachlorophenol	16.845	266	195318	34.395	ng/ul	97
72) Phenanthrene	17.233	178	1881546	39.145	ng/ul	99
74) Anthracene	17.322	178	1880739	38.913	ng/ul	97
75) 1,2,3,4-Tetrachloroben...	13.251	216	443131	36.333	ng/uL	98
76) Pentachlorobenzene	14.775	250	427255	39.636	ng/uL	95
77) Carbazole	17.592	167	1786367	40.628	ng/ul	97
78) Di-n-butylphthalate	18.163	149	2290553	42.011	ng/ul	99
80) Fluoranthene	19.251	202	2126387	42.416	ng/ul	100
82) Pyrene	19.610	202	2124537	41.307	ng/ul#	94
83) Butylbenzylphthalate	20.516	149	1068012	46.141	ng/ul	97
84) 3,3'-Dichlorobenzidine	21.298	252	537320	36.742	ng/ul	98
85) Benzo(a)anthracene	21.363	228	1895057	39.195	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.304	149	1474341	45.144	ng/ul	98
87) Chrysene	21.416	228	1856889	39.256	ng/ul	99
89) Di-n-octyl phthalate	22.204	149	2639049	46.396	ng/ul	100
90) Benzo(b)fluoranthene	23.010	252	1833916	41.811	ng/ul#	97
91) Benzo(k)fluoranthene	23.051	252	1718796	41.165	ng/ul	97
93) Benzo(a)pyrene	23.610	252	1694223	39.904	ng/ul	95
94) Indeno(1,2,3-cd)pyrene	26.098	276	1632522	37.372	ng/ul#	92
95) Dibenzo(a,h)anthracene	26.121	278	1377950	36.931	ng/ul	95
96) Benzo(g,h,i)perylene	26.833	276	1376599	37.317	ng/ul#	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 4.861 ng/uL m
RT: 3.316 min Scan# 39
Delta R.T. 0.000 min
Lab File: BN020099.D
Acq: 10 Jun 2022 14:17

Instrument :
BNA_N
ClientSampleId :
EXYW2MSMS





Tgt Ion: 88 Resp: 18215
Ion Ratio Lower Upper
88 100
43 39.7 22.4 33.6
58 81.9 61.0 91.4

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




06/13/2022
Supervised By :mohammad
ahmed


06/14/2022



#5
Pyridine
Concen: 10.868 ng/uL
RT: 3.711 min Scan# 106
Delta R.T. 0.000 min
Lab File: BN020099.D
Acq: 10 Jun 2022 14:17


Tgt Ion: 79 Resp: 110537
Ion Ratio Lower Upper
79 100
52 86.3 65.0 97.4
51 42.9 31.6 47.4





#6
Benzaldehyde
Concen: 47.275 ng/ul m
RT: 6.957 min Scan# 658
Delta R.T. -0.006 min
Lab File: BN020099.D
Acq: 10 Jun 2022 14:17

Instrument :
BNA_N
ClientSampleId :
EXYW2MSMS





Tgt Ion: 77 Resp: 28338
Ion Ratio Lower Upper
77 100
105 100.3 74.7 112.1
106 94.2 70.2 105.4

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




06/13/2022
Supervised By :mohammad
ahmed


06/14/2022



#8
Phenol
Concen: 8.107 ng/ul
RT: 7.016 min Scan# 668
Delta R.T. 0.000 min
Lab File: BN020099.D
Acq: 10 Jun 2022 14:17

Tgt Ion: 94 Resp: 109829
Ion Ratio Lower Upper
94 100
65 28.4 23.6 35.4
66 40.2 31.2 46.8






#10
Bis(2-Chloroethyl)ether
Concen: 33.490 ng/ul
RT: 7.246 min Scan# 706
Delta R.T. 0.000 min
Lab File: BN020099.D
Acq: 10 Jun 2022 14:17

Instrument :

BNA_N

ClientSampleId :

EXYW2MSMS





Tgt Ion: 93 Resp: 355080
Ion Ratio Lower Upper
93 100
63 78.2 63.4 95.0
95 32.7 26.2 39.2

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




06/13/2022
Supervised By :mohammad
ahmed


06/14/2022



#12
2-Chlorophenol
Concen: 26.683 ng/ul
RT: 7.387 min Scan# 731
Delta R.T. 0.000 min
Lab File: BN020099.D
Acq: 10 Jun 2022 14:17


Tgt Ion:128 Resp: 288302
Ion Ratio Lower Upper
128 100
64 47.9 38.6 57.8
130 31.9 25.0 37.4





#13
2-Methylphenol
Concen: 21.256 ng/ul
RT: 8.263 min Scan# 880
Delta R.T. 0.000 min
Lab File: BN020099.D
Acq: 10 Jun 2022 14:17

Instrument :
BNA_N
ClientSampleId :
EXYW2MSMS





Tgt Ion:108 Resp: 22136
Ion Ratio Lower Upper
108 100
107 89.6 71.5 107.3

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




06/13/2022
Supervised By :mohammad
ahmed


06/14/2022



#14
2,2'-oxybis(1-Chloropropane)
Concen: 32.879 ng/ul
RT: 8.351 min Scan# 895
Delta R.T. -0.006 min
Lab File: BN020099.D
Acq: 10 Jun 2022 14:17


Tgt Ion: 45 Resp: 546652
Ion Ratio Lower Upper
45 100
77 12.5 10.9 16.3
79 10.0 8.6 13.0





#16
Acetophenone
Concen: 34.148 ng/ul
RT: 8.640 min Scan# 943
Delta R.T. 0.000 min
Lab File: BN020099.D
Acq: 10 Jun 2022 14:17

Instrument :
BNA_N
ClientSampleId :
EXYW2MSMS






Tgt Ion:105 Resp: 571572
Ion Ratio Lower Upper
105 100
77 77.2 63.5 95.3
51 29.5 23.5 35.3

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




06/13/2022
Supervised By :mohammad
ahmed


06/14/2022



#17
N-Nitroso-di-n-propylamine
Concen: 34.712 ng/ul
RT: 8.628 min Scan# 942
Delta R.T. 0.000 min
Lab File: BN020099.D
Acq: 10 Jun 2022 14:17


Tgt Ion: 70 Resp: 296802
Ion Ratio Lower Upper
70 100
42 64.3 47.8 71.6
101 10.5 7.8 11.6
130 23.1 15.8 23.8





#18
4-Methylphenol
Concen: 18.511 ng/ul
RT: 8.593 min Scan# 911
Delta R.T. 0.000 min
Lab File: BN020099.D
Acq: 10 Jun 2022 14:17

Instrument :
BNA_N
ClientSampleId :
EXYW2MSMS





Tgt Ion:108 Resp: 212764
Ion Ratio Lower Upper
108 100
107 109.8 88.0 132.0

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




06/13/2022
Supervised By :mohammad
ahmed


06/14/2022



#19
Hexachloroethane
Concen: 31.197 ng/ul
RT: 8.899 min Scan# 988
Delta R.T. -0.006 min
Lab File: BN020099.D
Acq: 10 Jun 2022 14:17


Tgt Ion:117 Resp: 135629
Ion Ratio Lower Upper
117 100
201 91.7 83.2 124.8
199 57.9 50.2 75.4





#22
Nitrobenzene
Concen: 33.896 ng/ul
RT: 9.016 min Scan# 1007
Delta R.T. 0.000 min
Lab File: BN020099.D
Acq: 10 Jun 2022 14:17

Instrument :
BNA_N
ClientSampleId :
EXYW2MSMS






Tgt Ion: 77 Resp: 42816
Ion Ratio Lower Upper
77 100
123 46.7 34.5 51.7
65 14.3 11.1 16.7

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




06/13/2022
Supervised By :mohammad
ahmed


06/14/2022



#23
Isophorone
Concen: 35.186 ng/ul
RT: 9.540 min Scan# 1097
Delta R.T. 0.000 min
Lab File: BN020099.D
Acq: 10 Jun 2022 14:17

Tgt Ion: 82 Resp: 877562
Ion Ratio Lower Upper
82 100
95 7.4 5.8 8.8
138 18.0 13.8 20.8





#25

2-Nitrophenol

Concen: 34.350 ng/ul

RT: 9.722 min Scan# 1140

Delta R.T. -0.006 min

Lab File: BN020099.D

Acq: 10 Jun 2022 14:17

Instrument :

BNA_N

ClientSampleId :

EXYW2MSMS

Tgt Ion:139	Resp: 21945
Ion Ratio	Lower Upper
139 100	
65 51.4	43.7 65.5
109 31.3	25.8 38.6

Manual Integrations





APPROVED

Reviewed By :Jagrut
Upadhyay

06/13/2022

Supervised By :mohammad
ahmed

06/14/2022



#26

2,4-Dimethylphenol

Concen: 26.709 ng/ul




RT: 9.793 min Scan# 1140


Delta R.T. 0.000 min

Lab File: BN020099.D

Acq: 10 Jun 2022 14:17


Tgt Ion:107	Resp: 353518
Ion Ratio	Lower Upper
107 100	
121 49.8	38.3 57.5
122 84.3	66.3 99.5





#27
 Bis(2-Chloroethoxy)methane
 Concen: 33.869 ng/ul
 RT: 10.022 min Scan# 1179
 Delta R.T. -0.006 min
 Lab File: BN020099.D
 Acq: 10 Jun 2022 14:17

Instrument :
 BNA_N
 ClientSampleId :
 EXYW2MSMS






Tgt Ion: 93 Resp: 51693
 Ion Ratio Lower Upper
 93 100
 95 33.3 25.4 38.0
 123 11.9 8.3 12.5

Manual Integrations
 APPROVED

Reviewed By :Jagrut
 Upadhyay




06/13/2022
 Supervised By :mohammad
 ahmed


06/14/2022



#29
 2,4-Dichlorophenol
 Concen: 32.411 ng/ul
 RT: 10.263 min Scan# 1220
 Delta R.T. 0.000 min
 Lab File: BN020099.D
 Acq: 10 Jun 2022 14:17


Tgt Ion:162 Resp: 338390
 Ion Ratio Lower Upper
 162 100
 164 64.1 52.9 79.3
 98 36.8 29.8 44.8





#30
Naphthalene
Concen: 33.566 ng/ul
RT: 10.663 min Scan# 1287
Delta R.T. 0.000 min
Lab File: BN020099.D
Acq: 10 Jun 2022 14:17

Instrument :
BNA_N
ClientSampleId :
EXYW2MSMS





Tgt Ion:128 Resp: 1245744
Ion Ratio Lower Upper
128 100
129 10.7 8.7 13.1
127 13.3 10.6 16.0

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




06/13/2022
Supervised By :mohammad
ahmed


06/14/2022



#32
4-Chloroaniline
Concen: 28.872 ng/ul
RT: 10.769 min Scan# 1306
Delta R.T. 0.000 min
Lab File: BN020099.D
Acq: 10 Jun 2022 14:17

Tgt Ion:127 Resp: 472020
Ion Ratio Lower Upper
127 100
129 32.1 26.1 39.1






#33
Hexachlorobutadiene
Concen: 31.578 ng/ul
RT: 10.957 min Scan# 1338
Delta R.T. 0.000 min
Lab File: BN020099.D
Acq: 10 Jun 2022 14:17

Instrument :

BNA_N

ClientSampleId :

EXYW2MSMS



Tgt Ion:225 Resp: 189290
Ion Ratio Lower Upper
225 100
223 60.6 51.8 77.6
227 63.6 50.6 76.0

Manual Integrations



APPROVED

Reviewed By :Jagrut
Upadhyay

06/13/2022




Supervised By :mohammad
ahmed


06/14/2022



#34
Caprolactam
Concen: 4.747 ng/ul m
RT: 11.534 min Scan# 1436
Delta R.T. 0.018 min
Lab File: BN020099.D
Acq: 10 Jun 2022 14:17


Tgt Ion:113 Resp: 19355
Ion Ratio Lower Upper
113 100
55 184.2 145.5 218.3
56 141.4 106.5 159.7





#35
4-Chloro-3-methylphenol
Concen: 31.903 ng/ul
RT: 11.893 min Scan# 1497
Delta R.T. -0.006 min
Lab File: BN020099.D
Acq: 10 Jun 2022 14:17

Instrument :
BNA_N
ClientSampleId :
EXYW2MSMS






Tgt Ion:107 Resp: 393302
Ion Ratio Lower Upper
107 100
144 24.8 20.6 31.0
142 76.6 62.2 93.4

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




06/13/2022
Supervised By :mohammad
ahmed


06/14/2022



#36
2-Methylnaphthalene
Concen: 34.658 ng/ul
RT: 12.275 min Scan# 1562
Delta R.T. 0.000 min
Lab File: BN020099.D
Acq: 10 Jun 2022 14:17

Tgt Ion:142 Resp: 869583
Ion Ratio Lower Upper
142 100
141 90.0 70.6 105.8






#37
1-Methylnaphthalene
Concen: 35.091 ng/ul
RT: 12.492 min Scan# 1599
Delta R.T. 0.000 min
Lab File: BN020099.D
Acq: 10 Jun 2022 14:17

Instrument :

BNA_N

ClientSampleId :

EXYW2MSMS





Tgt Ion:142 Resp: 89849
Ion Ratio Lower Upper
142 100
141 90.5 70.9 106.3
116 3.8 3.1 4.7

Manual Integrations
APPROVED


Reviewed By :Jagrut
Upadhyay

06/13/2022
Supervised By :mohammad
ahmed



06/14/2022




#39
1,2,4,5-Tetrachlorobenzene
Concen: 34.957 ng/ul
RT: 12.645 min Scan# 1625
Delta R.T. -0.006 min
Lab File: BN020099.D
Acq: 10 Jun 2022 14:17




Tgt Ion:216 Resp: 401766
Ion Ratio Lower Upper
216 100
214 77.1 63.4 95.0
179 23.1 17.8 26.8
108 20.4 15.1 22.7





#40
Hexachlorocyclopentadiene
Concen: 26.396 ng/ul
RT: 12.628 min Scan# 1622
Delta R.T. -0.006 min
Lab File: BN020099.D
Acq: 10 Jun 2022 14:17

Instrument :
BNA_N
ClientSampleId :
EXYW2MSMS





Tgt Ion:237 Resp: 184542
Ion Ratio Lower Upper
237 100
235 62.5 48.8 73.2
272 10.9 9.3 13.9

Manual Integrations
APPROVED


Reviewed By :Jagrut
Upadhyay

06/13/2022
Supervised By :mohammad
ahmed



06/14/2022




#41
2,4,6-Trichlorophenol
Concen: 38.280 ng/ul
RT: 12.881 min Scan# 1665
Delta R.T. -0.006 min
Lab File: BN020099.D
Acq: 10 Jun 2022 14:17



Tgt Ion:196 Resp: 303360
Ion Ratio Lower Upper
196 100
198 94.1 79.8 119.8
200 29.1 25.0 37.6






#42
2,4,5-Trichlorophenol
Concen: 37.970 ng/ul
RT: 12.951 min Scan# 1677
Delta R.T. -0.006 min
Lab File: BN020099.D
Acq: 10 Jun 2022 14:17

Instrument :

BNA_N

ClientSampleId :

EXYW2MSMS



Tgt Ion:196 Resp: 33325
Ion Ratio Lower Upper
196 100
198 100.7 77.6 116.4
200 31.1 24.4 36.6

Manual Integrations




APPROVED

Reviewed By :Jagrut
Upadhyay

06/13/2022




Supervised By :mohammad
ahmed


06/14/2022



#43
1,1'-Biphenyl
Concen: 35.769 ng/ul
RT: 13.287 min Scan# 1734
Delta R.T. 0.000 min
Lab File: BN020099.D
Acq: 10 Jun 2022 14:17


Tgt Ion:154 Resp: 1193544
Ion Ratio Lower Upper
154 100
153 38.3 31.5 47.3
76 12.1 10.5 15.7





#44
2-Chloronaphthalene
Concen: 36.222 ng/ul
RT: 13.322 min Scan# 1740
Delta R.T. -0.006 min
Lab File: BN020099.D
Acq: 10 Jun 2022 14:17

Instrument :
BNA_N
ClientSampleId :
EXYW2MSMS





Tgt Ion:162 Resp: 916940
Ion Ratio Lower Upper
162 100
164 33.6 25.8 38.6
127 42.9 31.4 47.2

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




06/13/2022
Supervised By :mohammad
ahmed


06/14/2022



#45
2-Nitroaniline
Concen: 41.794 ng/ul
RT: 13.522 min Scan# 1774
Delta R.T. -0.006 min
Lab File: BN020099.D
Acq: 10 Jun 2022 14:17

Tgt Ion: 65 Resp: 329138
Ion Ratio Lower Upper
65 100
92 67.1 54.1 81.1
138 100.7 77.0 115.4





#47

Dimethylphthalate

Concen: 37.896 ng/ul

RT: 13.910 min Scan# 1839

Delta R.T. 0.000 min

Lab File: BN020099.D


Acq: 10 Jun 2022 14:17

Instrument :

BNA_N

ClientSampleId :

EXYW2MSMS



Tgt Ion:163 Resp: 1250889

Ion Ratio Lower Upper

163 100

194 3.9 3.4 5.0

164 10.5 7.7 11.5

Manual Integrations




APPROVED

Reviewed By :Jagrut
Upadhyay

06/13/2022

Supervised By :mohammad
ahmed

06/14/2022



#48

2,6-Dinitrotoluene

Concen: 41.881 ng/ul

RT: 14.028 min Scan# 1860

Delta R.T. 0.000 min

Lab File: BN020099.D

Acq: 10 Jun 2022 14:17




Tgt Ion:165 Resp: 270291


Ion Ratio Lower Upper

165 100

89 58.1 50.6 75.8


121 21.4 17.9 26.9





#50
Acenaphthylene
Concen: 36.305 ng/ul
RT: 14.169 min Scan# 1884
Delta R.T. 0.000 min
Lab File: BN020099.D
Acq: 10 Jun 2022 14:17

Instrument :
BNA_N
ClientSampleId :
EXYW2MSMS






Tgt Ion:152 Resp: 1528428
Ion Ratio Lower Upper
152 100
151 19.7 16.0 24.0
153 12.7 10.7 16.1

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




06/13/2022
Supervised By :mohammad
ahmed


06/14/2022



#51
3-Nitroaniline
Concen: 40.121 ng/ul
RT: 14.351 min Scan# 1915
Delta R.T. 0.000 min
Lab File: BN020099.D
Acq: 10 Jun 2022 14:17


Tgt Ion:138 Resp: 273733
Ion Ratio Lower Upper
138 100
108 10.8 8.6 12.8
92 110.8 92.9 139.3





#52
Acenaphthene
Concen: 36.550 ng/ul
RT: 14.510 min Scan# 1942
Delta R.T. -0.006 min
Lab File: BN020099.D
Acq: 10 Jun 2022 14:17

Instrument :
BNA_N
ClientSampleId :
EXYW2MSMS






Tgt Ion:153 Resp: 100508
Ion Ratio Lower Upper
153 100
152 44.9 38.4 57.6
154 86.8 69.7 104.5

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




06/13/2022
Supervised By :mohammad
ahmed


06/14/2022



#53
2,4-Dinitrophenol
Concen: 29.275 ng/ul
RT: 14.557 min Scan# 1950
Delta R.T. 0.000 min
Lab File: BN020099.D
Acq: 10 Jun 2022 14:17


Tgt Ion:184 Resp: 120961
Ion Ratio Lower Upper
184 100
63 77.9 55.9 83.9
154 73.2 53.1 79.7





#55
4-Nitrophenol
Concen: 8.356 ng/ul
RT: 14.657 min Scan# 1967
Delta R.T. 0.000 min
Lab File: BN020099.D
Acq: 10 Jun 2022 14:17

Instrument :
BNA_N
ClientSampleId :
EXYW2MSMS






Tgt Ion:109 Resp: 46864
Ion Ratio Lower Upper
109 100
139 130.0 100.1 150.1
65 135.9 104.6 157.0

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




06/13/2022
Supervised By :mohammad
ahmed


06/14/2022



#56
Dibenzofuran
Concen: 36.723 ng/ul
RT: 14.845 min Scan# 1999
Delta R.T. -0.006 min
Lab File: BN020099.D
Acq: 10 Jun 2022 14:17


Tgt Ion:168 Resp: 1422724
Ion Ratio Lower Upper
168 100
139 42.5 31.1 46.7





#57
2,4-Dinitrotoluene
Concen: 41.478 ng/ul
RT: 14.810 min Scan# 1992
Delta R.T. 0.000 min
Lab File: BN020099.D
Acq: 10 Jun 2022 14:17

Instrument :
BNA_N
ClientSampleId :
EXYW2MSMS





Tgt Ion:165 Resp: 396348
Ion Ratio Lower Upper
165 100
63 48.0 36.3 54.5
182 2.9 2.3 3.5

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




06/13/2022
Supervised By :mohammad
ahmed


06/14/2022



#58
2,3,4,6-Tetrachlorophenol
Concen: 38.567 ng/ul
RT: 15.075 min Scan# 2038
Delta R.T. 0.000 min
Lab File: BN020099.D
Acq: 10 Jun 2022 14:17

Tgt Ion:232 Resp: 265714
Ion Ratio Lower Upper
232 100
131 56.6 38.5 57.7
130 3.4 2.2 3.2#
166 38.0 25.8 38.8






#59
Diethylphthalate
Concen: 39.144 ng/ul
RT: 15.275 min Scan# 2071
Delta R.T. 0.000 min
Lab File: BN020099.D
Acq: 10 Jun 2022 14:17

Instrument :

BNA_N

ClientSampleId :

EXYW2MSMS






Tgt Ion:149 Resp: 1330248
Ion Ratio Lower Upper
149 100
177 20.4 17.8 26.8
150 12.6 10.5 15.7

Manual Integrations
APPROVED


Reviewed By :Jagrut
Upadhyay

06/13/2022
Supervised By :mohammad
ahmed



06/14/2022




#61
Fluorene
Concen: 36.951 ng/ul
RT: 15.498 min Scan# 2110
Delta R.T. 0.000 min
Lab File: BN020099.D
Acq: 10 Jun 2022 14:17



Tgt Ion:166 Resp: 1113958
Ion Ratio Lower Upper
166 100
165 96.1 77.7 116.5
167 13.4 11.1 16.7





#62
4-Chlorophenyl-phenylether
Concen: 36.297 ng/ul
RT: 15.492 min Scan# 2109
Delta R.T. 0.000 min
Lab File: BN020099.D
Acq: 10 Jun 2022 14:17

Instrument :

BNA_N

ClientSampleId :

EXYW2MSMS





Tgt Ion:204 Resp: 50392
Ion Ratio Lower Upper
204 100
206 31.0 26.7 40.1
141 68.1 50.2 75.2

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




06/13/2022
Supervised By :mohammad
ahmed


06/14/2022



#63
4-Nitroaniline
Concen: 44.044 ng/ul
RT: 15.510 min Scan# 2112
Delta R.T. 0.000 min
Lab File: BN020099.D
Acq: 10 Jun 2022 14:17


Tgt Ion:138 Resp: 280800
Ion Ratio Lower Upper
138 100
92 52.9 41.0 61.6
108 83.1 65.3 97.9





#66
4,6-Dinitro-2-methylphenol
Concen: 38.575 ng/ul
RT: 15.575 min Scan# 2122
Delta R.T. 0.000 min
Lab File: BN020099.D
Acq: 10 Jun 2022 14:17

Instrument :
BNA_N
ClientSampleId :
EXYW2MSMS






Tgt Ion:198 Resp: 206340
Ion Ratio Lower Upper
198 100
182 4.3 3.0 4.4
77 30.4 20.1 30.1

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




06/13/2022
Supervised By :mohammad
ahmed


06/14/2022



#67
N-Nitrosodiphenylamine
Concen: 39.002 ng/ul
RT: 15.704 min Scan# 2145
Delta R.T. 0.000 min
Lab File: BN020099.D
Acq: 10 Jun 2022 14:17


Tgt Ion:169 Resp: 1018381
Ion Ratio Lower Upper
169 100
168 65.9 52.4 78.6
167 36.7 27.5 41.3





#68
4-Bromophenyl-phenylether
Concen: 41.212 ng/ul
RT: 16.386 min Scan# 2161
Delta R.T. 0.000 min
Lab File: BN020099.D
Acq: 10 Jun 2022 14:17

Instrument :
BNA_N
ClientSampleId :
EXYW2MSMS





Tgt Ion:248 Resp: 326882
Ion Ratio Lower Upper
248 100
250 98.5 77.8 116.8
141 85.3 62.6 93.8

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




06/13/2022
Supervised By :mohammad
ahmed


06/14/2022



#69
Hexachlorobenzene
Concen: 39.087 ng/ul
RT: 16.504 min Scan# 2281
Delta R.T. 0.000 min
Lab File: BN020099.D
Acq: 10 Jun 2022 14:17


Tgt Ion:284 Resp: 356374
Ion Ratio Lower Upper
284 100
142 46.3 27.8 41.8#
249 34.3 25.7 38.5





#70
Atrazine
Concen: 40.354 ng/ul
RT: 16.657 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN020099.D
Acq: 10 Jun 2022 14:17

Instrument :
BNA_N
ClientSampleId :
EXYW2MSMS






Tgt Ion:200 Resp: 367648
Ion Ratio Lower Upper
200 100
173 28.6 20.5 30.7
215 43.7 38.9 58.3

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




06/13/2022
Supervised By :mohammad
ahmed


06/14/2022



#71
Pentachlorophenol
Concen: 34.395 ng/ul
RT: 16.845 min Scan# 2339
Delta R.T. 0.000 min
Lab File: BN020099.D
Acq: 10 Jun 2022 14:17


Tgt Ion:266 Resp: 195318
Ion Ratio Lower Upper
266 100
264 62.1 50.2 75.2
268 65.0 49.3 73.9





#72
Phenanthrene
Concen: 39.145 ng/ul
RT: 17.233 min Scan# 2405
Delta R.T. 0.000 min
Lab File: BN020099.D
Acq: 10 Jun 2022 14:17

Instrument :
BNA_N
ClientSampleId :
EXYW2MSMS





Tgt Ion:178 Resp: 1881540
Ion Ratio Lower Upper
178 100
179 15.1 12.2 18.4
176 18.4 15.2 22.8

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




06/13/2022
Supervised By :mohammad
ahmed


06/14/2022



#74
Anthracene
Concen: 38.913 ng/ul
RT: 17.322 min Scan# 2420
Delta R.T. -0.006 min
Lab File: BN020099.D
Acq: 10 Jun 2022 14:17


Tgt Ion:178 Resp: 1880739
Ion Ratio Lower Upper
178 100
179 15.1 13.0 19.4
176 18.2 15.5 23.3





#75
1,2,3,4-Tetrachlorobenzene
Concen: 36.333 ng/uL
RT: 13.251 min Scan# 1728
Delta R.T. 0.000 min
Lab File: BN020099.D
Acq: 10 Jun 2022 14:17

Instrument :
BNA_N
ClientSampleId :
EXYW2MSMS





Tgt Ion:216 Resp: 44313
Ion Ratio Lower Upper
216 100
214 76.9 63.2 94.8
218 48.6 37.9 56.9

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




06/13/2022
Supervised By :mohammad
ahmed


06/14/2022



#76
Pentachlorobenzene
Concen: 39.636 ng/uL
RT: 14.775 min Scan# 1987
Delta R.T. 0.000 min
Lab File: BN020099.D
Acq: 10 Jun 2022 14:17


Tgt Ion:250 Resp: 427255
Ion Ratio Lower Upper
250 100
248 62.1 53.0 79.4
252 62.7 53.4 80.2





#77
 Carbazole
 Concen: 40.628 ng/ul
 RT: 17.592 min Scan# 2466
 Delta R.T. 0.000 min
 Lab File: BN020099.D
 Acq: 10 Jun 2022 14:17

Instrument :
 BNA_N
 ClientSampleId :
 EXYW2MSMS






Tgt Ion:167 Resp: 178636
 Ion Ratio Lower Upper
 167 100
 166 21.3 16.1 24.1
 139 13.7 10.2 15.2

Manual Integrations
 APPROVED

Reviewed By :Jagrut
 Upadhyay




06/13/2022
 Supervised By :mohammad
 ahmed


06/14/2022



#78
 Di-n-butylphthalate
 Concen: 42.011 ng/ul
 RT: 18.163 min Scan# 2563
 Delta R.T. 0.000 min
 Lab File: BN020099.D
 Acq: 10 Jun 2022 14:17


Tgt Ion:149 Resp: 2290553
 Ion Ratio Lower Upper
 149 100
 150 9.4 7.1 10.7
 104 5.4 4.1 6.1





#80
Fluoranthene
Concen: 42.416 ng/u1
RT: 19.251 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN020099.D
Acq: 10 Jun 2022 14:17

Instrument :
BNA_N
ClientSampleId :
EXYW2MSMS





Tgt Ion:202 Resp: 212638
Ion Ratio Lower Upper
202 100
101 11.3 8.9 13.3
100 8.6 6.7 10.1

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




06/13/2022
Supervised By :mohammad
ahmed


06/14/2022



#82
Pyrene
Concen: 41.307 ng/u1
RT: 19.610 min Scan# 2809
Delta R.T. -0.006 min
Lab File: BN020099.D
Acq: 10 Jun 2022 14:17


Tgt Ion:202 Resp: 2124537
Ion Ratio Lower Upper
202 100
101 14.7 9.7 14.5#
100 12.1 8.0 12.0#





#83
Butylbenzylphthalate
Concen: 46.141 ng/ul
RT: 20.516 min Scan# 2963
Delta R.T. -0.006 min
Lab File: BN020099.D
Acq: 10 Jun 2022 14:17

Instrument :
BNA_N
ClientSampleId :
EXYW2MSMS






Tgt Ion:149 Resp: 1068012
Ion Ratio Lower Upper
149 100
91 71.7 58.8 88.2
206 19.0 18.1 27.1

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




06/13/2022
Supervised By :mohammad
ahmed


06/14/2022



#84
3,3'-Dichlorobenzidine
Concen: 36.742 ng/ul
RT: 21.298 min Scan# 3096
Delta R.T. 0.000 min
Lab File: BN020099.D
Acq: 10 Jun 2022 14:17


Tgt Ion:252 Resp: 537320
Ion Ratio Lower Upper
252 100
254 64.8 50.8 76.2
126 12.3 8.2 12.4





#85
Benzo(a)anthracene
Concen: 39.195 ng/ul
RT: 21.363 min Scan# 3107
Delta R.T. 0.000 min
Lab File: BN020099.D
Acq: 10 Jun 2022 14:17

Instrument :
BNA_N
ClientSampleId :
EXYW2MSMS






Tgt Ion:228 Resp: 189505
Ion Ratio Lower Upper
228 100
229 19.4 16.3 24.5
226 27.6 23.2 34.8

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




06/13/2022
Supervised By :mohammad
ahmed


06/14/2022



#86
Bis(2-ethylhexyl)phthalate
Concen: 45.144 ng/ul
RT: 21.304 min Scan# 3097
Delta R.T. 0.000 min
Lab File: BN020099.D
Acq: 10 Jun 2022 14:17


Tgt Ion:149 Resp: 1474341
Ion Ratio Lower Upper
149 100
167 27.0 22.5 33.7
279 3.8 3.5 5.3





#87
Chrysene
Concen: 39.256 ng/ul
RT: 21.416 min Scan# 3116
Delta R.T. -0.006 min
Lab File: BN020099.D
Acq: 10 Jun 2022 14:17

Instrument :
BNA_N
ClientSampleId :
EXYW2MSMS






Tgt Ion:228 Resp: 1856889
Ion Ratio Lower Upper
228 100
226 30.0 24.2 36.2
229 19.1 16.2 24.2

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




06/13/2022
Supervised By :mohammad
ahmed


06/14/2022



#89
Di-n-octyl phthalate
Concen: 46.396 ng/ul
RT: 22.204 min Scan# 3250
Delta R.T. -0.006 min
Lab File: BN020099.D
Acq: 10 Jun 2022 14:17


Tgt Ion:149 Resp: 2639049





#90
Benzo(b)fluoranthene
Concen: 41.811 ng/ul
RT: 23.010 min Scan# 3386
Delta R.T. 0.006 min
Lab File: BN020099.D
Acq: 10 Jun 2022 14:17

Instrument :
BNA_N
ClientSampleId :
EXYW2MSMS






Tgt Ion:252 Resp: 1833910
Ion Ratio Lower Upper
252 100
253 21.3 16.5 24.7
125 11.8 7.8 11.8

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




06/13/2022
Supervised By :mohammad
ahmed


06/14/2022



#91
Benzo(k)fluoranthene
Concen: 41.165 ng/ul
RT: 23.051 min Scan# 3394
Delta R.T. 0.000 min
Lab File: BN020099.D
Acq: 10 Jun 2022 14:17

Tgt Ion:252 Resp: 1718796
Ion Ratio Lower Upper
252 100
253 21.7 18.2 27.4
125 12.1 8.2 12.4






#93
Benzo(a)pyrene
Concen: 39.904 ng/ul
RT: 23.610 min Scan# 3488
Delta R.T. 0.000 min
Lab File: BN020099.D
Acq: 10 Jun 2022 14:17

Instrument :

BNA_N

ClientSampleId :

EXYW2MSMS



Tgt Ion:252 Resp: 169422
Ion Ratio Lower Upper
252 100
253 20.7 18.7 28.1
125 12.7 9.2 13.8

Manual Integrations




APPROVED

Reviewed By :Jagrut
Upadhyay

06/13/2022




Supervised By :mohammad
ahmed


06/14/2022



#94
Indeno(1,2,3-cd)pyrene
Concen: 37.372 ng/ul
RT: 26.098 min Scan# 3912
Delta R.T. 0.000 min
Lab File: BN020099.D
Acq: 10 Jun 2022 14:17


Tgt Ion:276 Resp: 163252
Ion Ratio Lower Upper
276 100
138 29.9 19.5 29.3#
277 23.7 21.4 32.0





#95
Dibenzo(a,h)anthracene
Concen: 36.931 ng/ul
RT: 26.121 min Scan# 3916
Delta R.T. 0.000 min
Lab File: BN020099.D
Acq: 10 Jun 2022 14:17

Instrument :
BNA_N
ClientSampleId :
EXYW2MSMS






Tgt Ion:278 Resp: 1377950
Ion Ratio Lower Upper
278 100
139 20.1 13.8 20.8
279 24.9 18.3 27.5

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay

06/13/2022
Supervised By :mohammad
ahmed

06/14/2022



#96
Benzo(g,h,i)perylene
Concen: 37.317 ng/ul
RT: 26.833 min Scan# 4037
Delta R.T. -0.012 min
Lab File: BN020099.D
Acq: 10 Jun 2022 14:17

Tgt Ion:276 Resp: 1376599
Ion Ratio Lower Upper
276 100
138 26.4 17.0 25.6#
277 25.0 17.8 26.6

