

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061122\
 Data File : BN020123.D
 Acq On : 11 Jun 2022 05:21
 Operator : CG/JU
 Sample : PB145445BS
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SLCS445

Quant Time: Jun 11 05:55:09 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN060622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jun 06 15:14:59 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 06/13/2022
 Supervised By :mohammad ahmed 06/14/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.816	152	157514	20.000	ng/u1	0.00
20) Naphthalene-d8	10.610	136	740165	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.445	164	476835	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.187	188	962517	20.000	ng/u1	0.00
79) Chrysene-d12	21.375	240	782412	20.000	ng/u1	0.00
88) Perylene-d12	23.698	264	727033	20.000	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.281	96	25591	7.037	ng/uL	0.00
4) Pyridine-d5	3.687	84	355125	35.727	ng/u1	0.00
7) Phenol-d5	6.987	99	507574	38.943	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.152	67	305965	37.573	ng/u1	0.00
11) 2-Chlorophenol-d4	7.352	132	413497	39.045	ng/u1	0.00
15) 4-Methylphenol-d8	8.528	113	441360	39.140	ng/u1	0.00
21) Nitrobenzene-d5	8.969	128	217810	38.680	ng/u1	0.00
24) 2-Nitrophenol-d4	9.693	143	239721	40.114	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.234	165	403628	38.689	ng/u1	0.00
31) 4-Chloroaniline-d4	10.740	131	627390	37.136	ng/u1	0.00
46) Dimethylphthalate-d6	13.857	166	1293026	37.268	ng/u1	0.00
49) Acenaphthylene-d8	14.134	160	1489956	37.031	ng/u1	-0.01
54) 4-Nitrophenol-d4	14.645	143	249180	35.841	ng/u1	0.00
60) Fluorene-d10	15.440	176	1047912	36.499	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.563	200	211519	37.937	ng/u1	0.00
73) Anthracene-d10	17.287	188	1575578	37.508	ng/u1	0.00
81) Pyrene-d10	19.581	212	1717480	40.491	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.551	264	1361375	38.271	ng/u1	-0.01
Target Compounds					Qvalue	
2) 1,4-Dioxane	3.317	88	51163	13.327	ng/uL#	88
5) Pyridine	3.705	79	368071	35.322	ng/u1	95
6) Benzaldehyde	6.958	77	297177	48.389	ng/u1	91
8) Phenol	7.016	94	531602	38.303	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.240	93	401686	36.978	ng/u1	97
12) 2-Chlorophenol	7.381	128	421683	38.093	ng/u1	98
13) 2-Methylphenol	8.258	108	406185	38.069	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.346	45	626591	36.785	ng/u1	97
16) Acetophenone	8.634	105	642864	37.487	ng/u1	96
17) N-Nitroso-di-n-propyla...	8.628	70	329948	37.664	ng/u1	95
18) 4-Methylphenol	8.593	108	449816	38.197	ng/u1	98
19) Hexachloroethane	8.893	117	164524	36.937	ng/u1	84
22) Nitrobenzene	9.010	77	481803	36.902	ng/u1	94
23) Isophorone	9.540	82	950904	36.887	ng/u1	99
25) 2-Nitrophenol	9.722	139	251966	38.156	ng/u1	98
26) 2,4-Dimethylphenol	9.787	107	483272	35.325	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.022	93	570488	36.163	ng/u1	99
29) 2,4-Dichlorophenol	10.257	162	412453	38.220	ng/u1	98
30) Naphthalene	10.657	128	1365227	35.590	ng/u1	100
32) 4-Chloroaniline	10.763	127	589239	34.870	ng/u1	100
33) Hexachlorobutadiene	10.952	225	223483	36.070	ng/u1	99
34) Caprolactam	11.522	113	157770m	37.438	ng/u1	
35) 4-Chloro-3-methylphenol	11.899	107	495382	38.876	ng/u1	100
36) 2-Methylnaphthalene	12.269	142	945813	36.470	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061122\
 Data File : BN020123.D
 Acq On : 11 Jun 2022 05:21
 Operator : CG/JU
 Sample : PB145445BS
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SLCS445

Quant Time: Jun 11 05:55:09 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN060622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jun 06 15:14:59 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 06/13/2022
 Supervised By :mohammad ahmed 06/14/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.487	142	966348	36.514	ng/ul	94
39) 1,2,4,5-Tetrachloroben...	12.646	216	438492	36.056	ng/ul	99
40) Hexachlorocyclopentadiene	12.628	237	241944	32.704	ng/ul	98
41) 2,4,6-Trichlorophenol	12.881	196	319617	38.115	ng/ul	97
42) 2,4,5-Trichlorophenol	12.951	196	352308	37.935	ng/ul	100
43) 1,1'-Biphenyl	13.281	154	1272714	36.046	ng/ul	99
44) 2-Chloronaphthalene	13.322	162	967711	36.127	ng/ul	96
45) 2-Nitroaniline	13.522	65	329397	39.528	ng/ul	95
47) Dimethylphthalate	13.904	163	1253557	35.890	ng/ul	98
48) 2,6-Dinitrotoluene	14.022	165	271185	39.711	ng/ul	93
50) Acenaphthylene	14.169	152	1607622	36.088	ng/ul	99
51) 3-Nitroaniline	14.345	138	286966	39.749	ng/ul	99
52) Acenaphthene	14.510	153	1023257	35.166	ng/ul	98
53) 2,4-Dinitrophenol	14.551	184	137070	31.350	ng/ul	95
55) 4-Nitrophenol	14.663	109	209304	35.269	ng/ul	96
56) Dibenzofuran	14.845	168	1446186	35.277	ng/ul	97
57) 2,4-Dinitrotoluene	14.810	165	391264	38.696	ng/ul	98
58) 2,3,4,6-Tetrachlorophenol	15.075	232	271454	37.235	ng/ul	93
59) Diethylphthalate	15.275	149	1294507	35.999	ng/ul	98
61) Fluorene	15.493	166	1107984	34.734	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.487	204	514028	34.991	ng/ul	91
63) 4-Nitroaniline	15.510	138	281384m	41.710	ng/ul	
66) 4,6-Dinitro-2-methylph...	15.575	198	209341	37.034	ng/ul	94
67) N-Nitrosodiphenylamine	15.698	169	1028008	37.255	ng/ul	99
68) 4-Bromophenyl-phenylether	16.381	248	314710	37.545	ng/ul#	89
69) Hexachlorobenzene	16.504	284	350954	36.424	ng/ul#	92
70) Atrazine	16.651	200	359604	37.351	ng/ul	96
71) Pentachlorophenol	16.839	266	207559	34.587	ng/ul	91
72) Phenanthrene	17.228	178	1837362	36.172	ng/ul	99
74) Anthracene	17.322	178	1847247	36.167	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.246	216	474425	36.809	ng/ul	99
76) Pentachlorobenzene	14.769	250	419555	36.830	ng/ul	98
77) Carbazole	17.587	167	1755710	37.785	ng/ul	97
78) Di-n-butylphthalate	18.157	149	2238267	38.846	ng/ul	99
80) Fluoranthene	19.245	202	2072161	40.450	ng/ul	98
82) Pyrene	19.610	202	2090281	39.772	ng/ul	97
83) Butylbenzylphthalate	20.510	149	1021118	43.171	ng/ul#	96
84) 3,3'-Dichlorobenzidine	21.292	252	572590	38.316	ng/ul#	95
85) Benzo(a)anthracene	21.357	228	1888378	38.221	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.298	149	1433082	42.942	ng/ul	99
87) Chrysene	21.410	228	1815080	37.551	ng/ul	99
89) Di-n-octyl phthalate	22.198	149	2519443	42.577	ng/ul	100
90) Benzo(b)fluoranthene	22.998	252	1863115m	40.831	ng/ul	
91) Benzo(k)fluoranthene	23.045	252	1618258	37.255	ng/ul	97
93) Benzo(a)pyrene	23.598	252	1665376	37.705	ng/ul	94
94) Indeno(1,2,3-cd)pyrene	26.098	276	1610167	35.432	ng/ul#	92
95) Dibenzo(a,h)anthracene	26.104	278	1365237	35.173	ng/ul	95
96) Benzo(g,h,i)perylene	26.827	276	1359920	35.437	ng/ul#	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061122\
 Data File : BN020123.D
 Acq On : 11 Jun 2022 05:21
 Operator : CG/JU
 Sample : PB145445BS
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :

BNA_N

ClientSampleId :

SLCS445

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 06/13/2022
 Supervised By :mohammad ahmed 06/14/2022


Quant Time: Jun 11 05:55:09 2022


Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN060622.M

Quant Title : SVOA CALIBRATION

QLast Update : Mon Jun 06 15:14:59 2022


Response via : Initial Calibration





#2
1,4-Dioxane
Concen: 13.327 ng/uL
RT: 3.317 min Scan# 39
Delta R.T. 0.001 min
Lab File: BN020123.D
Acq: 11 Jun 2022 05:21




Instrument :
BNA_N
ClientSampleId :
SLCS445



Tgt Ion: 88 Resp: 5116
Ion Ratio Lower Upper
88 100
43 34.7 22.4 33.6
58 86.3 61.0 91.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 06/13/2022
Supervised By :mohammad ahmed 06/14/2022



#5
Pyridine
Concen: 35.322 ng/uL
RT: 3.705 min Scan# 105
Delta R.T. -0.006 min
Lab File: BN020123.D
Acq: 11 Jun 2022 05:21


Tgt Ion: 79 Resp: 368071
Ion Ratio Lower Upper
79 100
52 87.2 65.0 97.4
51 40.7 31.6 47.4





#6
Benzaldehyde
Concen: 48.389 ng/ul
RT: 6.958 min Scan# 658
Delta R.T. -0.006 min
Lab File: BN020123.D
Acq: 11 Jun 2022 05:21




Instrument :
BNA_N
ClientSampleId :
SLCS445



Tgt Ion: 77 Resp: 29717
Ion Ratio Lower Upper
77 100
105 101.4 74.7 112.1
106 97.4 70.2 105.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 06/13/2022
Supervised By :mohammad ahmed 06/14/2022



#8
Phenol
Concen: 38.303 ng/ul
RT: 7.016 min Scan# 668
Delta R.T. 0.000 min
Lab File: BN020123.D
Acq: 11 Jun 2022 05:21

Tgt Ion: 94 Resp: 531602
Ion Ratio Lower Upper
94 100
65 29.2 23.6 35.4
66 40.4 31.2 46.8





#10

Bis(2-Chloroethyl)ether

Concen: 36.978 ng/ul

RT: 7.240 min Scan# 706

Delta R.T. -0.006 min

Lab File: BN020123.D


Acq: 11 Jun 2022 05:21

Instrument :

BNA_N

ClientSampleId :

SLCS445



Tgt Ion: 93 Resp: 401680

Ion Ratio Lower Upper

93 100

63 82.8 63.4 95.0




95 33.5 26.2 39.2

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 06/13/2022

Supervised By :mohammad ahmed 06/14/2022



#12

2-Chlorophenol

Concen: 38.093 ng/ul

RT: 7.381 min Scan# 730

Delta R.T. -0.006 min

Lab File: BN020123.D

Acq: 11 Jun 2022 05:21




Tgt Ion:128 Resp: 421683


Ion Ratio Lower Upper

128 100

64 49.7 38.6 57.8


130 31.8 25.0 37.4





#13
2-Methylphenol
Concen: 38.069 ng/ul
RT: 8.258 min Scan# 81
Delta R.T. -0.006 min
Lab File: BN020123.D
Acq: 11 Jun 2022 05:21




Instrument :
BNA_N
ClientSampleId :
SLCS445



Tgt Ion:108 Resp: 40618
Ion Ratio Lower Upper
108 100
107 90.2 71.5 107.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 06/13/2022
Supervised By :mohammad ahmed 06/14/2022



#14
2,2'-oxybis(1-Chloropropane)
Concen: 36.785 ng/ul
RT: 8.346 min Scan# 894
Delta R.T. -0.012 min
Lab File: BN020123.D
Acq: 11 Jun 2022 05:21


Tgt Ion: 45 Resp: 626591
Ion Ratio Lower Upper
45 100
77 12.6 10.9 16.3
79 9.5 8.6 13.0





#16
Acetophenone
Concen: 37.487 ng/ul
RT: 8.634 min Scan# 943
Delta R.T. -0.006 min
Lab File: BN020123.D
Acq: 11 Jun 2022 05:21




Instrument :
BNA_N
ClientSampleId :
SLCS445




Tgt Ion:105 Resp: 642864
Ion Ratio Lower Upper
105 100
77 75.4 63.5 95.3
51 30.7 23.5 35.3

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/13/2022
Supervised By :mohammad ahmed 06/14/2022




#17
N-Nitroso-di-n-propylamine
Concen: 37.664 ng/ul
RT: 8.628 min Scan# 942
Delta R.T. 0.000 min
Lab File: BN020123.D
Acq: 11 Jun 2022 05:21




Tgt Ion: 70 Resp: 329948
Ion Ratio Lower Upper
70 100
42 64.4 47.8 71.6
101 10.0 7.8 11.6
130 22.2 15.8 23.8





#18
4-Methylphenol
Concen: 38.197 ng/ul
RT: 8.593 min Scan# 91
Delta R.T. 0.000 min
Lab File: BN020123.D
Acq: 11 Jun 2022 05:21




Instrument :
BNA_N
ClientSampleId :
SLCS445



Tgt Ion:108 Resp: 449810
Ion Ratio Lower Upper
108 100
107 107.4 88.0 132.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 06/13/2022
Supervised By :mohammad ahmed 06/14/2022



#19
Hexachloroethane
Concen: 36.937 ng/ul
RT: 8.893 min Scan# 987
Delta R.T. -0.012 min
Lab File: BN020123.D
Acq: 11 Jun 2022 05:21


Tgt Ion:117 Resp: 164524
Ion Ratio Lower Upper
117 100
201 85.2 83.2 124.8
199 54.4 50.2 75.4





#22
Nitrobenzene
Concen: 36.902 ng/uI
RT: 9.010 min Scan# 1007
Delta R.T. -0.006 min
Lab File: BN020123.D
Acq: 11 Jun 2022 05:21




Instrument :
BNA_N
ClientSampleId :
SLCS445




Tgt Ion: 77 Resp: 48180
Ion Ratio Lower Upper
77 100
123 47.9 34.5 51.7
65 14.8 11.1 16.7

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/13/2022
Supervised By :mohammad ahmed 06/14/2022




#23
Isophorone
Concen: 36.887 ng/uI
RT: 9.540 min Scan# 1097
Delta R.T. 0.000 min
Lab File: BN020123.D
Acq: 11 Jun 2022 05:21




Tgt Ion: 82 Resp: 950904
Ion Ratio Lower Upper
82 100
95 7.7 5.8 8.8
138 17.7 13.8 20.8





#25
2-Nitrophenol
Concen: 38.156 ng/ul
RT: 9.722 min Scan# 1128
Delta R.T. -0.006 min
Lab File: BN020123.D
Acq: 11 Jun 2022 05:21




Instrument :
BNA_N
ClientSampleId :
SLCS445




Tgt Ion:139 Resp: 251960
Ion Ratio Lower Upper
139 100
65 52.7 43.7 65.5
109 32.2 25.8 38.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/13/2022
Supervised By :mohammad ahmed 06/14/2022




#26
2,4-Dimethylphenol
Concen: 35.325 ng/ul
RT: 9.787 min Scan# 1139
Delta R.T. -0.006 min
Lab File: BN020123.D
Acq: 11 Jun 2022 05:21




Tgt Ion:107 Resp: 483272
Ion Ratio Lower Upper
107 100
121 48.1 38.3 57.5
122 81.1 66.3 99.5





#27
Bis(2-Chloroethoxy)methane
Concen: 36.163 ng/ul
RT: 10.022 min Scan# 1179
Delta R.T. -0.006 min
Lab File: BN020123.D
Acq: 11 Jun 2022 05:21




Instrument :
BNA_N
ClientSampleId :
SLCS445




Tgt Ion: 93 Resp: 570488
Ion Ratio Lower Upper
93 100
95 32.0 25.4 38.0
123 11.6 8.3 12.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/13/2022
Supervised By :mohammad ahmed 06/14/2022




#29
2,4-Dichlorophenol
Concen: 38.220 ng/ul
RT: 10.257 min Scan# 1219
Delta R.T. -0.006 min
Lab File: BN020123.D
Acq: 11 Jun 2022 05:21




Tgt Ion:162 Resp: 412453
Ion Ratio Lower Upper
162 100
164 65.9 52.9 79.3
98 40.3 29.8 44.8





#30
Naphthalene
Concen: 35.590 ng/ul
RT: 10.657 min Scan# 1287
Delta R.T. -0.006 min
Lab File: BN020123.D
Acq: 11 Jun 2022 05:21




Instrument :
BNA_N
ClientSampleId :
SLCS445



Tgt Ion:128 Resp: 1365227
Ion Ratio Lower Upper
128 100
129 10.8 8.7 13.1
127 13.2 10.6 16.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 06/13/2022
Supervised By :mohammad ahmed 06/14/2022



#32
4-Chloroaniline
Concen: 34.870 ng/ul
RT: 10.763 min Scan# 1305
Delta R.T. -0.006 min
Lab File: BN020123.D
Acq: 11 Jun 2022 05:21


Tgt Ion:127 Resp: 589239
Ion Ratio Lower Upper
127 100
129 32.5 26.1 39.1





#33
Hexachlorobutadiene
Concen: 36.070 ng/ul
RT: 10.952 min Scan# 11
Delta R.T. -0.006 min
Lab File: BN020123.D
Acq: 11 Jun 2022 05:21




Instrument :
BNA_N
ClientSampleId :
SLCS445



Tgt Ion:225 Resp: 22348
Ion Ratio Lower Upper
225 100
223 62.6 51.8 77.6
227 63.2 50.6 76.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 06/13/2022
Supervised By :mohammad ahmed 06/14/2022



#34
Caprolactam
Concen: 37.438 ng/ul m
RT: 11.522 min Scan# 1434
Delta R.T. 0.006 min
Lab File: BN020123.D
Acq: 11 Jun 2022 05:21


Tgt Ion:113 Resp: 157770
Ion Ratio Lower Upper
113 100
55 180.4 145.5 218.3
56 134.3 106.5 159.7





#35
4-Chloro-3-methylphenol
Concen: 38.876 ng/ul
RT: 11.899 min Scan# 1498
Delta R.T. 0.000 min
Lab File: BN020123.D
Acq: 11 Jun 2022 05:21




Instrument :
BNA_N
ClientSampleId :
SLCS445



Tgt Ion:107 Resp: 49538
Ion Ratio Lower Upper
107 100
144 25.2 20.6 31.0
142 78.0 62.2 93.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 06/13/2022
Supervised By :mohammad ahmed 06/14/2022



#36
2-Methylnaphthalene
Concen: 36.470 ng/ul
RT: 12.269 min Scan# 1561
Delta R.T. -0.006 min
Lab File: BN020123.D
Acq: 11 Jun 2022 05:21


Tgt Ion:142 Resp: 945813
Ion Ratio Lower Upper
142 100
141 89.5 70.6 105.8





#37
1-Methylnaphthalene
Concen: 36.514 ng/ul
RT: 12.487 min Scan# 1598
Delta R.T. -0.006 min
Lab File: BN020123.D
Acq: 11 Jun 2022 05:21




Instrument :
BNA_N
ClientSampleId :
SLCS445




Tgt Ion:142 Resp: 966348
Ion Ratio Lower Upper
142 100
141 94.1 70.9 106.3
116 4.0 3.1 4.7

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/13/2022
Supervised By :mohammad ahmed 06/14/2022




#39
1,2,4,5-Tetrachlorobenzene
Concen: 36.056 ng/ul
RT: 12.646 min Scan# 1625
Delta R.T. -0.006 min
Lab File: BN020123.D
Acq: 11 Jun 2022 05:21




Tgt Ion:216 Resp: 438492
Ion Ratio Lower Upper
216 100
214 79.4 63.4 95.0
179 22.3 17.8 26.8
108 20.1 15.1 22.7





#40
Hexachlorocyclopentadiene
Concen: 32.704 ng/ul
RT: 12.628 min Scan# 1622
Delta R.T. -0.006 min
Lab File: BN020123.D
Acq: 11 Jun 2022 05:21




Instrument :
BNA_N
ClientSampleId :
SLCS445




Tgt Ion:237 Resp: 241944
Ion Ratio Lower Upper
237 100
235 62.7 48.8 73.2
272 11.9 9.3 13.9

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/13/2022
Supervised By :mohammad ahmed 06/14/2022




#41
2,4,6-Trichlorophenol
Concen: 38.115 ng/ul
RT: 12.881 min Scan# 1665
Delta R.T. -0.006 min
Lab File: BN020123.D
Acq: 11 Jun 2022 05:21




Tgt Ion:196 Resp: 319617
Ion Ratio Lower Upper
196 100
198 96.4 79.8 119.8
200 30.8 25.0 37.6





#42
2,4,5-Trichlorophenol
Concen: 37.935 ng/ul
RT: 12.951 min Scan# 1677
Delta R.T. -0.006 min
Lab File: BN020123.D
Acq: 11 Jun 2022 05:21




Instrument :
BNA_N
ClientSampleId :
SLCS445




Tgt Ion:196 Resp: 352308
Ion Ratio Lower Upper
196 100
198 97.1 77.6 116.4
200 30.8 24.4 36.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/13/2022
Supervised By :mohammad ahmed 06/14/2022




#43
1,1'-Biphenyl
Concen: 36.046 ng/ul
RT: 13.281 min Scan# 1733
Delta R.T. -0.006 min
Lab File: BN020123.D
Acq: 11 Jun 2022 05:21




Tgt Ion:154 Resp: 1272714
Ion Ratio Lower Upper
154 100
153 38.8 31.5 47.3
76 12.0 10.5 15.7





#44
2-Chloronaphthalene
Concen: 36.127 ng/ul
RT: 13.322 min Scan# 1740
Delta R.T. -0.006 min
Lab File: BN020123.D
Acq: 11 Jun 2022 05:21




Instrument :
BNA_N
ClientSampleId :
SLCS445




Tgt Ion:162 Resp: 96771
Ion Ratio Lower Upper
162 100
164 32.3 25.8 38.6
127 43.6 31.4 47.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/13/2022
Supervised By :mohammad ahmed 06/14/2022




#45
2-Nitroaniline
Concen: 39.528 ng/ul
RT: 13.522 min Scan# 1774
Delta R.T. -0.006 min
Lab File: BN020123.D
Acq: 11 Jun 2022 05:21




Tgt Ion: 65 Resp: 329397
Ion Ratio Lower Upper
65 100
92 67.3 54.1 81.1
138 104.7 77.0 115.4





#47
Dimethylphthalate
Concen: 35.890 ng/ul
RT: 13.904 min Scan# 1839
Delta R.T. -0.006 min
Lab File: BN020123.D
Acq: 11 Jun 2022 05:21




Instrument :
BNA_N
ClientSampleId :
SLCS445




Tgt Ion:163 Resp: 125355
Ion Ratio Lower Upper
163 100
194 3.9 3.4 5.0
164 10.5 7.7 11.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/13/2022
Supervised By :mohammad ahmed 06/14/2022




#48
2,6-Dinitrotoluene
Concen: 39.711 ng/ul
RT: 14.022 min Scan# 1859
Delta R.T. -0.006 min
Lab File: BN020123.D
Acq: 11 Jun 2022 05:21




Tgt Ion:165 Resp: 271185
Ion Ratio Lower Upper
165 100
89 57.0 50.6 75.8
121 20.7 17.9 26.9





#50
Acenaphthylene
Concen: 36.088 ng/ul
RT: 14.169 min Scan# 1884
Delta R.T. 0.000 min
Lab File: BN020123.D
Acq: 11 Jun 2022 05:21




Instrument :
BNA_N
ClientSampleId :
SLCS445




Tgt Ion:152 Resp: 160762
Ion Ratio Lower Upper
152 100
151 20.2 16.0 24.0
153 13.0 10.7 16.1

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/13/2022
Supervised By :mohammad ahmed 06/14/2022




#51
3-Nitroaniline
Concen: 39.749 ng/ul
RT: 14.345 min Scan# 1914
Delta R.T. -0.006 min
Lab File: BN020123.D
Acq: 11 Jun 2022 05:21




Tgt Ion:138 Resp: 286966
Ion Ratio Lower Upper
138 100
108 11.0 8.6 12.8
92 115.1 92.9 139.3





#52
Acenaphthene
Concen: 35.166 ng/ul
RT: 14.510 min Scan# 1942
Delta R.T. -0.006 min
Lab File: BN020123.D
Acq: 11 Jun 2022 05:21




Instrument :
BNA_N
ClientSampleId :
SLCS445




Tgt Ion:153 Resp: 102325
Ion Ratio Lower Upper
153 100
152 45.5 38.4 57.6
154 88.5 69.7 104.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/13/2022
Supervised By :mohammad ahmed 06/14/2022




#53
2,4-Dinitrophenol
Concen: 31.350 ng/ul
RT: 14.551 min Scan# 1949
Delta R.T. -0.006 min
Lab File: BN020123.D
Acq: 11 Jun 2022 05:21




Tgt Ion:184 Resp: 137070
Ion Ratio Lower Upper
184 100
63 75.9 55.9 83.9
154 68.5 53.1 79.7





#55
4-Nitrophenol
Concen: 35.269 ng/ul
RT: 14.663 min Scan# 1967
Delta R.T. 0.006 min
Lab File: BN020123.D
Acq: 11 Jun 2022 05:21




Instrument :
BNA_N
ClientSampleId :
SLCS445



Tgt Ion:109 Resp: 209304
Ion Ratio Lower Upper
109 100
139 121.8 100.1 150.1
65 124.6 104.6 157.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 06/13/2022
Supervised By :mohammad ahmed 06/14/2022



#56
Dibenzofuran
Concen: 35.277 ng/ul
RT: 14.845 min Scan# 1999
Delta R.T. -0.006 min
Lab File: BN020123.D
Acq: 11 Jun 2022 05:21


Tgt Ion:168 Resp: 1446186
Ion Ratio Lower Upper
168 100
139 40.6 31.1 46.7





#57
2,4-Dinitrotoluene
Concen: 38.696 ng/ul
RT: 14.810 min Scan# 1993
Delta R.T. 0.000 min
Lab File: BN020123.D
Acq: 11 Jun 2022 05:21




Instrument :
BNA_N
ClientSampleId :
SLCS445




Tgt Ion:165 Resp: 391264
Ion Ratio Lower Upper
165 100
63 46.7 36.3 54.5
182 3.0 2.3 3.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/13/2022
Supervised By :mohammad ahmed 06/14/2022




#58
2,3,4,6-Tetrachlorophenol
Concen: 37.235 ng/ul
RT: 15.075 min Scan# 2038
Delta R.T. 0.000 min
Lab File: BN020123.D
Acq: 11 Jun 2022 05:21




Tgt Ion:232 Resp: 271454
Ion Ratio Lower Upper
232 100
131 53.6 38.5 57.7
130 2.9 2.2 3.2
166 35.7 25.8 38.8





#59
Diethylphthalate
Concen: 35.999 ng/ul
RT: 15.275 min Scan# 2071
Delta R.T. 0.000 min
Lab File: BN020123.D
Acq: 11 Jun 2022 05:21




Instrument :
BNA_N
ClientSampleId :
SLCS445



Tgt Ion:149 Resp: 129450
Ion Ratio Lower Upper
149 100
177 21.4 17.8 26.8
150 12.5 10.5 15.7




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 06/13/2022
Supervised By :mohammad ahmed 06/14/2022



#61
Fluorene
Concen: 34.734 ng/ul
RT: 15.493 min Scan# 2109
Delta R.T. -0.006 min
Lab File: BN020123.D
Acq: 11 Jun 2022 05:21


Tgt Ion:166 Resp: 1107984
Ion Ratio Lower Upper
166 100
165 97.7 77.7 116.5
167 13.8 11.1 16.7





#62
4-Chlorophenyl-phenylether
Concen: 34.991 ng/ul
RT: 15.487 min Scan# 2108
Delta R.T. -0.006 min
Lab File: BN020123.D
Acq: 11 Jun 2022 05:21




Instrument :
BNA_N
ClientSampleId :
SLCS445




Tgt Ion:204 Resp: 514028
Ion Ratio Lower Upper
204 100
206 32.8 26.7 40.1
141 72.4 50.2 75.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/13/2022
Supervised By :mohammad ahmed 06/14/2022




#63
4-Nitroaniline
Concen: 41.710 ng/ul m
RT: 15.510 min Scan# 2112
Delta R.T. 0.000 min
Lab File: BN020123.D
Acq: 11 Jun 2022 05:21



Tgt Ion:138 Resp: 281384
Ion Ratio Lower Upper
138 100
92 48.7 41.0 61.6
108 83.9 65.3 97.9





#66
4,6-Dinitro-2-methylphenol
Concen: 37.034 ng/ul
RT: 15.575 min Scan# 2123
Delta R.T. 0.000 min
Lab File: BN020123.D
Acq: 11 Jun 2022 05:21

Instrument :

BNA_N





ClientSampleId :

SLCS445

Tgt Ion:198 Resp: 20934
Ion Ratio Lower Upper
198 100
182 3.8 3.0 4.4
77 28.4 20.1 30.1




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 06/13/2022
Supervised By :mohammad ahmed 06/14/2022



#67
N-Nitrosodiphenylamine
Concen: 37.255 ng/ul
RT: 15.698 min Scan# 2144
Delta R.T. -0.006 min
Lab File: BN020123.D
Acq: 11 Jun 2022 05:21


Tgt Ion:169 Resp: 1028008
Ion Ratio Lower Upper
169 100
168 65.8 52.4 78.6
167 35.7 27.5 41.3





#68
4-Bromophenyl-phenylether
Concen: 37.545 ng/ul
RT: 16.381 min Scan# 211
Delta R.T. -0.006 min
Lab File: BN020123.D
Acq: 11 Jun 2022 05:21




Instrument :
BNA_N
ClientSampleId :
SLCS445




Tgt Ion:248 Resp: 314710
Ion Ratio Lower Upper
248 100
250 93.7 77.8 116.8
141 95.2 62.6 93.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/13/2022
Supervised By :mohammad ahmed 06/14/2022




#69
Hexachlorobenzene
Concen: 36.424 ng/ul
RT: 16.504 min Scan# 2281
Delta R.T. 0.000 min
Lab File: BN020123.D
Acq: 11 Jun 2022 05:21




Tgt Ion:284 Resp: 350954
Ion Ratio Lower Upper
284 100
142 41.9 27.8 41.8#
249 34.0 25.7 38.5





#70
Atrazine
Concen: 37.351 ng/ul
RT: 16.651 min Scan# 21
Delta R.T. -0.006 min
Lab File: BN020123.D
Acq: 11 Jun 2022 05:21




Instrument :
BNA_N
ClientSampleId :
SLCS445




Tgt Ion:200 Resp: 359604
Ion Ratio Lower Upper
200 100
173 29.7 20.5 30.7
215 47.1 38.9 58.3

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/13/2022
Supervised By :mohammad ahmed 06/14/2022




#71
Pentachlorophenol
Concen: 34.587 ng/ul
RT: 16.839 min Scan# 2338
Delta R.T. -0.006 min
Lab File: BN020123.D
Acq: 11 Jun 2022 05:21




Tgt Ion:266 Resp: 207559
Ion Ratio Lower Upper
266 100
264 69.2 50.2 75.2
268 68.4 49.3 73.9





#72
Phenanthrene
Concen: 36.172 ng/ul
RT: 17.228 min Scan# 2404
Delta R.T. -0.006 min
Lab File: BN020123.D
Acq: 11 Jun 2022 05:21




Instrument :
BNA_N
ClientSampleId :
SLCS445




Tgt Ion:178 Resp: 1837362
Ion Ratio Lower Upper
178 100
179 15.7 12.2 18.4
176 19.5 15.2 22.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/13/2022
Supervised By :mohammad ahmed 06/14/2022




#74
Anthracene
Concen: 36.167 ng/ul
RT: 17.322 min Scan# 2420
Delta R.T. -0.006 min
Lab File: BN020123.D
Acq: 11 Jun 2022 05:21




Tgt Ion:178 Resp: 1847247
Ion Ratio Lower Upper
178 100
179 15.2 13.0 19.4
176 18.5 15.5 23.3





#75
1,2,3,4-Tetrachlorobenzene
Concen: 36.809 ng/uL
RT: 13.246 min Scan# 1727
Delta R.T. -0.006 min
Lab File: BN020123.D
Acq: 11 Jun 2022 05:21




Instrument :
BNA_N
ClientSampleId :
SLCS445




Tgt Ion:216 Resp: 474429
Ion Ratio Lower Upper
216 100
214 77.5 63.2 94.8
218 48.1 37.9 56.9

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/13/2022
Supervised By :mohammad ahmed 06/14/2022




#76
Pentachlorobenzene
Concen: 36.830 ng/uL
RT: 14.769 min Scan# 1986
Delta R.T. -0.006 min
Lab File: BN020123.D
Acq: 11 Jun 2022 05:21




Tgt Ion:250 Resp: 419555
Ion Ratio Lower Upper
250 100
248 70.0 53.0 79.4
252 66.9 53.4 80.2





#77
 Carbazole
 Concen: 37.785 ng/ul
 RT: 17.587 min Scan# 2465
 Delta R.T. -0.006 min
 Lab File: BN020123.D
 Acq: 11 Jun 2022 05:21

Instrument :
 BNA_N
 ClientSampleId :
 SLCS445






Tgt Ion:167 Resp: 1755710


Ion	Ratio	Lower	Upper
167	100		
166	21.1	16.1	24.1
139	14.2	10.2	15.2

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 06/13/2022
 Supervised By :mohammad ahmed 06/14/2022






#78
 Di-n-butylphthalate
 Concen: 38.846 ng/ul
 RT: 18.157 min Scan# 2562
 Delta R.T. -0.006 min
 Lab File: BN020123.D
 Acq: 11 Jun 2022 05:21



Tgt Ion:149 Resp: 2238267


Ion	Ratio	Lower	Upper
149	100		
150	9.5	7.1	10.7
104	5.4	4.1	6.1





#80
Fluoranthene
Concen: 40.450 ng/u1
RT: 19.245 min Scan# 2747
Delta R.T. -0.006 min
Lab File: BN020123.D
Acq: 11 Jun 2022 05:21




Instrument :
BNA_N
ClientSampleId :
SLCS445




Tgt Ion:202 Resp: 207216
Ion Ratio Lower Upper
202 100
101 12.3 8.9 13.3
100 9.0 6.7 10.1

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/13/2022
Supervised By :mohammad ahmed 06/14/2022




#82
Pyrene
Concen: 39.772 ng/u1
RT: 19.610 min Scan# 2809
Delta R.T. -0.006 min
Lab File: BN020123.D
Acq: 11 Jun 2022 05:21



Tgt Ion:202 Resp: 2090281
Ion Ratio Lower Upper
202 100
101 13.6 9.7 14.5
100 11.1 8.0 12.0





#83

Butylbenzylphthalate

Concen: 43.171 ng/ul

RT: 20.510 min Scan# 2962

Delta R.T. -0.012 min

Lab File: BN020123.D


Acq: 11 Jun 2022 05:21

Instrument :

BNA_N

ClientSampleId :

SLCS445



Tgt Ion:149 Resp: 1021118

Ion Ratio Lower Upper

149 100

91 71.7 58.8 88.2



206 17.9 18.1 27.1

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 06/13/2022

Supervised By :mohammad ahmed 06/14/2022



#84

3,3'-Dichlorobenzidine

Concen: 38.316 ng/ul

RT: 21.292 min Scan# 3095

Delta R.T. -0.006 min

Lab File: BN020123.D

Acq: 11 Jun 2022 05:21




Tgt Ion:252 Resp: 572590


Ion Ratio Lower Upper

252 100

254 67.3 50.8 76.2

126 13.2 8.2 12.4





#85

Benzo(a)anthracene

Concen: 38.221 ng/ul

RT: 21.357 min Scan# 3106

Delta R.T. -0.006 min

Lab File: BN020123.D

Acq: 11 Jun 2022 05:21

Instrument :

BNA_N

ClientSampleId :

SLCS445

Tgt Ion:228 Resp: 1888378

Ion Ratio Lower Upper

228 100





229 20.1 16.3 24.5

226 27.3 23.2 34.8

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 06/13/2022
Supervised By :mohammad ahmed 06/14/2022



#86

Bis(2-ethylhexyl)phthalate

Concen: 42.942 ng/ul

RT: 21.298 min Scan# 3096

Delta R.T. -0.006 min

Lab File: BN020123.D

Acq: 11 Jun 2022 05:21




Tgt Ion:149 Resp: 1433082


Ion Ratio Lower Upper

149 100

167 27.6 22.5 33.7


279 3.5 3.5 5.3





#87
Chrysene
Concen: 37.551 ng/ul
RT: 21.410 min Scan# 3115
Delta R.T. -0.012 min
Lab File: BN020123.D
Acq: 11 Jun 2022 05:21




Instrument :
BNA_N
ClientSampleId :
SLCS445



Tgt Ion:228 Resp: 1815080
Ion Ratio Lower Upper
228 100
226 31.4 24.2 36.2
229 20.0 16.2 24.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 06/13/2022
Supervised By :mohammad ahmed 06/14/2022



#89
Di-n-octyl phthalate
Concen: 42.577 ng/ul
RT: 22.198 min Scan# 3249
Delta R.T. -0.012 min
Lab File: BN020123.D
Acq: 11 Jun 2022 05:21


Tgt Ion:149 Resp: 2519443





#90
Benzo(b)fluoranthene
Concen: 40.831 ng/ul m
RT: 22.998 min Scan# 3385
Delta R.T. -0.006 min
Lab File: BN020123.D
Acq: 11 Jun 2022 05:21




Instrument :
BNA_N
ClientSampleId :
SLCS445



Tgt Ion:252 Resp: 186311
Ion Ratio Lower Upper
252 100
253 20.8 16.5 24.7
125 12.1 7.8 11.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 06/13/2022
Supervised By :mohammad ahmed 06/14/2022



#91
Benzo(k)fluoranthene
Concen: 37.255 ng/ul
RT: 23.045 min Scan# 3393
Delta R.T. -0.006 min
Lab File: BN020123.D
Acq: 11 Jun 2022 05:21


Tgt Ion:252 Resp: 1618258
Ion Ratio Lower Upper
252 100
253 21.6 18.2 27.4
125 11.7 8.2 12.4





#93
Benzo(a)pyrene
Concen: 37.705 ng/u1
RT: 23.598 min Scan# 3487
Delta R.T. -0.012 min
Lab File: BN020123.D
Acq: 11 Jun 2022 05:21




Instrument :
BNA_N
ClientSampleId :
SLCS445




Tgt Ion:252 Resp: 1665370
Ion Ratio Lower Upper
252 100
253 19.7 18.7 28.1
125 12.4 9.2 13.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/13/2022
Supervised By :mohammad ahmed 06/14/2022




#94
Indeno(1,2,3-cd)pyrene
Concen: 35.432 ng/u1
RT: 26.098 min Scan# 3912
Delta R.T. 0.000 min
Lab File: BN020123.D
Acq: 11 Jun 2022 05:21



Tgt Ion:276 Resp: 1610167
Ion Ratio Lower Upper
276 100
138 30.8 19.5 29.3#
277 25.1 21.4 32.0









#95
Dibenzo(a,h)anthracene
Concen: 35.173 ng/ul
RT: 26.104 min Scan# 3914
Delta R.T. -0.017 min
Lab File: BN020123.D
Acq: 11 Jun 2022 05:21

Instrument :
BNA_N
ClientSampleId :
SLCS445

Tgt Ion:278 Resp: 136523
Ion Ratio Lower Upper
278 100
139 20.5 13.8 20.8
279 24.1 18.3 27.5

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 06/13/2022
Supervised By :mohammad ahmed 06/14/2022



#96
Benzo(g,h,i)perylene
Concen: 35.437 ng/ul
RT: 26.827 min Scan# 4036
Delta R.T. -0.017 min
Lab File: BN020123.D
Acq: 11 Jun 2022 05:21

Tgt Ion:276 Resp: 1359920
Ion Ratio Lower Upper
276 100
138 26.8 17.0 25.6#
277 25.0 17.8 26.6

