

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061122\
 Data File : BN020124.D
 Acq On : 11 Jun 2022 05:57
 Operator : CG/JU
 Sample : PB145446BS
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SLCS446

Quant Time: Jun 11 06:30:59 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN060622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jun 06 15:14:59 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 06/13/2022
 Supervised By :mohammad ahmed 06/14/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.816	152	165575	20.000	ng/u1	0.00
20) Naphthalene-d8	10.610	136	769263	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.445	164	495612	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.186	188	1034878	20.000	ng/u1	0.00
79) Chrysene-d12	21.374	240	897763	20.000	ng/u1	0.00
88) Perylene-d12	23.704	264	830227	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.281	96	22971	6.009	ng/uL	0.00
4) Pyridine-d5	3.681	84	326396	31.238	ng/u1	0.00
7) Phenol-d5	6.987	99	459040	33.504	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.152	67	277233	32.387	ng/u1	0.00
11) 2-Chlorophenol-d4	7.352	132	378101	33.964	ng/u1	0.00
15) 4-Methylphenol-d8	8.528	113	399610	33.712	ng/u1	0.00
21) Nitrobenzene-d5	8.969	128	200412	34.244	ng/u1	0.00
24) 2-Nitrophenol-d4	9.693	143	218594	35.195	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.228	165	371868	34.297	ng/u1	0.00
31) 4-Chloroaniline-d4	10.740	131	554798	31.597	ng/u1	0.00
46) Dimethylphthalate-d6	13.857	166	1213636	33.655	ng/u1	0.00
49) Acenaphthylene-d8	14.139	160	1396837	33.401	ng/u1	0.00
54) 4-Nitrophenol-d4	14.645	143	240272	33.250	ng/u1	0.00
60) Fluorene-d10	15.439	176	982040	32.909	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.557	200	202907	33.848	ng/u1	0.00
73) Anthracene-d10	17.286	188	1510274	33.439	ng/u1	0.00
81) Pyrene-d10	19.580	212	1670352	34.320	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.551	264	1410701	34.728	ng/u1	-0.01
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.317	88	46377	11.492	ng/uL#	90
5) Pyridine	3.705	79	337519	30.813	ng/u1	95
6) Benzaldehyde	6.957	77	266318	41.253	ng/u1	92
8) Phenol	7.016	94	482296	33.058	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.240	93	363017	31.791	ng/u1	97
12) 2-Chlorophenol	7.381	128	381152	32.756	ng/u1	97
13) 2-Methylphenol	8.257	108	369776	32.969	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.346	45	562922	31.438	ng/u1	97
16) Acetophenone	8.634	105	584920	32.448	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.628	70	297557	32.313	ng/u1	94
18) 4-Methylphenol	8.593	108	410755	33.182	ng/u1	99
19) Hexachloroethane	8.899	117	147744	31.555	ng/u1	89
22) Nitrobenzene	9.010	77	441274	32.520	ng/u1	95
23) Isophorone	9.540	82	858471	32.042	ng/u1	99
25) 2-Nitrophenol	9.722	139	233707	34.053	ng/u1	98
26) 2,4-Dimethylphenol	9.787	107	430585	30.283	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.022	93	518001	31.594	ng/u1	98
29) 2,4-Dichlorophenol	10.257	162	375894	33.515	ng/u1	97
30) Naphthalene	10.657	128	1245349	31.237	ng/u1	100
32) 4-Chloroaniline	10.763	127	544672	31.013	ng/u1	98
33) Hexachlorobutadiene	10.951	225	202008	31.371	ng/u1	98
34) Caprolactam	11.522	113	149962m	34.239	ng/u1	
35) 4-Chloro-3-methylphenol	11.893	107	455203	34.372	ng/u1	97
36) 2-Methylnaphthalene	12.269	142	873289	32.400	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061122\
 Data File : BN020124.D
 Acq On : 11 Jun 2022 05:57
 Operator : CG/JU
 Sample : PB145446BS
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SLCS446


Quant Time: Jun 11 06:30:59 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN060622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jun 06 15:14:59 2022
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 06/13/2022
 Supervised By :mohammad ahmed 06/14/2022


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.487	142	887551	32.268	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.645	216	406904	32.191	ng/ul	99
40) Hexachlorocyclopentadiene	12.628	237	219622	28.562	ng/ul	98
41) 2,4,6-Trichlorophenol	12.881	196	292118	33.516	ng/ul	98
42) 2,4,5-Trichlorophenol	12.951	196	326835	33.859	ng/ul	95
43) 1,1'-Biphenyl	13.281	154	1161376	31.646	ng/ul	99
44) 2-Chloronaphthalene	13.322	162	895728	32.173	ng/ul	98
45) 2-Nitroaniline	13.522	65	312138	36.038	ng/ul	95
47) Dimethylphthalate	13.904	163	1170916	32.254	ng/ul	98
48) 2,6-Dinitrotoluene	14.022	165	252873	35.627	ng/ul	95
50) Acenaphthylene	14.169	152	1465284	31.646	ng/ul	100
51) 3-Nitroaniline	14.345	138	265651	35.402	ng/ul	98
52) Acenaphthene	14.510	153	963220	31.848	ng/ul	97
53) 2,4-Dinitrophenol	14.551	184	128222	28.216	ng/ul	94
55) 4-Nitrophenol	14.657	109	201256	32.628	ng/ul	98
56) Dibenzofuran	14.845	168	1349158	31.664	ng/ul	98
57) 2,4-Dinitrotoluene	14.810	165	371291	35.329	ng/ul	100
58) 2,3,4,6-Tetrachlorophenol	15.069	232	254048	33.527	ng/ul	93
59) Diethylphthalate	15.269	149	1233365	32.999	ng/ul	98
61) Fluorene	15.492	166	1055232	31.827	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.486	204	481349	31.525	ng/ul	90
63) 4-Nitroaniline	15.510	138	268401	38.279	ng/ul	96
66) 4,6-Dinitro-2-methylph...	15.575	198	197369	32.474	ng/ul	96
67) N-Nitrosodiphenylamine	15.698	169	962987	32.459	ng/ul	96
68) 4-Bromophenyl-phenylether	16.381	248	291252	32.317	ng/ul#	88
69) Hexachlorobenzene	16.504	284	331603	32.009	ng/ul#	92
70) Atrazine	16.651	200	351847	33.990	ng/ul	94
71) Pentachlorophenol	16.839	266	199521	30.922	ng/ul	99
72) Phenanthrene	17.228	178	1752313	32.085	ng/ul	98
74) Anthracene	17.322	178	1750120	31.869	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.245	216	434591	31.360	ng/ul	99
76) Pentachlorobenzene	14.769	250	403620	32.954	ng/ul	91
77) Carbazole	17.586	167	1684593	33.719	ng/ul	97
78) Di-n-butylphthalate	18.157	149	2139410	34.535	ng/ul	99
80) Fluoranthene	19.245	202	2025615	34.460	ng/ul	97
82) Pyrene	19.610	202	2034955	33.744	ng/ul	96
83) Butylbenzylphthalate	20.510	149	993616	36.611	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.292	252	572642	33.396	ng/ul#	96
85) Benzo(a)anthracene	21.357	228	1833231	32.338	ng/ul	97
86) Bis(2-ethylhexyl)phtha...	21.298	149	1372853	35.851	ng/ul	100
87) Chrysene	21.410	228	1848801	33.334	ng/ul	98
89) Di-n-octyl phthalate	22.198	149	2488812	36.832	ng/ul	100
90) Benzo(b)fluoranthene	22.998	252	1871995	35.926	ng/ul	98
91) Benzo(k)fluoranthene	23.045	252	1700651m	34.286	ng/ul	
93) Benzo(a)pyrene	23.604	252	1678935	33.287	ng/ul	94
94) Indeno(1,2,3-cd)pyrene	26.092	276	1644100	31.682	ng/ul#	93
95) Dibenzo(a,h)anthracene	26.109	278	1406323	31.728	ng/ul	98
96) Benzo(g,h,i)perylene	26.827	276	1401210	31.974	ng/ul	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 11.492 ng/uL
RT: 3.317 min Scan# 39
Delta R.T. 0.001 min
Lab File: BN020124.D
Acq: 11 Jun 2022 05:57




Instrument :
BNA_N
ClientSampleId :
SLCS446



Tgt Ion: 88 Resp: 4637
Ion Ratio Lower Upper
88 100
43 34.7 22.4 33.6
58 84.0 61.0 91.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 06/13/2022
Supervised By :mohammad ahmed 06/14/2022



#5
Pyridine
Concen: 30.813 ng/uL
RT: 3.705 min Scan# 105
Delta R.T. -0.006 min
Lab File: BN020124.D
Acq: 11 Jun 2022 05:57


Tgt Ion: 79 Resp: 337519
Ion Ratio Lower Upper
79 100
52 86.4 65.0 97.4
51 41.4 31.6 47.4





#6
Benzaldehyde
Concen: 41.253 ng/ul
RT: 6.957 min Scan# 658
Delta R.T. -0.006 min
Lab File: BN020124.D
Acq: 11 Jun 2022 05:57




Instrument :
BNA_N
ClientSampleId :
SLCS446




Tgt Ion: 77 Resp: 266318
Ion Ratio Lower Upper
77 100
105 99.5 74.7 112.1
106 96.0 70.2 105.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/13/2022
Supervised By :mohammad ahmed 06/14/2022




#8
Phenol
Concen: 33.058 ng/ul
RT: 7.016 min Scan# 668
Delta R.T. 0.000 min
Lab File: BN020124.D
Acq: 11 Jun 2022 05:57




Tgt Ion: 94 Resp: 482296
Ion Ratio Lower Upper
94 100
65 29.0 23.6 35.4
66 39.4 31.2 46.8





#10
Bis(2-Chloroethyl)ether
Concen: 31.791 ng/ul
RT: 7.240 min Scan# 706
Delta R.T. -0.006 min
Lab File: BN020124.D
Acq: 11 Jun 2022 05:57




Instrument :
BNA_N
ClientSampleId :
SLCS446



Tgt Ion: 93 Resp: 36301
Ion Ratio Lower Upper
93 100
63 82.5 63.4 95.0
95 31.5 26.2 39.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 06/13/2022
Supervised By :mohammad ahmed 06/14/2022



#12
2-Chlorophenol
Concen: 32.756 ng/ul
RT: 7.381 min Scan# 730
Delta R.T. -0.006 min
Lab File: BN020124.D
Acq: 11 Jun 2022 05:57


Tgt Ion:128 Resp: 381152
Ion Ratio Lower Upper
128 100
64 50.5 38.6 57.8
130 32.6 25.0 37.4





#13
2-Methylphenol
Concen: 32.969 ng/ul
RT: 8.257 min Scan# 81
Delta R.T. -0.006 min
Lab File: BN020124.D
Acq: 11 Jun 2022 05:57




Instrument :
BNA_N
ClientSampleId :
SLCS446



Tgt Ion:108 Resp: 369770
Ion Ratio Lower Upper
108 100
107 88.2 71.5 107.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 06/13/2022
Supervised By :mohammad ahmed 06/14/2022



#14
2,2'-oxybis(1-Chloropropane)
Concen: 31.438 ng/ul
RT: 8.346 min Scan# 894
Delta R.T. -0.012 min
Lab File: BN020124.D
Acq: 11 Jun 2022 05:57


Tgt Ion: 45 Resp: 562922
Ion Ratio Lower Upper
45 100
77 12.6 10.9 16.3
79 9.4 8.6 13.0





#16
Acetophenone
Concen: 32.448 ng/ul
RT: 8.634 min Scan# 943
Delta R.T. -0.006 min
Lab File: BN020124.D
Acq: 11 Jun 2022 05:57




Instrument :
BNA_N
ClientSampleId :
SLCS446



Tgt Ion: 105 Resp: 584920
Ion Ratio Lower Upper
105 100
77 77.8 63.5 95.3
51 30.3 23.5 35.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 06/13/2022
Supervised By :mohammad ahmed 06/14/2022



#17
N-Nitroso-di-n-propylamine
Concen: 32.313 ng/ul
RT: 8.628 min Scan# 942
Delta R.T. 0.000 min
Lab File: BN020124.D
Acq: 11 Jun 2022 05:57


Tgt Ion: 70 Resp: 297557
Ion Ratio Lower Upper
70 100
42 65.0 47.8 71.6
101 9.6 7.8 11.6
130 22.2 15.8 23.8





#18
4-Methylphenol
Concen: 33.182 ng/ul
RT: 8.593 min Scan# 91
Delta R.T. 0.000 min
Lab File: BN020124.D
Acq: 11 Jun 2022 05:57




Instrument :
BNA_N
ClientSampleId :
SLCS446



Tgt Ion:108 Resp: 41075
Ion Ratio Lower Upper
108 100
107 108.8 88.0 132.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 06/13/2022
Supervised By :mohammad ahmed 06/14/2022



#19
Hexachloroethane
Concen: 31.555 ng/ul
RT: 8.899 min Scan# 988
Delta R.T. -0.006 min
Lab File: BN020124.D
Acq: 11 Jun 2022 05:57


Tgt Ion:117 Resp: 147744
Ion Ratio Lower Upper
117 100
201 91.5 83.2 124.8
199 56.0 50.2 75.4





#22
Nitrobenzene
Concen: 32.520 ng/ul
RT: 9.010 min Scan# 1007
Delta R.T. -0.006 min
Lab File: BN020124.D
Acq: 11 Jun 2022 05:57




Instrument :
BNA_N
ClientSampleId :
SLCS446




Tgt Ion: 77 Resp: 441274
Ion Ratio Lower Upper
77 100
123 46.9 34.5 51.7
65 14.4 11.1 16.7

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/13/2022
Supervised By :mohammad ahmed 06/14/2022




#23
Isophorone
Concen: 32.042 ng/ul
RT: 9.540 min Scan# 1097
Delta R.T. 0.000 min
Lab File: BN020124.D
Acq: 11 Jun 2022 05:57



Tgt Ion: 82 Resp: 858471
Ion Ratio Lower Upper
82 100
95 7.4 5.8 8.8
138 17.9 13.8 20.8





#25

2-Nitrophenol

Concen: 34.053 ng/ul

RT: 9.722 min Scan# 1128

Delta R.T. -0.006 min

Lab File: BN020124.D

Acq: 11 Jun 2022 05:57

Instrument :

BNA_N

ClientSampleId :





SLCS446

Tgt Ion:	139	Resp:	23370
Ion Ratio	Lower	Upper	
139	100		
65	53.8	43.7	65.5
109	33.7	25.8	38.6

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 06/13/2022
Supervised By :mohammad ahmed 06/14/2022



#26

2,4-Dimethylphenol

Concen: 30.283 ng/ul




RT: 9.787 min Scan# 1139


Delta R.T. -0.006 min

Lab File: BN020124.D

Acq: 11 Jun 2022 05:57


Tgt Ion:	107	Resp:	430585
Ion Ratio	Lower	Upper	
107	100		
121	50.1	38.3	57.5
122	81.7	66.3	99.5





#27
Bis(2-Chloroethoxy)methane
Concen: 31.594 ng/ul
RT: 10.022 min Scan# 1179
Delta R.T. -0.006 min
Lab File: BN020124.D
Acq: 11 Jun 2022 05:57




Instrument :
BNA_N
ClientSampleId :
SLCS446



Tgt Ion: 93 Resp: 51800
Ion Ratio Lower Upper
93 100
95 32.6 25.4 38.0
123 11.4 8.3 12.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 06/13/2022
Supervised By :mohammad ahmed 06/14/2022



#29
2,4-Dichlorophenol
Concen: 33.515 ng/ul
RT: 10.257 min Scan# 1219
Delta R.T. -0.006 min
Lab File: BN020124.D
Acq: 11 Jun 2022 05:57


Tgt Ion:162 Resp: 375894
Ion Ratio Lower Upper
162 100
164 63.4 52.9 79.3
98 38.0 29.8 44.8





#30
Naphthalene
Concen: 31.237 ng/ul
RT: 10.657 min Scan# 1287
Delta R.T. -0.006 min
Lab File: BN020124.D
Acq: 11 Jun 2022 05:57




Instrument :
BNA_N
ClientSampleId :
SLCS446



Tgt Ion:128 Resp: 1245349
Ion Ratio Lower Upper
128 100
129 10.9 8.7 13.1
127 13.5 10.6 16.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 06/13/2022
Supervised By :mohammad ahmed 06/14/2022



#32
4-Chloroaniline
Concen: 31.013 ng/ul
RT: 10.763 min Scan# 1305
Delta R.T. -0.006 min
Lab File: BN020124.D
Acq: 11 Jun 2022 05:57


Tgt Ion:127 Resp: 544672
Ion Ratio Lower Upper
127 100
129 31.5 26.1 39.1





#33
Hexachlorobutadiene
Concen: 31.371 ng/ul
RT: 10.951 min Scan# 11
Delta R.T. -0.006 min
Lab File: BN020124.D
Acq: 11 Jun 2022 05:57




Instrument :
BNA_N
ClientSampleId :
SLCS446




Tgt Ion:225 Resp: 202008
Ion Ratio Lower Upper
225 100
223 62.1 51.8 77.6
227 62.5 50.6 76.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/13/2022
Supervised By :mohammad ahmed 06/14/2022




#34
Caprolactam
Concen: 34.239 ng/ul m
RT: 11.522 min Scan# 1434
Delta R.T. 0.006 min
Lab File: BN020124.D
Acq: 11 Jun 2022 05:57




Tgt Ion:113 Resp: 149962
Ion Ratio Lower Upper
113 100
55 177.8 145.5 218.3
56 135.6 106.5 159.7





#35
4-Chloro-3-methylphenol
Concen: 34.372 ng/ul
RT: 11.893 min Scan# 1497
Delta R.T. -0.006 min
Lab File: BN020124.D
Acq: 11 Jun 2022 05:57




Instrument :
BNA_N
ClientSampleId :
SLCS446




Tgt Ion:107 Resp: 45520
Ion Ratio Lower Upper
107 100
144 24.4 20.6 31.0
142 75.5 62.2 93.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/13/2022
Supervised By :mohammad ahmed 06/14/2022




#36
2-Methylnaphthalene
Concen: 32.400 ng/ul
RT: 12.269 min Scan# 1561
Delta R.T. -0.006 min
Lab File: BN020124.D
Acq: 11 Jun 2022 05:57




Tgt Ion:142 Resp: 873289
Ion Ratio Lower Upper
142 100
141 89.1 70.6 105.8





#37
1-Methylnaphthalene
Concen: 32.268 ng/ul
RT: 12.487 min Scan# 1598
Delta R.T. -0.006 min
Lab File: BN020124.D
Acq: 11 Jun 2022 05:57




Instrument :
BNA_N
ClientSampleId :
SLCS446




Tgt Ion:142 Resp: 88755
Ion Ratio Lower Upper
142 100
141 90.6 70.9 106.3
116 4.0 3.1 4.7

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/13/2022
Supervised By :mohammad ahmed 06/14/2022




#39
1,2,4,5-Tetrachlorobenzene
Concen: 32.191 ng/ul
RT: 12.645 min Scan# 1625
Delta R.T. -0.006 min
Lab File: BN020124.D
Acq: 11 Jun 2022 05:57




Tgt Ion:216 Resp: 406904
Ion Ratio Lower Upper
216 100
214 80.3 63.4 95.0
179 23.0 17.8 26.8
108 20.0 15.1 22.7





#40
Hexachlorocyclopentadiene
Concen: 28.562 ng/ul
RT: 12.628 min Scan# 1622
Delta R.T. -0.006 min
Lab File: BN020124.D
Acq: 11 Jun 2022 05:57




Instrument :
BNA_N
ClientSampleId :
SLCS446




Tgt Ion:237 Resp: 21962
Ion Ratio Lower Upper
237 100
235 63.0 48.8 73.2
272 11.9 9.3 13.9

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/13/2022
Supervised By :mohammad ahmed 06/14/2022




#41
2,4,6-Trichlorophenol
Concen: 33.516 ng/ul
RT: 12.881 min Scan# 1665
Delta R.T. -0.006 min
Lab File: BN020124.D
Acq: 11 Jun 2022 05:57




Tgt Ion:196 Resp: 292118
Ion Ratio Lower Upper
196 100
198 97.0 79.8 119.8
200 30.5 25.0 37.6





#42
2,4,5-Trichlorophenol
Concen: 33.859 ng/ul
RT: 12.951 min Scan# 1677
Delta R.T. -0.006 min
Lab File: BN020124.D
Acq: 11 Jun 2022 05:57




Instrument :
BNA_N
ClientSampleId :
SLCS446



Tgt Ion:196 Resp: 32683
Ion Ratio Lower Upper
196 100
198 90.4 77.6 116.4
200 30.8 24.4 36.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 06/13/2022
Supervised By :mohammad ahmed 06/14/2022



#43
1,1'-Biphenyl
Concen: 31.646 ng/ul
RT: 13.281 min Scan# 1733
Delta R.T. -0.006 min
Lab File: BN020124.D
Acq: 11 Jun 2022 05:57


Tgt Ion:154 Resp: 1161376
Ion Ratio Lower Upper
154 100
153 39.4 31.5 47.3
76 12.3 10.5 15.7





#44
2-Chloronaphthalene
Concen: 32.173 ng/ul
RT: 13.322 min Scan# 1740
Delta R.T. -0.006 min
Lab File: BN020124.D
Acq: 11 Jun 2022 05:57




Instrument :
BNA_N
ClientSampleId :
SLCS446




Tgt Ion:162 Resp: 895728
Ion Ratio Lower Upper
162 100
164 31.7 25.8 38.6
127 41.4 31.4 47.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/13/2022
Supervised By :mohammad ahmed 06/14/2022




#45
2-Nitroaniline
Concen: 36.038 ng/ul
RT: 13.522 min Scan# 1774
Delta R.T. -0.006 min
Lab File: BN020124.D
Acq: 11 Jun 2022 05:57




Tgt Ion: 65 Resp: 312138
Ion Ratio Lower Upper
65 100
92 68.3 54.1 81.1
138 104.4 77.0 115.4





#47
Dimethylphthalate
Concen: 32.254 ng/ul
RT: 13.904 min Scan# 1839
Delta R.T. -0.006 min
Lab File: BN020124.D
Acq: 11 Jun 2022 05:57




Instrument :
BNA_N
ClientSampleId :
SLCS446




Tgt Ion:163 Resp: 1170910
Ion Ratio Lower Upper
163 100
194 4.1 3.4 5.0
164 10.5 7.7 11.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/13/2022
Supervised By :mohammad ahmed 06/14/2022




#48
2,6-Dinitrotoluene
Concen: 35.627 ng/ul
RT: 14.022 min Scan# 1859
Delta R.T. -0.006 min
Lab File: BN020124.D
Acq: 11 Jun 2022 05:57




Tgt Ion:165 Resp: 252873
Ion Ratio Lower Upper
165 100
89 58.5 50.6 75.8
121 21.7 17.9 26.9





#50
Acenaphthylene
Concen: 31.646 ng/ul
RT: 14.169 min Scan# 1884
Delta R.T. 0.000 min
Lab File: BN020124.D
Acq: 11 Jun 2022 05:57




Instrument :
BNA_N
ClientSampleId :
SLCS446




Tgt Ion:152 Resp: 1465284
Ion Ratio Lower Upper
152 100
151 20.1 16.0 24.0
153 13.3 10.7 16.1

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/13/2022
Supervised By :mohammad ahmed 06/14/2022




#51
3-Nitroaniline
Concen: 35.402 ng/ul
RT: 14.345 min Scan# 1914
Delta R.T. -0.006 min
Lab File: BN020124.D
Acq: 11 Jun 2022 05:57




Tgt Ion:138 Resp: 265651
Ion Ratio Lower Upper
138 100
108 11.3 8.6 12.8
92 118.6 92.9 139.3





#52
Acenaphthene
Concen: 31.848 ng/u1
RT: 14.510 min Scan# 1942
Delta R.T. -0.006 min
Lab File: BN020124.D
Acq: 11 Jun 2022 05:57




Instrument :
BNA_N
ClientSampleId :
SLCS446




Tgt Ion:153 Resp: 963220
Ion Ratio Lower Upper
153 100
152 45.0 38.4 57.6
154 89.4 69.7 104.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/13/2022
Supervised By :mohammad ahmed 06/14/2022




#53
2,4-Dinitrophenol
Concen: 28.216 ng/u1
RT: 14.551 min Scan# 1949
Delta R.T. -0.006 min
Lab File: BN020124.D
Acq: 11 Jun 2022 05:57




Tgt Ion:184 Resp: 128222
Ion Ratio Lower Upper
184 100
63 78.1 55.9 83.9
154 68.0 53.1 79.7





#55
4-Nitrophenol
Concen: 32.628 ng/ul
RT: 14.657 min Scan# 1967
Delta R.T. 0.000 min
Lab File: BN020124.D
Acq: 11 Jun 2022 05:57




Instrument :
BNA_N
ClientSampleId :
SLCS446




Tgt Ion:109 Resp: 201250
Ion Ratio Lower Upper
109 100
139 120.6 100.1 150.1
65 130.0 104.6 157.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/13/2022
Supervised By :mohammad ahmed 06/14/2022




#56
Dibenzofuran
Concen: 31.664 ng/ul
RT: 14.845 min Scan# 1999
Delta R.T. -0.006 min
Lab File: BN020124.D
Acq: 11 Jun 2022 05:57




Tgt Ion:168 Resp: 1349158
Ion Ratio Lower Upper
168 100
139 40.1 31.1 46.7





#57
2,4-Dinitrotoluene
Concen: 35.329 ng/ul
RT: 14.810 min Scan# 1993
Delta R.T. 0.000 min
Lab File: BN020124.D
Acq: 11 Jun 2022 05:57




Instrument :
BNA_N
ClientSampleId :
SLCS446




Tgt Ion:165 Resp: 37129
Ion Ratio Lower Upper
165 100
63 45.2 36.3 54.5
182 2.7 2.3 3.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/13/2022
Supervised By :mohammad ahmed 06/14/2022




#58
2,3,4,6-Tetrachlorophenol
Concen: 33.527 ng/ul
RT: 15.069 min Scan# 2037
Delta R.T. -0.006 min
Lab File: BN020124.D
Acq: 11 Jun 2022 05:57




Tgt Ion:232 Resp: 254048
Ion Ratio Lower Upper
232 100
131 53.3 38.5 57.7
130 3.1 2.2 3.2
166 35.5 25.8 38.8





#59
Diethylphthalate
Concen: 32.999 ng/ul
RT: 15.269 min Scan# 2071
Delta R.T. -0.006 min
Lab File: BN020124.D
Acq: 11 Jun 2022 05:57




Instrument :
BNA_N
ClientSampleId :
SLCS446




Tgt Ion:149 Resp: 123336
Ion Ratio Lower Upper
149 100
177 20.9 17.8 26.8
150 12.9 10.5 15.7

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/13/2022
Supervised By :mohammad ahmed 06/14/2022




#61
Fluorene
Concen: 31.827 ng/ul
RT: 15.492 min Scan# 2109
Delta R.T. -0.006 min
Lab File: BN020124.D
Acq: 11 Jun 2022 05:57




Tgt Ion:166 Resp: 1055232
Ion Ratio Lower Upper
166 100
165 96.4 77.7 116.5
167 13.3 11.1 16.7





#62
4-Chlorophenyl-phenylether
Concen: 31.525 ng/ul
RT: 15.486 min Scan# 2108
Delta R.T. -0.006 min
Lab File: BN020124.D
Acq: 11 Jun 2022 05:57




Instrument :
BNA_N
ClientSampleId :
SLCS446




Tgt Ion:204 Resp: 481349
Ion Ratio Lower Upper
204 100
206 32.5 26.7 40.1
141 73.7 50.2 75.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/13/2022
Supervised By :mohammad ahmed 06/14/2022




#63
4-Nitroaniline
Concen: 38.279 ng/ul
RT: 15.510 min Scan# 2112
Delta R.T. 0.000 min
Lab File: BN020124.D
Acq: 11 Jun 2022 05:57




Tgt Ion:138 Resp: 268401
Ion Ratio Lower Upper
138 100
92 50.6 41.0 61.6
108 86.2 65.3 97.9





#66
4,6-Dinitro-2-methylphenol
Concen: 32.474 ng/ul
RT: 15.575 min Scan# 2123
Delta R.T. 0.000 min
Lab File: BN020124.D
Acq: 11 Jun 2022 05:57




Instrument :
BNA_N
ClientSampleId :
SLCS446




Tgt Ion:198 Resp: 197369
Ion Ratio Lower Upper
198 100
182 3.6 3.0 4.4
77 27.2 20.1 30.1

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/13/2022
Supervised By :mohammad ahmed 06/14/2022




#67
N-Nitrosodiphenylamine
Concen: 32.459 ng/ul
RT: 15.698 min Scan# 2144
Delta R.T. -0.006 min
Lab File: BN020124.D
Acq: 11 Jun 2022 05:57




Tgt Ion:169 Resp: 962987
Ion Ratio Lower Upper
169 100
168 69.7 52.4 78.6
167 36.1 27.5 41.3





#68
4-Bromophenyl-phenylether
Concen: 32.317 ng/ul
RT: 16.381 min Scan# 211
Delta R.T. -0.006 min
Lab File: BN020124.D
Acq: 11 Jun 2022 05:57




Instrument :
BNA_N
ClientSampleId :
SLCS446




Tgt Ion:248 Resp: 291252
Ion Ratio Lower Upper
248 100
250 99.7 77.8 116.8
141 98.4 62.6 93.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/13/2022
Supervised By :mohammad ahmed 06/14/2022




#69
Hexachlorobenzene
Concen: 32.009 ng/ul
RT: 16.504 min Scan# 2281
Delta R.T. 0.000 min
Lab File: BN020124.D
Acq: 11 Jun 2022 05:57



Tgt Ion:284 Resp: 331603
Ion Ratio Lower Upper
284 100
142 43.0 27.8 41.8#
249 30.8 25.7 38.5





#70

Atrazine

Concen: 33.990 ng/ul

RT: 16.651 min Scan# 21

Delta R.T. -0.006 min

Lab File: BN020124.D

Acq: 11 Jun 2022 05:57

Instrument :

BNA_N

ClientSampleId :





SLCS446

Tgt Ion:200 Resp: 35184
 Ion Ratio Lower Upper
 200 100
 173 27.5 20.5 30.7
 215 43.9 38.9 58.3

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 06/13/2022
 Supervised By :mohammad ahmed 06/14/2022



#71

Pentachlorophenol

Concen: 30.922 ng/ul




RT: 16.839 min Scan# 2338


Delta R.T. -0.006 min

Lab File: BN020124.D

Acq: 11 Jun 2022 05:57


Tgt Ion:266 Resp: 199521
 Ion Ratio Lower Upper
 266 100
 264 63.1 50.2 75.2
 268 63.1 49.3 73.9





#72
Phenanthrene
Concen: 32.085 ng/u1
RT: 17.228 min Scan# 2404
Delta R.T. -0.006 min
Lab File: BN020124.D
Acq: 11 Jun 2022 05:57




Instrument :
BNA_N
ClientSampleId :
SLCS446




Tgt Ion:178 Resp: 1752311
Ion Ratio Lower Upper
178 100
179 16.0 12.2 18.4
176 20.0 15.2 22.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/13/2022
Supervised By :mohammad ahmed 06/14/2022




#74
Anthracene
Concen: 31.869 ng/u1
RT: 17.322 min Scan# 2420
Delta R.T. -0.006 min
Lab File: BN020124.D
Acq: 11 Jun 2022 05:57




Tgt Ion:178 Resp: 1750120
Ion Ratio Lower Upper
178 100
179 15.7 13.0 19.4
176 19.0 15.5 23.3





#75
1,2,3,4-Tetrachlorobenzene
Concen: 31.360 ng/uL
RT: 13.245 min Scan# 1727
Delta R.T. -0.006 min
Lab File: BN020124.D
Acq: 11 Jun 2022 05:57




Instrument :
BNA_N
ClientSampleId :
SLCS446




Tgt Ion:216 Resp: 43459
Ion Ratio Lower Upper
216 100
214 77.7 63.2 94.8
218 47.0 37.9 56.9

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/13/2022
Supervised By :mohammad ahmed 06/14/2022




#76
Pentachlorobenzene
Concen: 32.954 ng/uL
RT: 14.769 min Scan# 1986
Delta R.T. -0.006 min
Lab File: BN020124.D
Acq: 11 Jun 2022 05:57



Tgt Ion:250 Resp: 403620
Ion Ratio Lower Upper
250 100
248 59.5 53.0 79.4
252 59.6 53.4 80.2





#77

Carbazole

Concen: 33.719 ng/ul

RT: 17.586 min Scan# 2465

Delta R.T. -0.006 min

Lab File: BN020124.D


Acq: 11 Jun 2022 05:57

Instrument :

BNA_N

ClientSampleId :

SLCS446



Tgt Ion:167 Resp: 168459

Ion Ratio Lower Upper

167 100




166 21.9 16.1 24.1

139 13.6 10.2 15.2

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 06/13/2022
Supervised By :mohammad ahmed 06/14/2022



#78

Di-n-butylphthalate

Concen: 34.535 ng/ul

RT: 18.157 min Scan# 2562

Delta R.T. -0.006 min

Lab File: BN020124.D

Acq: 11 Jun 2022 05:57




Tgt Ion:149 Resp: 2139410


Ion Ratio Lower Upper

149 100

150 9.2 7.1 10.7


104 5.4 4.1 6.1





#80
 Fluoranthene
 Concen: 34.460 ng/u1
 RT: 19.245 min Scan# 2747
 Delta R.T. -0.006 min
 Lab File: BN020124.D
 Acq: 11 Jun 2022 05:57

Instrument :
 BNA_N
 ClientSampleId :
 SLCS446






Tgt Ion:202 Resp: 2025615

Ion	Ratio	Lower	Upper
202	100		
101	12.1	8.9	13.3
100	9.2	6.7	10.1

Manual Integrations
 APPROVED




Reviewed By :Jagrut Upadhyay 06/13/2022
 Supervised By :mohammad ahmed 06/14/2022




#82
 Pyrene
 Concen: 33.744 ng/u1
 RT: 19.610 min Scan# 2809
 Delta R.T. -0.006 min
 Lab File: BN020124.D
 Acq: 11 Jun 2022 05:57

Tgt Ion:202 Resp: 2034955


Ion	Ratio	Lower	Upper
202	100		
101	13.9	9.7	14.5
100	11.4	8.0	12.0





#83
Butylbenzylphthalate
Concen: 36.611 ng/ul
RT: 20.510 min Scan# 2962
Delta R.T. -0.012 min
Lab File: BN020124.D
Acq: 11 Jun 2022 05:57




Instrument :
BNA_N
ClientSampleId :
SLCS446




Tgt Ion:149 Resp: 993610
Ion Ratio Lower Upper
149 100
91 73.0 58.8 88.2
206 19.1 18.1 27.1

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/13/2022
Supervised By :mohammad ahmed 06/14/2022




#84
3,3'-Dichlorobenzidine
Concen: 33.396 ng/ul
RT: 21.292 min Scan# 3095
Delta R.T. -0.006 min
Lab File: BN020124.D
Acq: 11 Jun 2022 05:57




Tgt Ion:252 Resp: 572642
Ion Ratio Lower Upper
252 100
254 66.3 50.8 76.2
126 13.7 8.2 12.4





#85
Benzo(a)anthracene
Concen: 32.338 ng/ul
RT: 21.357 min Scan# 3106
Delta R.T. -0.006 min
Lab File: BN020124.D
Acq: 11 Jun 2022 05:57




Instrument :
BNA_N
ClientSampleId :
SLCS446




Tgt Ion:228 Resp: 183323
Ion Ratio Lower Upper
228 100
229 19.6 16.3 24.5
226 27.4 23.2 34.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/13/2022
Supervised By :mohammad ahmed 06/14/2022




#86
Bis(2-ethylhexyl)phthalate
Concen: 35.851 ng/ul
RT: 21.298 min Scan# 3096
Delta R.T. -0.006 min
Lab File: BN020124.D
Acq: 11 Jun 2022 05:57




Tgt Ion:149 Resp: 1372853
Ion Ratio Lower Upper
149 100
167 28.0 22.5 33.7
279 3.7 3.5 5.3





#87
Chrysene
Concen: 33.334 ng/ul
RT: 21.410 min Scan# 3115
Delta R.T. -0.012 min
Lab File: BN020124.D
Acq: 11 Jun 2022 05:57




Instrument :
BNA_N
ClientSampleId :
SLCS446



Tgt Ion:228 Resp: 1848800
Ion Ratio Lower Upper
228 100
226 29.2 24.2 36.2
229 19.0 16.2 24.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 06/13/2022
Supervised By :mohammad ahmed 06/14/2022



#89
Di-n-octyl phthalate
Concen: 36.832 ng/ul
RT: 22.198 min Scan# 3249
Delta R.T. -0.012 min
Lab File: BN020124.D
Acq: 11 Jun 2022 05:57


Tgt Ion:149 Resp: 2488812





#90
Benzo(b)fluoranthene
Concen: 35.926 ng/ul
RT: 22.998 min Scan# 3385
Delta R.T. -0.006 min
Lab File: BN020124.D
Acq: 11 Jun 2022 05:57




Instrument :
BNA_N
ClientSampleId :
SLCS446




Tgt Ion:252 Resp: 1871998
Ion Ratio Lower Upper
252 100
253 20.9 16.5 24.7
125 11.6 7.8 11.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/13/2022
Supervised By :mohammad ahmed 06/14/2022




#91
Benzo(k)fluoranthene
Concen: 34.286 ng/ul m
RT: 23.045 min Scan# 3393
Delta R.T. -0.006 min
Lab File: BN020124.D
Acq: 11 Jun 2022 05:57




Tgt Ion:252 Resp: 1700651
Ion Ratio Lower Upper
252 100
253 23.5 18.2 27.4
125 12.6 8.2 12.4





#93
Benzo(a)pyrene
Concen: 33.287 ng/ul
RT: 23.604 min Scan# 3488
Delta R.T. -0.006 min
Lab File: BN020124.D
Acq: 11 Jun 2022 05:57




Instrument :
BNA_N
ClientSampleId :
SLCS446




Tgt Ion:252 Resp: 1678935
Ion Ratio Lower Upper
252 100
253 19.9 18.7 28.1
125 13.1 9.2 13.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/13/2022
Supervised By :mohammad ahmed 06/14/2022




#94
Indeno(1,2,3-cd)pyrene
Concen: 31.682 ng/ul
RT: 26.092 min Scan# 3911
Delta R.T. -0.006 min
Lab File: BN020124.D
Acq: 11 Jun 2022 05:57



Tgt Ion:276 Resp: 1644100
Ion Ratio Lower Upper
276 100
138 30.2 19.5 29.3#
277 25.6 21.4 32.0





#95
Dibenzo(a,h)anthracene
Concen: 31.728 ng/ul
RT: 26.109 min Scan# 3914
Delta R.T. -0.012 min
Lab File: BN020124.D
Acq: 11 Jun 2022 05:57

Instrument :

BNA_N





ClientSampleId :

SLCS446

Tgt Ion:278 Resp: 140632
Ion Ratio Lower Upper
278 100
139 19.4 13.8 20.8
279 22.9 18.3 27.5

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 06/13/2022
Supervised By :mohammad ahmed 06/14/2022



#96
Benzo(g,h,i)perylene
Concen: 31.974 ng/ul
RT: 26.827 min Scan# 4036
Delta R.T. -0.018 min
Lab File: BN020124.D
Acq: 11 Jun 2022 05:57

Tgt Ion:276 Resp: 1401210
Ion Ratio Lower Upper
276 100
138 25.5 17.0 25.6
277 23.6 17.8 26.6

