


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061122\
 Data File : BN020093.D
 Acq On : 10 Jun 2022 10:05
 Operator : CG/JU
 Sample : DFTPP243
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP243

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN060622.M
 Title : SVOA CALIBRATION
 Last Update : Mon Jun 06 15:14:59 2022



AutoFind: Scans 2474, 2475, 2476; Background Corrected with Scan 2467

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	40.0	168240	PASS
68	69	0.00	2	1.8	2883	PASS
69	198	0.00	100	38.0	159424	PASS
70	69	0.00	2	0.6	914	PASS
127	198	10	80	50.4	211563	PASS
197	198	0.00	2	0.2	757	PASS
198	198	100	100	100.0	420075	PASS
199	198	5	9	6.6	27643	PASS
275	198	10	60	25.0	105029	PASS
365	198	1	100	3.3	13934	PASS
441	198	0.01	100	13.7	57645	PASS
442	198	50	100	91.7	385024	PASS
443	442	15	24	19.0	73216	PASS

m/z	Abundance
40.00	1901.0
41.00	182.0
43.00	197.0
44.00	821.0
48.90	331.0
54.90	195.0
57.00	242.0
73.00	299.0
83.90	224.0
96.10	155.0
207.00	892.0
207.90	182.0
281.00	171.0

Instrument :

BNA_N

ClientSampleId :

DFTPP243

m/z	Abundance
35.90	154.0
37.00	805.0
38.00	2381.0
39.00	14650.0
40.00	1260.0
41.00	580.0
42.90	228.0
44.00	601.0
45.00	463.0
49.10	1536.0
50.00	53880.0
51.00	212224.0
52.00	10833.0
52.90	505.0
55.00	1155.0
55.90	6134.0
57.00	14135.0
57.90	596.0
58.80	155.0
61.00	3091.0
61.90	3100.0
63.00	8089.0
63.90	1133.0
65.00	4133.0
66.00	309.0
66.90	239.0
68.10	3573.0
68.90	201024.0
69.90	1061.0
73.00	1712.0
74.00	21280.0
75.00	32368.0
76.00	11136.0
77.00	224576.0
78.00	15467.0
79.00	15585.0
80.00	12020.0
81.00	16760.0
82.00	4230.0
83.00	3900.0
84.00	537.0
85.00	3249.0
86.00	4523.0
86.90	2400.0
88.00	794.0
89.00	384.0
90.90	3871.0
91.90	4474.0
93.00	28128.0
94.00	1783.0
94.80	519.0
95.90	1323.0
98.00	21648.0
99.00	18120.0
100.00	1535.0
101.00	9491.0
102.00	476.0
102.90	3097.0
104.00	6201.0
105.00	6072.0
106.00	2101.0
107.00	75352.0
108.00	11163.0
109.10	2395.0
110.00	132096.0
111.00	18896.0
112.00	2695.0
112.90	631.0
113.90	265.0
114.90	353.0
116.00	4433.0
117.00	64456.0
118.00	4545.0
119.00	585.0
120.00	1013.0
120.90	299.0
121.90	4439.0
123.00	7352.0
123.90	3364.0
125.00	3027.0
126.10	1536.0
127.00	256576.0
128.00	19808.0
128.90	101024.0
129.90	8563.0
131.00	1692.0
132.00	925.0
132.70	353.0
133.90	2966.0
135.00	8383.0
136.00	3539.0
137.00	4750.0

Instrument :
BNA_N
ClientSampleId :
DFTPP243

137.90	1025.0
138.90	570.0
140.00	1312.0
141.00	13074.0
142.00	3888.0
142.90	3129.0
143.90	738.0
145.00	798.0
146.00	2178.0
147.00	7161.0
147.90	15356.0
148.90	3233.0
149.90	914.0
151.10	1688.0
151.80	940.0
153.00	4176.0
153.90	3240.0
155.00	7695.0
156.00	10872.0
157.10	2349.0
157.90	2246.0
159.00	1704.0
160.00	3806.0
161.00	6196.0
162.00	1603.0
163.00	432.0
163.90	637.0
165.00	4731.0
166.00	3975.0
167.00	26040.0
168.00	13585.0
169.00	2188.0
170.00	846.0
170.90	992.0
171.90	2325.0
173.00	2822.0
173.90	5861.0
175.00	10150.0
176.00	3028.0
177.00	4869.0
178.00	1665.0
178.90	18632.0
180.00	13200.0
180.90	5786.0
181.90	1012.0
183.10	448.0
184.00	1689.0
185.00	8336.0
186.00	69472.0
187.00	21480.0
188.00	2053.0
188.90	4344.0
189.90	881.0
191.00	1933.0
192.00	5742.0
193.00	7151.0
193.90	1249.0
195.10	757.0
196.00	13065.0
197.90	476992.0
198.90	33320.0
199.90	2802.0
201.40	1913.0
202.90	3421.0
204.00	17728.0
205.00	30280.0
206.00	125648.0
207.00	17136.0
208.00	4530.0
209.00	1695.0
210.00	1850.0
210.90	4978.0
212.90	427.0
213.90	196.0
215.00	1205.0
216.00	2890.0
216.90	30936.0
217.90	4454.0
218.90	450.0
221.00	17712.0
221.90	3882.0
223.00	7327.0
224.00	67712.0
225.00	17704.0
226.10	2072.0
227.00	29368.0
228.00	4640.0
228.90	6297.0
230.00	881.0
231.00	2651.0
232.00	538.0
233.10	506.0
234.00	1860.0
235.00	2022.0
235.90	1588.0
237.00	2519.0

Instrument :
BNA_N
ClientSampleId :
DFTPP243

237.90	423.0
239.00	1224.0
239.90	1116.0
241.00	1748.0
241.90	3382.0
243.00	3795.0
244.00	49744.0
245.00	7290.0
246.00	11191.0
247.00	2032.0
248.00	521.0
248.90	1651.0
249.80	355.0
251.00	447.0
251.90	496.0
252.90	1251.0
255.00	267584.0
256.00	38584.0
257.00	3411.0
258.00	15251.0
259.00	2519.0
260.00	461.0
261.00	488.0
264.10	387.0
265.00	6512.0
265.90	929.0
269.80	190.0
270.70	623.0
272.00	860.0
273.00	7938.0
274.00	21952.0
275.00	110368.0
276.00	16608.0
277.00	10490.0
277.90	1679.0
278.90	346.0
281.10	203.0
281.70	242.0
283.00	1073.0
284.00	892.0
285.00	1543.0
285.90	437.0
288.90	479.0
289.90	299.0
291.00	171.0
292.00	526.0
293.00	2216.0
293.90	561.0
295.90	32856.0
297.00	4670.0
297.90	307.0
300.80	375.0
301.90	587.0
303.00	4104.0
303.90	1248.0
308.00	742.0
308.90	354.0
309.90	447.0
312.00	172.0
312.90	197.0
314.00	1444.0
314.90	3949.0
316.00	2092.0
317.00	464.0
321.00	1159.0
322.00	479.0
323.00	10017.0
324.00	2085.0
325.00	203.0
326.00	169.0
326.90	2029.0
327.80	914.0
332.00	839.0
332.90	1068.0
334.00	7119.0
335.00	1859.0
336.00	217.0
338.90	261.0
339.90	188.0
341.00	1234.0
341.90	384.0
345.90	2172.0
347.00	524.0
352.00	2867.0
353.00	2239.0
354.00	3729.0
355.00	655.0
358.90	211.0
364.10	264.0
364.90	13691.0
365.90	1860.0
370.10	303.0
370.90	775.0
372.00	5383.0
373.00	1106.0
376.80	160.0

Instrument :
BNA_N
ClientSampleId :
DFTPP243

382.90	1444.0
383.90	392.0
389.90	560.0
390.80	438.0
391.90	332.0
400.90	281.0
402.00	2522.0
402.90	3210.0
403.90	1368.0
404.90	215.0
414.90	247.0
420.90	2539.0
422.00	3250.0
423.00	19728.0
424.00	6021.0
424.90	816.0
439.70	311.0
441.00	54736.0
441.90	355968.0
443.00	74760.0
444.00	7801.0
445.00	393.0

Instrument :
BNA_N
ClientSampleId :
DFTPP243

m/z	Abundance
37.00	883.0
38.00	2163.0
39.00	13316.0
39.90	1407.0
40.90	392.0
43.10	187.0
44.00	708.0
45.00	434.0
49.00	1472.0
50.00	47896.0
51.00	186944.0
52.00	9333.0
52.90	312.0
55.00	1039.0
55.90	5122.0
57.00	12070.0
57.90	517.0
58.90	209.0
59.80	177.0
61.00	2596.0
62.00	2547.0
63.00	7360.0
64.00	1045.0
65.00	3754.0
66.00	231.0
66.90	253.0
68.00	3351.0
69.00	177792.0
70.00	1047.0
70.90	212.0
72.10	167.0
73.00	1702.0
74.00	18824.0
75.00	29736.0
76.00	9861.0
77.00	209024.0
78.00	14250.0
79.00	14313.0
80.00	11033.0
81.00	15887.0
82.00	4148.0
83.00	3679.0
83.90	482.0
85.00	3175.0
86.00	4533.0
86.90	2223.0
87.90	630.0
88.90	470.0
91.00	3865.0
92.00	3967.0
93.00	24912.0
94.00	1605.0
94.90	320.0
96.00	1051.0
97.00	383.0
98.00	19824.0
99.00	14953.0
99.90	1338.0
101.00	8425.0
102.00	536.0
103.00	2884.0
104.00	5854.0
105.00	5240.0
105.90	1770.0
107.00	70056.0
108.00	10485.0
109.00	2019.0
110.00	122296.0
110.90	19176.0
111.90	1948.0
113.00	724.0
113.80	181.0
114.90	425.0
116.00	3742.0
116.90	58192.0
117.90	3920.0
118.90	523.0
119.90	845.0
121.00	287.0
122.00	4753.0
123.00	7456.0
123.90	3174.0
124.90	2827.0
127.00	237696.0
128.00	18712.0
129.00	95144.0
129.90	7199.0
131.00	1418.0
132.00	863.0
132.90	399.0
133.90	3126.0
134.90	7740.0

Instrument :
BNA_N
ClientSampleId :
DFTPP243

135.90	3273.0
137.00	4115.0
138.00	817.0
138.90	590.0
139.90	1406.0
140.90	12191.0
142.00	4221.0
142.90	2620.0
144.00	701.0
145.00	843.0
145.90	2414.0
147.00	6344.0
148.00	14723.0
149.00	2708.0
150.00	978.0
151.00	1632.0
151.80	1001.0
153.00	3923.0
153.90	2835.0
155.00	6357.0
156.00	11001.0
157.00	1598.0
157.90	2163.0
158.90	1733.0
160.00	3368.0
161.00	6302.0
161.90	1786.0
162.90	417.0
164.00	534.0
165.00	4404.0
165.90	3528.0
167.00	24416.0
168.00	13288.0
169.00	1945.0
169.90	914.0
170.80	1009.0
171.90	2315.0
172.90	2574.0
174.00	5330.0
175.00	9323.0
175.90	2458.0
177.00	4691.0
177.90	1738.0
179.00	18216.0
180.00	12484.0
180.90	5766.0
181.90	777.0
182.90	373.0
183.90	1439.0
185.00	8490.0
186.00	67952.0
187.00	21056.0
188.00	2169.0
189.00	4338.0
190.00	895.0
191.00	1932.0
192.00	6022.0
193.00	6527.0
194.00	1198.0
194.90	701.0
196.00	12856.0
198.00	482944.0
198.90	30712.0
200.00	2338.0
201.60	1816.0
202.90	3189.0
204.00	17584.0
205.00	28960.0
206.00	129784.0
207.00	17936.0
208.00	4540.0
209.00	1553.0
210.00	2095.0
211.00	4739.0
212.80	347.0
214.00	181.0
215.00	1558.0
215.90	2907.0
216.90	33336.0
217.90	4063.0
218.90	212.0
221.00	16243.0
221.70	4141.0
222.90	7227.0
224.00	67880.0
225.00	17472.0
226.00	1925.0
227.00	28864.0
228.00	3584.0
229.00	6348.0
229.90	795.0
231.00	2654.0
232.00	421.0
232.90	528.0
234.00	1719.0
235.00	2220.0

Instrument :
BNA_N
ClientSampleId :
DFTPP243

235.90	1638.0
236.90	2606.0
238.00	440.0
238.90	1290.0
240.00	1023.0
240.90	1454.0
242.00	3322.0
243.00	3548.0
244.00	51696.0
245.00	6844.0
246.00	10986.0
247.00	2138.0
248.00	513.0
248.90	1778.0
249.90	289.0
250.80	498.0
251.90	611.0
253.00	1239.0
255.00	277504.0
256.00	38392.0
257.00	3355.0
257.90	15756.0
258.90	2434.0
259.90	402.0
260.90	498.0
262.90	178.0
264.00	593.0
265.00	6985.0
265.90	705.0
267.10	182.0
269.90	311.0
270.90	763.0
272.10	998.0
273.00	8393.0
274.00	22112.0
275.00	125104.0
276.00	15608.0
277.00	10028.0
278.00	1918.0
279.00	487.0
281.00	166.0
281.90	416.0
282.90	1204.0
284.00	1015.0
285.00	1904.0
286.00	308.0
287.90	215.0
288.90	552.0
290.10	373.0
291.00	203.0
292.00	662.0
292.90	2391.0
294.00	649.0
295.10	882.0
296.00	35608.0
296.90	4930.0
297.90	299.0
300.90	407.0
302.00	673.0
303.00	3759.0
303.90	1207.0
307.90	463.0
309.00	410.0
310.00	539.0
313.00	290.0
314.00	1692.0
314.90	4031.0
316.00	2084.0
317.00	422.0
321.00	1109.0
322.10	672.0
323.00	10261.0
324.00	2342.0
324.80	187.0
326.00	263.0
326.90	2124.0
327.90	1090.0
332.00	1236.0
332.90	976.0
334.00	6126.0
335.00	1757.0
336.10	256.0
339.10	172.0
341.00	1273.0
341.90	327.0
345.90	2398.0
346.90	450.0
351.00	245.0
352.00	3125.0
352.90	2636.0
354.00	3806.0
355.00	745.0
358.90	317.0
364.90	16432.0
365.90	1999.0
367.10	163.0

Instrument :
BNA_N
ClientSampleId :
DFTPP243

369.90	323.0
370.90	802.0
372.00	5889.0
373.00	1244.0
374.00	258.0
382.90	1381.0
383.90	614.0
384.90	189.0
390.00	639.0
391.00	613.0
391.90	480.0
401.00	353.0
402.00	2257.0
403.00	3905.0
404.10	1490.0
404.90	264.0
421.00	2829.0
422.00	3307.0
423.00	23912.0
424.00	6033.0
424.90	681.0
441.00	67992.0
442.00	464320.0
443.00	84520.0
444.00	6071.0
445.00	503.0

Instrument :
BNA_N
ClientSampleId :
DFTPP243

m/z	Abundance
36.90	359.0
38.00	1168.0
39.00	7131.0
39.90	1249.0
41.00	399.0
43.10	231.0
44.00	867.0
44.90	215.0
49.00	904.0
50.00	25904.0
51.00	105552.0
52.00	4866.0
53.00	186.0
54.90	750.0
56.00	3192.0
57.00	7226.0
57.90	296.0
59.00	182.0
60.90	1143.0
62.00	1839.0
63.00	3952.0
63.90	616.0
65.00	2007.0
67.00	199.0
68.00	1726.0
68.90	99456.0
70.00	633.0
71.00	194.0
73.00	1135.0
74.00	10495.0
75.00	16196.0
76.00	5780.0
77.00	113936.0
78.00	7759.0
79.00	7830.0
80.00	6200.0
81.00	8935.0
82.00	1929.0
82.90	1939.0
83.90	310.0
84.90	1733.0
86.00	2425.0
86.90	1135.0
87.90	373.0
88.90	308.0
90.90	2100.0
92.00	2006.0
93.00	13784.0
94.00	909.0
95.10	171.0
96.00	808.0
96.90	257.0
98.00	11232.0
99.00	9235.0
100.00	695.0
100.90	5061.0
101.90	259.0
102.90	1819.0
103.90	3180.0
104.90	2969.0
106.00	1103.0
107.00	40032.0
108.00	5771.0
108.90	1036.0
110.00	69024.0
110.90	10832.0
112.00	1126.0
112.90	427.0
114.90	206.0
115.90	2249.0
117.00	32912.0
118.00	1933.0
118.90	300.0
119.80	485.0
121.00	218.0
121.90	2952.0
122.90	4097.0
123.90	1778.0
124.90	1549.0
126.10	894.0
127.00	140416.0
128.00	10186.0
128.90	53400.0
130.00	4643.0
131.00	819.0
131.90	296.0
132.90	392.0
133.90	1535.0
135.00	4778.0
136.00	1848.0
137.00	2391.0
138.00	528.0

Instrument :
BNA_N
ClientSampleId :
DFTPP243

139.00	248.0
140.00	782.0
141.00	7074.0
141.90	2308.0
142.90	1601.0
144.00	488.0
144.90	388.0
146.00	1300.0
147.00	4015.0
148.00	8873.0
149.00	1721.0
150.00	585.0
151.00	918.0
151.90	508.0
152.90	2328.0
153.90	1786.0
155.00	4102.0
156.00	6473.0
157.00	984.0
157.80	1670.0
159.00	1135.0
159.90	2347.0
161.00	2812.0
162.00	1090.0
163.00	410.0
163.90	339.0
164.90	2763.0
166.00	2299.0
167.00	15540.0
168.00	7430.0
168.90	1232.0
169.90	482.0
171.00	642.0
172.00	1115.0
173.00	1509.0
174.00	3212.0
175.00	5657.0
176.00	1698.0
177.00	2584.0
178.00	1013.0
178.90	11210.0
179.90	7661.0
180.90	3539.0
181.90	655.0
183.10	387.0
183.90	882.0
185.00	5686.0
186.00	42392.0
187.00	11647.0
187.90	1447.0
188.90	2729.0
190.00	444.0
190.90	1470.0
191.90	3566.0
193.00	4520.0
194.00	919.0
195.00	551.0
196.00	8393.0
197.00	2271.0
197.90	300288.0
199.00	18896.0
200.00	1368.0
201.40	1499.0
202.90	2191.0
204.00	11599.0
205.00	18184.0
206.00	75600.0
207.00	10726.0
208.00	2857.0
208.90	1125.0
210.10	1348.0
211.00	2911.0
213.00	205.0
215.00	797.0
216.00	1669.0
216.90	21144.0
217.90	2648.0
218.90	339.0
220.10	668.0
221.00	9641.0
221.80	2732.0
223.00	4443.0
224.00	43336.0
225.00	10441.0
226.00	1401.0
226.90	17744.0
227.90	2445.0
229.00	4269.0
230.00	481.0
231.00	1587.0
231.90	198.0
232.90	366.0
233.90	1526.0
235.00	1170.0
235.90	916.0
236.90	1657.0

Instrument :
BNA_N
ClientSampleId :
DFTPP243

238.00	255.0
238.90	744.0
240.00	702.0
241.00	849.0
242.00	2485.0
243.00	2547.0
244.00	34440.0
245.00	5063.0
246.00	7028.0
246.90	1376.0
247.90	473.0
248.90	1380.0
249.90	179.0
250.90	293.0
251.90	415.0
252.90	638.0
254.10	2033.0
255.00	170304.0
256.00	24816.0
257.00	2157.0
257.90	10270.0
258.90	1779.0
259.90	372.0
260.90	242.0
264.10	459.0
265.00	4366.0
265.90	630.0
266.80	168.0
270.00	152.0
270.90	381.0
271.90	535.0
273.00	5206.0
273.90	15068.0
275.00	79616.0
276.00	10248.0
276.90	7711.0
277.90	1269.0
279.00	229.0
280.90	210.0
281.80	244.0
283.00	807.0
284.00	556.0
285.00	1235.0
286.00	274.0
288.00	151.0
288.90	240.0
289.80	240.0
291.00	228.0
291.90	400.0
292.90	1566.0
294.10	441.0
294.70	590.0
296.00	24712.0
297.00	3267.0
298.00	255.0
300.90	289.0
301.90	432.0
303.00	2839.0
304.00	690.0
308.00	422.0
309.00	174.0
309.90	283.0
313.00	157.0
313.90	1198.0
314.90	2745.0
315.90	1331.0
317.00	207.0
321.00	817.0
322.10	371.0
323.00	7606.0
324.00	1294.0
324.80	159.0
325.90	174.0
326.90	1504.0
327.90	765.0
331.90	577.0
332.90	700.0
334.00	5234.0
335.00	1516.0
340.90	913.0
341.80	201.0
345.90	1600.0
346.80	315.0
352.00	2390.0
353.00	1818.0
354.00	2627.0
355.00	473.0
358.90	223.0
364.90	11679.0
365.90	1477.0
370.90	514.0
372.00	4754.0
373.00	1212.0
382.90	992.0
383.90	283.0
389.90	611.0

Instrument :
BNA_N
ClientSampleId :
DFTPP243

391.00	474.0
392.00	293.0
400.90	302.0
402.00	1770.0
403.00	2180.0
404.00	1103.0
405.10	193.0
420.90	2514.0
422.00	2521.0
423.00	17640.0
424.00	4031.0
425.00	619.0
440.10	495.0
441.00	50208.0
442.00	334784.0
443.00	60368.0
444.00	5850.0
445.10	352.0

Instrument :
BNA_N
ClientSampleId :
DFTPP243