

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061220\
 Data File : BN010841.D
 Acq On : 12 Jun 2020 09:51
 Operator : CG/JU
 Sample : PB129499BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB129499BL

Quant Time: Jun 12 11:15:46 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 12 11:11:17 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.56	152	2602	0.40	ng	0.00
7) Naphthalene-d8	10.32	136	8882	0.40	ng	0.01
13) Acenaphthene-d10	14.19	164	4505	0.40	ng	0.00
19) Phenanthrene-d10	16.94	188	9890	0.40	ng	0.02
27) Chrysene-d12	21.13	240	9955	0.40	ng	0.00
34) Perylene-d12	23.29	264	7997	0.40	ng	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Fluorophenol	5.21	112	2521	0.50	ng	0.00
5) Phenol-d6	6.76	99	2486	0.42	ng	0.00
8) Nitrobenzene-d5	8.71	82	2706	0.41	ng	0.01
11) 2-Methylnaphthalene-d10	11.91	152	4889	0.35	ng	0.01
14) 2,4,6-Tribromophenol	15.70	330	633	0.39	ng	0.02
15) 2-Fluorobiphenyl	12.80	172	7181	0.40	ng	0.00
25) Fluoranthene-d10	18.97	212	10759	0.39	ng	0.00
29) Terphenyl-d14	19.58	244	10120	0.43	ng	0.00

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061220\
Data File : BN010841.D
Acq On : 12 Jun 2020 09:51
Operator : CG/JU
Sample : PB129499BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB129499BL

Quant Time: Jun 12 11:15:46 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 12 11:11:17 2020
Response via : Initial Calibration

