

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061220\
 Data File : BN010840.D
 Acq On : 12 Jun 2020 09:14
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 12 11:12:49 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 12 11:11:17 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	176#	0.00
2	1,4-Dioxane	0.392	0.417	-6.4	179#	0.00
3	n-Nitrosodimethylamine	0.228	0.225	1.3	198#	0.00
4 S	2-Fluorophenol	0.772	1.063	-37.7#	237#	0.00
5 S	Phenol-d6	0.910	1.268	-39.3#	246#	0.00
6	bis(2-Chloroethyl)ether	0.887	1.099	-23.9	218#	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	194#	0.00
8 S	Nitrobenzene-d5	0.296	0.309	-4.4	195#	0.01
9	Naphthalene	1.005	1.063	-5.8	196#	0.01
10	Hexachlorobutadiene	0.248	0.244	1.6	177#	0.00
11 SURR	2-Methylnaphthalene-d10	0.620	0.711	-14.7	212#	0.00
12	2-Methylnaphthalene	0.673	0.721	-7.1	203#	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	201#	0.01
14 S	2,4,6-Tribromophenol	0.143	0.159	-11.2	231#	0.01
15 S	2-Fluorobiphenyl	1.578	1.614	-2.3	193#	0.01
16	Acenaphthylene	1.639	1.777	-8.4	216#	0.01
17	Acenaphthene	1.117	1.176	-5.3	204#	0.00
18	Fluorene	1.441	1.500	-4.1	204#	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	202#	0.00
20	4-Bromophenyl-phenylether	0.240	0.233	2.9	191#	0.00
21	Hexachlorobenzene	0.251	0.261	-4.0	200#	0.01
22	Pentachlorophenol	0.084	0.097	-15.5	240#	0.01
23	Phenanthrene	1.093	1.164	-6.5	210#	0.00
24	Anthracene	0.912	1.006	-10.3	224#	0.01
25 SURR	Fluoranthene-d10	1.104	1.324	-19.9	239#	0.00
26	Fluoranthene	1.240	1.382	-11.5	225#	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	233#	0.00
28	Pvrene	1.323	1.382	-4.5	234#	0.00
29 S	Terphenyl-d14	0.936	0.959	-2.5	228#	0.00
30	Benzo(a)anthracene	1.176	1.342	-14.1	268#	0.01
31	Chrysene	1.370	1.434	-4.7	225#	0.00
32	Bis(2-ethylhexyl)phthalate	0.392	0.584	-49.0#	355#	0.00
33	Indeno(1,2,3-cd)pyrene	1.531	1.643	-7.3	239#	0.01
34 I	Perylene-d12	1.000	1.000	0.0	265#	0.00
35	Benzo(b)fluoranthene	1.360	1.309	3.8	246#	0.00
36	Benzo(k)fluoranthene	1.449	1.335	7.9	232#	0.00
37 C	Benzo(a)pyrene	1.167	1.201	-2.9	272#	0.00
38	Dibenzo(a,h)anthracene	1.330	1.230	7.5	241#	0.00
39	Benzo(a,h,i)perylene	1.391	1.252	10.0	224#	0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061220\
Data File : BN010840.D
Acq On : 12 Jun 2020 09:14
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 12 11:12:49 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 12 11:11:17 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			