

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061220\
 Data File : BN010849.D
 Acq On : 12 Jun 2020 18:20
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 12 18:51:50 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 12 11:11:17 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	133	0.00
2	1,4-Dioxane	0.392	0.427	-8.9	138	0.00
3	n-Nitrosodimethylamine	0.228	0.215	5.7	142	0.00
4 S	2-Fluorophenol	0.772	0.980	-26.9#	165#	0.00
5 S	Phenol-d6	0.910	1.133	-24.5	165#	0.00
6	bis(2-Chloroethyl)ether	0.887	1.065	-20.1	159#	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	142	0.00
8 S	Nitrobenzene-d5	0.296	0.306	-3.4	141	0.01
9	Naphthalene	1.005	1.082	-7.7	146	0.01
10	Hexachlorobutadiene	0.248	0.250	-0.8	133	0.00
11 SURR	2-Methylnaphthalene-d10	0.620	0.657	-6.0	144	0.00
12	2-Methylnaphthalene	0.673	0.710	-5.5	146	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	145	0.00
14 S	2,4,6-Tribromophenol	0.143	0.144	-0.7	151#	0.00
15 S	2-Fluorobiphenyl	1.578	1.619	-2.6	140	0.00
16	Acenaphthylene	1.639	1.717	-4.8	151#	0.00
17	Acenaphthene	1.117	1.160	-3.8	145	0.00
18	Fluorene	1.441	1.460	-1.3	143	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	142	0.00
20	4-Bromophenyl-phenylether	0.240	0.231	3.7	133	0.00
21	Hexachlorobenzene	0.251	0.259	-3.2	140	0.00
22	Pentachlorophenol	0.084	0.080	4.8	140	0.00
23	Phenanthrene	1.093	1.122	-2.7	142	0.00
24	Anthracene	0.912	0.941	-3.2	148	0.00
25 SURR	Fluoranthene-d10	1.104	1.146	-3.8	146	0.00
26	Fluoranthene	1.240	1.280	-3.2	147	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	152#	0.00
28	Pvrene	1.323	1.348	-1.9	149	0.00
29 S	Terphenyl-d14	0.936	0.948	-1.3	147	0.00
30	Benzo(a)anthracene	1.176	1.267	-7.7	165#	0.00
31	Chrysene	1.370	1.367	0.2	140	0.00
32	Bis(2-ethylhexyl)phthalate	0.392	0.522	-33.2#	207#	0.00
33	Indeno(1,2,3-cd)pyrene	1.531	1.713	-11.9	162#	0.00
34 I	Perylene-d12	1.000	1.000	0.0	169#	0.00
35	Benzo(b)fluoranthene	1.360	1.341	1.4	161#	0.00
36	Benzo(k)fluoranthene	1.449	1.396	3.7	155#	0.00
37 C	Benzo(a)pyrene	1.167	1.206	-3.3	174#	0.00
38	Dibenzo(a,h)anthracene	1.330	1.267	4.7	159#	0.00
39	Benzo(a,h,i)perylene	1.391	1.334	4.1	153#	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061220\
Data File : BN010849.D
Acq On : 12 Jun 2020 18:20
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 12 18:51:50 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 12 11:11:17 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			