

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061220\
 Data File : BN010851.D
 Acq On : 12 Jun 2020 20:07
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 12 20:50:28 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 12 11:11:17 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	136	0.00
2	1,4-Dioxane	0.392	0.423	-7.9	141	0.00
3	n-Nitrosodimethylamine	0.228	0.224	1.8	153#	0.02
4 S	2-Fluorophenol	0.772	0.943	-22.2	163#	0.00
5 S	Phenol-d6	0.910	1.137	-24.9	171#	0.00
6	bis(2-Chloroethyl)ether	0.887	1.053	-18.7	162#	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	146	0.00
8 S	Nitrobenzene-d5	0.296	0.302	-2.0	144	0.01
9	Naphthalene	1.005	1.052	-4.7	146	0.01
10	Hexachlorobutadiene	0.248	0.247	0.4	135	0.00
11 SURR	2-Methylnaphthalene-d10	0.620	0.651	-5.0	147	0.00
12	2-Methylnaphthalene	0.673	0.705	-4.8	149	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	148	0.00
14 S	2,4,6-Tribromophenol	0.143	0.142	0.7	151#	0.01
15 S	2-Fluorobiphenyl	1.578	1.622	-2.8	142	0.01
16	Acenaphthylene	1.639	1.693	-3.3	151#	0.01
17	Acenaphthene	1.117	1.162	-4.0	148	0.00
18	Fluorene	1.441	1.475	-2.4	147	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	139	0.00
20	4-Bromophenyl-phenylether	0.240	0.236	1.7	134	0.00
21	Hexachlorobenzene	0.251	0.269	-7.2	142	0.01
22	Pentachlorophenol	0.084	0.085	-1.2	146	0.01
23	Phenanthrene	1.093	1.140	-4.3	142	0.00
24	Anthracene	0.912	0.972	-6.6	150	0.01
25 SURR	Fluoranthene-d10	1.104	1.166	-5.6	146	0.00
26	Fluoranthene	1.240	1.278	-3.1	143	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	148	0.00
28	Pvrene	1.323	1.369	-3.5	148	0.00
29 S	Terphenyl-d14	0.936	0.953	-1.8	144	0.00
30	Benzo(a)anthracene	1.176	1.241	-5.5	158#	0.00
31	Chrysene	1.370	1.427	-4.2	143	0.00
32	Bis(2-ethylhexyl)phthalate	0.392	0.486	-24.0	188#	0.00
33	Indeno(1,2,3-cd)pyrene	1.531	1.710	-11.7	158#	0.00
34 I	Perylene-d12	1.000	1.000	0.0	166#	0.00
35	Benzo(b)fluoranthene	1.360	1.298	4.6	153#	0.00
36	Benzo(k)fluoranthene	1.449	1.410	2.7	154#	0.00
37 C	Benzo(a)pyrene	1.167	1.166	0.1	165#	0.00
38	Dibenzo(a,h)anthracene	1.330	1.281	3.7	157#	0.00
39	Benzo(a,h,i)perylene	1.391	1.315	5.5	148	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061220\
Data File : BN010851.D
Acq On : 12 Jun 2020 20:07
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 12 20:50:28 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 12 11:11:17 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		