

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061220\
 Data File : BN010851.D
 Acq On : 12 Jun 2020 20:07
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 12 20:50:28 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 12 11:11:17 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	136	0.00
2	1,4-Dioxane	0.400	0.432	-8.0	141	0.00
3	n-Nitrosodimethylamine	0.400	0.394	1.5	153	0.02
4 S	2-Fluorophenol	0.400	0.489	-22.2	163	0.00
5 S	Phenol-d6	0.400	0.500	-25.0	171	0.00
6	bis(2-Chloroethyl)ether	0.400	0.475	-18.7	162	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	146	0.00
8 S	Nitrobenzene-d5	0.400	0.409	-2.2	144	0.01
9	Naphthalene	0.400	0.419	-4.7	146	0.01
10	Hexachlorobutadiene	0.400	0.398	0.5	135	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.420	-5.0	147	0.00
12	2-Methylnaphthalene	0.400	0.419	-4.7	149	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	148	0.00
14 S	2,4,6-Tribromophenol	0.400	0.397	0.8	151	0.01
15 S	2-Fluorobiphenyl	0.400	0.411	-2.7	142	0.01
16	Acenaphthylene	0.400	0.413	-3.2	151	0.01
17	Acenaphthene	0.400	0.416	-4.0	148	0.00
18	Fluorene	0.400	0.409	-2.2	147	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	139	0.00
20	4-Bromophenyl-phenylether	0.400	0.392	2.0	134	0.00
21	Hexachlorobenzene	0.400	0.429	-7.2	142	0.01
22	Pentachlorophenol	0.400	0.405	-1.3	146	0.01
23	Phenanthrene	0.400	0.417	-4.2	142	0.00
24	Anthracene	0.400	0.426	-6.5	150	0.01
25 SURR	Fluoranthene-d10	0.400	0.423	-5.7	146	0.00
26	Fluoranthene	0.400	0.412	-3.0	143	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	148	0.00
28	Pvrene	0.400	0.414	-3.5	148	0.00
29 S	Terphenyl-d14	0.400	0.407	-1.7	144	0.00
30	Benzo(a)anthracene	0.400	0.422	-5.5	158	0.00
31	Chrysene	0.400	0.416	-4.0	143	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.496	-24.0	188	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.447	-11.7	158	0.00
34 I	Perylene-d12	0.400	0.400	0.0	166	0.00
35	Benzo(b)fluoranthene	0.400	0.382	4.5	153	0.00
36	Benzo(k)fluoranthene	0.400	0.389	2.8	154	0.00
37 C	Benzo(a)pyrene	0.400	0.400	0.0	165	0.00
38	Dibenzo(a,h)anthracene	0.400	0.385	3.8	157	0.00
39	Benzo(a,h,i)perylene	0.400	0.378	5.5	148	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061220\
Data File : BN010851.D
Acq On : 12 Jun 2020 20:07
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 12 20:50:28 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 12 11:11:17 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			