

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061224\
 Data File : BN031933.D
 Acq On : 12 Jun 2024 09:29
 Operator : MA/JU
 Sample : PB161332BL
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK332

Quant Time: Jun 12 22:50:49 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN052924.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jun 11 22:28:10 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.652	152	372934	20.000	ng/u1	0.00
20) Naphthalene-d8	10.434	136	1707187	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.298	164	1165749	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.051	188	2374802	20.000	ng/u1	0.00
79) Chrysene-d12	21.292	240	2015982	20.000	ng/u1	0.00
88) Perylene-d12	24.215	264	1834569	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.170	96	46858	5.198	ng/uL	0.00
4) Pyridine-d5	3.575	84	617488	24.334	ng/u1	0.00
7) Phenol-d5	6.828	99	889690	27.828	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.993	67	518285	27.715	ng/u1	-0.01
11) 2-Chlorophenol-d4	7.187	132	716351	29.365	ng/u1	0.00
15) 4-Methylphenol-d8	8.363	113	747698	28.690	ng/u1	0.00
21) Nitrobenzene-d5	8.810	128	380550	29.154	ng/u1	0.00
24) 2-Nitrophenol-d4	9.528	143	403284	29.909	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.063	165	691130	27.456	ng/u1	0.00
31) 4-Chloroaniline-d4	10.581	131	1059984	28.361	ng/u1	0.00
46) Dimethylphthalate-d6	13.716	166	2279080	28.063	ng/u1	0.00
49) Acenaphthylene-d8	13.992	160	2605089	27.897	ng/u1	0.00
54) 4-Nitrophenol-d4	14.516	143	421667	25.932	ng/u1	0.00
60) Fluorene-d10	15.298	176	1932234	28.119	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.428	200	297762	24.945	ng/u1	0.00
73) Anthracene-d10	17.151	188	3004278	29.549	ng/u1	0.00
81) Pyrene-d10	19.451	212	3344647	31.003	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.015	264	2465927	28.149	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061224\
Data File : BN031933.D
Acq On : 12 Jun 2024 09:29
Operator : MA/JU
Sample : PB161332BL
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK332

Quant Time: Jun 12 22:50:49 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN052924.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Jun 11 22:28:10 2024
Response via : Initial Calibration

