

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061224\
 Data File : BN031913.D
 Acq On : 11 Jun 2024 16:28
 Operator : MA/JU
 Sample : P2642-18
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 A4DG2

Quant Time: Jun 11 17:11:51 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN052924.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed May 29 15:13:07 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.657	152	272438	20.000	ng/u1	-0.03
20) Naphthalene-d8	10.445	136	1290767	20.000	ng/u1	-0.02
38) Acenaphthene-d10	14.310	164	842194	20.000	ng/u1	-0.02
64) Phenanthrene-d10	17.057	188	1761742	20.000	ng/u1	-0.01
79) Chrysene-d12	21.304	240	1498517	20.000	ng/u1	0.00
88) Perylene-d12	24.221	264	1684547	20.000	ng/u1	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.175	96	29881	4.538	ng/uL	-0.02
4) Pyridine-d5	3.604	84	36107	1.948	ng/u1	0.00
7) Phenol-d5	6.840	99	737712	31.586	ng/u1	-0.01
9) Bis-(2-Chloroethyl)eth...	7.004	67	422967	30.961	ng/u1	-0.02
11) 2-Chlorophenol-d4	7.192	132	567121	31.824	ng/u1	-0.02
15) 4-Methylphenol-d8	8.381	113	614389	32.271	ng/u1	0.00
21) Nitrobenzene-d5	8.822	128	310055	31.416	ng/u1	-0.02
24) 2-Nitrophenol-d4	9.534	143	322204	31.605	ng/u1	-0.02
28) 2,4-Dichlorophenol-d3	10.069	165	641749	33.719	ng/u1	-0.02
31) 4-Chloroaniline-d4	10.586	131	276956	9.801	ng/u1	-0.02
46) Dimethylphthalate-d6	13.733	166	2211113	37.686	ng/u1	-0.01
49) Acenaphthylene-d8	14.004	160	2364002	35.041	ng/u1	-0.01
54) 4-Nitrophenol-d4	14.539	143	391322	33.311	ng/u1	0.01
60) Fluorene-d10	15.304	176	1845868	37.182	ng/u1	-0.02
65) 4,6-Dinitro-2-methylph...	15.439	200	301182	34.011	ng/u1	0.00
73) Anthracene-d10	17.157	188	2956707	39.201	ng/u1	-0.01
81) Pyrene-d10	19.457	212	3389146	42.264	ng/u1	-0.01
92) Benzo(a)pyrene-d12	24.027	264	3206706	39.866	ng/u1	-0.01
Target Compounds						
					Qvalue	
8) Phenol	6.869	94	3882958	163.360	ng/u1	96
10) Bis(2-Chloroethyl)ether	7.098	93	1050259	53.437	ng/u1	97
12) 2-Chlorophenol	7.228	128	1942028	104.137	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.475	70	2200293	147.165	ng/u1	96
19) Hexachloroethane	8.728	117	1089940	138.017	ng/u1	94
22) Nitrobenzene	8.863	77	1167648	49.658	ng/u1	97
25) 2-Nitrophenol	9.569	139	821933	73.794	ng/u1	97
27) Bis(2-Chloroethoxy)met...	9.869	93	2308884	76.805	ng/u1	98
29) 2,4-Dichlorophenol	10.098	162	1272149	67.695	ng/u1	97
30) Naphthalene	10.498	128	10725364	160.554	ng/u1	99
35) 4-Chloro-3-methylphenol	11.739	107	2621446	119.341	ng/u1	99
41) 2,4,6-Trichlorophenol	12.733	196	1252292	82.835	ng/u1	97
42) 2,4,5-Trichlorophenol	12.804	196	2794137	170.080	ng/u1	97
44) 2-Chloronaphthalene	13.175	162	3625718	78.145	ng/u1	98
48) 2,6-Dinitrotoluene	13.910	165	3028374	253.903	ng/u1	99
50) Acenaphthylene	14.033	152	12012218	157.549	ng/u1	99
52) Acenaphthene	14.374	153	6248625	123.863	ng/u1	100
55) 4-Nitrophenol	14.563	109	1671890	191.852	ng/u1	85
57) 2,4-Dinitrotoluene	14.698	165	2497807	142.723	ng/u1	100
59) Diethylphthalate	15.157	149	10964186	177.078	ng/u1	99
61) Fluorene	15.363	166	2554180	46.277	ng/u1	98
62) 4-Chlorophenyl-phenyle...	15.363	204	3760534	140.669	ng/u1	93
69) Hexachlorobenzene	16.363	284	4142899	214.141	ng/u1	99
71) Pentachlorophenol	16.716	266	2578172	208.289	ng/u1	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061224\
 Data File : BN031913.D
 Acq On : 11 Jun 2024 16:28
 Operator : MA/JU
 Sample : P2642-18
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 A4DG2

Quant Time: Jun 11 17:11:51 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN052924.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed May 29 15:13:07 2024
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
72) Phenanthrene	17.104	178	5905065	65.798	ng/ul	98
74) Anthracene	17.198	178	7308604	80.370	ng/ul	98
78) Di-n-butylphthalate	18.033	149	15201039	142.493	ng/ul#	94
80) Fluoranthene	19.133	202	14724618	153.648	ng/ul	98
82) Pyrene	19.492	202	8877856	90.291	ng/ul	96
85) Benzo(a)anthracene	21.286	228	11486555	114.519	ng/ul	97
86) Bis(2-ethylhexyl)phtha...	21.215	149	11728857	173.256	ng/ul#	98
87) Chrysene	21.356	228	13064346	139.443	ng/ul	96
91) Benzo(k)fluoranthene	23.386	252	16044536	162.389	ng/ul	97
95) Dibenzo(a,h)anthracene	27.603	278	9777140	110.201	ng/ul	95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061224\
Data File : BN031913.D
Acq On : 11 Jun 2024 16:28
Operator : MA/JU
Sample : P2642-18
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
A4DG2

Quant Time: Jun 11 17:11:51 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN052924.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed May 29 15:13:07 2024
Response via : Initial Calibration









#8
Phenol
Concen: 163.360 ng/ul
RT: 6.869 min Scan# 60
Delta R.T. -0.012 min
Lab File: BN031913.D
Acq: 11 Jun 2024 16:28

Instrument :
BNA_N
ClientSampleId :
A4DG2

Tgt Ion: 94 Resp: 3882958




Ion	Ratio	Lower	Upper
94	100		
65	26.2	20.3	30.5
66	35.8	25.8	38.8




#10
Bis(2-Chloroethyl)ether
Concen: 53.437 ng/ul
RT: 7.098 min Scan# 699
Delta R.T. -0.018 min
Lab File: BN031913.D
Acq: 11 Jun 2024 16:28

Tgt Ion: 93 Resp: 1050259

Ion	Ratio	Lower	Upper
93	100		
63	68.6	56.8	85.2
95	31.5	26.0	39.0









#12
2-Chlorophenol
Concen: 104.137 ng/ul
RT: 7.228 min Scan# 721
Delta R.T. -0.024 min
Lab File: BN031913.D
Acq: 11 Jun 2024 16:28




Instrument :
BNA_N
ClientSampleId :
A4DG2


Tgt Ion:128 Resp: 1942028
Ion Ratio Lower Upper
128 100
64 42.6 33.9 50.9
130 32.3 26.3 39.5



#17
N-Nitroso-di-n-propylamine
Concen: 147.165 ng/ul
RT: 8.475 min Scan# 933
Delta R.T. -0.018 min
Lab File: BN031913.D
Acq: 11 Jun 2024 16:28

Tgt Ion: 70 Resp: 2200293
Ion Ratio Lower Upper
70 100
42 48.7 37.3 55.9
101 10.4 7.8 11.6
130 23.1 16.4 24.6





#19

Hexachloroethane

Concen: 138.017 ng/ul

RT: 8.728 min Scan# 91

Delta R.T. -0.030 min

Lab File: BN031913.D


Acq: 11 Jun 2024 16:28

Instrument :

BNA_N

ClientSampleId :

A4DG2






Tgt Ion:117 Resp: 1089940

Ion Ratio Lower Upper

117 100

201 94.7 71.6 107.4

199 61.6 45.6 68.4



#22

Nitrobenzene

Concen: 49.658 ng/ul

RT: 8.863 min Scan# 999

Delta R.T. -0.018 min

Lab File: BN031913.D

Acq: 11 Jun 2024 16:28




Tgt Ion: 77 Resp: 1167648


Ion Ratio Lower Upper

77 100

123 47.9 36.6 55.0

65 12.9 10.2 15.4









#25
2-Nitrophenol
Concen: 73.794 ng/ul
RT: 9.569 min Scan# 1118
Delta R.T. -0.018 min
Lab File: BN031913.D
Acq: 11 Jun 2024 16:28




Instrument :
BNA_N
ClientSampleId :
A4DG2


Tgt Ion:139 Resp: 821933
Ion Ratio Lower Upper
139 100
65 46.3 38.3 57.5
109 30.1 22.1 33.1



#27
Bis(2-Chloroethoxy)methane
Concen: 76.805 ng/ul
RT: 9.869 min Scan# 1170
Delta R.T. -0.024 min
Lab File: BN031913.D
Acq: 11 Jun 2024 16:28

Tgt Ion: 93 Resp: 2308884
Ion Ratio Lower Upper
93 100
95 34.0 26.1 39.1
123 11.8 9.0 13.6





#29

2,4-Dichlorophenol

Concen: 67.695 ng/ul

RT: 10.098 min Scan# 1208

Delta R.T. -0.018 min

Lab File: BN031913.D

Acq: 11 Jun 2024 16:28

Instrument :

BNA_N

ClientSampleId :

A4DG2





Tgt Ion:162 Resp: 1272149

Ion Ratio Lower Upper

162 100

164 62.8 52.7 79.1

98 37.7 30.5 45.7



#30

Napthalene

Concen: 160.554 ng/ul

RT: 10.498 min Scan# 1277

Delta R.T. -0.018 min

Lab File: BN031913.D

Acq: 11 Jun 2024 16:28




Tgt Ion:128 Resp:10725364


Ion Ratio Lower Upper

128 100

129 11.2 8.5 12.7

127 13.4 10.3 15.5









#35
4-Chloro-3-methylphenol
Concen: 119.341 ng/ul
RT: 11.739 min Scan# 1488
Delta R.T. -0.012 min
Lab File: BN031913.D
Acq: 11 Jun 2024 16:28




Instrument :
BNA_N
ClientSampleId :
A4DG2


Tgt Ion:107 Resp: 2621446
Ion Ratio Lower Upper
107 100
144 25.9 20.5 30.7
142 79.1 64.0 96.0



#41
2,4,6-Trichlorophenol
Concen: 82.835 ng/ul
RT: 12.733 min Scan# 1657
Delta R.T. -0.012 min
Lab File: BN031913.D
Acq: 11 Jun 2024 16:28

Tgt Ion:196 Resp: 1252292
Ion Ratio Lower Upper
196 100
198 98.2 76.4 114.6
200 31.7 24.2 36.4









#42
2,4,5-Trichlorophenol
Concen: 170.080 ng/ul
RT: 12.804 min Scan# 1669
Delta R.T. -0.012 min
Lab File: BN031913.D
Acq: 11 Jun 2024 16:28




Instrument :
BNA_N
ClientSampleId :
A4DG2


Tgt Ion:196 Resp: 2794137
Ion Ratio Lower Upper
196 100
198 95.4 73.8 110.6
200 31.0 25.6 38.4



#44
2-Chloronaphthalene
Concen: 78.145 ng/ul
RT: 13.175 min Scan# 1732
Delta R.T. -0.024 min
Lab File: BN031913.D
Acq: 11 Jun 2024 16:28

Tgt Ion:162 Resp: 3625718
Ion Ratio Lower Upper
162 100
164 33.9 27.0 40.4
127 41.4 31.8 47.6









#48
 2,6-Dinitrotoluene
 Concen: 253.903 ng/ul
 RT: 13.910 min Scan# 1857
 Delta R.T. -0.000 min
 Lab File: BN031913.D
 Acq: 11 Jun 2024 16:28

Instrument :
 BNA_N
 ClientSampleId :
 A4DG2

Tgt Ion:165 Resp: 3028374




Ion	Ratio	Lower	Upper
165	100		
89	56.1	44.6	67.0
121	20.6	17.3	25.9




#50
 Acenaphthylene
 Concen: 157.549 ng/ul
 RT: 14.033 min Scan# 1878
 Delta R.T. -0.012 min
 Lab File: BN031913.D
 Acq: 11 Jun 2024 16:28

Tgt Ion:152 Resp:12012218

Ion	Ratio	Lower	Upper
152	100		
151	20.6	15.9	23.9
153	13.6	10.4	15.6





#52

Acenaphthene

Concen: 123.863 ng/ul

RT: 14.374 min Scan# 1936

Delta R.T. -0.012 min

Lab File: BN031913.D

Acq: 11 Jun 2024 16:28

Instrument :

BNA_N

ClientSampleId :

A4DG2





Tgt Ion:153 Resp: 6248625

Ion Ratio Lower Upper

153 100

152 46.8 37.8 56.6

154 87.4 69.8 104.8



#55

4-Nitrophenol

Concen: 191.852 ng/ul

RT: 14.563 min Scan# 1968

Delta R.T. 0.023 min

Lab File: BN031913.D

Acq: 11 Jun 2024 16:28




Tgt Ion:109 Resp: 1671890


Ion Ratio Lower Upper

109 100

139 130.0 113.4 170.2

65 109.0 106.2 159.2









#57
2,4-Dinitrotoluene
Concen: 142.723 ng/ul
RT: 14.698 min Scan# 1989
Delta R.T. 0.005 min
Lab File: BN031913.D
Acq: 11 Jun 2024 16:28




Instrument :
BNA_N
ClientSampleId :
A4DG2


Tgt Ion:165 Resp: 2497807
Ion Ratio Lower Upper
165 100
63 40.0 32.2 48.2
182 3.1 2.3 3.5



#59
Diethylphthalate
Concen: 177.078 ng/ul
RT: 15.157 min Scan# 2069
Delta R.T. -0.000 min
Lab File: BN031913.D
Acq: 11 Jun 2024 16:28

Tgt Ion:149 Resp:10964186
Ion Ratio Lower Upper
149 100
177 22.0 17.1 25.7
150 12.7 10.0 15.0





#61

Fluorene

Concen: 46.277 ng/ul

RT: 15.363 min Scan# 2104

Delta R.T. -0.012 min

Lab File: BN031913.D

Acq: 11 Jun 2024 16:28

Instrument :

BNA_N

ClientSampleId :

A4DG2





Tgt Ion:166 Resp: 2554180

Ion Ratio Lower Upper

166 100

165 98.8 80.3 120.5

167 13.8 11.3 16.9



#62

4-Chlorophenyl-phenylether

Concen: 140.669 ng/ul

RT: 15.363 min Scan# 2104

Delta R.T. -0.012 min

Lab File: BN031913.D

Acq: 11 Jun 2024 16:28




Tgt Ion:204 Resp: 3760534


Ion Ratio Lower Upper

204 100

206 33.4 25.7 38.5

141 67.1 47.7 71.5









#69
Hexachlorobenzene
Concen: 214.141 ng/ul
RT: 16.363 min Scan# 21
Delta R.T. -0.006 min
Lab File: BN031913.D
Acq: 11 Jun 2024 16:28




Instrument :
BNA_N
ClientSampleId :
A4DG2


Tgt Ion:284 Resp: 4142899
Ion Ratio Lower Upper
284 100
142 38.0 31.1 46.7
249 32.5 25.7 38.5



#71
Pentachlorophenol
Concen: 208.289 ng/ul
RT: 16.716 min Scan# 2334
Delta R.T. -0.000 min
Lab File: BN031913.D
Acq: 11 Jun 2024 16:28

Tgt Ion:266 Resp: 2578172
Ion Ratio Lower Upper
266 100
264 67.1 54.7 82.1
268 64.2 53.4 80.2









#72
 Phenanthrene
 Concen: 65.798 ng/ul
 RT: 17.104 min Scan# 2410
 Delta R.T. -0.012 min
 Lab File: BN031913.D
 Acq: 11 Jun 2024 16:28

Instrument :
 BNA_N
 ClientSampleId :
 A4DG2

Tgt Ion:178 Resp: 5905065




Ion	Ratio	Lower	Upper
178	100		
179	15.2	12.6	19.0
176	18.7	15.8	23.6




#74
 Anthracene
 Concen: 80.370 ng/ul
 RT: 17.198 min Scan# 2416
 Delta R.T. -0.006 min
 Lab File: BN031913.D
 Acq: 11 Jun 2024 16:28

Tgt Ion:178 Resp: 7308604

Ion	Ratio	Lower	Upper
178	100		
179	15.6	11.8	17.6
176	19.2	14.8	22.2





#78

Di-n-butylphthalate

Concen: 142.493 ng/ul

RT: 18.033 min Scan# 21

Delta R.T. -0.012 min

Lab File: BN031913.D

Acq: 11 Jun 2024 16:28

Instrument :

BNA_N

ClientSampleId :

A4DG2





Tgt Ion:149 Resp:15201039

Ion Ratio Lower Upper

149 100

150 11.3 7.4 11.0#

104 7.1 4.1 6.1#



#80

Fluoranthene

Concen: 153.648 ng/ul

RT: 19.133 min Scan# 2745

Delta R.T. -0.000 min

Lab File: BN031913.D

Acq: 11 Jun 2024 16:28




Tgt Ion:202 Resp:14724618


Ion Ratio Lower Upper

202 100

101 12.1 10.3 15.5

100 9.5 8.0 12.0









#82
 Pyrene
 Concen: 90.291 ng/ul
 RT: 19.492 min Scan# 2111
 Delta R.T. -0.000 min
 Lab File: BN031913.D
 Acq: 11 Jun 2024 16:28

Instrument :
 BNA_N
 ClientSampleId :
 A4DG2

Tgt Ion:202 Resp: 8877856




Ion	Ratio	Lower	Upper
202	100		
101	13.8	12.6	18.8
100	11.5	10.2	15.2




#85
 Benzo(a)anthracene
 Concen: 114.519 ng/ul
 RT: 21.286 min Scan# 3111
 Delta R.T. -0.000 min
 Lab File: BN031913.D
 Acq: 11 Jun 2024 16:28

Tgt Ion:228 Resp:11486555

Ion	Ratio	Lower	Upper
228	100		
229	19.9	16.7	25.1
226	30.6	22.8	34.2









#86
 Bis(2-ethylhexyl)phthalate
 Concen: 173.256 ng/ul
 RT: 21.215 min Scan# 3099
 Delta R.T. -0.006 min
 Lab File: BN031913.D
 Acq: 11 Jun 2024 16:28




Instrument :
 BNA_N
 ClientSampleId :
 A4DG2


Tgt Ion:149 Resp:11728857
 Ion Ratio Lower Upper
 149 100
 167 27.6 21.4 32.0
 279 5.3 3.3 4.9#



#87
 Chrysene
 Concen: 139.443 ng/ul
 RT: 21.356 min Scan# 3123
 Delta R.T. 0.005 min
 Lab File: BN031913.D
 Acq: 11 Jun 2024 16:28

Tgt Ion:228 Resp:13064346
 Ion Ratio Lower Upper
 228 100
 226 32.0 23.7 35.5
 229 20.0 15.1 22.7





#91

Benzo(k)fluoranthene

Concen: 162.389 ng/u1

RT: 23.386 min Scan# 3468

Delta R.T. 0.011 min

Lab File: BN031913.D

Acq: 11 Jun 2024 16:28

Instrument :

BNA_N

ClientSampleId :

A4DG2





Tgt Ion:252 Resp:16044536

Ion Ratio Lower Upper

252 100

253 24.2 17.6 26.4

125 11.6 9.4 14.2



#95

Dibenzo(a,h)anthracene

Concen: 110.201 ng/u1

RT: 27.603 min Scan# 4185

Delta R.T. -0.000 min

Lab File: BN031913.D

Acq: 11 Jun 2024 16:28

Tgt Ion:278 Resp: 9777140

Ion Ratio Lower Upper

278 100

139 17.9 15.0 22.6

279 25.5 17.4 26.0

