

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061325\
 Data File : BN037277.D
 Acq On : 15 Jun 2025 00:54
 Operator : RC/JU
 Sample : Q2314-06
 Misc :
 ALS Vial : 56 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 VPB182-HYD-20250612

Quant Time: Jun 16 05:33:45 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN061325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 13 18:43:34 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.575	152	1176	0.400	ng	0.00
7) Naphthalene-d8	10.351	136	2844	0.400	ng	#-0.01
13) Acenaphthene-d10	14.224	164	1497	0.400	ng	0.00
19) Phenanthrene-d10	16.971	188	2633	0.400	ng	0.00
29) Chrysene-d12	21.171	240	1998	0.400	ng	0.00
35) Perylene-d12	23.360	264	1328	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.177	112	23	0.008	ng	0.00
5) Phenol-d6	0.000	99	0d	0.000	ng	
8) Nitrobenzene-d5	8.728	82	775	0.276	ng	0.00
11) 2-Methylnaphthalene-d10	11.960	152	102	0.027	ng	0.00
14) 2,4,6-Tribromophenol	15.742	330	17	0.027	ng	0.01
15) 2-Fluorobiphenyl	12.843	172	2056	0.327	ng	0.00
27) Fluoranthene-d10	19.012	212	2003	0.291	ng	0.00
31) Terphenyl-d14	19.621	244	1791	0.397	ng	0.00
Target Compounds						
6) bis(2-Chloroethyl)ether	6.983	93	10173	3.731	ng	# 37
34) Bis(2-ethylhexyl)phtha...	21.099	149	2175	0.433	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061325\
Data File : BN037277.D
Acq On : 15 Jun 2025 00:54
Operator : RC/JU
Sample : Q2314-06
Misc :
ALS Vial : 56 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
VPB182-HYD-20250612

Quant Time: Jun 16 05:33:45 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN061325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 13 18:43:34 2025
Response via : Initial Calibration

