


Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN061325\  
Data File : BN037272.D  
Acq On : 14 Jun 2025 21:51  
Operator : RC/JU  
Sample : DFTPP  
Misc :  
ALS Vial : 51 Sample Multiplier: 1

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA\_N\Methods\8270E-Tune.M  
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
Last Update : Thu Jun 19 02:50:12 2025



AutoFind: Scans 2472, 2473, 2474; Background Corrected with Scan 2466

| Target<br>Mass | Rel. to<br>Mass | Lower<br>Limit% | Upper<br>Limit% | Rel.<br>Abn% | Raw<br>Abn | Result<br>Pass/Fail |
|----------------|-----------------|-----------------|-----------------|--------------|------------|---------------------|
| 68             | 69              | 0.00            | 2               | 1.5          | 741        | PASS                |
| 69             | 69              | 100             | 100             | 100.0        | 49899      | PASS                |
| 70             | 69              | 0.00            | 2               | 0.6          | 291        | PASS                |
| 197            | 198             | 0.00            | 2               | 0.0          | 0          | PASS                |
| 198            | 198             | 100             | 100             | 100.0        | 76024      | PASS                |
| 199            | 198             | 5               | 9               | 7.0          | 5292       | PASS                |
| 365            | 198             | 1               | 100             | 5.1          | 3876       | PASS                |
| 441            | 443             | 0.01            | 150             | 82.8         | 7543       | PASS                |
| 442            | 442             | 100             | 100             | 100.0        | 45411      | PASS                |
| 443            | 442             | 15              | 24              | 20.1         | 9108       | PASS                |

# DDT Breakdown




| Date          | Instrument Name  | DFTPP Data File    |
|---------------|------------------|--------------------|
| 6/13/2025     | BNA_N            | BN037272.D         |
|               |                  |                    |
|               |                  |                    |
| Compound Name | Response         | Retention Time     |
|               |                  |                    |
| DDT           | 228732           | 20.475             |
| DDD           | 3511             | 20.033             |
| DDE           | 0                | 0                  |
|               |                  |                    |
|               |                  |                    |
| SUM(DDD+DDE)  | SUM(DDT+DDD+DDE) | % Breakdown Of DDT |
|               |                  |                    |
| 3511          | 232243           | 1.51               |
|               |                  |                    |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN061325\  
Data File : BN037272.D  
Acq On : 14 Jun 2025 21:51  
Operator : RC/JU  
Sample : DFTPP  
Misc :  
ALS Vial : 51 Sample Multiplier: 1

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP

Quant Time: Jun 16 03:00:33 2025  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_N\Methods\8270E-Tune.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Mon Jun 16 03:00:26 2025  
Response via : Initial Calibration



TIC: BN037272.D\data.ms

**(70) Pentachlorophenol (C)**

**16.628min ( 0.000) 15244.65 ng**




**response 68073**

| Ion    | Exp%   | Act%   |
|--------|--------|--------|
| 265.70 | 100.00 | 100.00 |
| 268.00 | 62.20  | 71.57  |
| 264.00 | 61.60  | 62.32  |
| 0.00   | 0.00   | 0.00   |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN061325\  
Data File : BN037272.D  
Acq On : 14 Jun 2025 21:51  
Operator : RC/JU  
Sample : DFTPP  
Misc :  
ALS Vial : 51 Sample Multiplier: 1

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP

Quant Time: Jun 16 03:00:33 2025  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_N\Methods\8270E-Tune.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Mon Jun 16 03:00:26 2025  
Response via : Initial Calibration



TIC: BN037272.D\data.ms

(77) Benzidine

19.233min ( 0.000) 0.00 ng

response 404622

| Ion    | Exp%   | Act%   |
|--------|--------|--------|
| 184.00 | 100.00 | 100.00 |
| 185.00 | 15.50  | 13.08  |
| 183.00 | 13.20  | 12.35  |
| 0.00   | 0.00   | 0.00   |