

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061422\
 Data File : BN020157.D
 Acq On : 14 Jun 2022 15:07
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN061422

Quant Time: Jun 14 17:10:14 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN061422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 14 16:17:22 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	105	0.00
2	1,4-Dioxane	0.471	0.497	-5.5	100	0.00
3	n-Nitrosodimethylamine	0.441	0.391	11.3	100	0.00
4 S	2-Fluorophenol	1.101	1.102	-0.1	103	0.00
5 S	Phenol-d6	1.256	1.204	4.1	102	0.00
6	bis(2-Chloroethyl)ether	1.092	1.059	3.0	104	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	105	0.00
8 S	Nitrobenzene-d5	0.286	0.271	5.2	102	0.00
9	Naphthalene	1.209	1.238	-2.4	104	0.00
10	Hexachlorobutadiene	0.214	0.219	-2.3	104	0.00
11 SURR	2-Methylnaphthalene-d10	0.715	0.733	-2.5	104	0.00
12	2-Methylnaphthalene	0.834	0.856	-2.6	104	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	103	0.00
14 S	2,4,6-Tribromophenol	0.152	0.125	17.8	94	0.00
15 S	2-Fluorobiphenyl	1.613	1.660	-2.9	103	0.00
16	Acenaphthylene	1.888	1.814	3.9	103	0.00
17	Acenaphthene	1.326	1.354	-2.1	103	0.00
18	Fluorene	1.778	1.835	-3.2	102	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	102	0.00
20	4,6-Dinitro-2-methylphenol	0.036	0.033	8.3	101	0.00
21	4-Bromophenyl-phenylether	0.235	0.232	1.3	103	0.00
22	Hexachlorobenzene	0.258	0.258	0.0	103	0.00
23	Atrazine	0.171	0.153	10.5	98	0.00
24	Pentachlorophenol	0.092	0.068	26.1#	91	0.00
25	Phenanthrene	1.370	1.366	0.3	102	0.00
26	Anthracene	1.168	1.123	3.9	101	0.01
27 SURR	Fluoranthene-d10	1.242	1.212	2.4	100	0.00
28	Fluoranthene	1.539	1.524	1.0	102	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	100	0.00
30	Pyrene	1.728	1.803	-4.3	100	0.00
31 S	Terphenyl-d14	0.971	1.013	-4.3	100	0.00
32	Benzo(a)anthracene	1.493	1.458	2.3	100	0.00
33	Chrysene	1.607	1.634	-1.7	99	0.00
34	Bis(2-ethylhexyl)phthalate	0.885	0.775	12.4	87	0.00
35 I	Perylene-d12	1.000	1.000	0.0	96	0.00
36	Indeno(1,2,3-cd)pyrene	1.684	1.780	-5.7	99	0.00
37	Benzo(b)fluoranthene	1.629	1.687	-3.6	100	0.00
38	Benzo(k)fluoranthene	1.651	1.732	-4.9	98	0.00
39 C	Benzo(a)pyrene	1.366	1.423	-4.2	97	0.00
40	Dibenzo(a,h)anthracene	1.351	1.443	-6.8	100	0.00
41	Benzo(g,h,i)perylene	1.479	1.595	-7.8	100	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN061422