

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN061424\  
 Data File : BN032010.D  
 Acq On : 14 Jun 2024 21:06  
 Operator : MA/JU  
 Sample : P2696-07  
 Misc :  
 ALS Vial : 10 Sample Multiplier: 1

Instrument :  
 BNA\_N  
 ClientSampleId :  
 A4BN5

Quant Time: Jun 14 21:53:14 2024  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_N\Methods\SFAM-EPA-BN052924.MA.M  
 Quant Title : SVOA CALIBRATION  
 QLast Update : Tue Jun 11 22:28:10 2024  
 Response via : Initial Calibration


| Compound                      | R.T.   | QIon | Response | Conc   | Units | Dev(Min) |
|-------------------------------|--------|------|----------|--------|-------|----------|
| Internal Standards            |        |      |          |        |       |          |
| 1) 1,4-Dichlorobenzene-d4     | 7.652  | 152  | 295851   | 20.000 | ng/ul | 0.00     |
| 20) Naphthalene-d8            | 10.434 | 136  | 1346768  | 20.000 | ng/ul | 0.00     |
| 38) Acenaphthene-d10          | 14.298 | 164  | 924688   | 20.000 | ng/ul | 0.00     |
| 64) Phenanthrene-d10          | 17.051 | 188  | 1921537  | 20.000 | ng/ul | 0.00     |
| 79) Chrysene-d12              | 21.280 | 240  | 1369420  | 20.000 | ng/ul | -0.01    |
| 88) Perylene-d12              | 24.210 | 264  | 1223996  | 20.000 | ng/ul | -0.01    |
| System Monitoring Compounds   |        |      |          |        |       |          |
| 3) 1,4-Dioxane-d8             | 3.176  | 96   | 20524    | 2.870  | ng/uL | 0.00     |
| 4) Pyridine-d5                | 3.581  | 84   | 298916   | 14.849 | ng/ul | 0.00     |
| 7) Phenol-d5                  | 6.834  | 99   | 416663   | 16.428 | ng/ul | 0.00     |
| 9) Bis-(2-Chloroethyl)eth...  | 6.993  | 67   | 241562   | 16.283 | ng/ul | -0.01    |
| 11) 2-Chlorophenol-d4         | 7.187  | 132  | 342749   | 17.711 | ng/ul | 0.00     |
| 15) 4-Methylphenol-d8         | 8.364  | 113  | 355080   | 17.175 | ng/ul | 0.00     |
| 21) Nitrobenzene-d5           | 8.811  | 128  | 174607   | 16.956 | ng/ul | 0.00     |
| 24) 2-Nitrophenol-d4          | 9.528  | 143  | 186478   | 17.531 | ng/ul | 0.00     |
| 28) 2,4-Dichlorophenol-d3     | 10.063 | 165  | 329085   | 16.572 | ng/ul | 0.00     |
| 31) 4-Chloroaniline-d4        | 10.581 | 131  | 506369   | 17.174 | ng/ul | 0.00     |
| 46) Dimethylphthalate-d6      | 13.716 | 166  | 1106493  | 17.176 | ng/ul | 0.00     |
| 49) Acenaphthylene-d8         | 13.993 | 160  | 1287016  | 17.375 | ng/ul | 0.00     |
| 54) 4-Nitrophenol-d4          | 14.522 | 143  | 183306   | 14.212 | ng/ul | 0.00     |
| 60) Fluorene-d10              | 15.298 | 176  | 948968   | 17.410 | ng/ul | 0.00     |
| 65) 4,6-Dinitro-2-methylph... | 15.428 | 200  | 107080   | 11.087 | ng/ul | 0.00     |
| 73) Anthracene-d10            | 17.145 | 188  | 1467229  | 17.835 | ng/ul | 0.00     |
| 81) Pyrene-d10                | 19.445 | 212  | 1600410  | 21.839 | ng/ul | 0.00     |
| 92) Benzo(a)pyrene-d12        | 24.010 | 264  | 999535   | 17.102 | ng/ul | -0.01    |
| Target Compounds              |        |      |          |        |       |          |
| 86) Bis(2-ethylhexyl)phtha... | 21.192 | 149  | 68252    | 1.103  | ng/ul | 96       |


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN061424\  
Data File : BN032010.D  
Acq On : 14 Jun 2024 21:06  
Operator : MA/JU  
Sample : P2696-07  
Misc :  
ALS Vial : 10 Sample Multiplier: 1

Instrument :  
BNA\_N  
ClientSampleId :  
A4BN5

Quant Time: Jun 14 21:53:14 2024  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_N\Methods\SFAM-EPA-BN052924.MA.M  
Quant Title : SVOA CALIBRATION  
QLast Update : Tue Jun 11 22:28:10 2024  
Response via : Initial Calibration





#86  
Bis(2-ethylhexyl)phthalate  
Concen: 1.103 ng/ul  
RT: 21.192 min Scan# 3095  
Delta R.T. -0.011 min  
Lab File: BN032010.D  
Acq: 14 Jun 2024 21:06

Instrument :  
BNA\_N  
ClientSampleId :  
A4BN5

Tgt Ion:149 Resp: 68252  
Ion Ratio Lower Upper  
149 100  
167 29.3 21.4 32.0  
279 4.3 3.3 4.9

