

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061520\
 Data File : BN010893.D
 Acq On : 16 Jun 2020 06:12
 Operator : CG/JU
 Sample : L2989-07
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE103D3-20200610

Quant Time: Jun 16 06:59:12 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 15 11:09:46 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.55	152	2364	0.40	ng	0.00
7) Naphthalene-d8	10.30	136	8232	0.40	ng	0.00
13) Acenaphthene-d10	14.17	164	4518	0.40	ng	0.00
19) Phenanthrene-d10	16.92	188	9029	0.40	ng	0.00
27) Chrysene-d12	21.12	240	8584	0.40	ng	-0.01
34) Perylene-d12	23.27	264	8930	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.20	112	754	0.17	ng	0.00
5) Phenol-d6	6.74	99	528	0.10	ng	0.00
8) Nitrobenzene-d5	8.68	82	2093	0.34	ng	0.00
11) 2-Methylnaphthalene-d10	11.89	152	3799	0.30	ng	0.00
14) 2,4,6-Tribromophenol	15.67	330	590	0.37	ng	0.00
15) 2-Fluorobiphenyl	12.79	172	6978	0.39	ng	0.00
25) Fluoranthene-d10	18.95	212	9881	0.40	ng	0.00
29) Terphenyl-d14	19.57	244	12054	0.60	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.22	88	830	0.358	ng	# 89
22) Pentachlorophenol	16.59	266	71	0.037	ng	# 84
32) Bis(2-ethylhexyl)phthalate	21.06	149	1171	0.139	ng	# 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061520\
Data File : BN010893.D
Acq On : 16 Jun 2020 06:12
Operator : CG/JU
Sample : L2989-07
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE103D3-20200610

Quant Time: Jun 16 06:59:12 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 15 11:09:46 2020
Response via : Initial Calibration

