

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061520\
 Data File : BN010867.D
 Acq On : 15 Jun 2020 09:48
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 15 11:11:29 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 15 11:09:46 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	161#	0.00
2	1,4-Dioxane	0.392	0.421	-7.4	165#	0.00
3	n-Nitrosodimethylamine	0.228	0.228	0.0	183#	0.00
4 S	2-Fluorophenol	0.772	0.971	-25.8#	198#	0.00
5 S	Phenol-d6	0.910	1.110	-22.0	197#	0.00
6	bis(2-Chloroethyl)ether	0.887	1.098	-23.8	200#	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	178#	0.00
8 S	Nitrobenzene-d5	0.296	0.308	-4.1	178#	0.00
9	Naphthalene	1.005	1.073	-6.8	181#	0.00
10	Hexachlorobutadiene	0.248	0.242	2.4	161#	0.00
11 SURR	2-Methylnaphthalene-d10	0.620	0.670	-8.1	183#	0.00
12	2-Methylnaphthalene	0.673	0.721	-7.1	186#	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	185#	0.00
14 S	2,4,6-Tribromophenol	0.143	0.130	9.1	174#	0.00
15 S	2-Fluorobiphenyl	1.578	1.629	-3.2	179#	0.00
16	Acenaphthylene	1.639	1.704	-4.0	191#	0.00
17	Acenaphthene	1.117	1.172	-4.9	186#	0.00
18	Fluorene	1.441	1.470	-2.0	184#	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	169#	0.00
20	4-Bromophenyl-phenylether	0.240	0.252	-5.0	173#	0.00
21	Hexachlorobenzene	0.251	0.272	-8.4	175#	0.00
22	Pentachlorophenol	0.084	0.088	-4.8	183#	0.00
23	Phenanthrene	1.093	1.170	-7.0	177#	0.00
24	Anthracene	0.912	0.971	-6.5	181#	0.00
25 SURR	Fluoranthene-d10	1.104	1.133	-2.6	172#	0.00
26	Fluoranthene	1.240	1.256	-1.3	171#	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	166#	0.00
28	Pvrene	1.323	1.416	-7.0	171#	0.00
29 S	Terphenyl-d14	0.936	0.998	-6.6	169#	0.00
30	Benzo(a)anthracene	1.176	1.256	-6.8	179#	0.00
31	Chrysene	1.370	1.398	-2.0	157#	0.00
32	Bis(2-ethylhexyl)phthalate	0.392	0.457	-16.6	198#	0.00
33	Indeno(1,2,3-cd)pyrene	1.531	1.631	-6.5	169#	0.00
34 I	Perylene-d12	1.000	1.000	0.0	181#	0.00
35	Benzo(b)fluoranthene	1.360	1.328	2.4	171#	0.00
36	Benzo(k)fluoranthene	1.449	1.434	1.0	170#	0.00
37 C	Benzo(a)pyrene	1.167	1.168	-0.1	181#	0.00
38	Dibenzo(a,h)anthracene	1.330	1.231	7.4	165#	0.00
39	Benzo(a,h,i)perylene	1.391	1.330	4.4	163#	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061520\
Data File : BN010867.D
Acq On : 15 Jun 2020 09:48
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 15 11:11:29 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 15 11:09:46 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			