

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061520\
 Data File : BN010867.D
 Acq On : 15 Jun 2020 09:48
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 15 11:11:29 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 15 11:09:46 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	161	0.00
2	1,4-Dioxane	0.400	0.430	-7.5	165	0.00
3	n-Nitrosodimethylamine	0.400	0.400	0.0	183	0.00
4 S	2-Fluorophenol	0.400	0.503	-25.7#	198	0.00
5 S	Phenol-d6	0.400	0.488	-22.0	197	0.00
6	bis(2-Chloroethyl)ether	0.400	0.495	-23.7	200	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	178	0.00
8 S	Nitrobenzene-d5	0.400	0.416	-4.0	178	0.00
9	Naphthalene	0.400	0.427	-6.7	181	0.00
10	Hexachlorobutadiene	0.400	0.390	2.5	161	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.432	-8.0	183	0.00
12	2-Methylnaphthalene	0.400	0.429	-7.2	186	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	185	0.00
14 S	2,4,6-Tribromophenol	0.400	0.365	8.8	174	0.00
15 S	2-Fluorobiphenyl	0.400	0.413	-3.2	179	0.00
16	Acenaphthylene	0.400	0.416	-4.0	191	0.00
17	Acenaphthene	0.400	0.420	-5.0	186	0.00
18	Fluorene	0.400	0.408	-2.0	184	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	169	0.00
20	4-Bromophenyl-phenylether	0.400	0.419	-4.7	173	0.00
21	Hexachlorobenzene	0.400	0.434	-8.5	175	0.00
22	Pentachlorophenol	0.400	0.420	-5.0	183	0.00
23	Phenanthrene	0.400	0.428	-7.0	177	0.00
24	Anthracene	0.400	0.426	-6.5	181	0.00
25 SURR	Fluoranthene-d10	0.400	0.411	-2.7	172	0.00
26	Fluoranthene	0.400	0.405	-1.3	171	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	166	0.00
28	Pvrene	0.400	0.428	-7.0	171	0.00
29 S	Terphenyl-d14	0.400	0.426	-6.5	169	0.00
30	Benzo(a)anthracene	0.400	0.427	-6.7	179	0.00
31	Chrysene	0.400	0.408	-2.0	157	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.466	-16.5	198	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.426	-6.5	169	0.00
34 I	Perylene-d12	0.400	0.400	0.0	181	0.00
35	Benzo(b)fluoranthene	0.400	0.390	2.5	171	0.00
36	Benzo(k)fluoranthene	0.400	0.396	1.0	170	0.00
37 C	Benzo(a)pyrene	0.400	0.400	0.0	181	0.00
38	Dibenzo(a,h)anthracene	0.400	0.370	7.5	165	0.00
39	Benzo(a,h,i)perylene	0.400	0.383	4.3	163	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061520\
Data File : BN010867.D
Acq On : 15 Jun 2020 09:48
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 15 11:11:29 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 15 11:09:46 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				