

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061520\
 Data File : BN010882.D
 Acq On : 15 Jun 2020 20:02
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 16 03:22:24 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 15 11:09:46 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	147	0.00
2	1,4-Dioxane	0.400	0.463	-15.8	163	0.00
3	n-Nitrosodimethylamine	0.400	0.367	8.3	154	0.00
4 S	2-Fluorophenol	0.400	0.488	-22.0	176	0.00
5 S	Phenol-d6	0.400	0.495	-23.7	182	0.00
6	bis(2-Chloroethyl)ether	0.400	0.487	-21.7	179	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	157	-0.01
8 S	Nitrobenzene-d5	0.400	0.415	-3.7	156	0.00
9	Naphthalene	0.400	0.427	-6.7	159	0.00
10	Hexachlorobutadiene	0.400	0.395	1.3	144	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.424	-6.0	158	0.00
12	2-Methylnaphthalene	0.400	0.422	-5.5	161	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	157	0.00
14 S	2,4,6-Tribromophenol	0.400	0.418	-4.5	170	0.00
15 S	2-Fluorobiphenyl	0.400	0.423	-5.7	156	0.00
16	Acenaphthylene	0.400	0.425	-6.2	166	0.00
17	Acenaphthene	0.400	0.427	-6.7	161	0.00
18	Fluorene	0.400	0.419	-4.7	161	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	156	0.00
20	4-Bromophenyl-phenylether	0.400	0.396	1.0	151	0.00
21	Hexachlorobenzene	0.400	0.423	-5.7	158	0.00
22	Pentachlorophenol	0.400	0.394	1.5	159	0.00
23	Phenanthrene	0.400	0.423	-5.7	161	0.00
24	Anthracene	0.400	0.424	-6.0	167	0.00
25 SURR	Fluoranthene-d10	0.400	0.430	-7.5	167	0.00
26	Fluoranthene	0.400	0.429	-7.2	168	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	172	0.00
28	Pvrene	0.400	0.412	-3.0	170	0.00
29 S	Terphenyl-d14	0.400	0.405	-1.3	166	0.00
30	Benzo(a)anthracene	0.400	0.437	-9.2	189	0.00
31	Chrysene	0.400	0.405	-1.3	161	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.495	-23.7	218	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.464	-16.0	190	0.00
34 I	Perylene-d12	0.400	0.400	0.0	198	0.00
35	Benzo(b)fluoranthene	0.400	0.375	6.3	179	0.00
36	Benzo(k)fluoranthene	0.400	0.404	-1.0	190	0.00
37 C	Benzo(a)pyrene	0.400	0.408	-2.0	201	0.00
38	Dibenzo(a,h)anthracene	0.400	0.388	3.0	188	0.00
39	Benzo(a,h,i)perylene	0.400	0.377	5.8	175	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061520\
Data File : BN010882.D
Acq On : 15 Jun 2020 20:02
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 16 03:22:24 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 15 11:09:46 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				