

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061520\
 Data File : BN010882.D
 Acq On : 15 Jun 2020 20:02
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 16 03:22:24 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 15 11:09:46 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	147	0.00
2	1,4-Dioxane	0.392	0.453	-15.6	163#	0.00
3	n-Nitrosodimethylamine	0.228	0.209	8.3	154#	0.00
4 S	2-Fluorophenol	0.772	0.942	-22.0	176#	0.00
5 S	Phenol-d6	0.910	1.126	-23.7	182#	0.00
6	bis(2-Chloroethyl)ether	0.887	1.080	-21.8	179#	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	157#	-0.01
8 S	Nitrobenzene-d5	0.296	0.307	-3.7	156#	0.00
9	Naphthalene	1.005	1.074	-6.9	159#	0.00
10	Hexachlorobutadiene	0.248	0.245	1.2	144	0.00
11 SURR	2-Methylnaphthalene-d10	0.620	0.657	-6.0	158#	0.00
12	2-Methylnaphthalene	0.673	0.710	-5.5	161#	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	157#	0.00
14 S	2,4,6-Tribromophenol	0.143	0.150	-4.9	170#	0.00
15 S	2-Fluorobiphenyl	1.578	1.667	-5.6	156#	0.00
16	Acenaphthylene	1.639	1.740	-6.2	166#	0.00
17	Acenaphthene	1.117	1.192	-6.7	161#	0.00
18	Fluorene	1.441	1.511	-4.9	161#	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	156#	0.00
20	4-Bromophenyl-phenylether	0.240	0.238	0.8	151#	0.00
21	Hexachlorobenzene	0.251	0.265	-5.6	158#	0.00
22	Pentachlorophenol	0.084	0.083	1.2	159#	0.00
23	Phenanthrene	1.093	1.155	-5.7	161#	0.00
24	Anthracene	0.912	0.966	-5.9	167#	0.00
25 SURR	Fluoranthene-d10	1.104	1.187	-7.5	167#	0.00
26	Fluoranthene	1.240	1.330	-7.3	168#	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	172#	0.00
28	Pvrene	1.323	1.362	-2.9	170#	0.00
29 S	Terphenyl-d14	0.936	0.949	-1.4	166#	0.00
30	Benzo(a)anthracene	1.176	1.284	-9.2	189#	0.00
31	Chrysene	1.370	1.386	-1.2	161#	0.00
32	Bis(2-ethylhexyl)phthalate	0.392	0.485	-23.7	218#	0.00
33	Indeno(1,2,3-cd)pyrene	1.531	1.774	-15.9	190#	0.00
34 I	Perylene-d12	1.000	1.000	0.0	198#	0.00
35	Benzo(b)fluoranthene	1.360	1.276	6.2	179#	0.00
36	Benzo(k)fluoranthene	1.449	1.462	-0.9	190#	0.00
37 C	Benzo(a)pyrene	1.167	1.189	-1.9	201#	0.00
38	Dibenzo(a,h)anthracene	1.330	1.291	2.9	188#	0.00
39	Benzo(a,h,i)perylene	1.391	1.312	5.7	175#	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061520\
Data File : BN010882.D
Acq On : 15 Jun 2020 20:02
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 16 03:22:24 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 15 11:09:46 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			