

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061520\
 Data File : BN010884.D
 Acq On : 16 Jun 2020 00:49
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 16 01:19:25 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 15 11:09:46 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	142	0.00
2	1,4-Dioxane	0.392	0.415	-5.9	144	0.00
3	n-Nitrosodimethylamine	0.228	0.192	15.8	137	0.00
4 S	2-Fluorophenol	0.772	0.911	-18.0	164#	0.00
5 S	Phenol-d6	0.910	1.072	-17.8	168#	0.00
6	bis(2-Chloroethyl)ether	0.887	0.973	-9.7	156#	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	149	-0.01
8 S	Nitrobenzene-d5	0.296	0.307	-3.7	148	0.00
9	Naphthalene	1.005	1.068	-6.3	151#	0.00
10	Hexachlorobutadiene	0.248	0.247	0.4	138	0.00
11 SURR	2-Methylnaphthalene-d10	0.620	0.653	-5.3	149	0.00
12	2-Methylnaphthalene	0.673	0.706	-4.9	152#	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	149	0.00
14 S	2,4,6-Tribromophenol	0.143	0.132	7.7	142	0.00
15 S	2-Fluorobiphenyl	1.578	1.665	-5.5	147	0.00
16	Acenaphthylene	1.639	1.683	-2.7	151#	0.00
17	Acenaphthene	1.117	1.152	-3.1	147	-0.01
18	Fluorene	1.441	1.471	-2.1	148	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	142	0.00
20	4-Bromophenyl-phenylether	0.240	0.240	0.0	139	0.00
21	Hexachlorobenzene	0.251	0.262	-4.4	141	0.00
22	Pentachlorophenol	0.084	0.083	1.2	144	0.01
23	Phenanthrene	1.093	1.114	-1.9	142	0.00
24	Anthracene	0.912	0.920	-0.9	145	0.00
25 SURR	Fluoranthene-d10	1.104	1.079	2.3	138	0.00
26	Fluoranthene	1.240	1.203	3.0	138	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	136	-0.01
28	Pvrene	1.323	1.404	-6.1	139	0.00
29 S	Terphenyl-d14	0.936	0.984	-5.1	137	0.00
30	Benzo(a)anthracene	1.176	1.187	-0.9	139	0.00
31	Chrysene	1.370	1.444	-5.4	133	0.00
32	Bis(2-ethylhexyl)phthalate	0.392	0.450	-14.8	160#	0.00
33	Indeno(1,2,3-cd)pyrene	1.531	1.745	-14.0	148	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	153#	0.00
35	Benzo(b)fluoranthene	1.360	1.307	3.9	142	0.00
36	Benzo(k)fluoranthene	1.449	1.424	1.7	143	0.00
37 C	Benzo(a)pyrene	1.167	1.194	-2.3	156#	-0.01
38	Dibenzo(a,h)anthracene	1.330	1.318	0.9	149	-0.01
39	Benzo(a,h,i)perylene	1.391	1.352	2.8	140	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061520\
Data File : BN010884.D
Acq On : 16 Jun 2020 00:49
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 16 01:19:25 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 15 11:09:46 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			