

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061520\
 Data File : BN010901.D
 Acq On : 16 Jun 2020 10:59
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 16 11:41:37 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 16 08:49:35 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	148	0.00
2	1,4-Dioxane	0.392	0.401	-2.3	145	0.00
3	n-Nitrosodimethylamine	0.228	0.210	7.9	155#	0.02
4 S	2-Fluorophenol	0.772	0.924	-19.7	173#	0.00
5 S	Phenol-d6	0.910	1.101	-21.0	179#	0.00
6	bis(2-Chloroethyl)ether	0.887	1.070	-20.6	178#	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	155#	-0.01
8 S	Nitrobenzene-d5	0.296	0.305	-3.0	154#	0.00
9	Naphthalene	1.005	1.061	-5.6	156#	0.00
10	Hexachlorobutadiene	0.248	0.246	0.8	143	-0.01
11 SURR	2-Methylnaphthalene-d10	0.620	0.653	-5.3	156#	0.00
12	2-Methylnaphthalene	0.673	0.706	-4.9	159#	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	159#	0.00
14 S	2,4,6-Tribromophenol	0.143	0.140	2.1	161#	0.00
15 S	2-Fluorobiphenyl	1.578	1.659	-5.1	157#	0.00
16	Acenaphthylene	1.639	1.705	-4.0	164#	0.00
17	Acenaphthene	1.117	1.166	-4.4	159#	-0.01
18	Fluorene	1.441	1.511	-4.9	162#	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	155#	0.00
20	4-Bromophenyl-phenylether	0.240	0.241	-0.4	152#	0.00
21	Hexachlorobenzene	0.251	0.262	-4.4	154#	0.00
22	Pentachlorophenol	0.084	0.085	-1.2	161#	0.00
23	Phenanthrene	1.093	1.165	-6.6	161#	-0.01
24	Anthracene	0.912	0.957	-4.9	164#	0.00
25 SURR	Fluoranthene-d10	1.104	1.176	-6.5	163#	0.00
26	Fluoranthene	1.240	1.315	-6.0	164#	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	167#	0.00
28	Pvrene	1.323	1.374	-3.9	166#	0.00
29 S	Terphenyl-d14	0.936	0.970	-3.6	165#	0.00
30	Benzo(a)anthracene	1.176	1.281	-8.9	183#	0.00
31	Chrysene	1.370	1.398	-2.0	157#	0.00
32	Bis(2-ethylhexyl)phthalate	0.392	0.459	-17.1	200#	0.00
33	Indeno(1,2,3-cd)pyrene	1.531	1.736	-13.4	181#	0.00
34 I	Perylene-d12	1.000	1.000	0.0	187#	0.00
35	Benzo(b)fluoranthene	1.360	1.308	3.8	174#	0.00
36	Benzo(k)fluoranthene	1.449	1.466	-1.2	180#	0.00
37 C	Benzo(a)pyrene	1.167	1.157	0.9	184#	0.00
38	Dibenzo(a,h)anthracene	1.330	1.300	2.3	179#	0.00
39	Benzo(a,h,i)perylene	1.391	1.345	3.3	170#	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061520\
Data File : BN010901.D
Acq On : 16 Jun 2020 10:59
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 36 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 16 11:41:37 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 16 08:49:35 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			