

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061621\
 Data File : BN014930.D
 Acq On : 16 Jun 2021 20:17
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN061621

Quant Time: Jun 17 01:30:06 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN061621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 17 01:26:44 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.763	152	18848	0.40	ng	# 0.00
7) Naphthalene-d8	10.521	136	85521	0.40	ng	# 0.00
13) Acenaphthene-d10	14.371	164	47231	0.40	ng	0.00
19) Phenanthrene-d10	17.116	188	89403	0.40	ng	0.00
29) Chrysene-d12	21.324	240	79801	0.40	ng	# 0.00
35) Perylene-d12	23.581	264	77195	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.384	112	22794	0.39	ng	0.00
5) Phenol-d6	6.924	99	28864	0.39	ng	0.00
8) Nitrobenzene-d5	8.899	82	24133	0.37	ng	0.00
11) 2-Methylnaphthalene-d10	12.115	152	53836	0.37	ng	0.00
14) 2,4,6-Tribromophenol	15.855	330	7451	0.34	ng	-0.01
15) 2-Fluorobiphenyl	13.001	172	68248	0.38	ng	0.00
27) Fluoranthene-d10	19.158	212	102224	0.36	ng	0.00
31) Terphenyl-d14	19.764	244	74833	0.38	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.365	88	2812	0.36	ng	# 55
3) n-Nitrosodimethylamine	3.715	42	6273	0.36	ng	# 67
6) bis(2-Chloroethyl)ether	7.191	93	22270	0.38	ng	# 83
9) Naphthalene	10.572	128	90101	0.38	ng	98
10) Hexachlorobutadiene	10.876	225	17838	0.37	ng	# 99
12) 2-Methylnaphthalene	12.190	142	61138	0.38	ng	# 90
16) Acenaphthylene	14.092	152	89275	0.37	ng	98
17) Acenaphthene	14.434	154	54833	0.38	ng	97
18) Fluorene	15.424	166	67304	0.37	ng	99
20) 4,6-Dinitro-2-methylph...	15.487	198	1698	0.41	ng	# 55
21) 4-Bromophenyl-phenylether	16.325	248	20167	0.35	ng	# 92
22) Hexachlorobenzene	16.422	284	22714	0.37	ng	# 93
23) Atrazine	16.592	200	18713	0.36	ng	93
24) Pentachlorophenol	16.763	266	7200	0.32	ng	99
25) Phenanthrene	17.164	178	102353	0.38	ng	99
26) Anthracene	17.249	178	99103	0.37	ng	99
28) Fluoranthene	19.188	202	118010	0.38	ng	# 97
30) Pyrene	19.550	202	121343	0.38	ng	99
32) Benzo(a)anthracene	21.302	228	112676	0.37	ng	98
33) Chrysene	21.356	228	108243	0.38	ng	98
34) Bis(2-ethylhexyl)phtha...	21.260	149	76053	0.36	ng	# 88
36) Indeno(1,2,3-cd)pyrene	25.864	276	89262	0.35	ng	# 86
37) Benzo(b)fluoranthene	22.903	252	105508	0.36	ng	# 96
38) Benzo(k)fluoranthene	22.951	252	110813	0.40	ng	# 93
39) Benzo(a)pyrene	23.482	252	95852	0.38	ng	# 95
40) Dibenzo(a,h)anthracene	25.881	278	72027	0.35	ng	# 90
41) Benzo(g,h,i)perylene	26.549	276	80263	0.36	ng	# 89


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061621\
 Data File : BN014930.D
 Acq On : 16 Jun 2021 20:17
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN061621

Quant Time: Jun 17 01:30:06 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN061621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 17 01:26:44 2021
 Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.763 min Scan# 584
 Delta R.T. -0.000 min
 Lab File: BN014930.D
 Acq: 16 Jun 2021 20:17




Instrument :
 BNA_N
 ClientSampleId :
 ICVBN061621


Tgt Ion:152 Resp: 18848
 Ion Ratio Lower Upper
 152 100
 150 154.9 119.0 178.6
 115 60.5 62.2 93.4#



#2
 1,4-Dioxane
 Concen: 0.36 ng
 RT: 3.365 min Scan# 107
 Delta R.T. 0.000 min
 Lab File: BN014930.D
 Acq: 16 Jun 2021 20:17

Tgt Ion: 88 Resp: 2812
 Ion Ratio Lower Upper
 88 100
 43 29.3 12.9 19.3#
 58 80.6 38.1 57.1#









#3
n-Nitrosodimethylamine
Concen: 0.36 ng
RT: 3.715 min Scan# 145
Delta R.T. 0.000 min
Lab File: BN014930.D
Acq: 16 Jun 2021 20:17




Instrument :
BNA_N
ClientSampleId :
ICVBN061621


Tgt Ion: 42 Resp: 6273
Ion Ratio Lower Upper
42 100
74 137.2 149.5 224.3#
44 5.4 4.2 6.2



#4
2-Fluorophenol
Concen: 0.39 ng
RT: 5.384 min Scan# 326
Delta R.T. 0.009 min
Lab File: BN014930.D
Acq: 16 Jun 2021 20:17

Tgt Ion: 112 Resp: 22794
Ion Ratio Lower Upper
112 100
64 58.3 39.0 58.6
63 29.8 20.6 30.8









#5
Phenol-d6
Concen: 0.39 ng
RT: 6.924 min Scan# 493
Delta R.T. 0.000 min
Lab File: BN014930.D
Acq: 16 Jun 2021 20:17




Instrument :
BNA_N
ClientSampleId :
ICVBN061621


Tgt Ion: 99 Resp: 28864
Ion Ratio Lower Upper
99 100
42 18.7 8.9 13.3#
71 31.9 35.6 53.4#



#6
bis(2-Chloroethyl)ether
Concen: 0.38 ng
RT: 7.191 min Scan# 522
Delta R.T. -0.009 min
Lab File: BN014930.D
Acq: 16 Jun 2021 20:17

Tgt Ion: 93 Resp: 22270
Ion Ratio Lower Upper
93 100
63 77.1 46.6 70.0#
95 33.2 28.6 42.8









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.521 min Scan# 818
Delta R.T. 0.000 min
Lab File: BN014930.D
Acq: 16 Jun 2021 20:17




Instrument :
BNA_N
ClientSampleId :
ICVBN061621


Tgt Ion:	136	Resp:	85521
Ion Ratio	Lower	Upper	
136	100		
137	11.1	9.8	14.8
54	9.6	5.4	8.2#
68	9.1	4.9	7.3#



#8
Nitrobenzene-d5
Concen: 0.37 ng
RT: 8.899 min Scan# 690
Delta R.T. 0.000 min
Lab File: BN014930.D
Acq: 16 Jun 2021 20:17

Tgt Ion:	82	Resp:	24133
Ion Ratio	Lower	Upper	
82	100		
128	46.7	31.6	47.4
54	46.2	29.0	43.4#









#9
Naphthalene
Concen: 0.38 ng
RT: 10.572 min Scan# 822
Delta R.T. 0.000 min
Lab File: BN014930.D
Acq: 16 Jun 2021 20:17




Instrument :
BNA_N
ClientSampleId :
ICVBN061621


Tgt Ion:	128	Resp:	90101
Ion Ratio	Lower	Upper	
128	100		
129	11.1	9.1	13.7
127	13.4	11.8	17.6



#10
Hexachlorobutadiene
Concen: 0.37 ng
RT: 10.876 min Scan# 846
Delta R.T. 0.000 min
Lab File: BN014930.D
Acq: 16 Jun 2021 20:17

Tgt Ion:	225	Resp:	17838
Ion Ratio	Lower	Upper	
225	100		
223	0.0	0.0	0.0
227	63.5	50.2	75.2









#11
2-Methylnaphthalene-d10
Concen: 0.37 ng
RT: 12.115 min Scan# 1032
Delta R.T. -0.004 min
Lab File: BN014930.D
Acq: 16 Jun 2021 20:17




Instrument :
BNA_N
ClientSampleId :
ICVBN061621


Tgt Ion:152 Resp: 53836
Ion Ratio Lower Upper
152 100
151 20.2 16.7 25.1



#12
2-Methylnaphthalene
Concen: 0.38 ng
RT: 12.190 min Scan# 1049
Delta R.T. 0.000 min
Lab File: BN014930.D
Acq: 16 Jun 2021 20:17

Tgt Ion:142 Resp: 61138
Ion Ratio Lower Upper
142 100
141 88.4 72.3 108.5
115 33.3 40.7 61.1#









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.371 min Scan# 1267
Delta R.T. 0.000 min
Lab File: BN014930.D
Acq: 16 Jun 2021 20:17




Instrument :
BNA_N
ClientSampleId :
ICVBN061621


Tgt Ion:164 Resp: 47231
Ion Ratio Lower Upper
164 100
162 102.0 83.8 125.6
160 45.6 37.4 56.0



#14
2,4,6-Tribromophenol
Concen: 0.34 ng
RT: 15.855 min Scan# 1384
Delta R.T. -0.013 min
Lab File: BN014930.D
Acq: 16 Jun 2021 20:17

Tgt Ion:330 Resp: 7451
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 33.4 22.4 33.6









#15
2-Fluorobiphenyl
Concen: 0.38 ng
RT: 13.001 min Scan# 1159
Delta R.T. 0.000 min
Lab File: BN014930.D
Acq: 16 Jun 2021 20:17




Instrument :
BNA_N
ClientSampleId :
ICVBN061621


Tgt Ion:172 Resp: 68248
Ion Ratio Lower Upper
172 100
171 33.9 30.6 45.8
170 22.0 20.2 30.4



#16
Acenaphthylene
Concen: 0.37 ng
RT: 14.092 min Scan# 1245
Delta R.T. 0.000 min
Lab File: BN014930.D
Acq: 16 Jun 2021 20:17

Tgt Ion:152 Resp: 89275
Ion Ratio Lower Upper
152 100
151 19.5 16.7 25.1
153 13.1 10.5 15.7









#17
Acenaphthene
Concen: 0.38 ng
RT: 14.434 min Scan# 1272
Delta R.T. 0.000 min
Lab File: BN014930.D
Acq: 16 Jun 2021 20:17




Instrument :
BNA_N
ClientSampleId :
ICVBN061621


Tgt Ion:154 Resp: 54833
Ion Ratio Lower Upper
154 100
153 111.6 90.9 136.3
152 53.0 45.8 68.8



#18
Fluorene
Concen: 0.37 ng
RT: 15.424 min Scan# 1350
Delta R.T. 0.000 min
Lab File: BN014930.D
Acq: 16 Jun 2021 20:17

Tgt Ion:166 Resp: 67304
Ion Ratio Lower Upper
166 100
165 97.7 79.2 118.8
167 13.6 10.7 16.1









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.116 min Scan# 1487
Delta R.T. 0.000 min
Lab File: BN014930.D
Acq: 16 Jun 2021 20:17




Instrument :
BNA_N
ClientSampleId :
ICVBN061621


Tgt Ion:188 Resp: 89403
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 11.6 9.0 13.4



#20
4,6-Dinitro-2-methylphenol
Concen: 0.41 ng
RT: 15.487 min Scan# 1355
Delta R.T. 0.000 min
Lab File: BN014930.D
Acq: 16 Jun 2021 20:17

Tgt Ion:198 Resp: 1698
Ion Ratio Lower Upper
198 100
182 14.5 34.5 51.7#
77 0.0 0.0 0.0









#21
4-Bromophenyl-phenylether
Concen: 0.35 ng
RT: 16.325 min Scan# 1422
Delta R.T. 0.000 min
Lab File: BN014930.D
Acq: 16 Jun 2021 20:17




Instrument :
BNA_N
ClientSampleId :
ICVBN061621


Tgt Ion:248 Resp: 20167
Ion Ratio Lower Upper
248 100
250 104.1 86.2 129.4
141 33.0 36.2 54.2#



#22
Hexachlorobenzene
Concen: 0.37 ng
RT: 16.422 min Scan# 1430
Delta R.T. 0.000 min
Lab File: BN014930.D
Acq: 16 Jun 2021 20:17

Tgt Ion:284 Resp: 22714
Ion Ratio Lower Upper
284 100
142 33.5 22.2 33.2#
249 30.8 26.6 39.8









#23
Atrazine
Concen: 0.36 ng
RT: 16.592 min Scan# 1444
Delta R.T. 0.000 min
Lab File: BN014930.D
Acq: 16 Jun 2021 20:17




Instrument :
BNA_N
ClientSampleId :
ICVBN061621


Tgt Ion:200 Resp: 18713
Ion Ratio Lower Upper
200 100
173 23.1 21.7 32.5
215 49.6 43.1 64.7



#24
Pentachlorophenol
Concen: 0.32 ng
RT: 16.763 min Scan# 1458
Delta R.T. 0.000 min
Lab File: BN014930.D
Acq: 16 Jun 2021 20:17

Tgt Ion:266 Resp: 7200
Ion Ratio Lower Upper
266 100
264 62.8 51.0 76.4
268 63.8 50.6 76.0









#25
Phenanthrene
Concen: 0.38 ng
RT: 17.164 min Scan# 1491
Delta R.T. 0.000 min
Lab File: BN014930.D
Acq: 16 Jun 2021 20:17




Instrument :
BNA_N
ClientSampleId :
ICVBN061621


Tgt Ion:178 Resp: 102353
Ion Ratio Lower Upper
178 100
176 18.9 15.6 23.4
179 15.4 12.1 18.1



#26
Anthracene
Concen: 0.37 ng
RT: 17.249 min Scan# 1498
Delta R.T. 0.000 min
Lab File: BN014930.D
Acq: 16 Jun 2021 20:17

Tgt Ion:178 Resp: 99103
Ion Ratio Lower Upper
178 100
176 18.3 15.1 22.7
179 15.3 12.2 18.2









#27
Fluoranthene-d10
Concen: 0.36 ng
RT: 19.158 min Scan# 1791
Delta R.T. 0.000 min
Lab File: BN014930.D
Acq: 16 Jun 2021 20:17




Instrument :
BNA_N
ClientSampleId :
ICVBN061621


Tgt Ion:212 Resp: 102224
Ion Ratio Lower Upper
212 100
106 13.9 7.0 10.6#
104 7.6 4.1 6.1#



#28
Fluoranthene
Concen: 0.38 ng
RT: 19.188 min Scan# 1797
Delta R.T. 0.000 min
Lab File: BN014930.D
Acq: 16 Jun 2021 20:17

Tgt Ion:202 Resp: 118010
Ion Ratio Lower Upper
202 100
101 11.6 6.6 10.0#
203 17.3 13.9 20.9









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.324 min Scan# 2086
Delta R.T. 0.000 min
Lab File: BN014930.D
Acq: 16 Jun 2021 20:17




Instrument :
BNA_N
ClientSampleId :
ICVBN061621


Tgt Ion:240 Resp: 79801
Ion Ratio Lower Upper
240 100
120 6.4 4.2 6.4#
236 25.5 21.4 32.2



#30
Pyrene
Concen: 0.38 ng
RT: 19.550 min Scan# 1870
Delta R.T. 0.000 min
Lab File: BN014930.D
Acq: 16 Jun 2021 20:17

Tgt Ion:202 Resp: 121343
Ion Ratio Lower Upper
202 100
200 20.5 16.9 25.3
203 17.5 13.9 20.9









#31
 Terphenyl-d14
 Concen: 0.38 ng
 RT: 19.764 min Scan# 1913
 Delta R.T. -0.005 min
 Lab File: BN014930.D
 Acq: 16 Jun 2021 20:17




Instrument :
 BNA_N
 ClientSampleId :
 ICVBN061621


Tgt Ion:244 Resp: 74833
 Ion Ratio Lower Upper
 244 100
 212 7.6 6.3 9.5
 122 13.3 6.2 9.2#



#32
 Benzo(a)anthracene
 Concen: 0.37 ng
 RT: 21.302 min Scan# 2084
 Delta R.T. 0.000 min
 Lab File: BN014930.D
 Acq: 16 Jun 2021 20:17

Tgt Ion:228 Resp: 112676
 Ion Ratio Lower Upper
 228 100
 226 26.9 23.0 34.6
 229 19.3 15.7 23.5









#33
Chrysene
Concen: 0.38 ng
RT: 21.356 min Scan# 2089
Delta R.T. 0.000 min
Lab File: BN014930.D
Acq: 16 Jun 2021 20:17




Instrument :
BNA_N
ClientSampleId :
ICVBN061621


Tgt Ion:228 Resp: 108243
Ion Ratio Lower Upper
228 100
226 29.6 25.0 37.6
229 19.6 15.9 23.9



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.36 ng
RT: 21.260 min Scan# 2080
Delta R.T. -0.010 min
Lab File: BN014930.D
Acq: 16 Jun 2021 20:17

Tgt Ion:149 Resp: 76053
Ion Ratio Lower Upper
149 100
167 28.2 28.9 43.3#
279 3.7 4.9 7.3#









#35
Perylene-d12
Concen: 0.40 ng
RT: 23.581 min Scan# 2531
Delta R.T. 0.000 min
Lab File: BN014930.D
Acq: 16 Jun 2021 20:17




Instrument :
BNA_N
ClientSampleId :
ICVBN061621


Tgt Ion:264 Resp: 77195
Ion Ratio Lower Upper
264 100
260 25.0 20.6 31.0
265 23.3 21.4 32.2



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.35 ng
RT: 25.864 min Scan# 3198
Delta R.T. 0.007 min
Lab File: BN014930.D
Acq: 16 Jun 2021 20:17

Tgt Ion:276 Resp: 89262
Ion Ratio Lower Upper
276 100
138 28.5 10.0 15.0#
277 25.0 20.4 30.6









#37
Benzo(b)fluoranthene
Concen: 0.36 ng
RT: 22.903 min Scan# 2333
Delta R.T. -0.003 min
Lab File: BN014930.D
Acq: 16 Jun 2021 20:17




Instrument :
BNA_N
ClientSampleId :
ICVBN061621


Tgt Ion:252 Resp: 105508
Ion Ratio Lower Upper
252 100
253 22.0 18.5 27.7
125 10.8 5.8 8.6#



#38
Benzo(k)fluoranthene
Concen: 0.40 ng
RT: 22.951 min Scan# 2347
Delta R.T. 0.000 min
Lab File: BN014930.D
Acq: 16 Jun 2021 20:17

Tgt Ion:252 Resp: 110813
Ion Ratio Lower Upper
252 100
253 22.3 19.1 28.7
125 13.5 5.4 8.0#









#39
Benzo(a)pyrene
Concen: 0.38 ng
RT: 23.482 min Scan# 2502
Delta R.T. 0.000 min
Lab File: BN014930.D
Acq: 16 Jun 2021 20:17




Instrument :
BNA_N
ClientSampleId :
ICVBN061621


Tgt Ion:252 Resp: 95852
Ion Ratio Lower Upper
252 100
253 22.0 19.4 29.2
125 12.1 7.4 11.2#



#40
Dibenzo(a,h)anthracene
Concen: 0.35 ng
RT: 25.881 min Scan# 3203
Delta R.T. 0.000 min
Lab File: BN014930.D
Acq: 16 Jun 2021 20:17

Tgt Ion:278 Resp: 72027
Ion Ratio Lower Upper
278 100
139 20.0 9.0 13.4#
279 24.1 21.0 31.4





#41
 Benzo(g,h,i)perylene
 Concen: 0.36 ng
 RT: 26.549 min Scan# 3398
 Delta R.T. -0.003 min
 Lab File: BN014930.D
 Acq: 16 Jun 2021 20:17

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN061621

Tgt Ion:	276	Resp:	80263
Ion Ratio	Lower	Upper	
276	100		
277	23.8	19.7	29.5
138	23.7	9.8	14.6#

