

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061621\
 Data File : BN014930.D
 Acq On : 16 Jun 2021 20:17
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN061621

Quant Time: Jun 17 01:44:22 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN061621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 17 01:26:44 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	112	0.00
2	1,4-Dioxane	0.168	0.149	11.3	110	0.00
3	n-Nitrosodimethylamine	0.372	0.333	10.5	111	0.00
4 S	2-Fluorophenol	1.237	1.209	2.3	112	0.00
5 S	Phenol-d6	1.567	1.531	2.3	112	0.00
6	bis(2-Chloroethyl)ether	1.232	1.182	4.1	110	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	111	0.00
8 S	Nitrobenzene-d5	0.302	0.282	6.6	114	0.00
9	Naphthalene	1.121	1.054	6.0	110	0.00
10	Hexachlorobutadiene	0.223	0.209	6.3	110	0.00
11 SURR	2-Methylnaphthalene-d10	0.686	0.630	8.2	110	0.00
12	2-Methylnaphthalene	0.754	0.715	5.2	109	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	108	0.00
14 S	2,4,6-Tribromophenol	0.188	0.158	16.0	102	-0.01
15 S	2-Fluorobiphenyl	1.519	1.445	4.9	110	0.00
16	Acenaphthylene	2.021	1.890	6.5	108	0.00
17	Acenaphthene	1.236	1.161	6.1	109	0.00
18	Fluorene	1.526	1.425	6.6	107	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	108	0.00
20	4,6-Dinitro-2-methylphenol	0.023	0.019	17.4	142	0.00
21	4-Bromophenyl-phenylether	0.258	0.226	12.4	100	0.00
22	Hexachlorobenzene	0.273	0.254	7.0	107	0.00
23	Atrazine	0.231	0.209	9.5	105	0.00
24	Pentachlorophenol	0.100	0.081	19.0	110	0.00
25	Phenanthrene	1.220	1.145	6.1	108	0.00
26	Anthracene	1.193	1.108	7.1	106	0.00
27 SURR	Fluoranthene-d10	1.261	1.143	9.4	107	0.00
28	Fluoranthene	1.383	1.320	4.6	108	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	107	0.00
30	Pyrene	1.584	1.521	4.0	107	0.00
31 S	Terphenyl-d14	0.991	0.938	5.3	107	0.00
32	Benzo(a)anthracene	1.508	1.412	6.4	107	0.00
33	Chrysene	1.445	1.356	6.2	106	0.00
34	Bis(2-ethylhexyl)phthalate	1.052	0.953	9.4	109	-0.01
35 I	Perylene-d12	1.000	1.000	0.0	107	0.00
36	Indeno(1,2,3-cd)pyrene	1.336	1.156	13.5	103	0.00
37	Benzo(b)fluoranthene	1.500	1.367	8.9	103	0.00
38	Benzo(k)fluoranthene	1.448	1.435	0.9	110	0.00
39 C	Benzo(a)pyrene	1.321	1.242	6.0	107	0.00
40	Dibenzo(a,h)anthracene	1.082	0.933	13.8	103	0.00
41	Benzo(g,h,i)perylene	1.160	1.040	10.3	106	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN061621