

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061621\
 Data File : BN014930.D
 Acq On : 16 Jun 2021 20:17
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN061621

Quant Time: Jun 17 01:44:22 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN061621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 17 01:26:44 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	112	0.00
2	1,4-Dioxane	0.400	0.356	11.0	110	0.00
3	n-Nitrosodimethylamine	0.400	0.358	10.5	111	0.00
4 S	2-Fluorophenol	0.400	0.391	2.3	112	0.00
5 S	Phenol-d6	0.400	0.391	2.3	112	0.00
6	bis(2-Chloroethyl)ether	0.400	0.384	4.0	110	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	111	0.00
8 S	Nitrobenzene-d5	0.400	0.373	6.8	114	0.00
9	Naphthalene	0.400	0.376	6.0	110	0.00
10	Hexachlorobutadiene	0.400	0.375	6.3	110	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.367	8.3	110	0.00
12	2-Methylnaphthalene	0.400	0.379	5.3	109	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	108	0.00
14 S	2,4,6-Tribromophenol	0.400	0.336	16.0	102	-0.01
15 S	2-Fluorobiphenyl	0.400	0.380	5.0	110	0.00
16	Acenaphthylene	0.400	0.374	6.5	108	0.00
17	Acenaphthene	0.400	0.376	6.0	109	0.00
18	Fluorene	0.400	0.374	6.5	107	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	108	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.406	-1.5	142	0.00
21	4-Bromophenyl-phenylether	0.400	0.349	12.8	100	0.00
22	Hexachlorobenzene	0.400	0.372	7.0	107	0.00
23	Atrazine	0.400	0.362	9.5	105	0.00
24	Pentachlorophenol	0.400	0.321	19.8	110	0.00
25	Phenanthrene	0.400	0.375	6.3	108	0.00
26	Anthracene	0.400	0.372	7.0	106	0.00
27 SURR	Fluoranthene-d10	0.400	0.363	9.3	107	0.00
28	Fluoranthene	0.400	0.382	4.5	108	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	107	0.00
30	Pyrene	0.400	0.384	4.0	107	0.00
31 S	Terphenyl-d14	0.400	0.378	5.5	107	0.00
32	Benzo(a)anthracene	0.400	0.374	6.5	107	0.00
33	Chrysene	0.400	0.376	6.0	106	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.362	9.5	109	-0.01
35 I	Perylene-d12	0.400	0.400	0.0	107	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.346	13.5	103	0.00
37	Benzo(b)fluoranthene	0.400	0.364	9.0	103	0.00
38	Benzo(k)fluoranthene	0.400	0.397	0.8	110	0.00
39 C	Benzo(a)pyrene	0.400	0.376	6.0	107	0.00
40	Dibenzo(a,h)anthracene	0.400	0.345	13.8	103	0.00
41	Benzo(g,h,i)perylene	0.400	0.359	10.3	106	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN061621