

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061621\
 Data File : BN014925.D
 Acq On : 16 Jun 2021 17:17
 Operator : CG/JU
 Sample : SSTDICCC0.4
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICCC0.4

Quant Time: Jun 16 19:01:10 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN061621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 16 18:59:52 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.763	152	16816	0.400	ng	# 0.00
7) Naphthalene-d8	10.521	136	77330	0.400	ng	# 0.00
13) Acenaphthene-d10	14.371	164	43730	0.400	ng	0.00
19) Phenanthrene-d10	17.115	188	83006	0.400	ng	0.00
29) Chrysene-d12	21.323	240	74699	0.400	ng	# 0.00
35) Perylene-d12	23.581	264	72137	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.375	112	20382	0.449	ng	0.00
5) Phenol-d6	6.924	99	25840	0.489	ng	0.00
8) Nitrobenzene-d5	8.899	82	21125	0.331	ng	0.00
11) 2-Methylnaphthalene-d10	12.119	152	49077	0.337	ng	0.00
14) 2,4,6-Tribromophenol	15.868	330	7338	0.400	ng	0.00
15) 2-Fluorobiphenyl	13.000	172	62172	0.376	ng	0.00
27) Fluoranthene-d10	19.158	212	95793	0.335	ng	0.00
31) Terphenyl-d14	19.768	244	70092	0.379	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.365	88	2550	0.522	ng	# 60
3) n-Nitrosodimethylamine	3.715	42	5671	1.094	ng	# 67
6) bis(2-Chloroethyl)ether	7.200	93	20185	0.504	ng	# 82
9) Naphthalene	10.572	128	81742	0.386	ng	98
10) Hexachlorobutadiene	10.876	225	16267	0.264	ng	# 98
12) 2-Methylnaphthalene	12.190	142	56010	0.376	ng	# 90
16) Acenaphthylene	14.092	152	82321	0.452	ng	98
17) Acenaphthene	14.434	154	50397	0.409	ng	97
18) Fluorene	15.424	166	62700	0.397	ng	99
20) 4,6-Dinitro-2-methylph...	15.487	198	1200	0.226	ng	# 63
21) 4-Bromophenyl-phenylether	16.324	248	20233	0.342	ng	# 92
22) Hexachlorobenzene	16.422	284	21140	0.315	ng	# 92
23) Atrazine	16.592	200	17743	0.553	ng	94
24) Pentachlorophenol	16.762	266	6527	0.432	ng	99
25) Phenanthrene	17.164	178	94770	0.402	ng	99
26) Anthracene	17.249	178	93136	0.446	ng	99
28) Fluoranthene	19.188	202	109380	0.366	ng	# 97
30) Pyrene	19.550	202	112902	0.508	ng	99
32) Benzo(a)anthracene	21.302	228	105720	0.498	ng	98
33) Chrysene	21.355	228	102040	0.409	ng	98
34) Bis(2-ethylhexyl)phtha...	21.270	149	70040	1.067	ng	# 88
36) Indeno(1,2,3-cd)pyrene	25.857	276	86979	0.401	ng	# 87
37) Benzo(b)fluoranthene	22.906	252	102388	0.445	ng	# 96
38) Benzo(k)fluoranthene	22.951	252	101048	0.369	ng	# 94
39) Benzo(a)pyrene	23.482	252	89416	0.410	ng	# 94
40) Dibenzo(a,h)anthracene	25.881	278	69928	0.422	ng	# 91
41) Benzo(g,h,i)perylene	26.552	276	75826	0.362	ng	# 89


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061621\
Data File : BN014925.D
Acq On : 16 Jun 2021 17:17
Operator : CG/JU
Sample : SSTDICCC0.4
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4

Quant Time: Jun 16 19:01:10 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN061621.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 16 18:59:52 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.763 min Scan# 584
Delta R.T. -0.000 min
Lab File: BN014925.D
Acq: 16 Jun 2021 17:17




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:152 Resp: 16816
Ion Ratio Lower Upper
152 100
150 153.2 119.0 178.6
115 62.0 62.2 93.4#



#2
1,4-Dioxane
Concen: 0.522 ng
RT: 3.365 min Scan# 107
Delta R.T. 0.000 min
Lab File: BN014925.D
Acq: 16 Jun 2021 17:17

Tgt Ion: 88 Resp: 2550
Ion Ratio Lower Upper
88 100
43 27.9 12.9 19.3#
58 77.1 38.1 57.1#









#3
n-Nitrosodimethylamine
Concen: 1.094 ng
RT: 3.715 min Scan# 145
Delta R.T. 0.000 min
Lab File: BN014925.D
Acq: 16 Jun 2021 17:17

Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4

Tgt Ion: 42 Resp: 5671




Ion	Ratio	Lower	Upper
42	100		
74	137.4	149.5	224.3#
44	8.1	4.2	6.2#




#4
2-Fluorophenol
Concen: 0.449 ng
RT: 5.375 min Scan# 325
Delta R.T. 0.000 min
Lab File: BN014925.D
Acq: 16 Jun 2021 17:17

Tgt Ion: 112 Resp: 20382

Ion	Ratio	Lower	Upper
112	100		
64	59.3	39.0	58.6#
63	30.1	20.6	30.8









#5
Phenol-d6
Concen: 0.489 ng
RT: 6.924 min Scan# 493
Delta R.T. 0.000 min
Lab File: BN014925.D
Acq: 16 Jun 2021 17:17




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion: 99 Resp: 25840
Ion Ratio Lower Upper
99 100
42 19.1 8.9 13.3#
71 31.7 35.6 53.4#



#6
bis(2-Chloroethyl)ether
Concen: 0.504 ng
RT: 7.200 min Scan# 523
Delta R.T. 0.000 min
Lab File: BN014925.D
Acq: 16 Jun 2021 17:17

Tgt Ion: 93 Resp: 20185
Ion Ratio Lower Upper
93 100
63 77.6 46.6 70.0#
95 32.5 28.6 42.8









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.521 min Scan# 818
Delta R.T. 0.000 min
Lab File: BN014925.D
Acq: 16 Jun 2021 17:17




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:	136	Resp:	77330
Ion Ratio	Lower	Upper	
136	100		
137	11.5	9.8	14.8
54	11.3	5.4	8.2#
68	11.2	4.9	7.3#



#8
Nitrobenzene-d5
Concen: 0.331 ng
RT: 8.899 min Scan# 690
Delta R.T. 0.000 min
Lab File: BN014925.D
Acq: 16 Jun 2021 17:17

Tgt Ion:	82	Resp:	21125
Ion Ratio	Lower	Upper	
82	100		
128	43.0	31.6	47.4
54	47.1	29.0	43.4#









#9
Naphthalene
Concen: 0.386 ng
RT: 10.572 min Scan# 822
Delta R.T. 0.000 min
Lab File: BN014925.D
Acq: 16 Jun 2021 17:17




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:	128	Resp:	81742
Ion	Ratio	Lower	Upper
128	100		
129	11.3	9.1	13.7
127	13.6	11.8	17.6



#10
Hexachlorobutadiene
Concen: 0.264 ng
RT: 10.876 min Scan# 846
Delta R.T. 0.000 min
Lab File: BN014925.D
Acq: 16 Jun 2021 17:17

Tgt Ion:	225	Resp:	16267
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	64.0	50.2	75.2









#11
2-Methylnaphthalene-d10
Concen: 0.337 ng
RT: 12.119 min Scan# 1033
Delta R.T. 0.000 min
Lab File: BN014925.D
Acq: 16 Jun 2021 17:17




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:152 Resp: 49077
Ion Ratio Lower Upper
152 100
151 20.3 16.7 25.1



#12
2-Methylnaphthalene
Concen: 0.376 ng
RT: 12.190 min Scan# 1049
Delta R.T. 0.000 min
Lab File: BN014925.D
Acq: 16 Jun 2021 17:17

Tgt Ion:142 Resp: 56010
Ion Ratio Lower Upper
142 100
141 88.6 72.3 108.5
115 34.1 40.7 61.1#









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.371 min Scan# 1267
Delta R.T. 0.000 min
Lab File: BN014925.D
Acq: 16 Jun 2021 17:17




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:164 Resp: 43730
Ion Ratio Lower Upper
164 100
162 101.9 83.8 125.6
160 45.3 37.4 56.0



#14
2,4,6-Tribromophenol
Concen: 0.400 ng
RT: 15.868 min Scan# 1385
Delta R.T. 0.000 min
Lab File: BN014925.D
Acq: 16 Jun 2021 17:17

Tgt Ion:330 Resp: 7338
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 33.7 22.4 33.6#









#15
2-Fluorobiphenyl
Concen: 0.376 ng
RT: 13.000 min Scan# 1159
Delta R.T. 0.000 min
Lab File: BN014925.D
Acq: 16 Jun 2021 17:17




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:172 Resp: 62172
Ion Ratio Lower Upper
172 100
171 33.9 30.6 45.8
170 22.5 20.2 30.4



#16
Acenaphthylene
Concen: 0.452 ng
RT: 14.092 min Scan# 1245
Delta R.T. 0.000 min
Lab File: BN014925.D
Acq: 16 Jun 2021 17:17

Tgt Ion:152 Resp: 82321
Ion Ratio Lower Upper
152 100
151 19.6 16.7 25.1
153 13.0 10.5 15.7









#17
Acenaphthene
Concen: 0.409 ng
RT: 14.434 min Scan# 1272
Delta R.T. 0.000 min
Lab File: BN014925.D
Acq: 16 Jun 2021 17:17




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:154 Resp: 50397
Ion Ratio Lower Upper
154 100
153 111.9 90.9 136.3
152 52.6 45.8 68.8



#18
Fluorene
Concen: 0.397 ng
RT: 15.424 min Scan# 1350
Delta R.T. 0.000 min
Lab File: BN014925.D
Acq: 16 Jun 2021 17:17

Tgt Ion:166 Resp: 62700
Ion Ratio Lower Upper
166 100
165 97.5 79.2 118.8
167 13.5 10.7 16.1









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.115 min Scan# 1487
Delta R.T. 0.000 min
Lab File: BN014925.D
Acq: 16 Jun 2021 17:17




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:188 Resp: 83006
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 12.4 9.0 13.4



#20
4,6-Dinitro-2-methylphenol
Concen: 0.226 ng
RT: 15.487 min Scan# 1355
Delta R.T. 0.000 min
Lab File: BN014925.D
Acq: 16 Jun 2021 17:17

Tgt Ion:198 Resp: 1200
Ion Ratio Lower Upper
198 100
182 19.3 34.5 51.7#
77 0.0 0.0 0.0









#21
4-Bromophenyl-phenylether
Concen: 0.342 ng
RT: 16.324 min Scan# 1422
Delta R.T. 0.000 min
Lab File: BN014925.D
Acq: 16 Jun 2021 17:17




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:	248	Resp:	20233
Ion Ratio	Lower	Upper	
248	100		
250	102.5	86.2	129.4
141	35.0	36.2	54.2#



#22
Hexachlorobenzene
Concen: 0.315 ng
RT: 16.422 min Scan# 1430
Delta R.T. 0.000 min
Lab File: BN014925.D
Acq: 16 Jun 2021 17:17

Tgt Ion:	284	Resp:	21140
Ion Ratio	Lower	Upper	
284	100		
142	33.5	22.2	33.2#
249	30.6	26.6	39.8









#23
Atrazine
Concen: 0.553 ng
RT: 16.592 min Scan# 1444
Delta R.T. 0.000 min
Lab File: BN014925.D
Acq: 16 Jun 2021 17:17




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:200 Resp: 17743
Ion Ratio Lower Upper
200 100
173 23.4 21.7 32.5
215 50.1 43.1 64.7



#24
Pentachlorophenol
Concen: 0.432 ng
RT: 16.762 min Scan# 1458
Delta R.T. 0.000 min
Lab File: BN014925.D
Acq: 16 Jun 2021 17:17

Tgt Ion:266 Resp: 6527
Ion Ratio Lower Upper
266 100
264 63.6 51.0 76.4
268 64.9 50.6 76.0









#25
Phenanthrene
Concen: 0.402 ng
RT: 17.164 min Scan# 1491
Delta R.T. 0.000 min
Lab File: BN014925.D
Acq: 16 Jun 2021 17:17




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:	178	Resp:	94770
Ion Ratio	Lower	Upper	
178	100		
176	18.9	15.6	23.4
179	15.5	12.1	18.1



#26
Anthracene
Concen: 0.446 ng
RT: 17.249 min Scan# 1498
Delta R.T. 0.000 min
Lab File: BN014925.D
Acq: 16 Jun 2021 17:17

Tgt Ion:	178	Resp:	93136
Ion Ratio	Lower	Upper	
178	100		
176	18.2	15.1	22.7
179	15.3	12.2	18.2









#27
Fluoranthene-d10
Concen: 0.335 ng
RT: 19.158 min Scan# 1791
Delta R.T. 0.000 min
Lab File: BN014925.D
Acq: 16 Jun 2021 17:17




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:212 Resp: 95793
Ion Ratio Lower Upper
212 100
106 13.9 7.0 10.6#
104 7.6 4.1 6.1#



#28
Fluoranthene
Concen: 0.366 ng
RT: 19.188 min Scan# 1797
Delta R.T. 0.000 min
Lab File: BN014925.D
Acq: 16 Jun 2021 17:17

Tgt Ion:202 Resp: 109380
Ion Ratio Lower Upper
202 100
101 11.7 6.6 10.0#
203 17.4 13.9 20.9









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.323 min Scan# 2086
Delta R.T. 0.000 min
Lab File: BN014925.D
Acq: 16 Jun 2021 17:17




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:240 Resp: 74699
Ion Ratio Lower Upper
240 100
120 8.2 4.2 6.4#
236 25.9 21.4 32.2



#30
Pyrene
Concen: 0.508 ng
RT: 19.550 min Scan# 1870
Delta R.T. 0.000 min
Lab File: BN014925.D
Acq: 16 Jun 2021 17:17

Tgt Ion:202 Resp: 112902
Ion Ratio Lower Upper
202 100
200 20.6 16.9 25.3
203 17.5 13.9 20.9









#31
 Terphenyl-d14
 Concen: 0.379 ng
 RT: 19.768 min Scan# 1914
 Delta R.T. 0.000 min
 Lab File: BN014925.D
 Acq: 16 Jun 2021 17:17




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.4


Tgt Ion:244 Resp: 70092
 Ion Ratio Lower Upper
 244 100
 212 7.5 6.3 9.5
 122 12.5 6.2 9.2#



#32
 Benzo(a)anthracene
 Concen: 0.498 ng
 RT: 21.302 min Scan# 2084
 Delta R.T. 0.000 min
 Lab File: BN014925.D
 Acq: 16 Jun 2021 17:17

Tgt Ion:228 Resp: 105720
 Ion Ratio Lower Upper
 228 100
 226 27.6 23.0 34.6
 229 19.2 15.7 23.5









#33
Chrysene
Concen: 0.409 ng
RT: 21.355 min Scan# 2089
Delta R.T. 0.000 min
Lab File: BN014925.D
Acq: 16 Jun 2021 17:17




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:228 Resp: 102040
Ion Ratio Lower Upper
228 100
226 30.1 25.0 37.6
229 19.3 15.9 23.9



#34
Bis(2-ethylhexyl)phthalate
Concen: 1.067 ng
RT: 21.270 min Scan# 2081
Delta R.T. 0.000 min
Lab File: BN014925.D
Acq: 16 Jun 2021 17:17

Tgt Ion:149 Resp: 70040
Ion Ratio Lower Upper
149 100
167 28.4 28.9 43.3#
279 3.8 4.9 7.3#









#35
Perylene-d12
Concen: 0.400 ng
RT: 23.581 min Scan# 2531
Delta R.T. 0.000 min
Lab File: BN014925.D
Acq: 16 Jun 2021 17:17

Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4

Tgt Ion:264 Resp: 72137




Ion	Ratio	Lower	Upper
264	100		
260	24.9	20.6	31.0
265	23.2	21.4	32.2




#36
Indeno(1,2,3-cd)pyrene
Concen: 0.401 ng
RT: 25.857 min Scan# 3196
Delta R.T. 0.000 min
Lab File: BN014925.D
Acq: 16 Jun 2021 17:17

Tgt Ion:276 Resp: 86979


Ion	Ratio	Lower	Upper
276	100		
138	27.3	10.0	15.0#
277	25.0	20.4	30.6







#37
Benzo(b)fluoranthene
Concen: 0.445 ng
RT: 22.906 min Scan# 2334
Delta R.T. 0.000 min
Lab File: BN014925.D
Acq: 16 Jun 2021 17:17

Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4







Tgt Ion:252 Resp: 102388
Ion Ratio Lower Upper
252 100
253 22.1 18.5 27.7
125 11.1 5.8 8.6#



#38
Benzo(k)fluoranthene
Concen: 0.369 ng
RT: 22.951 min Scan# 2347
Delta R.T. 0.000 min
Lab File: BN014925.D
Acq: 16 Jun 2021 17:17


Tgt Ion:252 Resp: 101048
Ion Ratio Lower Upper
252 100
253 22.1 19.1 28.7
125 11.5 5.4 8.0#







#39
Benzo(a)pyrene
Concen: 0.410 ng
RT: 23.482 min Scan# 2502
Delta R.T. 0.000 min
Lab File: BN014925.D
Acq: 16 Jun 2021 17:17

Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4







Tgt Ion:252 Resp: 89416
Ion Ratio Lower Upper
252 100
253 22.1 19.4 29.2
125 12.5 7.4 11.2#



#40
Dibenzo(a,h)anthracene
Concen: 0.422 ng
RT: 25.881 min Scan# 3203
Delta R.T. 0.000 min
Lab File: BN014925.D
Acq: 16 Jun 2021 17:17

Tgt Ion:278 Resp: 69928
Ion Ratio Lower Upper
278 100
139 18.7 9.0 13.4#
279 23.9 21.0 31.4





#41
 Benzo(g,h,i)perylene
 Concen: 0.362 ng
 RT: 26.552 min Scan# 3399
 Delta R.T. 0.000 min
 Lab File: BN014925.D
 Acq: 16 Jun 2021 17:17

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICCC0.4

Tgt Ion:	276	Resp:	75826
Ion Ratio	Lower	Upper	
276	100		
277	23.8	19.7	29.5
138	23.6	9.8	14.6#

