

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061621\
 Data File : BN014929.D
 Acq On : 16 Jun 2021 19:41
 Operator : CG/JU
 Sample : SSTDICC5.0
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0

Quant Time: Jun 17 01:06:45 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN061621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 16 18:59:52 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.763	152	19405	0.400	ng	# 0.00
7) Naphthalene-d8	10.522	136	77713	0.400	ng	# 0.00
13) Acenaphthene-d10	14.371	164	42816	0.400	ng	0.00
19) Phenanthrene-d10	17.116	188	83062	0.400	ng	0.00
29) Chrysene-d12	21.324	240	75324	0.400	ng	# 0.00
35) Perylene-d12	23.581	264	78367	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.384	112	254522	4.859	ng	0.00
5) Phenol-d6	6.924	99	319436	5.238	ng	0.00
8) Nitrobenzene-d5	8.899	82	302416	4.711	ng	0.00
11) 2-Methylnaphthalene-d10	12.119	152	602868	4.121	ng	0.00
14) 2,4,6-Tribromophenol	15.868	330	102821	5.723	ng	0.00
15) 2-Fluorobiphenyl	13.001	172	765993	4.733	ng	0.00
27) Fluoranthene-d10	19.158	212	1193093	4.164	ng	0.00
31) Terphenyl-d14	19.769	244	866944	4.652	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.356	88	35462	6.289	ng	# 72
3) n-Nitrosodimethylamine	3.697	42	88701	14.829	ng	# 67
6) bis(2-Chloroethyl)ether	7.191	93	258650	5.596	ng	# 83
9) Naphthalene	10.572	128	1014117	4.766	ng	97
10) Hexachlorobutadiene	10.876	225	202203	3.259	ng	# 99
12) 2-Methylnaphthalene	12.190	142	658579	4.397	ng	# 89
16) Acenaphthylene	14.092	152	1023585	5.734	ng	98
17) Acenaphthene	14.435	154	632409	5.236	ng	94
18) Fluorene	15.424	166	764512	4.948	ng	98
21) 4-Bromophenyl-phenylether	16.325	248	246072	4.162	ng	# 92
22) Hexachlorobenzene	16.422	284	266481	3.965	ng	93
23) Atrazine	16.593	200	235877	7.350	ng	93
24) Pentachlorophenol	16.763	266	126318	8.350	ng	99
25) Phenanthrene	17.164	178	1193021	5.055	ng	99
26) Anthracene	17.250	178	1151901	5.512	ng	99
28) Fluoranthene	19.188	202	1304599	4.368	ng	# 97
30) Pyrene	19.550	202	1380809	6.158	ng	99
32) Benzo(a)anthracene	21.303	228	1349291	6.299	ng	98
33) Chrysene	21.356	228	1287750	5.123	ng	98
34) Bis(2-ethylhexyl)phtha...	21.260	149	996108	15.053	ng	# 89
36) Indeno(1,2,3-cd)pyrene	25.864	276	1357379	5.766	ng	# 87
37) Benzo(b)fluoranthene	22.907	252	1369220	5.475	ng	# 96
38) Benzo(k)fluoranthene	22.951	252	1306991	4.389	ng	# 94
39) Benzo(a)pyrene	23.482	252	1251819	5.281	ng	# 95
40) Dibenzo(a,h)anthracene	25.885	278	1111209	6.168	ng	# 91
41) Benzo(g,h,i)perylene	26.556	276	1170701	5.141	ng	# 90


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061621\
Data File : BN014929.D
Acq On : 16 Jun 2021 19:41
Operator : CG/JU
Sample : SSTDICC5.0
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Quant Time: Jun 17 01:06:45 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN061621.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 16 18:59:52 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.763 min Scan# 584
Delta R.T. 0.000 min
Lab File: BN014929.D
Acq: 16 Jun 2021 19:41




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:152 Resp: 19405
Ion Ratio Lower Upper
152 100
150 158.0 119.0 178.6
115 60.1 62.2 93.4#



#2
1,4-Dioxane
Concen: 6.289 ng
RT: 3.356 min Scan# 106
Delta R.T. -0.009 min
Lab File: BN014929.D
Acq: 16 Jun 2021 19:41

Tgt Ion: 88 Resp: 35462
Ion Ratio Lower Upper
88 100
43 25.7 12.9 19.3#
58 67.8 38.1 57.1#









#3
n-Nitrosodimethylamine
Concen: 14.829 ng
RT: 3.697 min Scan# 143
Delta R.T. -0.018 min
Lab File: BN014929.D
Acq: 16 Jun 2021 19:41




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion: 42 Resp: 88701
Ion Ratio Lower Upper
42 100
74 137.5 149.5 224.3#
44 7.1 4.2 6.2#



#4
2-Fluorophenol
Concen: 4.859 ng
RT: 5.384 min Scan# 326
Delta R.T. 0.010 min
Lab File: BN014929.D
Acq: 16 Jun 2021 19:41

Tgt Ion: 112 Resp: 254522
Ion Ratio Lower Upper
112 100
64 58.6 39.0 58.6#
63 29.7 20.6 30.8









#5
Phenol-d6
Concen: 5.238 ng
RT: 6.924 min Scan# 493
Delta R.T. 0.000 min
Lab File: BN014929.D
Acq: 16 Jun 2021 19:41




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion: 99 Resp: 319436
Ion Ratio Lower Upper
99 100
42 18.5 8.9 13.3#
71 31.7 35.6 53.4#



#6
bis(2-Chloroethyl)ether
Concen: 5.596 ng
RT: 7.191 min Scan# 522
Delta R.T. -0.009 min
Lab File: BN014929.D
Acq: 16 Jun 2021 19:41

Tgt Ion: 93 Resp: 258650
Ion Ratio Lower Upper
93 100
63 76.6 46.6 70.0#
95 32.7 28.6 42.8









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.522 min Scan# 818
Delta R.T. 0.000 min
Lab File: BN014929.D
Acq: 16 Jun 2021 19:41




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:	136	Resp:	77713
Ion	Ratio	Lower	Upper
136	100		
137	11.2	9.8	14.8
54	10.4	5.4	8.2#
68	9.9	4.9	7.3#



#8
Nitrobenzene-d5
Concen: 4.711 ng
RT: 8.899 min Scan# 690
Delta R.T. 0.000 min
Lab File: BN014929.D
Acq: 16 Jun 2021 19:41

Tgt Ion:	82	Resp:	302416
Ion	Ratio	Lower	Upper
82	100		
128	48.2	31.6	47.4#
54	44.9	29.0	43.4#









#9
Naphthalene
Concen: 4.766 ng
RT: 10.572 min Scan# 822
Delta R.T. 0.000 min
Lab File: BN014929.D
Acq: 16 Jun 2021 19:41




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:128 Resp: 1014117
Ion Ratio Lower Upper
128 100
129 11.1 9.1 13.7
127 13.1 11.8 17.6



#10
Hexachlorobutadiene
Concen: 3.259 ng
RT: 10.876 min Scan# 846
Delta R.T. 0.000 min
Lab File: BN014929.D
Acq: 16 Jun 2021 19:41

Tgt Ion:225 Resp: 202203
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.8 50.2 75.2









#11
2-Methylnaphthalene-d10
Concen: 4.121 ng
RT: 12.119 min Scan# 1033
Delta R.T. 0.000 min
Lab File: BN014929.D
Acq: 16 Jun 2021 19:41




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:152 Resp: 602868
Ion Ratio Lower Upper
152 100
151 20.1 16.7 25.1



#12
2-Methylnaphthalene
Concen: 4.397 ng
RT: 12.190 min Scan# 1049
Delta R.T. 0.000 min
Lab File: BN014929.D
Acq: 16 Jun 2021 19:41

Tgt Ion:142 Resp: 658579
Ion Ratio Lower Upper
142 100
141 88.3 72.3 108.5
115 32.8 40.7 61.1#









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.371 min Scan# 1267
Delta R.T. 0.000 min
Lab File: BN014929.D
Acq: 16 Jun 2021 19:41




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:164 Resp: 42816
Ion Ratio Lower Upper
164 100
162 102.5 83.8 125.6
160 45.8 37.4 56.0



#14
2,4,6-Tribromophenol
Concen: 5.723 ng
RT: 15.868 min Scan# 1385
Delta R.T. 0.000 min
Lab File: BN014929.D
Acq: 16 Jun 2021 19:41

Tgt Ion:330 Resp: 102821
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 33.2 22.4 33.6









#15
2-Fluorobiphenyl
Concen: 4.733 ng
RT: 13.001 min Scan# 1159
Delta R.T. 0.000 min
Lab File: BN014929.D
Acq: 16 Jun 2021 19:41




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:172 Resp: 765993
Ion Ratio Lower Upper
172 100
171 33.7 30.6 45.8
170 21.9 20.2 30.4



#16
Acenaphthylene
Concen: 5.734 ng
RT: 14.092 min Scan# 1245
Delta R.T. 0.000 min
Lab File: BN014929.D
Acq: 16 Jun 2021 19:41

Tgt Ion:152 Resp: 1023585
Ion Ratio Lower Upper
152 100
151 19.4 16.7 25.1
153 13.1 10.5 15.7









#17
Acenaphthene
Concen: 5.236 ng
RT: 14.435 min Scan# 1272
Delta R.T. 0.000 min
Lab File: BN014929.D
Acq: 16 Jun 2021 19:41




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:154 Resp: 632409
Ion Ratio Lower Upper
154 100
153 109.1 90.9 136.3
152 51.2 45.8 68.8



#18
Fluorene
Concen: 4.948 ng
RT: 15.424 min Scan# 1350
Delta R.T. 0.000 min
Lab File: BN014929.D
Acq: 16 Jun 2021 19:41

Tgt Ion:166 Resp: 764512
Ion Ratio Lower Upper
166 100
165 96.3 79.2 118.8
167 13.6 10.7 16.1









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.116 min Scan# 1487
Delta R.T. 0.000 min
Lab File: BN014929.D
Acq: 16 Jun 2021 19:41




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:188 Resp: 83062
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 12.3 9.0 13.4



#21
4-Bromophenyl-phenylether
Concen: 4.162 ng
RT: 16.325 min Scan# 1422
Delta R.T. 0.000 min
Lab File: BN014929.D
Acq: 16 Jun 2021 19:41

Tgt Ion:248 Resp: 246072
Ion Ratio Lower Upper
248 100
250 103.6 86.2 129.4
141 33.3 36.2 54.2#









#22
Hexachlorobenzene
Concen: 3.965 ng
RT: 16.422 min Scan# 1430
Delta R.T. 0.000 min
Lab File: BN014929.D
Acq: 16 Jun 2021 19:41




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:284 Resp: 266481
Ion Ratio Lower Upper
284 100
142 32.9 22.2 33.2
249 30.6 26.6 39.8



#23
Atrazine
Concen: 7.350 ng
RT: 16.593 min Scan# 1444
Delta R.T. 0.000 min
Lab File: BN014929.D
Acq: 16 Jun 2021 19:41

Tgt Ion:200 Resp: 235877
Ion Ratio Lower Upper
200 100
173 24.1 21.7 32.5
215 48.9 43.1 64.7









#24
Pentachlorophenol
Concen: 8.350 ng
RT: 16.763 min Scan# 1458
Delta R.T. 0.000 min
Lab File: BN014929.D
Acq: 16 Jun 2021 19:41




Instrument :
BNA_N
ClientSampled :
SSTDICC5.0


Tgt Ion:266 Resp: 126318
Ion Ratio Lower Upper
266 100
264 62.9 51.0 76.4
268 63.9 50.6 76.0



#25
Phenanthrene
Concen: 5.055 ng
RT: 17.164 min Scan# 1491
Delta R.T. 0.000 min
Lab File: BN014929.D
Acq: 16 Jun 2021 19:41

Tgt Ion:178 Resp: 1193021
Ion Ratio Lower Upper
178 100
176 18.8 15.6 23.4
179 15.4 12.1 18.1









#26
 Anthracene
 Concen: 5.512 ng
 RT: 17.250 min Scan# 1498
 Delta R.T. 0.000 min
 Lab File: BN014929.D
 Acq: 16 Jun 2021 19:41

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0

Tgt Ion:178 Resp: 1151901




Ion	Ratio	Lower	Upper
178	100		
176	18.2	15.1	22.7
179	15.5	12.2	18.2




#27
 Fluoranthene-d10
 Concen: 4.164 ng
 RT: 19.158 min Scan# 1791
 Delta R.T. 0.000 min
 Lab File: BN014929.D
 Acq: 16 Jun 2021 19:41

Tgt Ion:212 Resp: 1193093

Ion	Ratio	Lower	Upper
212	100		
106	13.9	7.0	10.6#
104	7.6	4.1	6.1#









#28
Fluoranthene
Concen: 4.368 ng
RT: 19.188 min Scan# 1797
Delta R.T. 0.000 min
Lab File: BN014929.D
Acq: 16 Jun 2021 19:41




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:202 Resp: 1304599
Ion Ratio Lower Upper
202 100
101 11.6 6.6 10.0#
203 17.5 13.9 20.9



#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.324 min Scan# 2086
Delta R.T. 0.000 min
Lab File: BN014929.D
Acq: 16 Jun 2021 19:41

Tgt Ion:240 Resp: 75324
Ion Ratio Lower Upper
240 100
120 7.3 4.2 6.4#
236 25.4 21.4 32.2









#30
Pyrene
Concen: 6.158 ng
RT: 19.550 min Scan# 1870
Delta R.T. 0.000 min
Lab File: BN014929.D
Acq: 16 Jun 2021 19:41

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Tgt Ion:202 Resp: 1380809




Ion	Ratio	Lower	Upper
202	100		
200	20.6	16.9	25.3
203	17.7	13.9	20.9




#31
Terphenyl-d14
Concen: 4.652 ng
RT: 19.769 min Scan# 1914
Delta R.T. 0.000 min
Lab File: BN014929.D
Acq: 16 Jun 2021 19:41

Tgt Ion:244 Resp: 866944

Ion	Ratio	Lower	Upper
244	100		
212	7.4	6.3	9.5
122	11.5	6.2	9.2#









#32
Benzo(a)anthracene
Concen: 6.299 ng
RT: 21.303 min Scan# 2084
Delta R.T. 0.000 min
Lab File: BN014929.D
Acq: 16 Jun 2021 19:41




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:228 Resp: 1349291
Ion Ratio Lower Upper
228 100
226 27.2 23.0 34.6
229 19.7 15.7 23.5



#33
Chrysene
Concen: 5.123 ng
RT: 21.356 min Scan# 2089
Delta R.T. 0.000 min
Lab File: BN014929.D
Acq: 16 Jun 2021 19:41

Tgt Ion:228 Resp: 1287750
Ion Ratio Lower Upper
228 100
226 29.7 25.0 37.6
229 19.7 15.9 23.9









#34
Bis(2-ethylhexyl)phthalate
Concen: 15.053 ng
RT: 21.260 min Scan# 2080
Delta R.T. -0.010 min
Lab File: BN014929.D
Acq: 16 Jun 2021 19:41




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:149 Resp: 996108
Ion Ratio Lower Upper
149 100
167 28.9 28.9 43.3#
279 3.9 4.9 7.3#



#35
Perylene-d12
Concen: 0.400 ng
RT: 23.581 min Scan# 2531
Delta R.T. 0.000 min
Lab File: BN014929.D
Acq: 16 Jun 2021 19:41

Tgt Ion:264 Resp: 78367
Ion Ratio Lower Upper
264 100
260 25.0 20.6 31.0
265 23.1 21.4 32.2









#36
Indeno(1,2,3-cd)pyrene
Concen: 5.766 ng
RT: 25.864 min Scan# 3198
Delta R.T. 0.007 min
Lab File: BN014929.D
Acq: 16 Jun 2021 19:41




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:276 Resp: 1357379
Ion Ratio Lower Upper
276 100
138 26.9 10.0 15.0#
277 25.1 20.4 30.6



#37
Benzo(b)fluoranthene
Concen: 5.475 ng
RT: 22.907 min Scan# 2334
Delta R.T. 0.000 min
Lab File: BN014929.D
Acq: 16 Jun 2021 19:41


Tgt Ion:252 Resp: 1369220
Ion Ratio Lower Upper
252 100
253 22.1 18.5 27.7
125 10.5 5.8 8.6#








#38
Benzo(k)fluoranthene
Concen: 4.389 ng
RT: 22.951 min Scan# 2347
Delta R.T. 0.000 min
Lab File: BN014929.D
Acq: 16 Jun 2021 19:41

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0







Tgt Ion:252 Resp: 1306991
Ion Ratio Lower Upper
252 100
253 21.9 19.1 28.7
125 10.7 5.4 8.0#



#39
Benzo(a)pyrene
Concen: 5.281 ng
RT: 23.482 min Scan# 2502
Delta R.T. 0.000 min
Lab File: BN014929.D
Acq: 16 Jun 2021 19:41

Tgt Ion:252 Resp: 1251819
Ion Ratio Lower Upper
252 100
253 22.0 19.4 29.2
125 11.6 7.4 11.2#









#40
Dibenzo(a,h)anthracene
Concen: 6.168 ng
RT: 25.885 min Scan# 3204
Delta R.T. 0.004 min
Lab File: BN014929.D
Acq: 16 Jun 2021 19:41

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Tgt Ion:278 Resp: 1111209
Ion Ratio Lower Upper
278 100
139 17.9 9.0 13.4#
279 23.5 21.0 31.4



#41
Benzo(g,h,i)perylene
Concen: 5.141 ng
RT: 26.556 min Scan# 3400
Delta R.T. 0.004 min
Lab File: BN014929.D
Acq: 16 Jun 2021 19:41

Tgt Ion:276 Resp: 1170701
Ion Ratio Lower Upper
276 100
277 23.8 19.7 29.5
138 23.2 9.8 14.6#

