

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061721\
 Data File : BN014933.D
 Acq On : 17 Jun 2021 08:42
 Operator : CG/JU
 Sample : SST0.103
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SST0.1003

Quant Time: Jun 17 11:03:38 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN061721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 17 11:01:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.739	152	6775	0.400	ng/ul	0.00
4) Naphthalene-d8	10.510	136	25727	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.371	164	14900	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.118	188	31266	0.400	ng/ul	0.00
17) Chrysene-d12	21.320	240	27402	0.400	ng/ul	0.00
23) Perylene-d12	23.577	264	24288	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	0.000	96	0d	0.000	ng/ul	
6) 2-Methylnaphthalene-d10	12.105	152	4341	0.098	ng/ul	0.00
18) Fluoranthene-d10	19.153	212	9219	0.133	ng/ul	0.00
Target Compounds						
					Qvalue	
5) Naphthalene	10.560	128	8079	0.108	ng/ul#	95
7) 2-Methylnaphthalene	12.182	142	5313	0.099	ng/ul	100
8) 1-Methylnaphthalene	12.402	142	5199	0.100	ng/ul	99
10) Acenaphthylene	14.084	152	7758	0.144	ng/ul	97
11) Acenaphthene	14.431	153	5499	0.107	ng/ul	98
12) Fluorene	15.421	166	6339	0.101	ng/ul	98
15) Phenanthrene	17.160	178	10466	0.108	ng/ul	98
16) Anthracene	17.249	178	9792	0.129	ng/ul	98
19) Fluoranthene	19.181	202	11420	0.134	ng/ul	96
20) Pyrene	19.544	202	11888	0.138	ng/ul	97
21) Benzo(a)anthracene	21.303	228	10301	0.119	ng/ul	97
22) Chrysene	21.356	228	11103	0.112	ng/ul	98
24) Benzo(b)fluoranthene	22.902	252	9356	0.088	ng/ul	89
25) Benzo(k)fluoranthene	22.948	252	11001	0.098	ng/ul#	88
26) Benzo(a)pyrene	23.483	252	8851	0.108	ng/ul#	87
27) Indeno(1,2,3-cd)pyrene	25.862	276	9244	0.080	ng/ul#	98
28) Dibenzo(a,h)anthracene	25.879	278	7225	0.078	ng/ul#	86
29) Benzo(g,h,i)perylene	26.546	276	8810	0.089	ng/ul	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061721\
Data File : BN014933.D
Acq On : 17 Jun 2021 08:42
Operator : CG/JU
Sample : SST0.103
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD0.1003

Quant Time: Jun 17 11:03:38 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN061721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 17 11:01:34 2021
Response via : Initial Calibration

