

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061725\
 Data File : BN037303.D
 Acq On : 17 Jun 2025 13:21
 Operator : RC/JU
 Sample : PB168433BL
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK433

Quant Time: Jun 17 15:12:50 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN061725.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 17 15:03:28 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.574	152	842	0.400	ng/u1	0.00
4) Naphthalene-d8	10.354	136	1924	0.400	ng/u1 #	0.00
9) Acenaphthene-d10	14.225	164	1506	0.400	ng/u1	0.00
13) Phenanthrene-d10	16.982	188	2886	0.400	ng/u1 #	0.00
17) Chrysene-d12	21.182	240	3218	0.400	ng/u1 #	0.00
23) Perylene-d12	23.368	264	3757	0.400	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.063	96	664	0.850	ng/u1	0.00
6) 2-Methylnaphthalene-d10	11.960	152	1000	0.312	ng/u1	0.00
18) Fluoranthene-d10	19.017	212	3060	0.298	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061725\
Data File : BN037303.D
Acq On : 17 Jun 2025 13:21
Operator : RC/JU
Sample : PB168433BL
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK433

Quant Time: Jun 17 15:12:50 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN061725.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 17 15:03:28 2025
Response via : Initial Calibration

