

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061820\
 Data File : BN010961.D
 Acq On : 18 Jun 2020 19:47
 Operator : CG/JU
 Sample : L3020-22MSD
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampled :
 MW-23MSD

Quant Time: Jun 19 04:04:40 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 18 14:14:42 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.54	152	2552	0.40	ng	0.00
7) Naphthalene-d8	10.29	136	9155	0.40	ng	0.00
13) Acenaphthene-d10	14.16	164	5246	0.40	ng	0.00
19) Phenanthrene-d10	16.91	188	10848	0.40	ng	0.00
27) Chrysene-d12	21.11	240	11583	0.40	ng	-0.01
34) Perylene-d12	23.26	264	10804	0.40	ng	0.00

System Monitoring Compounds

4) 2-Fluorophenol	5.19	112	1279	0.26	ng	0.00
5) Phenol-d6	6.74	99	943	0.16	ng	0.00
8) Nitrobenzene-d5	8.67	82	2978	0.44	ng	0.00
11) 2-Methylnaphthalene-d10	11.88	152	5744	0.40	ng	0.00
14) 2,4,6-Tribromophenol	15.66	330	1019	0.54	ng	0.00
15) 2-Fluorobiphenyl	12.78	172	10131	0.49	ng	0.00
25) Fluoranthene-d10	18.95	212	11663	0.39	ng	0.00
29) Terphenyl-d14	19.56	244	13222	0.49	ng	0.00

Target Compounds


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) 1,4-Dioxane	3.22	88	428	0.171	ng	# 62
3) n-Nitrosodimethylamine	3.53	42	202	0.139	ng	# 65
6) bis(2-Chloroethyl)ether	6.99	93	2113	0.373	ng	97
9) Naphthalene	10.33	128	8731	0.379	ng	99
10) Hexachlorobutadiene	10.63	225	1716	0.302	ng	# 98
12) 2-Methylnaphthalene	11.96	142	5566	0.361	ng	98
16) Acenaphthylene	13.87	152	7904	0.368	ng	97
17) Acenaphthene	14.22	154	5240	0.358	ng	98
18) Fluorene	15.22	166	7152	0.378	ng	99
20) 4-Bromophenyl-phenylether	16.11	248	2243	0.344	ng	93
21) Hexachlorobenzene	16.22	284	2315	0.340	ng	96
22) Pentachlorophenol	16.57	266	2596	1.137	ng	95
23) Phenanthrene	16.94	178	12185	0.411	ng	99
24) Anthracene	17.04	178	10602	0.429	ng	98
26) Fluoranthene	18.98	202	13926	0.414	ng	# 97
28) Pyrene	19.34	202	14259	0.372	ng	99
30) Benzo(a)anthracene	21.10	228	14995	0.440	ng	98
31) Chrysene	21.15	228	15049	0.379	ng	98
32) Bis(2-ethylhexyl)phthalate	21.06	149	8048	0.709	ng	96
33) Indeno(1,2,3-cd)pyrene	25.37	276	14430	0.325	ng	# 94
35) Benzo(b)fluoranthene	22.63	252	13093	0.356	ng	# 97
36) Benzo(k)fluoranthene	22.67	252	13056	0.334	ng	97
37) Benzo(a)pyrene	23.17	252	11310	0.359	ng	97
38) Dibenzo(a,h)anthracene	25.38	278	12054	0.335	ng	# 96
39) Benzo(g,h,i)perylene	26.00	276	12865	0.343	ng	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061820\
 Data File : BN010961.D
 Acq On : 18 Jun 2020 19:47
 Operator : CG/JU
 Sample : L3020-22MSD
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 Client Sampled :
 MW-23MSD

Quant Time: Jun 19 04:04:40 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 18 14:14:42 2020
 Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.54 min Scan# 585
 Delta R.T. 0.01 min
 Lab File: BN010961.D
 Acq: 18 Jun 2020 19:47




Instrument :
 BNA_N
 ClientSampled :
 MW-23MSD


Tgt Ion	Resp	Lower	Upper
152	100		
150	146.6	123.8	185.6
115	59.5	49.2	73.8



#2
 1,4-Dioxane
 Concen: 0.171 ng
 RT: 3.22 min Scan# 49
 Delta R.T. 0.01 min
 Lab File: BN010961.D
 Acq: 18 Jun 2020 19:47

Tgt Ion	Resp	Lower	Upper
88	100		
58	64.5	38.2	57.4#
43	51.6	15.1	22.7#









#3
 n-Nitrosodimethylamine
 Concen: 0.139 ng
 RT: 3.53 min Scan# 87
 Delta R.T. -0.03 min
 Lab File: BN010961.D
 Acq: 18 Jun 2020 19:47




Instrument :
 BNA_N
 ClientSampled :
 MW-23MSD


Tgt Ion	Resp	Lower	Upper
42	100		
74	224.8	140.7	211.1#
44	0.0	1.4	2.2#



#4
 2-Fluorophenol
 Concen: 0.260 ng
 RT: 5.19 min Scan# 293
 Delta R.T. -0.00 min
 Lab File: BN010961.D
 Acq: 18 Jun 2020 19:47

Tgt Ion	Resp	Lower	Upper
112	100		
64	44.6	37.4	56.0
63	25.2	22.2	33.4









#5
 Phenol-d6
 Concen: 0.162 ng
 RT: 6.74 min Scan# 485
 Delta R.T. -0.00 min
 Lab File: BN010961.D
 Acq: 18 Jun 2020 19:47




Instrument :
 BNA_N
 ClientSampled :
 MW-23MSD


Tgt Ion	Resp	Lower	Upper
99	943		
42	17.4	11.1	16.7#
71	53.4	31.4	47.2#



#6
 bis(2-Chloroethyl)ether
 Concen: 0.373 ng
 RT: 6.99 min Scan# 516
 Delta R.T. -0.00 min
 Lab File: BN010961.D
 Acq: 18 Jun 2020 19:47

Tgt Ion	Resp	Lower	Upper
93	2113		
63	58.8	49.4	74.2
95	33.8	26.7	40.1









#7
 Naphthalene-d8
 Concen: 0.400 ng
 RT: 10.29 min Scan# 837
 Delta R.T. -0.00 min
 Lab File: BN010961.D
 Acq: 18 Jun 2020 19:47




Instrument :
 BNA_N
ClientSampled :
 MW-23MSD


Tgt Ion	Resp	Lower	Upper
136	100		
137	12.0	9.7	14.5
54	3.7	4.9	7.3#
68	4.9	4.7	7.1



#8
 Nitrobenzene-d5
 Concen: 0.440 ng
 RT: 8.67 min Scan# 709
 Delta R.T. -0.00 min
 Lab File: BN010961.D
 Acq: 18 Jun 2020 19:47

Tgt Ion	Resp	Lower	Upper
82	100		
128	42.0	35.4	53.0
54	34.9	28.2	42.4









#9
 Naphthalene
 Concen: 0.379 ng
 RT: 10.33 min Scan# 840
 Delta R.T. -0.00 min
 Lab File: BN010961.D
 Acq: 18 Jun 2020 19:47




Instrument :
 BNA_N
ClientSampled :
 MW-23MSD


Tgt Ion	Resp	Lower	Upper
128	100		
129	12.2	9.6	14.4
127	14.5	11.4	17.2



#10
 Hexachlorobutadiene
 Concen: 0.302 ng
 RT: 10.63 min Scan# 864
 Delta R.T. -0.00 min
 Lab File: BN010961.D
 Acq: 18 Jun 2020 19:47

Tgt Ion	Resp	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	61.8	50.5	75.7








#11
 2-Methylnaphthalene-d10
 Concen: 0.404 ng
 RT: 11.88 min Scan# 1016
 Delta R.T. -0.01 min
 Lab File: BN010961.D
 Acq: 18 Jun 2020 19:47





Instrument :
 BNA_N
 ClientSampled :
 MW-23MSD


Tgt Ion	Resp	Lower	Upper
152	100		
151	16.3	15.3	22.9



#12
 2-Methylnaphthalene
 Concen: 0.361 ng
 RT: 11.96 min Scan# 1034
 Delta R.T. -0.00 min
 Lab File: BN010961.D
 Acq: 18 Jun 2020 19:47

Tgt Ion	Resp	Lower	Upper
142	100		
141	89.3	70.8	106.2
115	37.2	31.5	47.3








#13
 Acenaphthene-d10
 Concen: 0.400 ng
 RT: 14.16 min Scan# 1306
 Delta R.T. -0.00 min
 Lab File: BN010961.D
 Acq: 18 Jun 2020 19:47





Instrument :
 BNA_N
 ClientSampled :
 MW-23MSD


Tgt Ion	Resp	Lower	Upper
164	100		
162	100.2	81.0	121.6
160	46.5	38.9	58.3



#14
 2,4,6-Tribromophenol
 Concen: 0.543 ng
 RT: 15.66 min Scan# 1441
 Delta R.T. -0.00 min
 Lab File: BN010961.D
 Acq: 18 Jun 2020 19:47

Tgt Ion	Resp	Lower	Upper
330	100		
332	0.0	0.0	0.0
141	69.5	33.2	49.8#









#15
 2-Fluorobiphenyl
 Concen: 0.489 ng
 RT: 12.78 min Scan# 1182
 Delta R.T. -0.00 min
 Lab File: BN010961.D
 Acq: 18 Jun 2020 19:47




Instrument :
 BNA_N
 ClientSampled :
 MW-23MSD


Tgt Ion	Resp	Lower	Upper
172	10131		
171	34.9	28.5	42.7
170	22.7	19.7	29.5



#16
 Acenaphthylene
 Concen: 0.368 ng
 RT: 13.87 min Scan# 1280
 Delta R.T. -0.01 min
 Lab File: BN010961.D
 Acq: 18 Jun 2020 19:47

Tgt Ion	Resp	Lower	Upper
152	7904		
151	20.8	15.6	23.4
153	14.8	10.6	16.0








#17
 Acenaphthene
 Concen: 0.358 ng
 RT: 14.22 min Scan# 1312
 Delta R.T. -0.00 min
 Lab File: BN010961.D
 Acq: 18 Jun 2020 19:47





Instrument :
 BNA_N
 ClientSampled :
 MW-23MSD


Tgt Ion	Resp	Lower	Upper
154	100		
153	116.3	91.8	137.6
152	56.4	47.6	71.4



#18
 Fluorene
 Concen: 0.378 ng
 RT: 15.22 min Scan# 1401
 Delta R.T. -0.00 min
 Lab File: BN010961.D
 Acq: 18 Jun 2020 19:47

Tgt Ion	Resp	Lower	Upper
166	100		
165	96.8	77.2	115.8
167	14.8	10.7	16.1








#19
 Phenanthrene-d10
 Concen: 0.400 ng
 RT: 16.91 min Scan# 1557
 Delta R.T. -0.00 min
 Lab File: BN010961.D
 Acq: 18 Jun 2020 19:47





Instrument :
 BNA_N
 ClientSampled :
 MW-23MSD


Tgt Ion	Resp	Lower	Upper
188	10848		
94	0.0	0.0	0.0
80	9.3	4.5	6.7#



#20
 4-Bromophenyl-phenylether
 Concen: 0.344 ng
 RT: 16.11 min Scan# 1482
 Delta R.T. -0.00 min
 Lab File: BN010961.D
 Acq: 18 Jun 2020 19:47

Tgt Ion	Resp	Lower	Upper
248	2243		
250	99.8	81.8	122.8
141	63.8	42.6	64.0









#21
 Hexachlorobenzene
 Concen: 0.340 ng
 RT: 16.22 min Scan# 1492
 Delta R.T. -0.01 min
 Lab File: BN010961.D
 Acq: 18 Jun 2020 19:47




Instrument :
 BNA_N
 ClientSampled :
 MW-23MSD


Tgt Ion	Resp	Lower	Upper
284	2315		
142	42.6	31.0	46.4
249	34.0	27.1	40.7



#22
 Pentachlorophenol
 Concen: 1.137 ng
 RT: 16.57 min Scan# 1525
 Delta R.T. -0.01 min
 Lab File: BN010961.D
 Acq: 18 Jun 2020 19:47

Tgt Ion	Resp	Lower	Upper
266	2596		
264	64.4	54.8	82.2
268	65.9	55.9	83.9









#23
 Phenanthrene
 Concen: 0.411 ng
 RT: 16.94 min Scan# 1560
 Delta R.T. -0.01 min
 Lab File: BN010961.D
 Acq: 18 Jun 2020 19:47




Instrument :
 BNA_N
 ClientSampled :
 MW-23MSD


Tgt Ion	Resp	Lower	Upper
178	12185		
176	18.6	15.3	22.9
179	16.4	12.6	19.0



#24
 Anthracene
 Concen: 0.429 ng
 RT: 17.04 min Scan# 1569
 Delta R.T. -0.01 min
 Lab File: BN010961.D
 Acq: 18 Jun 2020 19:47

Tgt Ion	Resp	Lower	Upper
178	10602		
176	19.6	14.6	22.0
179	14.9	12.3	18.5









#25
 Fluoranthene-d10
 Concen: 0.390 ng
 RT: 18.95 min Scan# 1850
 Delta R.T. -0.01 min
 Lab File: BN010961.D
 Acq: 18 Jun 2020 19:47




Instrument :
 BNA_N
 ClientSampleID :
 MW-23MSD


Tgt Ion	Resp	Lower	Upper
212	11663		
106	10.5	6.6	10.0#
104	5.9	4.0	6.0



#26
 Fluoranthene
 Concen: 0.414 ng
 RT: 18.98 min Scan# 1856
 Delta R.T. -0.00 min
 Lab File: BN010961.D
 Acq: 18 Jun 2020 19:47

Tgt Ion	Resp	Lower	Upper
202	13926		
101	9.2	5.8	8.6#
203	17.7	13.7	20.5









#27
 Chrysene-d12
 Concen: 0.400 ng
 RT: 21.11 min Scan# 2167
 Delta R.T. -0.01 min
 Lab File: BN010961.D
 Acq: 18 Jun 2020 19:47




Instrument :
 BNA_N
 ClientSampled :
 MW-23MSD


Tgt Ion	Resp	Lower	Upper
240	11583		
120	16.1	7.5	11.3#
236	30.2	22.3	33.5



#28
 Pyrene
 Concen: 0.372 ng
 RT: 19.34 min Scan# 1929
 Delta R.T. -0.01 min
 Lab File: BN010961.D
 Acq: 18 Jun 2020 19:47

Tgt Ion	Resp	Lower	Upper
202	14259		
200	20.5	16.4	24.6
203	18.4	14.2	21.4









#29
 Terphenyl-d14
 Concen: 0.488 ng
 RT: 19.56 min Scan# 1973
 Delta R.T. -0.01 min
 Lab File: BN010961.D
 Acq: 18 Jun 2020 19:47




Instrument :
 BNA_N
ClientSampled :
 MW-23MSD


Tgt Ion	Resp	Lower	Upper
244	13222		
212	8.9	7.4	11.2
122	13.2	7.5	11.3#



#30
 Benzo(a)anthracene
 Concen: 0.440 ng
 RT: 21.10 min Scan# 2166
 Delta R.T. -0.00 min
 Lab File: BN010961.D
 Acq: 18 Jun 2020 19:47

Tgt Ion	Resp	Lower	Upper
228	14995		
226	27.4	21.3	31.9
229	21.6	16.6	24.8









#31
 Chrysene
 Concen: 0.379 ng
 RT: 21.15 min Scan# 2171
 Delta R.T. -0.00 min
 Lab File: BN010961.D
 Acq: 18 Jun 2020 19:47




Instrument :
 BNA_N
 ClientSampled :
 MW-23MSD


Tgt Ion	Resp	Lower	Upper
228	15049		
226	29.9	25.0	37.6
229	21.2	16.1	24.1



#32
 Bis(2-ethylhexyl)phthalate
 Concen: 0.709 ng
 RT: 21.06 min Scan# 2162
 Delta R.T. -0.00 min
 Lab File: BN010961.D
 Acq: 18 Jun 2020 19:47

Tgt Ion	Resp	Lower	Upper
149	8048		
167	27.4	24.1	36.1
279	5.5	4.6	7.0









#33
 Indeno(1,2,3-cd)pyrene
 Concen: 0.325 ng
 RT: 25.37 min Scan# 3154
 Delta R.T. -0.00 min
 Lab File: BN010961.D
 Acq: 18 Jun 2020 19:47




Instrument :
 BNA_N
 ClientSampled :
 MW-23MSD


Tgt Ion	Resp	Lower	Upper
276	14430		
138	21.3	13.0	19.6#
277	25.4	19.7	29.5



#34
 Perylene-d12
 Concen: 0.400 ng
 RT: 23.26 min Scan# 2538
 Delta R.T. -0.01 min
 Lab File: BN010961.D
 Acq: 18 Jun 2020 19:47

Tgt Ion	Resp	Lower	Upper
264	10804		
260	24.9	20.4	30.6
265	69.0	75.0	112.4#









#35
 Benzo(b)fluoranthene
 Concen: 0.356 ng
 RT: 22.63 min Scan# 2354
 Delta R.T. -0.00 min
 Lab File: BN010961.D
 Acq: 18 Jun 2020 19:47




Instrument :
 BNA_N
 ClientSampleId :
 MW-23MSD


Tgt Ion	Resp	Lower	Upper
252	13093		
253	29.0	22.4	33.6
125	14.3	9.4	14.0#



#36
 Benzo(k)fluoranthene
 Concen: 0.334 ng
 RT: 22.67 min Scan# 2366
 Delta R.T. -0.00 min
 Lab File: BN010961.D
 Acq: 18 Jun 2020 19:47

Tgt Ion	Resp	Lower	Upper
252	13056		
253	28.7	22.2	33.4
125	14.3	9.8	14.8









#37
 Benzo(a)pyrene
 Concen: 0.359 ng
 RT: 23.17 min Scan# 2512
 Delta R.T. -0.00 min
 Lab File: BN010961.D
 Acq: 18 Jun 2020 19:47




Instrument :
 BNA_N
 ClientSampled :
 MW-23MSD


Tgt Ion	Resp	Lower	Upper
252	11310		
253	30.6	25.8	38.8
125	17.1	12.4	18.6



#38
 Dibenzo(a,h)anthracene
 Concen: 0.335 ng
 RT: 25.38 min Scan# 3157
 Delta R.T. -0.01 min
 Lab File: BN010961.D
 Acq: 18 Jun 2020 19:47

Tgt Ion	Resp	Lower	Upper
278	12054		
139	16.6	10.6	15.8#
279	28.1	23.6	35.4





#39
 Benzo(a,h,i)perylene
 Concen: 0.343 ng
 RT: 26.00 min Scan# 3338
 Delta R.T. -0.01 min
 Lab File: BN010961.D
 Acq: 18 Jun 2020 19:47

Instrument :
 BNA_N
 ClientSampleId :
 MW-23MSD

Tot Ion	276	Resp	12865
Ion	Ratio	Lower	Upper
276	100		
277	25.6	20.5	30.7
138	20.4	13.7	20.5

