

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN061820\
 Data File : BN010954.D
 Acq On : 18 Jun 2020 13:15
 Operator : CG/JU
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDC00.4

Quant Time: Jun 18 14:16:09 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 18 14:14:42 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	110	0.00
2	1,4-Dioxane	0.392	0.472	-20.4	126	0.00
3	n-Nitrosodimethylamine	0.228	0.253	-11.0	138	0.04
4 S	2-Fluorophenol	0.772	0.856	-10.9	119	0.00
5 S	Phenol-d6	0.910	1.000	-9.9	120	0.00
6	bis(2-Chloroethyl)ether	0.887	0.902	-1.7	112	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	113	0.01
8 S	Nitrobenzene-d5	0.296	0.338	-14.2	124	0.00
9	Naphthalene	1.005	1.063	-5.8	114	0.00
10	Hexachlorobutadiene	0.248	0.255	-2.8	108	0.01
11 SURR	2-Methylnaphthalene-d10	0.620	0.642	-3.5	112	0.00
12	2-Methylnaphthalene	0.673	0.699	-3.9	114	0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	112	0.00
14 S	2,4,6-Tribromophenol	0.143	0.137	4.2	111	0.01
15 S	2-Fluorobiphenyl	1.578	1.682	-6.6	112	0.00
16	Acenaphthylene	1.639	1.828	-11.5	124	0.01
17	Acenaphthene	1.117	1.170	-4.7	113	0.01
18	Fluorene	1.441	1.478	-2.6	112	0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	109	0.01
20	4-Bromophenyl-phenylether	0.240	0.242	-0.8	107	0.01
21	Hexachlorobenzene	0.251	0.267	-6.4	110	0.01
22	Pentachlorophenol	0.084	0.091	-8.3	121	0.01
23	Phenanthrene	1.093	1.109	-1.5	108	0.01
24	Anthracene	0.912	0.920	-0.9	111	0.01
25 SURR	Fluoranthene-d10	1.104	1.102	0.2	107	0.00
26	Fluoranthene	1.240	1.239	0.1	109	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	110	0.00
28	Pvrene	1.323	1.379	-4.2	110	0.00
29 S	Terphenyl-d14	0.936	0.976	-4.3	109	0.00
30	Benzo(a)anthracene	1.176	1.187	-0.9	112	0.00
31	Chrysene	1.370	1.437	-4.9	106	0.00
32	Bis(2-ethylhexyl)phthalate	0.392	0.420	-7.1	120	0.00
33	Indeno(1,2,3-cd)pyrene	1.531	1.631	-6.5	112	0.00
34 I	Perylene-d12	1.000	1.000	0.0	120	0.00
35	Benzo(b)fluoranthene	1.360	1.320	2.9	112	0.00
36	Benzo(k)fluoranthene	1.449	1.477	-1.9	116	0.00
37 C	Benzo(a)pyrene	1.167	1.140	2.3	117	0.00
38	Dibenzo(a,h)anthracene	1.330	1.237	7.0	110	0.00
39	Benzo(a,h,i)perylene	1.391	1.335	4.0	108	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN061820\
Data File : BN010954.D
Acq On : 18 Jun 2020 13:15
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 18 14:16:09 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 18 14:14:42 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			