

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061820\
 Data File : BN010957.D
 Acq On : 18 Jun 2020 17:25
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 18 17:59:31 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 18 14:14:42 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	103	0.00
2	1,4-Dioxane	0.392	0.464	-18.4	116	0.00
3	n-Nitrosodimethylamine	0.228	0.248	-8.8	127	0.03
4 S	2-Fluorophenol	0.772	0.852	-10.4	111	0.00
5 S	Phenol-d6	0.910	1.003	-10.2	113	0.00
6	bis(2-Chloroethyl)ether	0.887	0.932	-5.1	108	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	109	0.01
8 S	Nitrobenzene-d5	0.296	0.323	-9.1	114	0.01
9	Naphthalene	1.005	1.062	-5.7	109	0.00
10	Hexachlorobutadiene	0.248	0.251	-1.2	102	0.01
11 SURR	2-Methylnaphthalene-d10	0.620	0.850	-37.1#	142	0.00
12	2-Methylnaphthalene	0.673	0.797	-18.4	125	0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	108	0.00
14 S	2,4,6-Tribromophenol	0.143	0.149	-4.2	117	0.02
15 S	2-Fluorobiphenyl	1.578	1.728	-9.5	112	0.01
16	Acenaphthylene	1.639	1.696	-3.5	111	0.01
17	Acenaphthene	1.117	1.173	-5.0	109	0.01
18	Fluorene	1.441	1.495	-3.7	110	0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	106	0.01
20	4-Bromophenyl-phenylether	0.240	0.271	-12.9	117	0.01
21	Hexachlorobenzene	0.251	0.286	-13.9	116	0.01
22	Pentachlorophenol	0.084	0.083	1.2	108	0.01
23	Phenanthrene	1.093	1.109	-1.5	105	0.01
24	Anthracene	0.912	0.911	0.1	107	0.01
25 SURR	Fluoranthene-d10	1.104	1.314	-19.0	125	0.00
26	Fluoranthene	1.240	1.242	-0.2	107	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	106	0.00
28	Pvrene	1.323	1.391	-5.1	107	0.00
29 S	Terphenyl-d14	0.936	1.092	-16.7	118	0.00
30	Benzo(a)anthracene	1.176	1.198	-1.9	109	0.00
31	Chrysene	1.370	1.459	-6.5	104	0.00
32	Bis(2-ethylhexyl)phthalate	0.392	0.435	-11.0	120	0.00
33	Indeno(1,2,3-cd)pyrene	1.531	1.745	-14.0	115	0.00
34 I	Perylene-d12	1.000	1.000	0.0	119	0.00
35	Benzo(b)fluoranthene	1.360	1.282	5.7	108	0.00
36	Benzo(k)fluoranthene	1.449	1.584	-9.3	124	0.00
37 C	Benzo(a)pyrene	1.167	1.208	-3.5	123	0.00
38	Dibenzo(a,h)anthracene	1.330	1.280	3.8	113	0.00
39	Benzo(a,h,i)perylene	1.391	1.334	4.1	107	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061820\
Data File : BN010957.D
Acq On : 18 Jun 2020 17:25
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 18 17:59:31 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 18 14:14:42 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			