

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061820\
 Data File : BN010957.D
 Acq On : 18 Jun 2020 17:25
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 18 17:59:31 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 18 14:14:42 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	103	0.00
2	1,4-Dioxane	0.400	0.474	-18.5	116	0.00
3	n-Nitrosodimethylamine	0.400	0.436	-9.0	127	0.03
4 S	2-Fluorophenol	0.400	0.441	-10.2	111	0.00
5 S	Phenol-d6	0.400	0.441	-10.2	113	0.00
6	bis(2-Chloroethyl)ether	0.400	0.420	-5.0	108	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	109	0.01
8 S	Nitrobenzene-d5	0.400	0.436	-9.0	114	0.01
9	Naphthalene	0.400	0.422	-5.5	109	0.00
10	Hexachlorobutadiene	0.400	0.404	-1.0	102	0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.548	-37.0#	142	0.00
12	2-Methylnaphthalene	0.400	0.474	-18.5	125	0.01
13 I	Acenaphthene-d10	0.400	0.400	0.0	108	0.00
14 S	2,4,6-Tribromophenol	0.400	0.417	-4.2	117	0.02
15 S	2-Fluorobiphenyl	0.400	0.438	-9.5	112	0.01
16	Acenaphthylene	0.400	0.414	-3.5	111	0.01
17	Acenaphthene	0.400	0.420	-5.0	109	0.01
18	Fluorene	0.400	0.415	-3.7	110	0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	106	0.01
20	4-Bromophenyl-phenylether	0.400	0.450	-12.5	117	0.01
21	Hexachlorobenzene	0.400	0.457	-14.2	116	0.01
22	Pentachlorophenol	0.400	0.393	1.8	108	0.01
23	Phenanthrene	0.400	0.406	-1.5	105	0.01
24	Anthracene	0.400	0.399	0.3	107	0.01
25 SURR	Fluoranthene-d10	0.400	0.476	-19.0	125	0.00
26	Fluoranthene	0.400	0.401	-0.3	107	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	106	0.00
28	Pvrene	0.400	0.421	-5.2	107	0.00
29 S	Terphenyl-d14	0.400	0.467	-16.8	118	0.00
30	Benzo(a)anthracene	0.400	0.407	-1.7	109	0.00
31	Chrysene	0.400	0.426	-6.5	104	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.444	-11.0	120	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.456	-14.0	115	0.00
34 I	Perylene-d12	0.400	0.400	0.0	119	0.00
35	Benzo(b)fluoranthene	0.400	0.377	5.8	108	0.00
36	Benzo(k)fluoranthene	0.400	0.437	-9.2	124	0.00
37 C	Benzo(a)pyrene	0.400	0.414	-3.5	123	0.00
38	Dibenzo(a,h)anthracene	0.400	0.385	3.8	113	0.00
39	Benzo(a,h,i)perylene	0.400	0.384	4.0	107	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061820\
Data File : BN010957.D
Acq On : 18 Jun 2020 17:25
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 18 17:59:31 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 18 14:14:42 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		