

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061920\  
 Data File : BN010970.D  
 Acq On : 19 Jun 2020 15:12  
 Operator : CG/JU  
 Sample : L3043-03  
 Misc :  
 ALS Vial : 8 Sample Multiplier: 1

Instrument :  
 BNA\_N  
 ClientSampleId :  
 DUP05-20200615

Quant Time: Jun 19 15:53:01 2020  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Fri Jun 19 12:43:49 2020  
 Response via : Initial Calibration

| Internal Standards             | R.T.  | QIon | Response | Conc  | Units | Dev(Min) |
|--------------------------------|-------|------|----------|-------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4      | 7.53  | 152  | 2038     | 0.40  | ng    | 0.00     |
| 7) Naphthalene-d8              | 10.28 | 136  | 7343     | 0.40  | ng    | 0.00     |
| 13) Acenaphthene-d10           | 14.16 | 164  | 3821     | 0.40  | ng    | 0.00     |
| 19) Phenanthrene-d10           | 16.91 | 188  | 8227     | 0.40  | ng    | 0.01     |
| 27) Chrysene-d12               | 21.12 | 240  | 7805     | 0.40  | ng    | 0.00     |
| 34) Perylene-d12               | 23.26 | 264  | 7814     | 0.40  | ng    | 0.00     |
| System Monitoring Compounds    |       |      |          |       |       |          |
| 4) 2-Fluorophenol              | 5.18  | 112  | 619      | 0.16  | ng    | 0.00     |
| 5) Phenol-d6                   | 6.73  | 99   | 413      | 0.09  | ng    | 0.00     |
| 8) Nitrobenzene-d5             | 8.67  | 82   | 2176     | 0.40  | ng    | 0.00     |
| 11) 2-Methylnaphthalene-d10    | 11.88 | 152  | 4787     | 0.42  | ng    | 0.00     |
| 14) 2,4,6-Tribromophenol       | 15.66 | 330  | 534      | 0.39  | ng    | 0.01     |
| 15) 2-Fluorobiphenyl           | 12.78 | 172  | 6105     | 0.40  | ng    | 0.00     |
| 25) Fluoranthene-d10           | 18.95 | 212  | 13189    | 0.58  | ng    | 0.00     |
| 29) Terphenyl-d14              | 19.56 | 244  | 9874     | 0.54  | ng    | 0.00     |
| Target Compounds               |       |      |          |       |       |          |
| 2) 1,4-Dioxane                 | 3.20  | 88   | 11432    | 5.727 | ng    | 98       |
| 6) bis(2-Chloroethyl)ether     | 6.98  | 93   | 338      | 0.075 | ng    | 98       |
| 32) Bis(2-ethylhexyl)phthalate | 21.06 | 149  | 468      | 0.061 | ng    | 93       |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061920\  
Data File : BN010970.D  
Acq On : 19 Jun 2020 15:12  
Operator : CG/JU  
Sample : L3043-03  
Misc :  
ALS Vial : 8 Sample Multiplier: 1

Instrument :  
BNA\_N  
ClientSampleId :  
DUP05-20200615

Quant Time: Jun 19 15:53:01 2020  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Fri Jun 19 12:43:49 2020  
Response via : Initial Calibration

