

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN061920\
 Data File : BN010964.D
 Acq On : 19 Jun 2020 11:37
 Operator : CG/JU
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDC00.4

Quant Time: Jun 19 12:46:20 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 19 12:43:49 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	179#	0.00
2	1,4-Dioxane	0.392	0.453	-15.6	198#	0.00
3	n-Nitrosodimethylamine	0.228	0.305	-33.8#	273#	0.00
4 S	2-Fluorophenol	0.772	0.976	-26.4#	222#	0.00
5 S	Phenol-d6	0.910	1.129	-24.1	222#	0.00
6	bis(2-Chloroethyl)ether	0.887	1.059	-19.4	214#	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	177#	0.00
8 S	Nitrobenzene-d5	0.296	0.330	-11.5	190#	0.00
9	Naphthalene	1.005	1.138	-13.2	191#	0.00
10	Hexachlorobutadiene	0.248	0.245	1.2	162#	0.00
11 SURR	2-Methylnaphthalene-d10	0.620	0.673	-8.5	183#	0.00
12	2-Methylnaphthalene	0.673	0.733	-8.9	188#	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	196#	0.00
14 S	2,4,6-Tribromophenol	0.143	0.172	-20.3	244#	0.00
15 S	2-Fluorobiphenyl	1.578	1.683	-6.7	197#	0.00
16	Acenaphthylene	1.639	1.768	-7.9	210#	0.00
17	Acenaphthene	1.117	1.137	-1.8	192#	0.00
18	Fluorene	1.441	1.522	-5.6	202#	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	202#	0.00
20	4-Bromophenyl-phenylether	0.240	0.243	-1.3	200#	0.00
21	Hexachlorobenzene	0.251	0.243	3.2	187#	0.00
22	Pentachlorophenol	0.084	0.111	-32.1#	277#	0.00
23	Phenanthrene	1.093	1.160	-6.1	210#	0.00
24	Anthracene	0.912	1.003	-10.0	225#	0.00
25 SURR	Fluoranthene-d10	1.104	1.207	-9.3	219#	0.00
26	Fluoranthene	1.240	1.350	-8.9	220#	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	224#	0.00
28	Pvrene	1.323	1.389	-5.0	226#	0.00
29 S	Terphenyl-d14	0.936	0.986	-5.3	225#	0.00
30	Benzo(a)anthracene	1.176	1.343	-14.2	258#	0.00
31	Chrysene	1.370	1.393	-1.7	210#	0.00
32	Bis(2-ethylhexyl)phthalate	0.392	0.590	-50.5#	344#	0.00
33	Indeno(1,2,3-cd)pyrene	1.531	1.663	-8.6	232#	0.00
34 I	Perylene-d12	1.000	1.000	0.0	252#	0.00
35	Benzo(b)fluoranthene	1.360	1.272	6.5	228#	0.00
36	Benzo(k)fluoranthene	1.449	1.357	6.3	224#	0.00
37 C	Benzo(a)pyrene	1.167	1.161	0.5	250#	0.00
38	Dibenzo(a,h)anthracene	1.330	1.258	5.4	234#	0.00
39	Benzo(a,h,i)perylene	1.391	1.276	8.3	217#	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN061920\
Data File : BN010964.D
Acq On : 19 Jun 2020 11:37
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 19 12:46:20 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 19 12:43:49 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		