

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN061920\
 Data File : BN010981.D
 Acq On : 19 Jun 2020 21:46
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 20 06:17:19 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 19 12:43:49 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	106	0.00
2	1,4-Dioxane	0.400	0.478	-19.5	121	0.00
3	n-Nitrosodimethylamine	0.400	0.506	-26.5#	152	0.02
4 S	2-Fluorophenol	0.400	0.468	-17.0	121	0.00
5 S	Phenol-d6	0.400	0.456	-14.0	121	0.00
6	bis(2-Chloroethyl)ether	0.400	0.473	-18.2	126	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	110	0.00
8 S	Nitrobenzene-d5	0.400	0.433	-8.2	115	0.00
9	Naphthalene	0.400	0.417	-4.2	110	0.00
10	Hexachlorobutadiene	0.400	0.399	0.3	102	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.406	-1.5	107	0.00
12	2-Methylnaphthalene	0.400	0.412	-3.0	111	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	114	0.00
14 S	2,4,6-Tribromophenol	0.400	0.413	-3.2	122	0.01
15 S	2-Fluorobiphenyl	0.400	0.408	-2.0	109	0.00
16	Acenaphthylene	0.400	0.409	-2.2	116	0.00
17	Acenaphthene	0.400	0.406	-1.5	111	0.00
18	Fluorene	0.400	0.421	-5.2	117	0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	113	0.00
20	4-Bromophenyl-phenylether	0.400	0.392	2.0	109	0.01
21	Hexachlorobenzene	0.400	0.391	2.3	106	0.00
22	Pentachlorophenol	0.400	0.395	1.3	116	0.00
23	Phenanthrene	0.400	0.404	-1.0	112	0.00
24	Anthracene	0.400	0.402	-0.5	115	0.00
25 SURR	Fluoranthene-d10	0.400	0.405	-1.3	114	0.00
26	Fluoranthene	0.400	0.402	-0.5	114	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	115	-0.01
28	Pvrene	0.400	0.417	-4.2	115	0.00
29 S	Terphenyl-d14	0.400	0.413	-3.2	113	0.00
30	Benzo(a)anthracene	0.400	0.403	-0.8	117	0.00
31	Chrysene	0.400	0.420	-5.0	111	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.481	-20.2	141	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.445	-11.2	122	0.00
34 I	Perylene-d12	0.400	0.400	0.0	128	0.00
35	Benzo(b)fluoranthene	0.400	0.379	5.3	117	0.00
36	Benzo(k)fluoranthene	0.400	0.389	2.8	118	0.00
37 C	Benzo(a)pyrene	0.400	0.400	0.0	127	0.00
38	Dibenzo(a,h)anthracene	0.400	0.386	3.5	121	0.00
39	Benzo(a,h,i)perylene	0.400	0.386	3.5	116	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN061920\
Data File : BN010981.D
Acq On : 19 Jun 2020 21:46
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 20 06:17:19 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 19 12:43:49 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				