

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN061920\
 Data File : BN010981.D
 Acq On : 19 Jun 2020 21:46
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 20 06:17:19 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 19 12:43:49 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	106	0.00
2	1,4-Dioxane	0.392	0.468	-19.4	121	0.00
3	n-Nitrosodimethylamine	0.228	0.288	-26.3#	152#	0.02
4 S	2-Fluorophenol	0.772	0.904	-17.1	121	0.00
5 S	Phenol-d6	0.910	1.037	-14.0	121	0.00
6	bis(2-Chloroethyl)ether	0.887	1.050	-18.4	126	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	110	0.00
8 S	Nitrobenzene-d5	0.296	0.320	-8.1	115	0.00
9	Naphthalene	1.005	1.047	-4.2	110	0.00
10	Hexachlorobutadiene	0.248	0.248	0.0	102	0.00
11 SURR	2-Methylnaphthalene-d10	0.620	0.630	-1.6	107	0.00
12	2-Methylnaphthalene	0.673	0.692	-2.8	111	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	114	0.00
14 S	2,4,6-Tribromophenol	0.143	0.148	-3.5	122	0.01
15 S	2-Fluorobiphenyl	1.578	1.610	-2.0	109	0.00
16	Acenaphthylene	1.639	1.676	-2.3	116	0.00
17	Acenaphthene	1.117	1.134	-1.5	111	0.00
18	Fluorene	1.441	1.519	-5.4	117	0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	113	0.00
20	4-Bromophenyl-phenylether	0.240	0.236	1.7	109	0.01
21	Hexachlorobenzene	0.251	0.245	2.4	106	0.00
22	Pentachlorophenol	0.084	0.083	1.2	116	0.00
23	Phenanthrene	1.093	1.103	-0.9	112	0.00
24	Anthracene	0.912	0.917	-0.5	115	0.00
25 SURR	Fluoranthene-d10	1.104	1.117	-1.2	114	0.00
26	Fluoranthene	1.240	1.246	-0.5	114	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	115	-0.01
28	Pvrene	1.323	1.379	-4.2	115	0.00
29 S	Terphenyl-d14	0.936	0.967	-3.3	113	0.00
30	Benzo(a)anthracene	1.176	1.186	-0.9	117	0.00
31	Chrysene	1.370	1.438	-5.0	111	0.00
32	Bis(2-ethylhexyl)phthalate	0.392	0.471	-20.2	141	0.00
33	Indeno(1,2,3-cd)pyrene	1.531	1.703	-11.2	122	0.00
34 I	Perylene-d12	1.000	1.000	0.0	128	0.00
35	Benzo(b)fluoranthene	1.360	1.288	5.3	117	0.00
36	Benzo(k)fluoranthene	1.449	1.408	2.8	118	0.00
37 C	Benzo(a)pyrene	1.167	1.166	0.1	127	0.00
38	Dibenzo(a,h)anthracene	1.330	1.285	3.4	121	0.00
39	Benzo(a,h,i)perylene	1.391	1.341	3.6	116	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN061920\
Data File : BN010981.D
Acq On : 19 Jun 2020 21:46
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 20 06:17:19 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 19 12:43:49 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			