

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN061920\
 Data File : BN010983.D
 Acq On : 20 Jun 2020 01:22
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 20 06:17:46 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 19 12:43:49 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	109	0.00
2	1,4-Dioxane	0.400	0.481	-20.2	125	0.00
3	n-Nitrosodimethylamine	0.400	0.463	-15.8	143	0.00
4 S	2-Fluorophenol	0.400	0.452	-13.0	121	0.00
5 S	Phenol-d6	0.400	0.487	-21.7	133	0.00
6	bis(2-Chloroethyl)ether	0.400	0.510	-27.5#	139	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	111	0.00
8 S	Nitrobenzene-d5	0.400	0.418	-4.5	112	0.00
9	Naphthalene	0.400	0.415	-3.7	110	0.00
10	Hexachlorobutadiene	0.400	0.402	-0.5	104	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.417	-4.2	110	0.00
12	2-Methylnaphthalene	0.400	0.415	-3.7	112	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	116	0.00
14 S	2,4,6-Tribromophenol	0.400	0.399	0.3	119	0.00
15 S	2-Fluorobiphenyl	0.400	0.418	-4.5	113	0.00
16	Acenaphthylene	0.400	0.413	-3.2	118	0.00
17	Acenaphthene	0.400	0.456	-14.0	127	0.00
18	Fluorene	0.400	0.410	-2.5	115	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	120	0.00
20	4-Bromophenyl-phenylether	0.400	0.390	2.5	115	0.00
21	Hexachlorobenzene	0.400	0.398	0.5	114	0.00
22	Pentachlorophenol	0.400	0.393	1.8	122	0.00
23	Phenanthrene	0.400	0.410	-2.5	120	0.00
24	Anthracene	0.400	0.407	-1.7	124	0.00
25 SURR	Fluoranthene-d10	0.400	0.410	-2.5	122	0.00
26	Fluoranthene	0.400	0.406	-1.5	122	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	133	0.00
28	Pvrene	0.400	0.390	2.5	125	0.00
29 S	Terphenyl-d14	0.400	0.391	2.3	124	0.00
30	Benzo(a)anthracene	0.400	0.399	0.3	134	0.00
31	Chrysene	0.400	0.417	-4.2	128	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.447	-11.7	152	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.451	-12.7	144	0.00
34 I	Perylene-d12	0.400	0.400	0.0	162	0.00
35	Benzo(b)fluoranthene	0.400	0.401	-0.3	157	0.00
36	Benzo(k)fluoranthene	0.400	0.392	2.0	151	0.00
37 C	Benzo(a)pyrene	0.400	0.401	-0.3	162	0.00
38	Dibenzo(a,h)anthracene	0.400	0.364	9.0	145	0.00
39	Benzo(a,h,i)perylene	0.400	0.357	10.8	136	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN061920\
Data File : BN010983.D
Acq On : 20 Jun 2020 01:22
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 20 06:17:46 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 19 12:43:49 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				